

Technical Information Report



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.


For a complete copy of this AAMI document, contact AAMI at +1-977-249-8226 or visit www.aami.org.

ANSI/AAMI/ ISO TIR17665 -2:2009/ (R)2016

Sterilization of health care products – Moist Heat – Part 2: Guidance on the application of ANSI/AAMI/ISO 17665-1

A Technical Report prepared by AAMI and registered with ANSI

ANSI/AAMI/ISO TIR17665-2:2009/(R)2016
(Revision of AAMI TIR13:1997, and with ANSI/
AAMI/ISO 17665-1:2006, revision of
ANSI/AAMI/ISO 11134:1993)



**Sterilization of health care products – Moist Heat –
Part 2: Guidance on the application of
ANSI/AAMI/ISO 17665-1**

PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at

Approved 18 June 2009 by **AAMI** +1-977-249-8226 or visit www.aami.org.

Registered 17 May 2009 and reaffirmed 19 June 2016 by
American National Standards Institute, Inc.

Abstract: AAMI/ISO TIR17665-2 provides guidance for validation and routine control of moist heat sterilization processes for medical devices.

Keywords: health care products, medical equipment, moist heat, sterilization, steam, air-steam mixtures, sterilizing agent, product definition, process definition, routine control, sterility, product release, validation, water-immersion



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

Published by

AAMI

4301 N. Fairfax Drive, Suite 301
Arlington, VA 22203-1633

www.aami.org

For a complete copy of this AAMI document, contact AAMI at 1-977-249-8226 or visit www.aami.org.

© 2009 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of ISO, ANSI, and AAMI. No part of this publication may be reproduced or distributed in any form, including an electronic retrieval system, without the prior written permission of AAMI. All requests pertaining to this document should be submitted to AAMI. It is illegal under federal law (17 U.S.C. § 101, et seq.) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, complete the reprint request form at www.aami.org or contact AAMI, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633. Phone: (703) 525-4890; Fax: (703) 276-0793.

Printed in the United States of America

ISBN 1-57020-349-0

AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

Although the material presented in a TIR might need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every five years but at least every 10 years. For a TIR, AAMI consults with a technical committee about five years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

CAUTION NOTICE: This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that might be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are voluntary, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633.

For a complete copy of this AAMI document, contact AAMI at
+1-977-249-8226 or visit www.aami.org.

ANSI Technical Report


This AAMI TIR has been registered by the American National Standards Institute as an ANSI Technical Report.

Publication of this ANSI Technical Report has been approved by the accredited standards developer (AAMI). This document is registered as a Technical Report series of publications according to the Procedures for the Registration of Technical Reports with ANSI. This document is not an American National Standard and the material contained herein is not normative in nature.

Comments on the content of this document should be sent to AAMI, Attn: Standards Department, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633.

Contents

	Page
Glossary of equivalent standards	vi
Committee representation.....	viii
Background of AAMI adoption of ISO TIR17665-2:2009	x
AAMI inclusion to ISO TIR17665-2:2009	xi
Foreword.....	xii
Introduction	xiii
1 Scope.....	1
2 Normative references	1
3 Terms and definitions.....	2
4 Quality management system elements.....	2
5 Sterilizing agent characterization.....	2
5.1 Sterilizing agent	2
5.2 Microbicidal effectiveness	2
5.3 Material effects	3
5.4 Environmental considerations	3
6 Process and equipment characterization.....	3
6.1 Process	3
6.2 Equipment.....	6
7 Product definition	8
8 Process definition.....	9
9 Validation.....	11
9.1 General	11
9.2 Installation qualification (IQ).....	12
9.3 Operational qualification (OQ).....	13
9.4 Performance qualification (PQ).....	14
9.5 Review and approval of the validation.....	16
10 Routine monitoring and control	17
11 Product release from sterilization	18
12 Maintaining process effectiveness	19
12.1 Demonstration of continued effectiveness	19
12.2 Recalibration	19
12.3 Maintenance of equipment.....	19
12.4 Requalification	19
12.5 Assessment of change.....	20
Annex A (informative) Evaluation of a sterilization process primarily based on the measurement of physical parameters	21
Annex B (informative) Evaluation of a sterilization process primarily based on biological inactivation and an accompanying mechanical air removal procedure	29
Annex C (informative) Temperature and pressure of saturated steam for use in moist heat sterilization	32
Annex D (informative) Special considerations for health care settings	35



Association for Advanced Medical Instrumentation in Healthcare Technology

PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-977-249-8226 or visit www.aami.org.

Annex E (informative) Index of normative clauses/subclauses of ISO 17665-1 and cited references or related guidance given in ISO 17665-1 and ISO/TS 17665-2.....	44
Bibliography	47



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-977-249-8226 or visit www.aami.org.

Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. For each International Standard that has been adopted by AAMI (and ANSI), the table below gives the corresponding U.S. designation and level of equivalency to the International Standard. NOTE: Documents are sorted by international designation.

Other normatively referenced International Standards may be under consideration for U.S. adoption by AAMI; therefore, this list should not be considered exhaustive.

International designation	U.S. designation	Equivalency
IEC 60601-1:2005	ANSI/AAMI ES60601-1:2005	Major technical variations
IEC 60601-1-2:2007	ANSI/AAMI/IEC 60601-1-2:2007	Identical
IEC 60601-2-2:2009	ANSI/AAMI/IEC 60601-2-2:2009	Identical
IEC 60601-2-4:2002	ANSI/AAMI DF80:2003	Major technical variations
IEC 60601-2-19:2009	ANSI/AAMI/IEC 60601-2-19:2009	Identical
IEC 60601-2-20:2009	ANSI/AAMI/IEC 60601-2-20:2009	Identical
IEC 60601-2-21:2009	ANSI/AAMI/IEC 60601-2-21:2009	Identical
IEC 60601-2-24:1998	ANSI/AAMI ID26:2004	Major technical variations
IEC 60601-2-47:2001	ANSI/AAMI EC38:2007	Major technical variations
IEC 60601-2-50:2009	ANSI/AAMI/IEC 60601-2-50:2009	Identical
IEC 80601-2-58:2008	ANSI/AAMI/IEC 80601-2-58:2008	Identical
IEC/TR 60878:2009	ANSI/AAMI/IEC TIR60878:2003	Identical
IEC/TR 62296:2009	ANSI/AAMI/IEC TIR62296:2009	Identical
IEC 62304:2006	ANSI/AAMI/IEC 62304:2006	Identical
IEC/TR 62348:2006	ANSI/AAMI/IEC TIR62348:2006	Identical
ISO 5840:2005	ANSI/AAMI/ISO 5840:2005	Identical
ISO 7198:1998	ANSI/AAMI/ISO 7198:1998/2001/(R)2004	Identical
ISO 7199:2009	ANSI/AAMI/ISO 7199:2009	Identical
ISO 8637:2004	ANSI/AAMI RD16:2007	Major technical variations
ISO 8638:2004	ANSI/AAMI RD17:2007	Major technical variations
ISO 10993-1:2003	ANSI/AAMI/ISO 10993-1:2003	Identical
ISO 10993-2:2006	ANSI/AAMI/ISO 10993-2:2006	Identical
ISO 10993-3:2003	ANSI/AAMI/ISO 10993-3:2003	Identical
ISO 10993-4:2002 and Amendment 1:2006	ANSI/AAMI/ISO 10993-4:2002/(R)2009 and Amendment 1:2006/(R)2009	Identical
ISO 10993-5:2009	ANSI/AAMI/ISO 10993-5:2009	Identical
ISO 10993-6:2007	ANSI/AAMI/ISO 10993-6:2007	Identical
ISO 10993-7:2008	ANSI/AAMI/ISO 10993-7:2008	Identical
ISO 10993-9:1999	ANSI/AAMI/ISO 10993-9:1999/(R)2005	Identical
ISO 10993-10:2002 and Amendment 1:2006	ANSI/AAMI BE78:2002/(R)2008 ANSI/AAMI BE78:2002/A1:2006/(R)2008	Minor technical variations Identical
ISO 10993-11:2006	ANSI/AAMI/ISO 10993-11:2006	Identical
ISO 10993-12:2007	ANSI/AAMI/ISO 10993-12:2007	Identical
ISO 10993-13:1998	ANSI/AAMI/ISO 10993-13:1999/(R)2004	Identical
ISO 10993-14:2001	ANSI/AAMI/ISO 10993-14:2001/(R)2006	Identical
ISO 10993-15:2000	ANSI/AAMI/ISO 10993-15:2000/(R)2006	Identical
ISO 10993-16:1997	ANSI/AAMI/ISO 10993-16:1997/(R)2009	Identical
ISO 10993-17:2002	ANSI/AAMI/ISO 10993-17:2002/(R)2008	Identical
ISO 10993-18:2005	ANSI/AAMI BE83:2006	Major technical variations
ISO/TS 10993-19:2006	ANSI/AAMI/ISO TIR10993-19:2006	Identical
ISO/TS 10993-20:2006	ANSI/AAMI/ISO TIR10993-20:2006	Identical
ISO 11135-1:2007	ANSI/AAMI/ISO 11135-1:2007	Identical

International designation	U.S. designation	Equivalency
ISO/TS 11135-2:2008	ANSI/AAMI/ISO TIR11135-2:2008	Identical
ISO 11137-1:2006	ANSI/AAMI/ISO 11137-1:2006	Identical
ISO 11137-2:2006 (2006-08-01 corrected version)	ANSI/AAMI/ISO 11137-2:2006	Identical
ISO 11137-3:2006	ANSI/AAMI/ISO 11137-3:2006	Identical
ISO 11138-1: 2006	ANSI/AAMI/ISO 11138-1:2006	Identical
ISO 11138-2: 2006	ANSI/AAMI/ISO 11138-2:2006	Identical
ISO 11138-3: 2006	ANSI/AAMI/ISO 11138-3:2006	Identical
ISO 11138-4: 2006	ANSI/AAMI/ISO 11138-4:2006	Identical
ISO 11138-5: 2006	ANSI/AAMI/ISO 11138-5:2006	Identical
ISO/TS 11139:2006	ANSI/AAMI/ISO 11139:2006	Identical
ISO 11140-1:2005	ANSI/AAMI/ISO 11140-1:2005	Identical
ISO 11140-3:2007	ANSI/AAMI/ISO 11140-3:2007	Identical
ISO 11140-4:2007	ANSI/AAMI/ISO 11140-4:2007	Identical
ISO 11140-5:2007	ANSI/AAMI/ISO 11140-5:2007	Identical
ISO 11607-1:2006	ANSI/AAMI/ISO 11607-1:2006	Identical
ISO 11607-2:2006	ANSI/AAMI/ISO 11607-2:2006	Identical
ISO 11737-1: 2006	ANSI/AAMI/ISO 11737-1:2006	Identical
ISO 11737-2:1998	ANSI/AAMI/ISO 11737-2:1998	Identical
ISO 13408-1:2008	ANSI/AAMI/ISO 13408-1:2008	Identical
ISO 13408-2:2003	ANSI/AAMI/ISO 13408-2:2003	Identical
ISO 13408-3:2006	ANSI/AAMI/ISO 13408-3:2006	Identical
ISO 13408-4:2005	ANSI/AAMI/ISO 13408-4:2005	Identical
ISO 13408-5:2006	ANSI/AAMI/ISO 13408-5:2006	Identical
ISO 13408-6:2006	ANSI/AAMI/ISO 13408-6:2006	Identical
ISO 13485:2003	ANSI/AAMI/ISO 13485:2003	Identical
ISO 14155-1:2003	ANSI/AAMI/ISO 14155-1:2003/(R)2008	Identical
ISO 14155-2:2003	ANSI/AAMI/ISO 14155-2:2003/(R)2008	Identical
ISO 14160:1998	ANSI/AAMI/ISO 14160:1998/(R)2008	Identical
ISO 14161:2000	ANSI/AAMI/ISO 14161:2000	Identical
ISO 14937:2000	ANSI/AAMI/ISO 14937:2000	Identical
ISO/TR 14969:2004	ANSI/AAMI/ISO TIR14969:2004	Identical
ISO 14971:2007	ANSI/AAMI/ISO 14971:2007	Identical
ISO 15223-1:2007 and A1:2008	ANSI/AAMI/ISO 15223-1:2007 and A1:2008	Identical
ISO 15225:2000 and A1:2004	ANSI/AAMI/ISO 15225:2000/(R)2006 and A1:2004/(R)2006	Identical
ISO 15674:2009	ANSI/AAMI/ISO 15674:2009	Identical
ISO 15675:2009	ANSI/AAMI/ISO 15675:2009	Identical
ISO 15882:2008	ANSI/AAMI/ISO 15882:2008	Identical
ISO/TR 16142:2006	ANSI/AAMI/ISO TIR16142:2005	Identical
ISO 17664:2004	ANSI/AAMI ST81:2004	Major technical variations
ISO 17665-1:2006	ANSI/AAMI/ISO 17665-1:2006	Identical (with inclusions)
ISO/TS 17665-2:2009	ANSI/AAMI/ISO TIR17665-2:2009	Identical
ISO 18472:2006	ANSI/AAMI/ISO 18472:2006	Identical
ISO/TS 19218:2005	ANSI/AAMI/ISO 19218:2005	Identical
ISO 22442-1:2007	ANSI/AAMI/ISO 22442-1:2007	Identical
ISO 22442-2:2007	ANSI/AAMI/ISO 22442-2:2007	Identical
ISO 22442-3:2007	ANSI/AAMI/ISO 22442-3:2007	Identical
ISO 25539-1:2003 and A1:2005	ANSI/AAMI/ISO 25539-1:2003 and A1:2005	Identical
ISO 25539-2:2008	ANSI/AAMI/ISO 25539-2:2008	Identical
ISO 81060-1:2007	ANSI/AAMI/ISO 81060-1:2007	Identical

Committee representation

Association for the Advancement of Medical Instrumentation

Industrial Moist Heat Sterilization Working Group

This Technical Information Report (TIR) was developed by the AAMI Industrial Moist Heat Sterilization Working Group under the auspices of the AAMI Sterilization Standards Committee. Working Group approval of the TIR does not necessarily imply that all committee members voted for its approval.

At the time this document was published, the **AAMI Industrial Moist Heat Sterilization Working Group** had the following members:

Cochairs Gregg A. Mosley
Donna Swenson

Members Richard Bancroft, Steris Corporation
Thomas J. Berger, PhD, Hospira Worldwide Inc.
Trabue D. Bryans, WuXi AppTec
Charles J. Buckle, Johnson & Johnson
Bradley J. Catalone, PhD, Olympus America Inc.
Douglas D. Davie, Sterilization Validation Services
Shawn A. Doyle, Sterilator Company Inc.
Joel R. Gorski, PhD, NAMSA
Charles Oren Hancock, H&W Technology LLC
John T. Ice, Covidien
Jim Kaiser, Bausch & Lomb Inc.
Steve Kirckhof, 3M Healthcare
Joseph J. Lasich, BS, Alcon Laboratories Inc.
Gregg A. Mosley, Biotest Laboratories Inc.
Russell R. Nyberg, Raven Biological Laboratories
Tory Peeler, Becton Dickinson & Company
Nancy Rakiewicz, Ethox International Inc.
Elizabeth A. Riegel, FDA/CDRH
Mike Sadowski, Baxter Healthcare Corporation
Barb Smith, Getinge USA
Donna Swenson, BS, CSPDM, West Suburban Medical Center
Dennis L. Wildes, St. Jude Medical Inc.

Alternates Annette Bojanski, Raven Biological Laboratories
Marjean Boyten, Fresenius Medical Care Renal Therapies Group
Charlie Christianson, St. Jude Medical Inc.
Angela Coon, Baxter Healthcare Corporation
Steven J. Elliott, WuXi AppTec
Catherine J. Finocchiaro, Bausch & Lomb Inc.
Thomas J. Frazar, Johnson & Johnson
John Grillo, PhD, Hospira Worldwide Inc.
Geetha C. Jayan, PhD, FDA/CDRH
Chris Marsh, Alcon Laboratories Inc.
Richard T. O'Donnell, Steris Corporation
Dave Parente, NAMSA
Laura Schmitt, Ethox International Inc.
Kelvin J. Witcher, 3M Healthcare
Cheryl Work, Becton Dickinson & Company

At the time this document was published, the **AAMI Sterilization Standards Committee** had the following members:

Cochairs Victoria M. Hitchins, PhD, FDA/CDRH
William E. Young, Boston Scientific Corporation

Members Trabue D. Bryans, WuXi AppTec Inc.
Peter A. Burke, PhD, Steris Corporation
Nancy Chobin, RN CSPDM, St Barnabas Healthcare System (Independent Expert)

Charles Cogdill, Boston Scientific Corporation
Ramona Conner, RN MSN CNOR, Association of Perioperative Registered Nurses
Jacqueline Daley, Association for Professionals in Infection Control and Epidemiology
Kimbrell Darnell, CR Bard
Lisa Foster, Sterigenics International
Joel R. Gorski, PhD, NAMSA
Deborah A. Havlik, Hospira Inc.
Victoria M. Hitchins, PhD, FDA/CDRH
Danny Hutson, Cardinal Health (MP&S)
Lois Atkinson Jones, MS, (Independent Expert)
Susan G. Klacik, CCSMC FCS ACE, IAHCMM
Byron J. Lambert, PhD, Abbott Laboratories
Colleen Patricia Landers, RN, Canadian Standards Association
David Liu, Johnson & Johnson
Lisa N. Macdonald, Becton Dickinson & Company
Jeff Martin, Alcon Laboratories Inc.
Patrick J. McCormick, PhD, Bausch & Lomb Inc.
Susie McDonald, American Society for Healthcare Central Service Professionals
Nancy Rakiewicz, Ethox International Inc.
Phil M. Schneider, 3M Healthcare
Michael H. Scholla, Dupont Nonwovens
Mark Seybold, Baxter Healthcare Corporation
Andrew Sharavara, PhD, Propper Manufacturing Co Inc.
Mark N. Smith, Getinge USA
William N. Thompson, Covidien
James L. Whitby, MA, MB, FRCP, (Independent Expert)
Martell Kress Winters, BS SM, Nelson Laboratories Inc.

Advancing Safety in Healthcare Technology

Alternates

Lloyd Brown, Covidien
Dave Dion, Cardinal Health (MP&S)
Thomas J. Frazar, Johnson & Johnson
Kathy Hoffman, Sterigenics International
Jim Kaiser, Bausch & Lomb Inc.
Joseph J. Lasich, BS, Alcon Laboratories Inc.
Chiu S. Lin, PhD, FDA/CDRH
Natalie Lind, IAHCMM
Ralph Makinen, Boston Scientific Corporation
Mary S. Mayo, CR Bard
David Ford McGoldrick, BS, Abbott Laboratories
Jerry R. Nelson, PhD, Nelson Laboratories Inc.
Karen Polkinghorne, Dupont Nonwovens
Janet M. Prust, 3M Healthcare
Mike Sadowski, Baxter Healthcare Corporation
John R. Scoville, Jr., Steris Corporation
Ralph Stick, WuXi AppTec Inc.
Jason Voisinet, Ethox International Inc.
Valerie Welter, Hospira Inc.
William E. Young, Boston Scientific Corporation

NOTE—Participation by federal agency representatives in the development of this document does not constitute endorsement by the federal government or any of its agencies.

Background of AAMI adoption of ISO/TS 17665-2:2009

The International Organization for Standardization (ISO) is a worldwide federation of national standards bodies. The United States is one of the ISO members that took an active role in the development of this standard.

The ISO 17665 series was developed by ISO Technical Committee 198 to fill a need for an international standard for moist heat sterilization of health care products. The standard combines and updates two separate ISO standards, ISO 11135:1993 and ISO 13683:1997, and also supersedes AAMI TIR13:1997, *Principles of industrial moist heat sterilization*. ISO/TC 198 has published two parts under the general title *Sterilization of health care products—Moist Heat*. The two parts are:

- Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices; and
- Part 2: Guidance on the application of ISO 17665-1

Concurrent with the development of the U.S. position on the ISO 17665 series, the AAMI Industrial Moist Heat Sterilization Working Group (AAMI ST/WG 03) decided to adopt the two parts verbatim; however, the Working Group did agree to add an inclusion to Part 2 to make it clear that section 10.5 does not require the use of a PCD or air detectors for every production cycle. In addition, an editorial correction was made to subclause 9.5.2 to make the formatting of the list consistent.

The requirements for validation and routine control are contained in the normative section of Part 1, with limited guidance in Annex A of that document, while Part 2 is a technical specification that provides the majority of the guidance related to compliance with the requirements in Part 1.

U.S. participation in ISO/TC 198 is organized through the U.S. Technical Advisory Group for ISO/TC 198, administered by the Association for the Advancement of Medical Instrumentation (AAMI). The United States made a considerable contribution to this technical specification.

This TIR contains guidelines that are not intended to be absolute or to be applicable in all circumstances. Judgment should be used in applying the information in this TIR.

As used within the context of this document, “shall” indicates requirements strictly to be followed to conform to the recommended practice. “Should” indicates that among several possibilities, one is recommended as particularly suitable, without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action should be avoided but is not prohibited. “May” is used to indicate that a course of action is permissible within the limits of the recommended practice. “Can” is used as a statement of possibility and capability. Finally, “must” is used only to describe “unavoidable” situations, including those mandated by government regulation.

of the document before making a purchasing decision.

NOTE—This background does not contain provisions of the AAMI/ISO Technical Information Report 17665-2, but does provide important information about the development and intended use of this TIR.

NOTE—Beginning with the ISO foreword on page xii, this AAMI Technical Information Report is identical to ISO/TS 17665-2:2009, with the exception of the minor clarifying inclusion in section 10.5 and the correction of inconsistent list numbering in subclause 9.5.2.

AAMI inclusion to ISO/TS 17665-2:2009

Subclause 10.5

Add the following after subclause 10.5.

While air removal should be assured for each operating cycle according to the ISO 17665-1, the above guidance does not mean that the use of a PCD or air detector(s) is required with every production cycle.

NOTE—Beginning with the ISO foreword on page xii, this AAMI Technical Information Report is identical to ISO/TS 17665-2:2009, with the exception of the minor clarifying inclusion in section 10.5 and the correction of inconsistent list numbering in subclause 9.5.2.



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-977-249-8226 or visit www.aami.org.

Foreword

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

In other circumstances, particularly when there is an urgent market requirement for such documents, a technical committee may decide to publish other types of document:

- an ISO Publicly Available Specification (ISO/PAS) represents an agreement between technical experts in an ISO working group and is accepted for publication if it is approved by more than 50 % of the members of the parent committee casting a vote;
- an ISO Technical Specification (ISO/TS) represents an agreement between the members of a technical committee and is accepted for publication if it is approved by 2/3 of the members of the committee casting a vote.

An ISO/PAS or ISO/TS is reviewed after three years in order to decide whether it will be confirmed for a further three years, revised to become an International Standard, or withdrawn. If the ISO/PAS or ISO/TS is confirmed, it is reviewed again after a further three years, at which time it must either be transformed into an International Standard or be withdrawn.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO/TS 17665-2 was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

ISO 17665 consists of the following parts, under the general title *Sterilization of health care products* — *Moist heat*:

- *Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices*
- *Part 2: Guidance on the application of ISO 17665-1* [Technical Specification]

Introduction

The guidance given in this Technical Specification is not intended as a checklist for assessing compliance with ISO 17665-1. This guidance is intended to assist in obtaining a uniform understanding and implementation of ISO 17665-1 by providing explanations and acceptable methods for achieving compliance with specified requirements. It highlights important aspects and provides examples. Methods other than those given in this guidance may be used. However, the use of alternative methods has to be demonstrated to be effective in achieving compliance with ISO 17665-1.

The main body of this document is applicable to all settings where moist heat sterilization is carried out. The annexes to this guidance document also specify detailed means of implementing the requirements of ISO 17665-1 and represent current best practices.

The numbering of the clauses in the main body of this Technical Specification corresponds to that in ISO 17665-1.

Medical devices reprocessed in health care facilities include a wide variety of product with varying levels of bioburden. Appropriate and thorough cleaning and, where necessary for safe handling, decontamination processes are essential prior to presenting product for sterilization. Mixed product loads are common in healthcare facilities with throughput volumes dictated by historical and predicted demand for sterile product.

Health care facilities do not normally specify sterilization processes for any individual medical device. Also, it is impractical for health care facilities to determine bioburden on a medical device. It is important that specified instruments be disassembled before decontamination and thoroughly inspected after completion of the sterilization process. Reassembly and assessment of functionality are also needed. Therefore, the medical device manufacturer's instructions (see ISO 17664^[23]) should be followed for all aspects of cleaning, disinfection, packaging and sterilization. Many devices can be fully immersed and can be washed and disinfected in automated equipment (see ISO 15883^[19-22]). For devices that cannot be fully immersed and that cannot tolerate thermal decontamination, alternative methods of disinfection should be used to ensure safe handling. Such procedures and policies should be in place to ensure that medical devices undergo appropriate reprocessing. Particular attention needs to be paid to the drying and storage of sterile medical devices. Requirements for packaging of medical devices are covered in ISO 11607-1^[8] and ISO 11607-2^[9].

If multiple sterilization cycles can lead to degradation and limit the useful life of a medical device, the manufacturer will specify the number of reprocessing cycles that can normally be tolerated.

When selecting a medical device, priority should be given to properties such as ease of cleaning and disassembly.

Additional guidance specific to health care is offered in Annex D of this Technical Specification.



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-977-249-8226 or visit www.aami.org.

Sterilization of health care products — Moist heat — Part 2: Guidance on the application of ISO 17665-1

1 Scope

This Technical Specification provides general guidance on the development, validation and routine control of moist heat sterilization processes and is intended to explain the requirements set forth in ISO 17665-1. The guidance given in this Technical Specification is provided to promote good practice related to moist heat sterilization processes and to assist those developing and validating a moist heat sterilization process according to ISO 17665-1.

NOTE 1 The structure of the main body of this ISO Technical Specification (Clauses 1 to 12) corresponds to the structure of ISO 17665-1, so that the guidance given under a particular clause or subclause of this part of ISO 17665 applies to the requirements given in the corresponding clause or subclause of ISO 17665-1. For example, guidance for subclause 5.2 of ISO 17665-1:2006 is given in 5.2. This guidance is provided in addition to the guidance given in ISO 17665-1:2006, Annex A. See also Annex E.

NOTE 2 Special considerations specific to sterilization processes performed in health care facilities are given in Annex D.

PREVIEW COPY

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

For a complete copy of this AAMI document, contact AAMI at 1-977-249-9226 or visit www.aami.org.
ISO 17665-1:2006, *Sterilization of health care products — Moist heat — Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices*

NOTE The normative references in ISO 17665-1 refer to published standards, the content of which should be used to assist in demonstrating compliance to the clause in which they are cited. Some are required mainly for moist heat sterilization in industry or for manufacturers of moist heat sterilizers and could go beyond typical practice for those performing sterilization in health care facilities.

ISO 17665-1 specifies a number of methods and procedures that can be used to monitor sterilization processes. The equipment required will normally be commercially available. A number of the normative references cited describe the specification and test methods used by commercial suppliers to qualify their products. The user of such products should ensure that purchased products comply with these standards, but will not normally need to refer to the standards.

ISO 17665-1 specifies the use of packaging complying with ISO 11607-1 and ISO 11607-2. Healthcare facilities should purchase packaging complying with these International Standards.