This is a preview of "AAMI TIR15:2016". Click here to purchase the full version from the ANSI store.

Technical Information Report



AAMITIR15: PREVIEW COPY This is a preview edition of an AAM2 Galacocument and is

This is a preview edition of an AAM guidalce Ocument and is intended to allow potential purchasers to evaluate the content of the document before makirPhysical aspects of Ear a complete copy of this AAM ethylene oxide sterilization



This is a preview of "AAMI TIR15:2016". Click here to purchase the full version from the ANSI store.



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

This is a preview of "AAMI TIR15:2016". Click here to purchase the full version from the ANSI store.

AAMI Technical Information Report

AAMI TIR15:2016 (Revision of AAMI TIR15:2009/(R)2013)



Physical aspects of ethylene oxide sterilization

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-977-249-8226 or visit www.aami.org.

Approved 18 November 2016 by AAMI

Abstract: This technical information report (TIR) provides additional guidance to augment the ANSI/AAMI/ISO 11135 series. Topics covered in this TIR are sterilization equipment, considerations for preconditioning, calculations for relative humidity, calculations of ethylene oxide concentration, and flammability.

Keywords: sterilization equipment, ethylene oxide, relative humidity, flammability, preconditioning

AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

This document is not an American National Standard, and the material contained herein is not normative in nature.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every 5 years but at least every 10 years. For a TIR, AAMI consults with a technical committee about 5 years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice, or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues that a standard issues that a standard practice is the inclusion of differing viewpoints on technical issues that a standard is the inclusion of differing viewpoints on technical issues that a standard is the inclusion of differing viewpoints on technical issues that a standard is the inclusion of differing viewpoints on technical issues that a standard is the inclusion of differing viewpoints on technical issues that a standard is the inclusion of differing viewpoints on technical issues that a standard is the inclusion of differing viewpoints on technical issues that a standard is the inclusion of differing viewpoints on technical issues that a standard is the inclusion of technical issues that a standard is the inclusion of technical issues that a standard is the inclusion of technical issues to the inclusion of technical issues that a standard is the inclusion of technical issues to the inclusion of technical issues to

CAUTION NOTICE: This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the documents Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

of the document before making a purchasing decision. Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 4301 N. Fairfax Dr., Suite 301, Arlington, VA 22203-1633.

For a complete copy of this AAMI document, contact AAMI at Published by

+1-977-249-8226 or visit www.aami.org.

AAMI 4301 N. Fairfax Dr., Suite 301 Arlington, VA 22203-1633 www.aami.org

© 2017 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, complete the reprint request form at www.aami.org or contact AAMI at 4301 N. Fairfax Dr., Suite 301, Arlington, VA 22203-1633. Phone: (703) 525-4890; Fax: (703) 276-0793.

Printed in the United States of America

ISBN 978-1-57020-653-5

Contents

| | | Page |
|-----------|--|------|
| Glossary | y of equivalent standards | iv |
| Committ | tee representation | v |
| Foreword | | |
| 1 | Scope | 1 |
| 2 | Definitions | 1 |
| 3 | EO sterilization processing equipment | 2 |
| 4 | Determination of minimum product temperature prior to preconditioning | 8 |
| 5 | Calculation of moisture content | |
| 6 | Calculation of EO concentration | |
| 7 | Calculations for use of flammability curve | 18 |
| 8 | Guide for statistical comparison to determine process equivalence | 22 |
| Bibliogra | aphyAdvancing Safety in Healthcare Technology | 26 |
| Figure | have been an example of the second of the se | |
| 1 | EO Cycle Flammability Diagram. PREVIEW COPY | 19 |
| Tables | PREVIEW COPY | |
| 1 | List of constants ลาเอาห่อโอะนเลา เมื่อเฐการ of some common good ligent combinations nt. and is | 12 |
| 2 | intended to allow potential purchasers to evaluate the content Gas constants (R) of the document before making a purchasing decision. | 13 |
| 3 | Properties of saturated steam—Commonly used temperatures | 16 |
| 4 | For a complete copy of this AAMI document, contact AAMI at Flammability review, partial pressures, and volume percentage +1-977-249-8226 or visit www.aami.org. | 21 |

Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. AAMI maintains a current list of each International Standard that has been adopted by AAMI (and ANSI). Available on the AAMI website at the address below, this list gives the corresponding U.S. designation and level of equivalency to the International Standard.

www.aami.org/standards/glossary.pdf



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

Committee representation

Association for the Advancement of Medical Instrumentation

Industrial Ethylene Oxide Sterilization Working Group

This technical information report (TIR) was developed by the AAMI Industrial Ethylene Oxide Sterilization Working Group under the auspices of the AAMI Sterilization Standards Committee. Working Group approval of the TIR does not necessarily imply that all committee members voted for its approval.

At the time this document was published, the **AAMI Industrial Ethylene Oxide Sterilization Working Group** had the following members:

| Cochairs: | Jonathan Bull Jeffrey Martin |
|-----------|---|
| Members: | Jeffrey Martin Erin Armstrong, B Braun Medical Edward Arscott, NAMSA David Ballard, Dynatec Scientific Labs Stacy Bohl, Boston Scientific Corporation Anne F. Booth, MS, Booth Scientific Corporation Carolyn Braithwaite-Nelson, Spectranetics Jonathan Bull, Johnson & Johnson Tim Carlson, BD Medical Sarah Chamberlain, Accuratus Labs Services Dennis Christenseen, BS, Sterlitzation Validation & Consulting, Gary N, Cranston, Consulting & Technical Services/PCS Elaine Daniell, Bard Medical Division Douglas D. Davie, Sterlitzation Validation Services Darci Diage, Direct Flow Medical David Dion, Cardinal Health COUNT COOPY Mary Ann Drosnock, MS, Healthmark Industries TIWilliamF chitzGerald, RE, fittzGerald, & Associates Littlance document and is Dan B, Flovd, RM, Dupont Protection Solutions In Lisal Poster, Aditivo S& & SA Consulting Matthewilfreeman, Theramb BCT Ore making a purchasing decision. Zory R, Glaser, PhD MPH CSPDM, Johns Hopkins University Michael Groendyk, Arthrex FbduglasHabbecht; Steihity Asstrance AMII document, contact AAMII at Arthur C. Harris, CioKA-249-8226 or visit www.aami.org. Deborah A. Havlik, Hospira, a Pfizer Company Jason Hedrick, Medtronic Clark W. Houghtling, Cosmed Group Krista Howard, WL Gore & Associates Naipur Jain, Intuitive Surgical Carolyn L. Kinsley, LexaMed Karen A. Kowalczyk, Centurion Sterilization Services Christine Loshbaugh, Edwards LifeSciences Molie Love, Smiths Medical Jeffrey Martin, Sterilization and Quality System Consulting Albert May, Andersen Products Inc. Patrick McCormick, PhD, Bausch & Lomb David Ford McGoldrick, BS, Abbott Vascular Devices Russell D. Mills, GE Healthcare Gerry A. O'Dell, |
| | Nancy Rakiewicz, IUVO BioScience Keith Reiner, Terumo Cardiovascular Systems |

Beth Ridgeway, Mesa Laboratories Manuel Saavedra Jr., Halvard Health Harry Shaffer, Sterilization Consulting Services Arnold Shechtman, BS, Validation Challenges Consulting David Silor, Zimmer Bill South, Isomedix Services Michael G. Sprague, Ethide Laboratories Sopheak Srun, MPH SM(NRCM), Quality Tech Services Fenil Sutaria, Medline Industries Mara Tafoya, WuXi AppTec Radhakrishna S. Tirumalai, US Pharmacopeial Convention Steven E. Turtil, FDA/CDRH Craig A. Wallace, 3M Healthcare P. Richard Warburton, ChemDAQ Richard L. Weisman, Fresenius Medical Care Beverly Whitaker, CQA RAB MBA, Indigo Consulting Group Dennis L . Wildes, St. Jude Medical William T. Young, Sterigenics International Roberto Zumbado, Sr., Philips August Baur, Centurion Sterilization Services MarJean Boyter, Fresenius Medical Care Robert Bradley, CBET, Mesa Laboratories Trabue D. Bryans, BryCor Greg Bush, Alcon Research Liza Chenette, Accuratus Labs Solutions Sean N. Colwell, WuXi Apptec Allison Diemert, Arthrexdvancing Safety in Healthcare Technology David Dominguez, Becton Dickinson Brian Drumheller, Brookhaven Medical

Venice Eldred, Medline Industries

Steven Elliott, FDA/CDRHD FY/LEW COPY Diane Faivre-Swiat, Cardinal Health EW COPY Mark Finley, Boston Scientific

TNaoni Game, St. Jude Medical of an AAMI guidance document and is Scott Giraud, Medtronic Cardiac Surgical Products In William R. Gordon, Steris Isomedix Services has ers to evaluate the content David M. Hilliker Chempagbefore making a purchasing decision. Brent Huberty, Boston Scientific

Nicole Jackson, Ecolab Fsaturking spectanetics y of this AAMI document, contact AAMI at Chris Kobus, GE Healthcare -8226 or visit www.aami.org. Ezra Koski, Terumo BCT James P. Kulla, BS MS, LexaMed Wesley Lantz, Zimmer John Lindley, Andersen Products Inc. Paul L. Littley, BSE, Nelson Laboratories Mauricio Martinez, Smiths Medical Nicole M. McLees, 3M Critical & Chronic Care Solutions Joseph Mello, Ethide Laboratories Laurie Nawrocki, NAMSA Koyejo Obadina, Abbott Vascular Ken Paddock, Baxter Healthcare Michael A. Padilla, SteriPro Labs Tyrone S. Rouse, Halyard Health Michael Shoene, Bausch & Lomb Kristen Spigiel, Stryker Instruments Elizabeth Steiner, IUVO BioScience Brian Wallace, Intuitive Surgical Scott Weiss, Ethicon Daryl Woodman, Andersen Products Casimir John Wos, PhD, FitzGerald & Associates

Alternates:

AAMI also acknowledges the **TIR15 Editing Task Group**, comprising the following members, for its special contribution in the development of this document:

Task group leaderGreg Bush, Alcon ResearchMembers:Gary Cranston, Consulting & Technical Services/PCS
William F. Fitzgerald, Fitzgerald & Associates
Arnold Shechtman, BS, Validation Challenges Consulting
Steven E. Turtil, FDA/CDRH

NOTE—Participation by federal agency representatives in the development of this technical information report does not constitute endorsement by the federal government or any of its agencies.



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

Foreword

This document is part of a series of technical information reports (TIRs) intended for use in conjunction with ANSI/AAMI/ISO 11135:2014. The other reports in the series are:

- AAMI TIR14:2009, Contract sterilization using ethylene oxide;
- AAMI TIR16:2009, Microbiological aspects of ethylene oxide sterilization;
- AAMI TIR28:2009, Product adoption and process equivalence for ethylene oxide sterilization;
- AAMI TIR56:2013, Guidance for the development, validation and routine control of an ethylene oxide sterilization process utilizing flexible bag systems for the sterilization of medical devices.

The original TIR15, along with other AAMI TIRs, provided additional guidance to the 1994 edition of the industrial ethylene oxide sterilization standard ANSI/AAMI/ISO 11135, which was revised in 2007 and 2014 under a new designation, ANSI/AAMI/ISO 11135:2014, *Sterilization of health care products—Ethylene oxide - Requirements for development, validation, and routine control of a sterilization process for medical devices.* In 2008 and 2014, the International Organization for Standardization (ISO) published its own guidance document for the 11135 standard, ISO/TR 11135-2:2008, *Sterilization of health care products—Ethylene oxide—Part 2: Guidance on the application of ISO 11135-1*, which was based to a great extent on the earlier AAMI TIRs. Correspondingly, the AAMI Industrial Ethylene Oxide Sterilization Working Group is updating its TIRs to take into account changes to the 11135 standard and to avoid redundancy with ANSI/AAMI/ISO 11135:2014.

This TIR contains guidelines that are not intended to be absolute or to apply in all circumstances. One should use judgment in applying the information in this TIR.

As used within the context of this document, "should" indicates that among several possibilities, one is recommended as particularly suitable, without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action should be avoided but is not prohibited. "May" is used to indicate that a course of action is permissible within the limits of the technical information report. "Can" is used as a statement of possibility and capability. Finally, "must" is used only to describe "unavoidable" situations, including those mandated by government regulation. See also the NOTE on Page 1.

Suggestions for improving this technical information report are invited. Comments and suggested revisions should be sent to AAMI, 4301sNsFairtaxDri_Suite dotti Artington, IVA22203-8263clance document and is

intended to allow potential purchasers to evaluate the content

NOTE—This foreword does not contain provisions of AAMICTIR15:2016 OPhysical aspects of guide sterilization, but it does provide important information about the development and intended use of the document.

AAMI Technical Information Report

AAMI TIR15:2016

Physical aspects of ethylene oxide sterilization

NOTE—This technical information report (TIR) is not a standard and the material contained herein is not normative in nature. In a few instances, the committee has used the term "shall" based on its knowledge of requirements contained in relevant standards, regulatory requirements, or both.

1 Scope

This AAMI TIR provides additional guidance to augment ANSI/AAMI/ISO 11135:2014. This TIR addresses some of the physical aspects of ethylene oxide (EO) sterilization and provides guidance on sterilization equipment, considerations for preconditioning, how to calculate relative humidity, EO concentration and flammability, as well as guidance on the use of statistics for process equivalence.

2 Definitions

For the purposes of this TIR, the terms and definitions in ANSI/AAMI/ISO 11135:2014 and the following apply:

2.1

dew point

temperature at which the saturation water vapor pressure is equal to the partial pressure of the water vapor in the atmosphere

NOTE Any cooling of the atmosphere below the dew point would produce water condensation.

2.2

humidity This is a preview edition of an AAMI guidance document and is A measure of the water vapor present in a gas. It is usually measured as absolute humidity, relative humidity, or dew point temperature.

of the document before making a purchasing decision.

2.3

preconditioning area

chamber or room where defined temperature and humidity conditions for the purpose of conditioning the product prior to entering the gas exposure phase of the process 26 or visit www.aami.org.

2.4

relative humidity (RH)

Ratio of water vapor partial pressure actually present to water vapor pressure required for saturation at a given temperature.

2.5

chamber humidification

phase of the sterilization process whereby moisture is added to the chamber to replace any environmental moisture potentially lost in the vacuum stages for air removal or inert gas dilution.

2.6

steam purity

measurement of the amount of solid, liquid, or vaporous contamination in the steam.

Note 1 to entry: Normally, steam purity is reported as the solids content. The steam purity needed for an EO sterilization system is defined by the equipment user and needs to be suitable for its intended use.