

This is a preview of "AAMI TIR16:2017". Click here to purchase the full version from the ANSI store.

Technical Information Report



AAMI TIR16: 2017

PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to review the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

This is a preview of "AAMI TIR16:2017". Click here to purchase the full version from the ANSI store.



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

AAMI Technical Information Report

AAMI TIR16:2017
(Revision of AAMI TIR16:2013)



Microbiological aspects of ethylene oxide sterilization **PREVIEW COPY**

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

Approved 17 November 2017 by
Association for the Advancement of Medical Instrumentation

Abstract: Addresses various microbiological aspects of the development and validation of an ethylene oxide sterilization process. Does not address the various factors that can have an effect on the bioburden of the product and on the sterilization process. Provides additional guidance to ANSI/AAMI/ISO 11135:2014 for medical device manufacturers, including those that use contract sterilization facilities or contract sterilization operations.

Keywords: sterilization, microbiological aspects, validation, ethylene oxide, bioburden, performance qualification

AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

This document is not an American National Standard, and the material contained herein is not normative in nature.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every five years but at least every 10 years. For a TIR, AAMI consults with a technical committee about five years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

CAUTION NOTICE: This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633.

Published by

AAMI
4301 N. Fairfax Drive, Suite 301
Arlington, VA 22203-1633
www.aami.org

© 2017 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, complete the reprint request form at www.aami.org or contact AAMI at 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633. Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

ISBN 978-1-57020-691-7

Contents		Page
Glossary of equivalent standards		iv
Committee representation		v
Foreword		viii
1	Scope	1
2	Terms and definitions	1
3	Process and equipment characterization	1
3.1	Sterilization equipment	1
3.2	Process characterization—Physical parameters	1
4	Process definition	3
4.1	Considerations for process definition	3
4.2	Methods for process definition	6
4.3	Sterilization process definition troubleshooting	12
4.4	Process challenge devices	14
5	Validation	16
5.1	Performance qualification	16
5.2	Placement and handling of PCDs, test samples, and sensors	16
5.3	Sterilization loads—general considerations	18
5.4	Simulation of anticipated process conditions	19
5.5	Release of validation loads	19
5.6	Release of small batches or lots	19
6	Maintaining process effectiveness	20
6.1	Assessment of change	20
6.2	Failure investigation	20
Bibliography		22

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. AAMI maintains a current list of each International Standard that has been adopted by AAMI (and ANSI). Available on the AAMI website at the address below, this list gives the corresponding U.S. designation and level of equivalency to the International Standard.

www.aami.org/standards/glossary.pdf



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

Committee representation

Association for the Advancement of Medical Instrumentation Industrial Ethylene Oxide Sterilization Working Group

This technical information report (TIR) was developed by the Association for the Advancement of Medical Instrumentation (AAMI) Industrial Ethylene Oxide Sterilization Working Group under the auspices of the AAMI Sterilization Standards Committee. Working Group approval of the TIR does not necessarily imply that all committee members voted for its approval.

At the time this document was published, the **AAMI Industrial Ethylene Oxide Sterilization Working Group** had the following members:

Cochairs: Jonathan Bull
Jeffrey Martin

Members: Erin Armstrong, B Braun Medical
Edward Arscott, NAMSA
David Ballard, Dynatec Scientific Labs
Erika Kitagawa Bawor, BS RM NRCM, Insulet
Stacy Bohl, Boston Scientific
Carolyn Braithwaite-Nelson, Spectranetics
David Brodersen, LexaMed
Jonathan Bull, Johnson & Johnson
Robb Calabro, Abbvie
Tim Carlson, BD Medical
Dennis Christensen, Sterilization Validation & Consulting
Gary N. Cranston, Consulting & Technical Services/PCS
Elaine Daniell, Bard Medical Division
Douglas D. Davie, Sterilization Validation Services
Darci Diage, Direct Flow Medical
David Dion, Cardinal Health
Mary Ann Drosnock, MS, Healthmark Industries
Trisha Fair, Crosstex/Cantel/SPS Medical
Paul Fioriti, Anthrex Manufacturing
William F. FitzGerald, PE, FitzGerald & Associates
Dan B. Floyd, RM, Nelson Laboratories Inc.
Lisa Foster, Aduvo QS & SA Consulting
Matthew Freeman, Terumo BCT
Zory R. Glaser, PhD MPH CSPDM, Johns Hopkins University
William Gordon, STERIS Isomedix Services
Douglas Harbrecht, Sterility Assurance
Arthur C. Harris, Cook
Deborah A. Havlik, Hospira Worldwide Inc.
Jason Hedrick, Medtronic
Clark W. Houghtling, Cosmed Group
Krista Howard, WL Gore & Associates
Nichole Jackson, Ecolab
Naipur Jain, Intuitive Surgical
Karen A. Kowalczyk, Centurion Sterilization Services
Christine Loshbaugh, Edwards LifeSciences
Jeffrey Martin, Sterilization and Quality System Consulting
Mauricio Martinez Sr., Smiths Healthcare Manufacturing
Ted May, Andersen Products Inc.
Patrick McCormick, PhD, Bausch & Lomb
David Ford McGoldrick, BS, Abbott Laboratories
Russell D. Mills, GE Healthcare
Gerry A. O'Dell, MS, Gerry O'Dell Consulting
Mike Padilla, SteriPro Labs
Michelle Peterson, Stryker Instruments
Andrew Porteous, Baxter Healthcare
Nancy Rakiewicz, MMC Sterilization Services Group

Keith Reiner, Terumo Cardiovascular Systems
Manuel Saavedra Jr., Halyard Health
Harry Shaffer, Sterilization Consulting Services
Arnold Shechtman, Validation Challenges Consulting, LLC
Michael G. Sprague, Ethide Laboratories
Sopheak Srun, MPH SM(NRCM), Quality Tech Services
Fenil Sutaria, Medline Industries
Mara Tafoya, WuXi Apptec
Radhakrishna S. Tirumalai, US Pharmacopeial Convention
Steven E. Turtill, FDA/CDRH
Craig A. Wallace, 3M Healthcare
P. Richard Warburton, ChemDAQ Inc.
Richard L. Weisman, Fresenius Medical Care Renal Therapies Group
Beverly Whitaker, CQA RAB MBA, Indigo Consulting Group
Dennis L. Wildes, St. Jude Medical
Roberto Zumbado, Philips

Alternates:

August Baur, Centurion Sterilization Services
MarJean Boyter, Fresenius Medical Care
Robert Bradley, CBET, Mesa Laboratories
Trabue D. Bryans, BryCor
Greg Bush, Alcon Research
Sarah Chamberlain, Avista Pharma Solutions
Sean Colwell, WuXi Apptec
David Dominguez, Carefusion
Zachary Dukerich, Anthrex Manufacturing
Venice Eldred, Medline Industries
Steven Elliott, FDA/CDRH
Diane Faivre-Swiat, Cardinal Health
Marc Finley, Boston Scientific
Naomi Gamm, St. Jude Medical
Scott Giraud, Medtronic
David M. Hilliker, ChemDaq
Satu King, Spectranetics
Chris Kobus, GE Healthcare
Ezra Koski, A, Terumo BCT
Kaumudi Kulkarni, Healthmark Industries
James P. Kulla, BS MS, LexaMed
John Lindley, Andersen Products Inc.
Paul E. Little, BSE, Nelson Laboratories
Nicole M. McLees, 3M Critical & Chronic Care Solutions
Joseph Mello, Ethide Laboratories
Laurie Nawrocki, NAMSA
Michael Nolan, SPS Medical Supply
Koyejo Obadina, Abbott Vascular
Kenn Paddock, Baxter Healthcare
Dave Parente, ECOLAB Healthcare
Tyrone S. Rouse, Halyard Health
Michael Schoene, Bausch & Lomb
Bill South, Isomedix Services
Kristen Spigiel, Stryker Instruments
Elizabeth Steiner, IUVO BioScience
Andrew Tuggles, Sr., Ethicon
Brian Wallace, Intuitive Surgical
Daryl Woodman, Andersen Products
Casimir John Wos, PhD, FitzGerald & Associates
William T. Young, Sterigenics International



Association for the Advancement of Medical Instrumentation

REVIEW COPY

This is a preview version of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at 1-800-854-6326 or visit www.aami.org.

AAMI also acknowledges the **Task Group**, comprised of the following members, for its special contribution in the development of this document:

Task Group Leader: Scott Giraud, Medtronic

Task Group Members: Tim Carlson, BD Medical
Elaine Daniell, Bard Medical Division
Dan B. Floyd, RM, Nelson Laboratories Inc.
David Ford McGoldrick, BS, Abbott Laboratories
Steven E. Turtill, FDA/CDRH

NOTE Participation by federal agency representatives in the development of this technical information report does not constitute endorsement by the federal government or any of its agencies.



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

Foreword

This document is part of a series of technical information reports (TIRs) intended for use in conjunction with ANSI/AAMI/ISO 11135:2014. The other reports in the series are listed below:

- AAMI TIR14:2009, *Contract sterilization using ethylene oxide*
- AAMI TIR15:2009, *Physical aspects of ethylene oxide sterilization*
- AAMI TIR28:2009/(R)2013, *Product adoption and process equivalence for ethylene oxide sterilization*
- AAMI TIR39:2009, *Guidance on selecting a microbial challenge and inoculation sites for sterilization validation of medical devices*
- AAMI TIR56:2013, *Guidance for the development, validation and routine control of an ethylene oxide sterilization process utilizing flexible bag systems for the sterilization of medical devices*

The original TIR16, along with other AAMI TIRs, provided additional guidance to the 1994 edition of the industrial EO sterilization standard 11135, which was revised in 2007 under a new designation, ANSI/AAMI/ISO 11135-1:2007, *Sterilization of health care products—Ethylene oxide—Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices*. In 2008, ISO published its own guidance document for the 11135 standard, ISO/TR 11135-2:2008, *Sterilization of health care products—Ethylene oxide—Part 2: Guidance on the application of ISO 11135-1*, which was based to a great extent on the earlier AAMI technical information reports. ANSI/AAMI/ISO 11135-1:2007 and ISO/TR 11135-2:2008 were revised into a single document: ISO 11135:2014.

This TIR provides guidance related to the microbiological aspects of EO sterilization that is typically not covered in depth, or at all, in the existing guidance documents for EO sterilization. It is designed to provide information that will assist in design, qualification, and routine processing of EO sterilization processes. This TIR condenses pertinent information that may be available in a variety of sources in one location and is based on practices that have been found to be used successfully within the United States. This TIR contains guidelines that are not intended to be absolute or to apply in all circumstances. One should use judgment in applying the information in this TIR.

As used within the context of this document, “should” indicates that among several possibilities one is recommended as particularly suitable without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action should be avoided but not prohibited. “May” is used to indicate that a course of action is permissible within the limits of the TIR. “Can” is used as a statement of possibility and capability. “Must” is used only to describe “unavoidable” situations, including those mandated by government regulations. See also the NOTE on Page 1.

Suggestions for improving this technical information report are invited. Comments and suggested revisions should be sent to AAMI, 4301 Fairfax Drive, Suite 301, Arlington, VA 22203.

+1-877-249-8226 or visit www.aami.org.

NOTE This foreword does not contain provisions of AAMI TIR16:2017, *Microbiological aspects of ethylene oxide sterilization*, but it does provide important information about the development and intended use of the document.

Microbiological aspects of ethylene oxide sterilization

NOTE This technical information report is not a standard, and the material contained herein is not normative in nature. The committee has used the term "shall" in a few instances, based on their knowledge of requirements contained in relevant standards and regulatory requirements.

1 Scope

This technical information report (TIR) addresses various microbiological aspects of the development and validation of an ethylene oxide (EO) sterilization process. It does not cover the various factors that can have an effect on the bioburden of the product and on the sterilization process. This TIR provides additional guidance to ANSI/AAMI/ISO 11135:2014 for medical device manufacturers, including those that use contract sterilization facilities or contract sterilization operations.

Although the information presented was developed for application to medical devices, the content of this guideline may also be applied to other relevant products or materials.

Products that have been used in a healthcare setting and are being presented for resterilization in accordance with the manufacturer's instructions (see ANSI/AAMI/ISO 17664) are a special case. There is the potential for such products to possess a wide range of contaminating microorganisms and residual inorganic and/or organic contamination in spite of the application of a cleaning process. Hence, it is important to pay particular attention to the validation and control of the cleaning and disinfection processes used during reprocessing. Healthcare facilities are encouraged to review AAMI ST35, TIR30, and TIR34 for additional information on handling reusable or non-sterile devices requiring sterilization processing.

2 Terms and definitions

For the purposes of this TIR, the terms and definitions in ANSI/AAMI/ISO 11135: 2014 and the following apply.

- 2.1 compromised tissue:** Skin or mucous membrane that has been intentionally or accidentally opened, exposed, or breached
- 2.2 inoculated carrier:** Supporting material on or in which a defined number of test organisms has been deposited

3 Process and equipment characterization

3.1 Sterilization equipment

Guidelines for equipment selection can be found in AAMI TIR15 and EN 1422. Careful selection of the sterilizing equipment and development of the facility design will enable a manufacturer to process a product safely and effectively.

3.2 Process characterization—Physical parameters

3.2.1 Introduction

The variables that have a significant effect on the lethality of an ethylene oxide (EO) sterilization process are EO concentration, temperature, EO exposure time, and often relative humidity (RH). These variables are interrelated and a change in one can often be compensated for by a change in another. Other aspects that might affect lethality are the actual depth and rate of evacuations during the sterilant-removal phases as evacuations can be affected by chamber temperature, humidity, sterilant levels, vacuum pump performance, and/or product load configuration.