

Technical Information Report

AAMI TIR60: 2014/(R)2019

Common mode rejection in
ECG monitoring

This is a preview of "AAMI TIR60:2014 (R20...)". [Click here to purchase the full version from the ANSI store.](#)

Common mode rejection in ECG monitoring

Approved 24 October 2014 and reaffirmed 11 October 2019 by
AAMI

Abstract: This technical information report (TIR) provides the details of how to correctly build, calibrate, and use the CMR test circuit specified in ECG performance standards. It also preserves the history, rationale, and performance requirements of the test method, as included in ANSI/AAMI EC13:2002, *Cardiac monitors, heart rate meters, and alarms*, which was revised by the adoption and finalization of ANSI/AAMI/IEC 60601-2-27:2011, *Medical electrical equipment - Part 2-27: Particular requirements for the basic safety and essential performance of electrocardiographic monitoring equipment*.

Keywords: line frequency, CMR, channel noise, cardiac monitor

AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every five years but at least every 10 years. For a TIR, AAMI consults with a technical committee about five years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice, or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

CAUTION NOTICE: This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 901 N. Glebe Road, Suite 300, Arlington, VA 22203.

Published by

AAMI
901 N. Glebe Road, Suite 300
Arlington, VA 22203
www.aami.org

© 2014 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact AAMI at 901 N. Glebe Road, Suite 300, Arlington, VA 22203. Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

ISBN 1-57020-567-1

Contents

Page

| | |
|--|-----|
| Committee representation | iiv |
| Acknowledgments | v |
| Foreword | vi |
| 1 Scope | 1 |
| 2 Background | 1 |
| 3 Applicability and use | 1 |
| 4 Common mode rejection requirement | 1 |
| 5 Common mode rejection testing | 2 |
| 6 History | 2 |
| Annex A (informative) CMR test fixture design & application notes for ECG devices | 4 |

Committee representation

Association for the Advancement of Medical Instrumentation

Electrocardiograph (ECG) Committee

This technical information report was developed by the ECG Committee of the Association for the Advancement of Medical Instrumentation. Committee approval of the standard does not necessarily imply that all committee members voted for its approval.

At the time this document was published, the **AAMI Electrocardiograph Committee** had the following members:

- Cochairs:* Richard A. Sunderland, Welch Allyn, Inc.
Ahmet Turkmen, BS MS PhD, University of Wisconsin-Stout
Brian J. Young, GE Healthcare
- Members:* Robert William Bain, CBET, Baltimore Medical Engineers & Technician Society
Scott Coggins, Covidien
Prakash C. Deedwania, MD, The VA Medical Center
Laura Dhatt, Physio-Control
Sreeram Dhurjaty, Dhurjaty Electronics Consulting LLC
Greg Downs, Spacelabs Medical Inc.
Mike Jaffe, Cardiorespiratory Consulting LLC
Janice M. Jenkins, PhD, University of Michigan College of Engineering
Carolyn Lall, Draeger Medical Systems Inc.
Walter G. Lloyd, Childrens Hospital Boston
Peter W. Macfarlane, Royal Infirmary
Luis A. Melendez, Partners Healthcare
George Moody, Massachusetts Institute of Technology
Cadathur Rajagopalan, PhD, SMIEEE, Mindray DS USA Inc.
Linda Ricci, FDA/CDRH
Johann-Jakob Schmid, Schiller AG
Jonathan Steinberg, MD, St Lukes Roosevelt Hospital Center
John J. Wang, Philips Electronics North America
Jim Welch, Sotera Wireless Inc.
Jeffrey Wiser, 3M Healthcare
Ted Yantsides, Conmed Corp.
- Alternates:* Mark J. Callahan, Covidien
Kejian Chen, 3M Healthcare
Yu Chen, PhD, Draeger Medical Systems Inc.
Steve Duke, Physio-Control
Ruth Durie, Spacelabs Medical Inc.
Richard Gregg, Philips Electronics North America
Charles S. Ho, PhD, FDA/CDRH
Shen Luo, PhD, Mindray Medical
Serkan Sezer, Schiller AG

NOTE—Participation by federal agency representatives in the development of this standard does not constitute endorsement by the federal government or any of its agencies.

Acknowledgments

The AAMI ECG Committee would like to thank Mr. Carl Pantiskas, formerly of Draeger, for providing the guidance and leadership and for insisting that the “Common mode rejection” section from the old ANSI/AAMI EC13:2002, *Cardiac monitors, heart rate meters, and alarms*, be preserved when this standard is replaced by ANSI/AAMI/IEC 60601-2-27:2011, *Medical electrical equipment – Part 2-27: Particular requirements for the basic safety and essential performance of electrocardiographic monitoring equipment*.

Foreword

The AAMI ECG Committee recently adopted and finalized ANSI/AAMI/IEC 60601-2-27:2011, *Medical electrical equipment - Part 2-27: Particular requirements for the basic safety and essential performance of electrocardiographic monitoring equipment*, which replaced ANSI/AAMI EC13:2002, *Cardiac monitors, heart rate meters, and alarms*. In so doing, the ECG Committee members discovered that there are some sections in the EC13 standard which are not present in ANSI/AAMI/IEC 60601-2-27:2011. After much discussion, the committee members decided to preserve the "Common mode rejection" section from the EC13 standard as an AAMI technical information report (TIR). It is hoped that this TIR will be sent to IEC so that this section may be included in the future 60601-2-27 or stand alone as an IEC technical report (TR).

The AAMI ECG Committee develops the U.S. position and votes on ECG-related standards under the purview of the the U.S. technical advisory group (TAG) for IEC/SC 62D, administered by the Association for the Advancement of Medical Instrumentation (AAMI) on behalf of the American National Standards Institute (ANSI). The United States made considerable contributions to the ECG-related standards.

AAMI TIR60:2014, *Common mode rejection for ECG monitoring*, is intended to provide guidance on a particular subject matter relating to the ECG monitors.

Common mode rejection for ECG monitoring

1 Scope

This technical information report (TIR) provides the details of how to correctly build, calibrate, and use the CMR test circuit specified in ECG performance standards. It also preserves the history, rationale, and performance requirements of the test method.

2 Background

This TIR was created principally to preserve what was formerly Annex C of ANSI/AAMI EC13:2002 before this standard was replaced and harmonized into ANSI/AAMI/ IEC 60601-2-27:2011 without this information included. It has also been created to preserve the rationale and historical background for minimum CMR performance requirements, to offer references on the range of CMR observed in the past, and to discuss why this fixture appropriately models the conditions encountered in clinical practice – none of which are preserved in the newer international ECG performance standards.

Many U.S. and international ECG performance standards specify the use of a CMR fixture of this design and include the same minimum device CMR performance, but offer little guidance for properly building, calibrating, and using this CMR fixture. This TIR fills that gap.

3 Applicability and use

This TIR can be used with ANSI/AAMI/IEC 60601-2-25, ANSI/AAMI/IEC 60601-2-27, ANSI/AAMI/IEC 60601-2-47, and any other ECG standards that require performing a CMR test.

4 Common mode rejection requirement

The cardiac monitor shall have the capability of rejecting line frequency signal mode interfering voltages as encountered on the surface of the body. A line frequency signal, with 10 V rms source and a 200 picofarad (pF) source capacitance, applied from power ground to all patient electrode connections connected to a common node and with a parallel combination of a 51 kilohm resistor and 47 nF capacitor imbalance impedance in series with each patient lead, including the RL or green lead, if supplied, shall not produce an output signal exceeding 1 mV peak-to-valley (p-v) referred-to-input (RTI) over a 60 second period. The test shall be done while any line frequency notch filter (if provided) is turned off, even if it requires special software or a special method of accessing the control over that filter. This requirement shall be met with sequential shorting of the series-impedance-simulating lead imbalance in each active lead and with a +/- 300 mV dc offset potential placed in series with any patient electrode connection. The manufacturer's recommended patient cable shall be used for verification testing.

NOTE—IEC ECG standards differ from their former ANSI/AAMI ECG counterparts by defining this testing to be done with the 51 kilohm/47 nF series-impedance-simulating lead imbalance in all but one of the active leads being shorted while one of switches S1 – Sn at a time is sequentially opened. Although this different test method gives different performance during individual CMR tests, the worst case performance of the group of all CMR tests is virtually the same with either test method.

NOTE—In conducting a test to verify this capability, undesirable stray and leakage capacitance may result from shielding components and from patient cable capacitances when the cable is connected to the voltage source. The test method takes these capacitances into consideration.