

American National Standard

ANSI/AAMI/ISO 11138-4:2006/(R)2010



This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document,
contact AAMI at (877) 249-8226

ANSI/AAMI/ISO 11138-4:2006/(R)2010

Sterilization of health care products— Biological indicators— Part 4: Biological indicators for dry heat sterilization processes



Association for the Advancement
of Medical Instrumentation

The Objectives and Uses of AAMI Standards and Recommended Practices

It is most important that the objectives and potential uses of an AAMI product standard or recommended practice are clearly understood. The objectives of AAMI's technical development program derive from AAMI's overall mission: the advancement of medical instrumentation. Essential to such advancement are (1) a continued increase in the safe and effective application of current technologies to patient care, and (2) the encouragement of new technologies. It is AAMI's view that standards and recommended practices can contribute significantly to the advancement of medical instrumentation, provided that they are drafted with attention to these objectives and provided that arbitrary and restrictive uses are avoided.

A voluntary *standard* for a *medical device* recommends to the manufacturer the information that should be provided with or on the product, basic safety and performance criteria that should be considered in qualifying the device for clinical use, and the measurement techniques that can be used to determine whether the device conforms with the safety and performance criteria and/or to compare the performance characteristics of different products. Some standards emphasize the information that should be provided with the device, including performance characteristics, instructions for use, warnings and precautions, and other data considered important in ensuring the safe and effective use of the device in the clinical environment. Recommending the disclosure of performance characteristics often necessitates the development of specialized test methods to facilitate uniformity in reporting; reaching consensus on these tests can represent a considerable part of committee work. When a drafting committee determines that clinical concerns warrant the establishment of *minimum* safety and performance criteria, referee tests must be provided and the reasons for establishing the criteria must be documented in the rationale.

A *recommended practice* provides guidelines for the use, care, and/or processing of a medical device or system. A recommended practice does not address device performance *per se*, but rather procedures and practices that will help ensure that a device is used safely and effectively and that its performance will be maintained.

Although a device standard is primarily directed to the manufacturer, it may also be of value to the potential purchaser or user of the device as a fume of reference for device evaluation. Similarly, even though a recommended practice is usually oriented towards health care professionals, it may be useful to the manufacturer in better understanding the environment in which a medical device will be used. Also, some recommended practices, while not addressing device performance criteria, provide guidelines to industrial personnel on such subjects as sterilization processing, methods of collecting data to establish safety and efficacy, human engineering, and other processing or evaluation techniques; such guidelines may be useful to health care professionals in understanding industrial practices.

In determining whether an AAMI standard or recommended practice is relevant to the specific needs of a potential user of the document, several important concepts must be recognized:

All AAMI standards and recommended practices are *voluntary* (unless, of course, they are adopted by government regulatory or procurement authorities). The application of a standard or recommended practice is solely within the discretion and professional judgment of the user of the document.

Each AAMI standard or recommended practice reflects the collective expertise of a committee of health care professionals and industrial representatives, whose work has been reviewed nationally (and sometimes internationally). As such, the consensus recommendations embodied in a standard or recommended practice are intended to respond to clinical needs and, ultimately, to help ensure patient safety. A standard or recommended practice is limited, however, in the sense that it responds generally to perceived risks and conditions that may not always be relevant to specific situations. A standard or recommended practice is an important *reference* in responsible decision-making, but it should never *replace* responsible decision-making.

Despite periodic review and revision (at least once every five years), a standard or recommended practice is necessarily a static document applied to a dynamic technology. Therefore, a standards user must carefully review the reasons why the document was initially developed and the specific rationale for each of its provisions. This review will reveal whether the document remains relevant to the specific needs of the user.

Particular care should be taken in applying a product standard to existing devices and equipment, and in applying a recommended practice to current procedures and practices. While observed or potential risks with existing equipment typically form the basis for the safety and performance criteria defined in a standard, professional judgment must be used in applying these criteria to existing equipment. No single source of information will serve to identify a particular product as "unsafe". A voluntary standard can be used as one resource, but the ultimate decision as to product safety and efficacy must take into account the specifics of its utilization and, of course, cost-benefit considerations. Similarly, a recommended practice should be analyzed in the context of the specific needs and resources of the individual institution or firm. Again, the rationale accompanying each AAMI standard and recommended practice is an excellent guide to the reasoning and data underlying its provision.

In summary, a standard or recommended practice is truly useful only when it is used in conjunction with other sources of information and policy guidance and in the context of professional experience and judgment.

INTERPRETATIONS OF AAMI STANDARDS AND RECOMMENDED PRACTICES

Requests for interpretations of AAMI standards and recommended practices must be made in writing, to the Manager for Technical Development. An official interpretation must be approved by letter ballot of the originating committee and subsequently reviewed and approved by the AAMI Standards Board. The interpretation will become official and representation of the Association only upon exhaustion of any appeals and upon publication of notice of interpretation in the "Standards Monitor" section of the *AAMI News*. The Association for the Advancement of Medical Instrumentation disclaims responsibility for any characterization or explanation of a standard or recommended practice which has not been developed and communicated in accordance with this procedure and which is not published, by appropriate notice, as an *official interpretation* in the *AAMI News*.

American National Standard

ANSI/AAMI/ISO 11138-4:2006/(R)2010



This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document,
contact AAMI at (877) 249-8226

**Sterilization of health care products—
Biological indicators—Part 4:
Biological indicators for dry heat
sterilization processes**

Approved 9 December 2005 by
Association for the Advancement of Medical Instrumentation

Approved 22 March 2006 and reaffirmed 22 April 2010 by
American National Standards Institute, Inc.

Abstract: Provides specific requirements for test organisms, inoculated carriers, and biological indicators intended for use in assessing the performance of sterilization processes employing dry heat as the sterilizing agent.

Keywords: organism, suspension, carrier, primary packaging, inoculated, resistance

AAMI Standard

This Association for the Advancement of Medical Instrumentation (AAMI) standard implies a consensus of those substantially concerned with its scope and provisions. The existence of an AAMI standard does not in any respect preclude anyone, whether they have approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. AAMI standards are subject to periodic review, and users are cautioned to obtain the latest editions.

CAUTION NOTICE: This AAMI standard may be revised or withdrawn at any time. AAMI procedures require that action be taken to reaffirm, revise, or withdraw this standard no later than five years from the date of publication. Interested parties may obtain current information on all AAMI standards by calling or writing AAMI.

All AAMI standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are voluntary, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

For a complete copy of this AAMI document,
contact AAMI at (877) 249-8226
or visit www.aami.org.

Published by

Association for the Advancement of Medical Instrumentation
1110 N. Glebe Road, Suite 220
Arlington, VA 22201-4795

© 2006 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of ISO, ANSI, and AAMI. No part of this publication may be reproduced or distributed in any form, including an electronic retrieval system, without the prior written permission of AAMI. All requests pertaining to this draft should be submitted to AAMI. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact AAMI, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795. Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

ISBN 1-57020-264-8

Contents

Page

Glossary of equivalent standards	iv
Committee representation.....	vi
Background of AAMI adoption of ISO 11138-4:2006	ix
Foreword.....	x
Introduction	xi
1 Scope.....	1
2 Normative references	1
3 Terms and definitions.....	1
4 General requirements.....	1
5 Test organism.....	1
6 Suspension.....	2
7 Carrier and primary packaging.....	2
8 Inoculated carriers and biological indicators	2
9 Population and resistance	2
Annex A (normative) Method for determination of resistance to dry heat sterilization	4
A.1 General.....	4
A.2 Method.....	4
A.3 Determination of resistance.....	4
Annex B (normative) Calculation of z value.....	5



Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. For each International Standard that has been adopted by AAMI (and ANSI), the table below gives the corresponding U.S. designation and level of equivalency to the International Standard. NOTE: Documents are sorted by international designation.

Other normatively referenced International Standards may be under consideration for U.S. adoption by AAMI; therefore, this list should not be considered exhaustive.

International designation	U.S. designation	Equivalency
IEC 60601-1:2005	ANSI/AAMI ES60601-1:2005	Major technical variations
IEC 60601-1-2:2001 and Amendment 1:2004	ANSI/AAMI/IEC 60601-1-2:2001 and Amendment 1:2004	Identical
IEC 60601-2-04:2002	ANSI/AAMI DF80:2003	Major technical variations
IEC 60601-2-19:1990 and Amendment 1:1996	ANSI/AAMI I136:2004	Major technical variations
IEC 60601-2-20:1990 and Amendment 1:1996	ANSI/AAMI I151:2004	Major technical variations
IEC 60601-2-21:1994 and Amendment 1:1996	ANSI/AAMI/IEC 60601-2-21 and Amendment 1:2000 (consolidated texts)	Identical
IEC 60601-2-24:1998	ANSI/AAMI ID26:2004	Major technical variations
IEC/TR 60878:2003	ANSI/AAMI/IEC TIR60878:2003	Identical
IEC/TR 62296:2003	ANSI/AAMI/IEC TIR62296:2003	Identical
IEC 62304:2006	ANSI/AAMI/IEC 62304:2006	Identical
IEC/TR 62348:2006	ANSI/AAMI/IEC TIR62348:2006	Identical
ISO 5840:2005	ANSI/AAMI/ISO 5840:2005	Identical
ISO 7198:1998	ANSI/AAMI/ISO 7198:1998/2001/(R)2004	Identical
ISO 7199:1996	ANSI/AAMI/ISO 7199:1996/(R)2002	Identical
ISO 10993-1:2003	ANSI/AAMI/ISO 10993-1:2003	Identical
ISO 10993-2:2006	ANSI/AAMI/ISO 10993-2:200x ²	Identical
ISO 10993-3:2003	ANSI/AAMI/ISO 10993-3:2003	Identical
ISO 10993-4:2002	ANSI/AAMI/ISO 10993-4:2002	Identical
ISO 10993-5:1999	ANSI/AAMI/ISO 10993-5:1999	Identical
ISO 10993-6:1994	ANSI/AAMI/ISO 10993-6:1995/(R)2001	Identical
ISO 10993-7:1995	ANSI/AAMI/ISO 10993-7:1995/(R)2001	Identical
ISO 10993-9:1999	ANSI/AAMI/ISO 10993-9:1999/(R)2005	Identical
ISO 10993-10:2002 and Amendment 1:2006	ANSI/AAMI BE78:2002 ANSI/AAMI BE78:2002/A1:2006	Minor technical variations Identical
ISO 10993-11:1993	ANSI/AAMI 10993-11:1993	Minor technical variations
ISO 10993-12:2002	ANSI/AAMI/ISO 10993-12:2002	Identical
ISO 10993-13:1998	ANSI/AAMI/ISO 10993-13:1999/(R)2004	Identical
ISO 10993-14:2001	ANSI/AAMI/ISO 10993-14:2001	Identical
ISO 10993-15:2000	ANSI/AAMI/ISO 10993-15:2000	Identical
ISO 10993-16:1997	ANSI/AAMI/ISO 10993-16:1997/(R)2003	Identical
ISO 10993-17:2002	ANSI/AAMI/ISO 10993-17:2002	Identical
ISO 10993-18:2005	ANSI/AAMI BE83:2006	Major technical variations

International designation	U.S. designation	Equivalency
ISO/TS 10993-19:2006	ANSI/AAMI/ISO TIR10993-19:2006	Identical
ISO/TS 10993-20:2006	ANSI/AAMI/ISO TIR10993-20:2006	Identical
ISO 11135:1994	ANSI/AAMI/ISO 11135:1994	Identical
ISO 11137-1:2006	ANSI/AAMI/ISO 11137-1:2006	Identical
ISO 11137-2:2006 (2006-08-01 corrected version)	ANSI/AAMI/ISO 11137-2:2006	Identical
ISO 11137-3:2006	ANSI/AAMI/ISO 11137-3:2006	Identical
ISO 11138-1: 2006	ANSI/AAMI/ISO 11138-1:2006	Identical
ISO 11138-2: 2006	ANSI/AAMI/ISO 11138-2:2006	Identical
ISO 11138-3: 2006	ANSI/AAMI/ISO 11138-3:2006	Identical
ISO 11138-4: 2006	ANSI/AAMI/ISO 11138-4:2006	Identical
ISO 11138-5: 2006	ANSI/AAMI/ISO 11138-5:2006	Identical
ISO/TS 11139:2006	ANSI/AAMI/ISO 11139:2006	Identical
ISO 11140-1:2005	ANSI/AAMI/ISO 11140-1:2005	Identical
ISO 11140-5:2000	ANSI/AAMI ST66:1999	Major technical variations
ISO 11607-1:2006	ANSI/AAMI/ISO 11607-1:2006	Identical
ISO 11607-2:2006	ANSI/AAMI/ISO 11607-2:2006	Identical
ISO 11737-1:2006	ANSI/AAMI/ISO 11737-1:2006	Identical
ISO 11737-2:1998	ANSI/AAMI/ISO 11737-2:1998	Identical
ISO 11737-3:2004	ANSI/AAMI/ISO 11737-3:2004	Identical
ISO 13485:2003	ANSI/AAMI/ISO 13485:2003	Identical
ISO 13488:1996	ANSI/AAMI/ISO 13488:1996	Identical
ISO 14155-1:2003	ANSI/AAMI/ISO 14155-1:2003	Identical
ISO 14155-2:2003	ANSI/AAMI/ISO 14155-2:2003	Identical
ISO 14160:1998	ANSI/AAMI/ISO 14160:1998	Identical
ISO 14161:2000	ANSI/AAMI/ISO 14161:2000	Identical
ISO 14937:2000	ANSI/AAMI/ISO 14937:2000	Identical
ISO/TR 14969:2004	ANSI/AAMI/ISO TIR14969:2004	Identical
ISO 14971:2000 and A1:2003	ANSI/AAMI/ISO 14971:2000 and A1:2003	Identical
ISO 15223:2000, A1:2002, and A2:2004	ANSI/AAMI/ISO 15223:2000, A1:2001, and A2:2004	Identical
ISO 15225:2000 and A1:2004	ANSI/AAMI/ISO 15225:2000/(R)2006 and A1:2004/(R)2006	Identical
ISO 15674:2001	ANSI/AAMI/ISO 15674:2001	Identical
ISO 15675:2001	ANSI/AAMI/ISO 15675:2001	Identical
ISO/TS 15843:2000	ANSI/AAMI/ISO TIR15843:2000	Identical
ISO 15882:2003	ANSI/AAMI/ISO 15882:2003	Identical
ISO/TR 16142:2006	ANSI/AAMI/ISO TIR16142:2006	Identical
ISO 17664:2004	ANSI/AAMI ST81:2004	Major technical variations
ISO 17665-1:200x ¹	ANSI/AAMI/ISO 17665-1:2006	Identical
ISO 18472:2006	ANSI/AAMI/ISO 18472:2006	Identical
ISO/TS 19218:2005	ANSI/AAMI/ISO 19218:2005	Identical
ISO 25539-1:2003 and A1:2005	ANSI/AAMI/ISO 25539-1:2003 and A1:2005	Identical

¹ In production

² Final approval pending

Committee representation

Association for the Advancement of Medical Instrumentation Biological Indicators Working Group

The adoption of ISO 11138-4:2006 as an American National Standard was initiated by the AAMI Biological Indicators Working Group of the AAMI Sterilization Standards Committee. The AAMI Biological Indicators Working Group also functions as a U.S. Technical Advisory Group to the relevant work in the International Organization for Standardization (ISO). U.S. representatives from the AAMI Biological Indicators Working Group (U.S. Sub-TAG for ISO/TC 198/WG 4) played an active part in developing the ISO standard.

At the time this document was published, the **AAMI Biological Indicators Working Group** had the following members:

- Cochairs:* Gregg Mosley
Phil Schneider
- Members:* Richard Bancroft, Esq., Albert Browne, Ltd.
Heidi L. Betti, CST, CRST, Mercy Medical Center, Springfield, MA
Trabue D. Bryans, Apptec
Virginia C. Chamberlain, PhD, VC Chamberlain and Assoc., Palm Harbor, FL
Carlos Chavez, PhD, Abbott Laboratories
Charles Cogdill, Boston Scientific Corporation
Joseph Connaghan, MS, Alcon Laboratories
Gary Cranston, Consulting and Technical Services/PCS
Kimbrell Darnell, Bard Medical Division
Kate Davenport, Northview Biosciences
Douglas Davie, Sterilization Validation Services
Shawn Doyle, Sterilator Company, Inc.
Sylvie Dufresne, TSO3, Inc.
Dan Floyd, RM, Nelson Laboratories, Inc
James Gibson, Jr., J.M. Gibson Associates, Odessa, FL
John Gillis, PhD, SGM Biotech, Inc.
Joel R. Gorski, PhD, NAMSA
John Grillo, PhD, Hospira, Inc.
Joyce Hansen, JM Hansen & Associates
Thomas L. Hansen, Terumo Medical Corporation
Arthur C. Harris, Cook Incorporated
John L. Holland, Becton Dickinson
Charles A. Hughes, SPS Medical Supply Corporation
Danny Hutson, Cardinal Health
Lois A. Jones, MS, Cary, NC
Linda Lavelle, Johnson & Johnson
Patrick McCormick, PhD, Bausch & Lomb, Inc.
James McGowan, Jr., BS MBA, Sterile Works, Inc.
Candace McManus, DrPH, Food & Drug Administration/Center for Devices and Radiological Health
Gregg Mosely, Biotest Laboratories, Inc.
Bobby Osburn, Department of Veteran Affairs

Wendy Royalty-Hann, Raven Biological Laboratories
Terri Rymer, Baxter Healthcare Corporation
Manuel Saavedra, Jr., Kimberly-Clark Corporation
Phil Schneider, 3M Healthcare
Zenius Seliokas, Stericon, Inc.
Andrew Sharavara, Propper Manufacturing Company, Inc.
Barb Smith, Getinge USA
Gayle Strahearn, STS Division of Ethox Corporation
Nuong Van Trinh, TYCO Healthcare/Kendall
Jonathan Wilder, H&W Technology LLC
Alternates: Solomon Alade, PhD, Alcon Laboratories, Inc.
Richard Alexander, Abbott Laboratories
Thomas Berger, PhD, Hospira, Inc.
William Boentges, BS, Cardinal Health
Greg Crego, STS Division of Ethox Corporation
Georgina Deloatch, Propper Manufacturing Company Inc.
Christophe A. Demetrius, U.S. Food and Drug Administration
Brian Drumheller, CR Bard Medical Division
Catherine Finocchiaro, Bausch & Lomb, Inc.
Douglas F. Harbrecht, Boston Scientific Corporation
Burt Kingsbury, Terumo Medical Corporation
Garrett Krusheski, SGM Biotech, Inc.
David Liu, Johnson and Johnson
Michael Mattison, Getinge USA
Richard T. O'Donnell, Steris Corporation
Timothy Ramsey, BS, Northview Biosciences
Mike Sadowski, Baxter Healthcare Corporation
Gary Socola, SPS Medical Supply Corporation
Ralph Stick, Aptec
Craig Wallace, 3M Healthcare
Julie Wheeler, NAMSA
David Woolley, BS, Nelson Laboratories, Inc.

NOTE—Participation by federal agency representatives in the development of this standard does not constitute endorsement by the federal government or any of its agencies.

AAMI Sterilization Standards Committee

Cochairs: Victoria M. Hitchins, PhD
William E. Young

Members: Trabue D. Bryans, AppTec
Virginia C. Chamberlain, PhD, VC Chamberlain & Associates (Independent Expert)
Nancy Chobin, RN, CSPDM, St. Barnabas Healthcare System (Independent Expert)
Anne M. Cofiell, CRCST, FCS, International Association of Healthcare Central Service
Materiel Management
Charles Cogdill, Boston Scientific Corporation
Ramona Conner, RN, MSN, CNOR, Association of Perioperative Registered Nurses
Jacqueline Daley, Association for Professionals in Infection Control and Epidemiology
Kimbrell Darnell, CR Bard
Lisa Foster, Sterigenics International
James M. Gibson, Jr., JM Gibson Associates
Barbara J. Goodman, RN, BS, CNOR (Independent Expert)
Joel R. Gorski, PhD, NAMSA
Deborah A. Havlik, Hospira Inc.

Victoria M. Hitchins, PhD, FDA/CDRH
Richard M. Johnson, MSc, BSc, Abbott Laboratories
Lois Atkinson Jones, MS (Independent Expert)
Byron J. Lambert, PhD, Guidant Corporation/Cardiac Rhythm Management
Colleen Patricia Landers, RN, Canadian Standards Association
David Liu, Johnson & Johnson
Jeff Martin, Alcon Laboratories Inc.
Patrick J. McCormick, PhD, Bausch & Lomb Inc.
Thomas K. Moore, Getinge USA
Barry F.J. Page, Barry Page Consulting (Independent Expert)
Nancy J. Rakiewicz, Ethox Corporation
Phil M. Schneider, 3M Healthcare
Michael H. Scholla, Dupont Nonwovens
Mark Seybold, Baxter Healthcare Corporation
Andrew Sharavara, Propper Manufacturing Co Inc.
Frank Sizemore, American Society for Healthcare Central Service Professionals
Gregory O. Stecklein, MS, MSM, Cardinal Health (MP&S)
William N. Thompson, TYCO Healthcare/Kendall
John W. Walker, Steris Corporation
James L. Whitby, MA, MB, FRCP, University of Western Ontario (Independent Expert)
Thelma Wilcott, Becton Dickinson & Company
Martell Kress Winters, BS, SM, Nelson Laboratories Inc.
William E. Young (Independent Expert)

Alternates:
Lloyd Brown, TYCO Healthcare/Kendall
Lina C. Bueno, Dupont Nonwovens
Craig M. Herring, Johnson & Johnson
Clark W. Houghtling, Steris Corporation
Danny Hutson, Cardinal Health (MP&S)
Jim Kaiser, Bausch & Lomb Inc.
Susan G. Klacik, AS, BS, International Association of Healthcare Central Service Materiel
Management
Joseph J. Lasich, BS, Alcon Laboratories Inc.
Chiu Lin, PhD, FDA/CDRH
Lisa N. Macdonald, Becton Dickinson & Company
Ralph Makinen, Guidant Corporation/Cardiac Rhythm Management
Mary S. Mayo, CR Bard
David Ford McGoldrick, BS, Abbott Laboratories
Jerry R. Nelson, MS, PhD, Nelson Laboratories Inc.
Jeff Peltier, Boston Scientific Corporation
Janet Prust, 3M Healthcare
Mike Sadowski, Baxter Healthcare Corporation
Ralph Stick, AppTec
Jason Voisinet, Ethox Corporation
Valerie Welter, Hospira Inc.
William T. Young, Sterigenics International

NOTE—Participation by federal agency representatives in the development of this document does not constitute endorsement by the federal government or any of its agencies.

Background of AAMI adoption of ISO 11138-4:2006

As indicated in the foreword to the main body of this document (page xi), the International Organization for Standardization (ISO) is a worldwide federation of national standards bodies. The United States is one of the ISO members that took an active role in the development of this standard.

ISO 11138-4:2006 was developed by ISO Technical Committee 198, *Sterilization of health care products*, to fill a need for an international standard for specifying the test organisms and performance requirements for biological indicators (including inoculated carriers and suspensions) used in assessing the performance of sterilization processes employing dry heat as the sterilizing agent.

U.S. participation in ISO/TC 198 is organized through the U.S. Technical Advisory Group (TAG) for ISO/TC 198, administered by the Association for the Advancement of Medical Instrumentation (AAMI) on behalf of the American National Standards Institute (ANSI). The U.S. TAG for ISO/TC 198 made considerable contributions to this standard and supports the requirements for biological indicators specified in this document.

AAMI encourages its committees to harmonize their work with international standards as much as possible. Upon review of the final Draft International Standard (DIS) of ISO 11138-4:2006, the AAMI Biological Indicator Working Group decided to adopt this document verbatim.

The ISO 11138:2006 biological indicator standards series was developed as the result of the joint revision of the ISO 11138:1994-1995 series of biological indicator standards (Parts 1-3) and the EN 866:1997-2000 series of biological indicator standards (Parts 1-8). The revised ISO 11138:2006 series of standards consist of the following parts:

ISO 11138-1, *Sterilization of health care products — Biological indicators — Part 1: General requirements*

ISO 11138-2, *Sterilization of health care products — Biological indicators — Part 2: Biological indicators for ethylene oxide sterilization processes*

ISO 11138-3, *Sterilization of health care products — Biological indicators — Part 3: Biological indicators for moist heat sterilization processes*

ISO 11138-4, *Sterilization of health care products — Biological indicators — Part 4: Biological indicators for dry heat sterilization processes*

ISO 11138-5, *Sterilization of health care products — Biological indicators — Part 5: Biological indicators for low-temperature steam and formaldehyde sterilization processes*

Major changes that were made to the predicate ISO and CEN series during the revision process which are incorporated into the revised ISO 11138:2006 series of standards include:

- a) Elimination of EN 866-4:1999, *Biological systems for testing sterilizers and sterilization processes — Part 4: Particular systems for use in irradiation sterilizers*. (Radiation biological indicators can demonstrate ISO 11138:2006 compliance by complying with the provisions of ISO 11138-1:2006 even though there is no specific subpart for irradiation).
- b) Elimination of EN 866-7:1999 and EN 866-8:1999, *Biological systems for testing sterilizers and sterilization processes — Part 7: Particular requirements for self-contained biological indicator*

systems for use in moist heat sterilizers and Part 8: Particular requirements for self-contained biological indicator systems for use in ethylene oxide sterilizers.

- c) Inclusion of specific information pertaining to self-contained biological indicators in ISO 11138-1:2006.
- d) Inclusion of a table with consolidated labeling requirements in ISO 11138-1:2006.
- e) Provision for use of biological indicators deviating from the specified minimum population and/or resistance criteria providing all other requirements of ISO 11138:2006 are met and the deviation is clearly indicated in the product labeling.
- f) Allowance for the calculation of D value by either the Holcomb-Spearman-Karber, Limited Holcomb-Spearman-Karber or Stumbo-Murphy-Cochran procedures as indicated in Annex D, 11138-1:2006.
- g) Allowance for the use of dual species biological indicators with appropriate documentation.
- h) Removal of the performance requirements for resistometers in ISO 11138:2006 Parts 2-4 (resistometer performance requirements are contained in ISO 18472:2006).
- i) Removal of the \log_{10} population \times D value ≥ 10 requirement in ISO 11138-3:2006 (moist heat) and 11138-4:2006 (dry heat).
- j) Provision for use of a liquid (rather than vapor phase) test method for characterization of biological indicators used in the low temperature steam and formaldehyde process in ISO 11138-5:2006.

AAMI and ANSI procedures require that standards be reviewed and, if necessary, revised to reflect technological advances that may have occurred since publication.

AAMI (and ANSI) have adopted other ISO standards. See the Glossary of Equivalent Standards for a list of ISO standards adopted by AAMI, which gives the corresponding U.S. designation and the level of equivalency with the ISO standard.

The concepts incorporated in this standard should not be considered inflexible or static. This standard, like any other, must be reviewed and updated periodically to assimilate progressive technological developments. To remain relevant, it must be modified as technological advances are made and as new data come to light.

Suggestions for improving this standard are invited. Comments and suggested revisions should be sent to Standards Department, AAMI, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795.

NOTE—Beginning with the foreword on page xi, this American National Standard is identical to ISO 11138-4:2006.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 11138-4 was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

ISO 11138 consists of the following parts, under the general title *Sterilization of health care products — Biological indicators*:

- *Part 1: General requirements*
- *Part 2: Biological indicators for ethylene oxide sterilization processes*
- *Part 3: Biological indicators for moist heat sterilization processes*
- *Part 4: Biological indicators for dry heat sterilization processes*
- *Part 5: Biological indicators for low-temperature steam and formaldehyde sterilization processes*

Introduction

ISO 11138-1 specifies production, labeling, test methods and performance requirements for the manufacture of biological indicators including inoculated carriers and suspensions intended for use in validation and monitoring sterilization processes. This part of ISO 11138 gives specific requirements for those biological indicators intended for use in dry heat sterilization processes.

The intent of providing requirements in the ISO 11138 series of International Standards is to provide general requirements and requirements for test methods. This series of International Standards represents the current "state-of-the-art" according to the experts representing manufacturers, users and regulatory authorities involved in developing the standard. The intent is not to promote the use of biological indicators where such use is not advised, but to provide common requirements for the production of those biological indicators that are known to be in use today.

Standards exist providing general requirements for the validation and control of general sterilization processes (see ISO 14937)¹⁾

NOTE Some countries or regions may have published standards covering requirements for sterilization or biological indicators.

Advice on selection, use, and interpretation of results when using biological indicators can be found in ISO 14161.

For a complete copy of this AAMI document,
contact AAMI at (877) 243-8226
or visit www.aami.org.

1) Although ISO/TC 198 has agreed to develop a standard applicable to dry heat processes, it was not available for reference at the time this document was prepared.

American National Standard

ANSI/AAMI/ISO 11138-4:2006/(R)2010

Sterilization of health care products — Biological Indicators — Part 4: Biological indicators for dry heat sterilization processes

1 Scope

This part of ISO 11138 provides specific requirements for test organisms, suspensions, inoculated carriers, biological indicators, and test methods intended for use in assessing the performance of sterilization processes employing dry heat as the sterilizing agent at sterilizing temperatures within the range of 120 °C to 180 °C.

NOTE 1 Requirements for validation and control of dry heat sterilization processes are provided by ISO 14937.

NOTE 2 Requirements for work place safety may be provided by national or regional regulations.

For a complete copy of this AAMI document,
contact AAMI at (877) 249-8226
or visit www.aami.org.