

# American National Standard



## ANSI/AAMI/ ISO 14971: 2019

Medical devices—  
Application of risk  
management to medical  
devices

This is a preview of "ANSI/AAMI/ISO 14971:...". [Click here to purchase the full version from the ANSI store.](#)

American National Standard

ANSI/AAMI/ISO 14971:2019  
(Revision of ANSI/AAMI/ISO 14971:2007)

# Medical devices—Application of risk management to medical devices

Approved 2 May 2019 by  
**AAMI**

Approved 10 May 2019 by  
**American National Standards Institute**

**Abstract:** Specifies a process for a manufacturer to identify the hazards associated with medical devices, including in vitro diagnostic (IVD) medical devices, to estimate and evaluate the associated risks, to control these risks, and to monitor the effectiveness of the controls.

**Keywords:** medical device, risk management

## AAMI Standard

This Association for the Advancement of Medical Instrumentation (AAMI) standard implies a consensus of those substantially concerned with its scope and provisions. The existence of an AAMI standard does not in any respect preclude anyone, whether they have approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. AAMI standards are subject to periodic review, and users are cautioned to obtain the latest editions.

**CAUTION NOTICE:** This AAMI standard may be revised or withdrawn at any time. AAMI procedures require that action be taken to reaffirm, revise, or withdraw this standard no later than 5 years from the date of publication. Interested parties may obtain current information on all AAMI standards by calling or writing AAMI.

All AAMI standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

*Published by*

AAMI  
901 N. Glebe Road, Suite 300  
Arlington, VA 22203

© 2019 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of ISO and AAMI. No part of this publication may be reproduced or distributed in any form, including an electronic retrieval system, without the prior written permission of AAMI. All requests pertaining to this draft should be submitted to AAMI. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact AAMI, 901 N. Glebe Road, Suite 300, Arlington, VA 22203 Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

**ISBN 978-1-57020-718-1**

## Contents

Page

Committee representation.....	iv
Background of ANSI/AAMI adoption of ISO 14971:2019.....	vi
Foreword.....	vii
Introduction .....	ix
1 Scope.....	1
2 Normative references .....	1
3 Terms and definitions.....	1
4 General requirements for <i>risk management system</i> .....	7
5 <i>Risk analysis</i> .....	11
6 <i>Risk evaluation</i> .....	13
7 <i>Risk control</i> .....	14
8 Evaluation of overall <i>residual risk</i> .....	16
9 <i>Risk management review</i> .....	16
10 Production and <i>post-production</i> activities.....	16
Annex A (informative) Rationale for requirements .....	19
Annex B (informative) <i>Risk management process for medical devices</i> .....	27
Annex C (informative) Fundamental <i>risk</i> concepts.....	31
Bibliography .....	37

## Committee representation

### Association for the Advancement of Medical Instrumentation Application of Risk Management to Medical Devices Working Group

The adoption of ISO 14971:2019 as an American National Standard was initiated by the AAMI Application of Risk Management to Medical Devices Working Group. AAMI QM/WG04 provides input to the Quality Management and Corresponding General Aspects for Medical Devices Committee which is the responsible group for providing the U.S. input to the relevant group in ISO/TC210. U.S. representatives from AAMI Quality Management and Corresponding General Aspects for Medical Devices Working Group and the TAG played an active part in developing the ISO document.

At the time this document was published, the committee **Application of Risk Management to Medical Devices Working Group** has the following members:

*Cochairs:* Christine Krenc  
Melissa Torres

*Members:* David Amor, Pear Therapeutics Inc  
Pat Baird, Philips  
Pierre Barbier, Alcon Laboratories Inc  
Edwin Bills, ELB Consulting  
Amar Chanani, Avanos Medical  
Roberto Del Cid, Conmed Corp  
Tushar Dharampal, Abbott Laboratories  
Elizabeth DiDonato, Department of Defense - Defense Blood Standard System Project Office  
Thomas Dold, Vista Regulatory Solutions  
Sherman Eagles, SoftwareCPR  
Hugo Felix, Owlet Baby Care  
Jason Fuller, Stryker Instruments Division  
David Geraghty, Spacelabs Healthcare  
Michael Glynn, Battelle Memorial Institute  
Richard Granquist, NAMSA  
Michael Groendyk, Arthrex Inc  
Michael Gustafson, Siemens Healthineers  
Casey Haley, LivaNova PLC  
John Hedley-Whyte, Harvard University  
Jose Justiniano, Johnson & Johnson  
Patricia Krantz-Zuppan, Medtronic Inc Campus  
Sonja Kraus, Draeger Medical Systems Inc.  
Christine Krenc, KTA Compliance Consulting  
Chad Kymal, Omnex  
Dan Laelle, Nonin Medical Inc  
Mark Leimbeck, UL LLC  
Eduardo Martinez, Abbott Laboratories  
Mercedes Massana, MDM Engineering Consultants  
Dino Mavromatis, Regulatory and Quality Solutions LLC  
Mike McAndrew, Getinge USA  
Michael McCarthy, Baxter Healthcare Corporation  
Kathleen Miller, Amgen Inc  
Enio Montenegro, Cadwell Laboratories Inc  
Pablo Montero, Boston Scientific Corporation  
Wade Munsch, MET Laboratories Inc  
Jorge Ochoa, Exponent  
B Park, Christine Park & Associates  
Anthony Piotrkowski, STERIS Corporation  
Frank Pokrop, Sotera Wireless Inc

Donald Powers, Powers Consulting Services  
Joseph Raciti, Tech Group North America dba West  
Pooja Roychoudhury, Regulatory and Quality Solutions LLC  
Bobbi Sharp, Northgate Technologies  
Mike Silvestri, Terumo Americas Corporate  
Molly Story, Sanofi  
Mark Swanson, H&M Consulting Group LLC  
Radhakrishna Tirumalai, US Pharmacopeia Convention Inc  
Marco Tognolo, Baxter Healthcare Corporation  
Melissa Torres, FDA/CDRH  
Jon Ward, Smiths Medical  
John Yager, DexCom Inc  
Daidi Zhong, Chongqing University  
Weiping Zhong, GE Healthcare  
Jennifer Zuba, WL Gore & Associates Inc  
Gottlieb Glauninger, Eli Lilly & Company  
Brodie Pedersen, Borderless Compliance LLC  
Kinza Shekhani, ICU Medical Inc  
Tim Croft, Hill-Rom Holdings

*Alternates:*

Dianne Batch, Becton Dickinson & Company  
Nicholas Bennett, Arthrex Inc  
William Brodbeck, STERIS Corporation  
Melissa Burns, FDA/CDRH  
Michael Caspers, Sanofi  
Charles Finch, ResMed Inc.  
Aaron Goodstein, Johnson & Johnson  
Lia Haley, Philips  
Kristen Hastings, Boston Scientific Corporation  
Marissa Koran, Smiths Medical  
Ken Link, NAMSA  
Joe Mroz, Stryker Instruments Division  
Susumu Nozawa, Siemens Healthineers  
Karen O'Neill, Draeger Medical Systems Inc.  
Shalini Pandya, Avanos Medical  
Luann Pendy, Medtronic Inc Campus  
Kanar Rabah, Sanofi  
Brian Ray, Hill-Rom Holdings  
James Shene, LivaNova PLC  
Mark Smith, Getinge USA  
Tomer Solomon, Tech Group North America dba West  
Tagore Somers, Eli Lilly & Company

---

NOTE—Participation by federal agency representatives in the development of this standard does not constitute endorsement by the federal government or any of its agencies.

---

## Background of ANSI/AAMI adoption of ISO 14971:2019

As indicated in the foreword to the main body of this document (page vii), the International Organization for Standardization (ISO) is a worldwide federation of national standards bodies. The United States is one of the ISO members that took an active role in the development of this standard, which was developed by ISO/TC210, Quality management and corresponding general aspects for medical devices to specify a process for a manufacturer to identify the hazards associated with medical devices, including in vitro diagnostic (IVD) medical devices, to estimate and evaluate the associated risks, to control these risks, and to monitor the effectiveness of the controls.

U.S. participation in ISO/TC210 is organized through the U.S. Technical Advisory Group to Quality Management and Corresponding General Aspects for Medical Devices Committee, administered by the Association for the Advancement of Medical Instrumentation. Experts from the United States made a considerable contribution to this standard.

AAMI encourages its committees to harmonize their work with International Standards in the area of quality management and corresponding general aspects for medical devices. Upon review of ISO 14971, the Quality Management and Corresponding General Aspects for Medical Devices Committee and the AAMI Application of Risk Management to Medical Devices Working Group decided to adopt it verbatim, as a revision of ANSI/AAMI/ISO 14971:2007.

AAMI and ANSI procedures require that standards be reviewed and, if necessary, revised every five years to reflect technological advances that may have occurred since publication.

As used within the context of this document, "shall" indicates requirements strictly to be followed to conform to the recommended practice. "Should" indicates that among several possibilities, one is recommended as particularly suitable, without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action should be avoided but is not prohibited.

"May" is used to indicate that a course of action is permissible within the limits of the standard. "Can" is used as a statement of possibility and capability. Finally, "must" is used only to describe "unavoidable" situations, including those mandated by government regulation.

---

NOTE Users of this standard are advised that this document is an AAMI identical adoption of an ISO document and that the following international conventions have been carried over to the AAMI publication:

- British English spelling (e.g. colour instead of color)
- Use of SI units (e.g. metres instead of feet, Celsius instead of Fahrenheit, etc.)
- Decimal comma instead of a decimal point (e.g. 1 000,15 instead of 1,000.15)

---

The concepts incorporated in this standard should not be considered inflexible or static. This standard, like any other, must be reviewed and updated periodically to assimilate progressive technological developments. To remain relevant, it must be modified as technological advances are made and as new data comes to light.

Suggestions for improving this standard are invited. Comments and suggested revisions should be sent to Standards Department, AAMI, 901 N. Glebe Road, Suite 300, Arlington, VA 22203

---

NOTE—Beginning with the ISO foreword on page vii, this American National Standard is identical to ISO 14971:2019.

---



## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 210, *Quality management and corresponding general aspects for medical devices*, and IEC/SC 62A, *Common aspects of electrical equipment used in medical practice*.

This third edition cancels and replaces the second edition (ISO 14971:2007), which has been technically revised. The main changes compared to the previous edition are as follows:

- A clause on normative references has been included, in order to respect the requirements for fixed in Clause 15 of ISO/IEC Directives, Part 2:2018.
- The defined terms are updated and many are derived from ISO/IEC Guide 63:2019. Defined terms are printed in italic to assist the reader in identifying them in the body of the document.
- Definitions of *benefit*, *reasonably foreseeable misuse* and *state of the art* have been introduced.
- More attention is given to the *benefits* that are expected from the use of the *medical device*. The term *benefit-risk* analysis has been aligned with terminology used in some regulations.
- It is explained that the *process* described in ISO 14971 can be used for managing *risks* associated with *medical devices*, including those related to data and systems security.
- The method for the evaluation of the overall *residual risk* and the criteria for its acceptability are required to be defined in the *risk management* plan. The method can include gathering and reviewing data and literature for the *medical device* and for similar *medical devices* and similar other products on the market. The criteria for the acceptability of the overall *residual risk* can be different from the criteria for acceptability of individual *risks*.

- The requirements to disclose *residual risks* have been moved and merged into one requirement, after the overall *residual risk* has been evaluated and judged acceptable.
- The review before commercial distribution of the *medical device* concerns the execution of the *risk management* plan. The results of the review are documented as the *risk management* report.
- The requirements for production and *post-production* activities have been clarified and restructured. More detail is given on the information to be collected and the actions to be taken when the collected information has been reviewed and determined to be relevant to *safety*.
- Several informative annexes are moved to the guidance in ISO/TR 24971, which has been revised in parallel. More information and a rationale for the requirements in this third edition of ISO 14971 have been provided in Annex A. The correspondence between the clauses of the second edition and those of this third edition is given in Annex B.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

The requirements contained in this document provide *manufacturers* with a framework within which experience, insight and judgment are applied systematically to manage the *risks* associated with the use of *medical devices*.

This document was developed specifically for *manufacturers of medical devices* on the basis of established principles of *risk management* that have evolved over many years. This document could be used as guidance in developing and maintaining a *risk management process* for other products that are not necessarily *medical devices* in some jurisdictions and for suppliers and other parties involved in the *medical device life cycle*.

This document deals with *processes* for managing *risks* associated with *medical devices*. *Risks* can be related to injury, not only to the patient, but also to the user and other persons. *Risks* can also be related to damage to property (for example objects, data, other equipment) or the environment.

*Risk management* is a complex subject because each stakeholder can place a different value on the acceptability of *risks* in relation to the anticipated *benefits*. The concepts of *risk management* are particularly important in relation to *medical devices* because of the variety of stakeholders including medical practitioners, the organizations providing health care, governments, industry, patients and members of the public.

It is generally accepted that the concept of *risk* has two key components:

- the probability of occurrence of *harm*; and
- the consequences of that *harm*, that is, how severe it might be.

All stakeholders need to understand that the use of a *medical device* involves an inherent degree of *risk*, even after the *risks* have been reduced to an acceptable level. It is well known that in the context of a clinical *procedure* some *residual risks* remain. The acceptability of a *risk* to a stakeholder is influenced by the key components listed above and by the stakeholder's perception of the *risk* and the *benefit*. Each stakeholder's perception can vary depending upon their cultural background, the socio-economic and educational background of the society concerned and the actual and perceived state of health of the patient. The way a *risk* is perceived also takes into account other factors, for example, whether exposure to the *hazard* or *hazardous situation* seems to be involuntary, avoidable, from a man-made source, due to negligence, arising from a poorly understood cause, or directed at a vulnerable group within society.

As one of the stakeholders, the *manufacturer* reduces *risks* and makes judgments relating to the *safety* of a *medical device*, including the acceptability of *residual risks*. The *manufacturer* takes into account the generally acknowledged *state of the art*, in order to determine the suitability of a *medical device* to be placed on the market for its *intended use*. This document specifies a *process* through which the *manufacturer* of a *medical device* can identify *hazards* associated with the *medical device*, estimate and evaluate the *risks* associated with these *hazards*, control these *risks*, and monitor the effectiveness of the controls throughout the *life cycle* of the *medical device*.

The decision to use a *medical device* in the context of a particular clinical *procedure* requires the *residual risks* to be balanced against the anticipated *benefits* of the *procedure*. Such decisions are beyond the scope of this document and take into account the *intended use*, the circumstances of use, the performance and *risks* associated with the *medical device*, as well as the *risks* and *benefits* associated with the clinical *procedure*. Some of these decisions can be made only by a qualified medical practitioner with knowledge of the state of health of an individual patient or the patient's own opinion.

For any particular *medical device*, other standards or regulations could require the application of specific methods for managing *risk*. In those cases, it is necessary to also follow the requirements outlined in those documents.

The verbal forms used in this document conform to the usage described in Clause 7 of the ISO/IEC Directives, Part 2:2018. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe permission (e.g. a permissible way to achieve compliance with a requirement or test);

- “can” is used to express possibility and capability; and
- “must” is used to express an external constraint that is not a requirement of the document.

# Medical devices—Application of risk management to medical devices

## 1 Scope

This document specifies terminology, principles and a *process* for *risk management* of *medical devices*, including software as a *medical device* and *in vitro diagnostic medical devices*. The *process* described in this document intends to assist *manufacturers* of *medical devices* to identify the *hazards* associated with the *medical device*, to estimate and evaluate the associated *risks*, to control these *risks*, and to monitor the effectiveness of the controls.

The requirements of this document are applicable to all phases of the *life cycle* of a *medical device*. The *process* described in this document applies to *risks* associated with a *medical device*, such as *risks* related to biocompatibility, data and systems security, electricity, moving parts, radiation, and usability.

The *process* described in this document can also be applied to products that are not necessarily *medical devices* in some jurisdictions and can also be used by others involved in the *medical device life cycle*.

This document does not apply to:

- decisions on the use of a *medical device* in the context of any particular clinical *procedure*; or
- business *risk management*.

This document requires *manufacturers* to establish objective criteria for *risk* acceptability but does not specify acceptable *risk* levels.

*Risk management* can be an integral part of a quality management system. However, this document does not require the *manufacturer* to have a quality management system in place.

NOTE Guidance on the application of this document can be found in ISO/TR 24971<sup>[9]</sup>.

## 2 Normative references

There are no normative references in this document.

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

### 3.1

#### *accompanying documentation*