

ANSI Z80.20-2010

American National Standard

*for Ophthalmics
Contact Lenses –
Standard Terminology,
Tolerances, Measurements and
Physicochemical Properties*

Developed by



Where IT all begins



This is a preview of "VC ANSI Z80.20-2010". [Click here to purchase the full version from the ANSI store.](#)

ANSI
Z80.20-2010
Revision of
ANSI Z80.20-2004

American National Standard
for Ophthalmics –

**Contact Lenses –
Standard Terminology, Tolerances,
Measurements and
Physicochemical Properties**

Secretariat
The Vision Council

Approved December 6, 2010
American National Standards Institute, Inc.

American National Standard

Approval of an American National Standard requires review by ANSI that the requirements for due process, consensus, and other criteria for approval have been met by the standards developer.

Consensus is established when, in the judgement of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made towards their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether he has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation of any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

CAUTION NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this standard. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

Developed by

The Accredited Committee Z80 for Ophthalmic Standards

The Vision Council
Z80 Secretariat
225 Reinekers Lane
Alexandria, VA 22314

Published by
The Vision Council
Z80 Secretariat
225 Reinekers Lane
Alexandria, VA 22314

Copyright © 2010 by The Vision Council
All rights reserved.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without prior written permission of the publisher.

Printed in the United States of America

Contents

	Page
Foreword	iii
1 Scope and Purpose	1
2 Normative References	2
3 Contact Lens Terminology.....	4
4 Nomenclature and Classification of Contact Lens Materials	16
5 Standard Tolerances of Finished Contact Lenses.....	21
6 Physicochemical Properties of Contact Lens Materials	27
7 Standard Measurement Methods for Contact Lenses	28

Tables

1 Contact Lens materials.....	1
2 Symbols for Parameters and Linear Dimensions	4
3 Tolerances for Corneal and Paralimbal Contact Lenses	22
4 Tolerances for Scleral (Haptic) Contact Lenses	24
5 Tolerance Limits of Material and Contact Lens Physical Properties	25
6 Physicochemical Properties: Repeatability, Test Methods, and Units of Measure	28
7 Summary of Sagittal Depth Equations in Terms of Apical Radius of Curvature (r), Eccentricity (e), Chord Diameter ($2h$), and Lens Diameter (D)	35
8 Specifications for a V-Groove Diameter Gauge	44

Figures

1 Linear Dimensions of a Corneal Contact Lens	5
2 Qualitative Specification of Edge Thickness and Shape for Rigid Contact Lenses	26
3 Optical System of a Microsphereometer	29
4 Optical System of an Ophthalmometer	33
5 Measurement of Sagittal Depth of a Soft Lens	36
6 Annular Cylinder Contact Lens Support for the Focimeter	37
7 Contact Lens Resting on an Annular Cylinder Contact Lens Stop	38
8 Paddle Contact Lens Support (Stop) for the Focimeter.....	39
9 Contact Lens Resting on a Paddle Contact Lens Stop	39
10 V-Groove Diameter Gauge	43
11 Principle of the Projection Comparator.....	45
12 Measurement of Center Thickness with a Dial Gauge	49
13 Apparatus for Determination of Spectral and Luminous Transmittance	54

	Page
14 Specimen Dimensions for Testing of Flexural Strength, Dimensions in Millimeters	59
15 Fixture Holding Specimen for Testing of Flexural Strength, Dimensions in Inches	60
16 Hydrogel Sample between Sensor and Retaining Mesh	66
17 Clamping Mechanism for Polarographic Measurement.....	66
18 Polarographic Cell Body	67
19 Nonhydrogel Sample between Sensor and Retaining Mesh.....	67
20 Coulometric Apparatus	71
21 Diffusion Cell for Coulometric Method	71
22 Measurement of Bulk Material with Prism-Coupling Device.....	77
Annexes	
A Conformance Criteria for Testing Products	82
B Measurement Systems Requirements	85
C Example of Calculating Gauge R&R	87
D Bibliography	94

Foreword (This foreword is not part of American National Standard ANSI Z80.20-2010.)

The Z80 Committee for Ophthalmics was organized in 1956, and the committee's initial standard on ophthalmic lenses was issued in 1964. In the ensuing years the committee expanded its scope and organization into a number of subcommittees, each charged with specific areas of responsibility. One of these, The Subcommittee for Contact Lenses, has published two standards (Insert names) which are dutifully reviewed on a five year basis to include any newly established information.

Since 1980, the Subcommittee for Contact Lenses has provided delegates to the International Standards Organization to represent the American Standards Institute's role (for the United States) in developing and approving contact lens and contact lens care products standards for the international community. In providing this service for ANSI it became apparent that the role of the delegates continued to require reaffirmation from the American community as a whole to support their input. Thus in addition to providing information to the international forum the Subcommittee for Contact Lenses continues to process all available information through its ANSI network.

This latest version of Z80.20 was written to include provisions for newer contact lenses and materials now on the public market and to insure that the US standard is in conformance with the International Standard. Thus this revision of Z80.20 and future reviews or revisions of this standard will ensure that the continued participation of American delegates to ISO will be supported by an America standard that has received a consensus by the Z80 Committee for Ophthalmic standards in addition to other American interests.

This standard contains four informative annexes, which are not considered part of this standard.

Suggestions for improvement of this standard will be welcome. They should be sent the Vision Council, 225 Reinekers Lane, Suite 700, Alexandria, VA 22314.

This standard was processed and approved for submittal to ANSI by the Accredited Standards Committee on Ophthalmic Standards, Z80. Committee approval of this standard does not necessarily imply that all committee members voted for its approval. At the time it approved this standard, the Z80 Committee had the following members

Thomas C. White, M.D., Chairman
Quido A. Cappelli, P.E., Vice-Chairman
Robert Rosenberg, O.D., Secretary

<i>Organization Represented</i>	<i>Name of Representative</i>
Advanced Medical Technologies Association	Douglas Fortunato Richard Courtney (Alt.) Glenn Davies (Alt.) Bernie Liebler (Alt.)
American Academy of Ophthalmology	Thomas White Carl Tubbs (Alt.) Pradeep Ramulu (Alt.) Rebecca Hyder (Alt.)
American Academy of Optometry	David S. Loshin
American Ceramic Society	Lyle Rubin Herbert Hoover (Alt.)
American Glaucoma Society	Steven Gedde Douglas Rhee (Alt.)

<i>Organization Represented</i>	<i>Name of Representative</i>
American Optometric Association	Jeffrey Weaver Robert Rosenberg (Alt.) William Benjamin (Alt.) William Brown (Alt.)
American Society of Cataract and Refractive Surgery Jack Holladay (Alt.)	Steven Klyce
Contact Lens Institute.....	Stan Rogaski Peter Mathers (Alt.)
Contact Lens Manufacturer's Association	Quido Cappelli Jan Svochak (Alt.) John Walfoort (Alt.)
Department of Veterans Affairs	John Townsend Sharon Atkin (Alt.)
Federated Cornea Societies/ASCRS	Michael Belin David Glasser (Alt.)
Food & Drug Administration	Don Calogero Robert James (Alt.)
National Association. of Optometrists & Opticians	Nick Mileti Joe Dezenzo (Alt.) Franklin Rozak (Alt.)
National Academy of Opticianry	James Elkins
Optical Laboratory Association.....	Daniel Torgersen Virginia (Susie) Leshner (Alt.) Richard Tinson (Alt.) Robert Dziuban (Alt.)
Opticians Association of America	Tom Hicks Chris Allen (Alt.)
Sunglass Association of America	Kenneth Frederick Scott MacGuffie (Alt.) Rick Van Arnham (Alt.)
US Leader to ISO TC 172/SC7	Charles E. Campbell
The Vision Council.....	Jeff Endres Steve Drake (Alt.) Neil Roche (Alt.) Dick Whitney (Alt.) Ken Wood (Alt.)

The Subcommittee for Contact Lenses, which developed this standard, is composed of two groups having areas of interest as follows:

- Contact lens parameters; Methods of measuring contact lens parameters;
Physical and chemical properties of contact lens materials; Methods of
measurement of physical and chemical properties.
- Contact lens care products.

The working group that coordinated the development of this revision and harmonization with ISO standards in progress was headed by Doug Fortunato.

The Subcommittee had the following members:

Quido A. Cappelli, Chairman
William Benjamin
Glen Davies
Doug Fortunato
Steve Galas
David Meadows
Mary Mowrey McKee
Ralph Stone

American National Standard for Ophthalmics –

Contact Lenses – Standard Terminology, Tolerances, Measurements and Physicochemical Properties

1 Scope and Purpose

1.1 Scope

This American National Standard applies to contact lenses worn over the front surface of the eye in contact with the precorneal tear film. The standard covers rigid intracorneal and haptic (scleral) contact lenses, as well as soft paralingual contact lenses. Table 1 provides a high-level list of materials used for both rigid and soft contact lenses.

Table 1 – Contact Lens Materials

Type	Subtype	Material
Rigid	Traditional	PMMA (polymethyl methacrylate), CAB
	Non-hydrogel	Group 1 Materials not containing silicone or fluorine
		Group 2 Materials containing silicone but not fluorine
		Group 3 Materials containing silicone and fluorine
		Group 4 Materials containing fluorine but not silicone
Soft	Hydrogel	Group 1 (low water (< 50%), non-ionic)
		Group 2 (high water (>= 50%), non-ionic)
		Group 3 (low water, (< 50%), ionic)
		Group 4 (high water, (>= 50%), ionic)
	Enhanced oxygen permeable materials (e.g., Silicone Hydrogels)	Group 5 Materials having a Dk greater than 40 Dk units (using mmHg) and having a Dk greater than that expected on the basis of the materials' water content alone.