

Reaffirmed by ANSI
March 21, 2006

Reaffirmed by ANSI
April 22, 2011

Reaffirmed by ANSI
March 31, 2016

ANSI S1.42-2001

AMERICAN NATIONAL STANDARD
**DESIGN RESPONSE OF WEIGHTING
NETWORKS
FOR ACOUSTICAL MEASUREMENTS**


ANSI S1.42-2001

Accredited Standards Committee S1, Acoustics

Standards Secretariat
Acoustical Society of America
35 Pinelawn Road, Suite 114E
Melville, New York 11747-3177

The American National Standards Institute, Inc. (ANSI) is the national coordinator of voluntary standards development and the clearinghouse in the U.S. for information on national and international standards.

The Acoustical Society of America (ASA) is an organization of scientists and engineers formed in 1929 to increase and diffuse the knowledge of acoustics and to promote its practical applications.



ANSI S1.42-2001

American National Standard

**Design Response of Weighting Networks
for Acoustical Measurements**

Secretariat

Acoustical Society of America

Approved 22 October 2001

American National Standards Institute, Inc.

Abstract

This Standard provides the design criteria for both the frequency-domain response (amplitude and phase) and time-domain of the A- and C-weighting networks used in acoustically related measurements. The poles and zeros for each weighting network are given, along with equations for computing the amplitude and phase responses as functions of frequency and impulse and step responses as functions of time. Other known weighting networks that had been standardized, such as the B-, D- and E-weightings, or weightings that were published in the past, are listed in the Annexes for reference.

AMERICAN NATIONAL STANDARDS IN ACOUSTICS

The Acoustical Society of America (ASA) provides the Secretariat for Accredited Standards Committees S1 on Acoustics, S2 on Mechanical Vibration and Shock, S3 on Bioacoustics, and S12 on Noise. These committees have wide representation from the technical community (manufacturers, consumers, and general-interest representatives). The standards are published by the Acoustical Society of America through the American Institute of Physics as American National Standards after approval by their respective standards committees and the American National Standards Institute.


These standards are developed and published as a public service to provide standards useful to the public, industry, and consumers, and to Federal, State, and local governments.

Each of the Accredited Standards Committees, operating in accordance with procedures approved by American National Standards Institute (ANSI), is responsible for developing, voting upon, and maintaining or revising its own standards. The ASA Standards Secretariat administers committee organization and activity, and provides liaison between the Accredited Standards Committees and ANSI. After the standards have been produced and adopted by the Accredited Standards Committees, and approved as American National Standards by ANSI, the ASA Standards Secretariat arranges for their publication and distribution.

An American National Standard implies a consensus of those substantially concerned with its scope and provisions. Consensus is established when, in the judgment of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made toward their resolution.

The use of American National Standards is completely voluntary. Their existence does not in any respect preclude anyone, whether he has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

CAUTION NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw a standard.



Standards Secretariat
Acoustical Society of America
35 Pinelawn Rd. Suite 114E
Melville, New York 11747
USA

Telephone: +1 631 390-0215
Telefax: +1 631 390-0217
E-mail: asastds@aip.org
Internet: <http://asa.aip.org>

Contents

	Page
Foreword	iii
1 Scope	1
2 Normative references	1
2.1 American National Standards	1
2.2 International Standards	1
3 Frequency-weighting characteristics	1
3.1 Frequency domain	1
4 Time-domain characteristics	4
4.1 Time domain	4

Annexes

A Other weighting networks	6
B G-weighting network for infrasound measurements	11
B.1 G-weighting network	11
C U-weighting network for measurement of audible sound in the presence of ultrasound	13
C.1 U-weighting network	13
D Bibliography	14

Figures

1 Relative amplitude response level of the A- and C-weighting networks to a steady-state sinusoid.	3
2 Phase response of the A- and C-weighting network to a steady-state sinusoid.	4
3 Time-domain responses of the A- and C-weighting networks to an impulse (delta function).	5
A1 Relative amplitude response level of the B-, D- and E-weighting networks to a steady-state sinusoid.	6
A2 Phase response of the B-, D- and E-weighting networks to a steady-state sinusoid.	8
B1 Relative amplitude response level of the G-weighting network to a steady-state sinusoid.	12
C1 Relative amplitude response level of the U-weighting networks to a steady-state sinusoid	14

Tables

1 Poles, zeros, and normalization constants of the transfer functions for the A-, and C-weighting networks	1
2 Relative amplitude response level (re: response at 1000 Hz) and phase response of the A-, and C-weighting networks to a steady-state sinusoid. Band numbers and nominal (or preferred) frequencies are as specified in ANSI S1.6-1984. Exact frequencies, shown to four significant figures, are given by $10^{0.1N}$, where N is the integer band number	2
3 Coefficients for the impulse responses of the A-, and C-weighting networks	4

	Page
A1 Relative amplitude response level (re: response at 1000 Hz) and phase response of the B-weighting networks to a steady-state sinusoid. Band Numbers and nominal (or preferred) frequencies are as specified in ANSI S1.6-1984. Exact frequencies, shown to four significant figures, are given by $10^{0.1N}$, where N is the integer band number.	7
A2 Poles, zeros, and normalization constants of the transfer functions for the B-, D- and E-weighting networks.	8
A3 Relative amplitude response level (re: response at 1000 Hz) and phase response of the D- and E-weighting networks to a steady-state sinusoid. Band numbers and nominal (or preferred) frequencies are as specified in ANSI S1.6-1984. Exact frequencies, shown to four significant figures, are given by $10^{0.1N}$, where N is the integer band number.	9
A4 Coefficients for the impulse responses of the D- and E-weighting networks.	10
B1 Nominal coordinates of the poles and zeros of the transfer function, in the complex frequency plane	11
B2 Poles, zeros, and normalization constants of the transfer functions for the G-weighting.	11
C1 Pole locations for U-weighting.	13
C2 U-weighting network relative frequency response level.	13

Foreword

[This foreword is for information only and is not an integral part of ANSI S1.42-2001 *American National Standard Design Response of Weighting Networks for Acoustical Measurements*.]

Weighting networks have been widely used for many years to provide standardized means to select preferred information from acoustical signals. For example, the A-weighting has the property to enhance information over the human hearing range. Where as, the C-weighting is designed to encompass information from a wider spectrum. There is no concrete data to prove that the weighting networks are perfectly in agreement with the intent, such as the human hearing range. However, over the years, the design goals of the previously standardized weighting networks have proven to be acceptable to a majority of users, and the design goals have not been changed. Therefore it is possible to compare measured data from a vast data pool that has been accumulated over a period of more than 40 years.

For acoustical measurements, the weighting networks and the measuring instruments play important roles in arriving at correct answers. In this standard, ANSI S1.42-2001, only the design goals of the weighting networks are given. The tolerances in implementation of the weighting networks are subjected to specifications given in instrument standards. It must be pointed out that in view of the specification tolerances of weighting networks in acoustical instruments, it is possible to have instruments of the same class specification to give different measured readings for the same acoustical signal input. This is particularly true in sound level meters with relatively wide weighting network tolerances at the high and the low frequency range.

For the sake of completeness, information on other weighting less popular seldom used networks such as B-, D- and E-weightings are listed in the Annexes.

This standard was developed under the jurisdiction of Accredited Standards Committee S1, Acoustics, which has the following scope:

Standards, specifications, methods of measurement and test, and terminology, in the fields of physical acoustics, including architectural acoustics, electroacoustics, sonics and ultrasonics, and underwater sound, but excluding those aspects which pertain to safety, human tolerance and comfort.

At the time this standard was submitted to Accredited Standards Committee S1, Acoustics, for final approval, the membership was as follows:

G. S. K. Wong, *Chair*

T. J. Kuemmel, *Vice Chair*

S. B. Blaeser, *Secretary*

Acoustical Society of America G. S. K. Wong
T. J. Kuemmel (*Alt.*)

Air Conditioning and Refrigeration Institute R. Comparin
M. Darbeau (*Alt.*)

American Industrial Hygiene Association	D. Driscoll L.H. Royster (<i>Alt.</i>)
Audio Engineering Society	D. Queen M.R. Chial (<i>Alt.</i>)
Bruel & Kjaer Instruments	M. Alexander J. Chou (<i>Alt.</i>)
Information Technology Industry Council (ITI)	R.D. Hellweg W.H. Johnson (<i>Alt.</i>)
Larson-Davis, Inc.	L. Davis G. Bock (<i>Alt.</i>)
Lucent Technologies	D. Quinlan R. Ruhala (<i>Alt.</i>)
National Council of Acoustical Consultants	G. E. Winzer
National Institute of Standards and Technology	V. Nedzelnitsky D. J. Evans (<i>Alt.</i>)
U.S. Army Human Research & Engineering Directorate	J. Kalb T.R. Letowski (<i>Alt.</i>)
U.S. Department of the Air Force	R. L. McKinley

Individual Experts of Accredited Standards Committee S1, Acoustics, were:

J. R. Bareham	D. L. Johnson	L.W. Sepmeyer
S .L. Ehrlich	T. J. Kuemmel	H. E. von Gierke
K. M. Eldred	W.W. Lang	G.S.K. Wong
W.J. Galloway	A.H. Marsh	R.W. Young
	P.D. Schomer	

Working Group S1-24, Design Response of Weighting Networks for Acoustical Measurements, which assisted Accredited Standards Committee S1, Acoustics, in the preparation of this Standard had the following membership:

G. S. K. Wong, *Chair*

D. J. Evans D. L. Johnson J. P. Seiler L. Wu

Suggestions for improvement will be welcomed. Send suggestions for improvement to Accredited Standards Committee S1, Acoustics, in care of the ASA Standards Secretariat, 35 Pinelawn Road, Suite 114 E, Melville, New York 11747-3177 USA.

Telephone: +1 631 390 0215
Fax: +1 631 390 0217
E-mail: asastds@aip.org

American National Standard

Design Response of Weighting Networks for Acoustical Measurements

1 Scope

The aim is to examine the design goal specifications of weighting networks for acoustical measurements. The scope is restricted to the design, or target, responses of weighting networks. Tolerances that belong in the instrument performance specifications are not included.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this Standard. At the time of publication, the editions indicated were valid. All normative documents are subject to revision, and parties to agreements based on this Standard are encouraged to investigate the possibility of applying the most recent editions of the normative documents listed below.

2.1 American National Standards

ANSI S1.4-1983 (R 2001) American National Standard Specification for Sound Level Meters.

ANSI S1.4A-1985 (R 2001) Amendment to ANSI S1.4-1983.

ANSI S1.6-1984 (R 2001) American National Standard Preferred Frequencies, Frequency Levels, and Band Numbers for Acoustical Measurements

ANSI S1.11-1986 (R 1998) American National Standard Specification for Octave-Band and Fractional-Octave-Band Analog and Digital Filters.

2.2 International Standards

IEC 60651:1979 Sound level meters, Amendment No. 1 (1993)

IEC 60804:1985 Integrating-averaging sound level meters.

Amendment No. 1 (1989)

Amendment No. 2 (1993)

IEC 61012:1990 Filters for the measurement of audible sound in the presence of ultrasound

IEC 61260:1995-08 Electroacoustics-Octave-band and fractional-octave-band filters.

IEC 61261:1976 Electro-acoustical measuring equipment for aircraft noise certification

ISO 7196:1995 Acoustics-Frequency-weighting characteristics for infrasound measurements

3 Frequency-weighting characteristics

3.1 Frequency domain

3.1.1 In the frequency domain, the complex transfer function $H(jf)$ of a filter driven by a continuous sinusoidal waveform of frequency f can be expressed as

Table 1 — Poles, zeros, and normalization constants of the transfer functions for the A-, and C-weighting networks.

		A-weighting		C-weighting	
Zeros	m	q_m	$\gamma_m + j\delta_m$	q_m	$\gamma_m + j\delta_m$
	1	4	$0 + j0$	2	$0 + j0$
Poles	n	r_n	$a_n + j\beta_n$	r_n	$a_n + j\beta_n$
	1	2	$-20.598997 + j0$	2	$-20.598997 + j0$
	2	1	$-107.65265 + j0$	2	$-12\,194.217 + j0$
	3	1	$-737.86223 + j0$	---	---
	4	2	$-12\,194.217 + j0$	---	---
K			3.5041384×10^{16}		2.2428792×10^{16}