

July 29, 1997

Reaffirmed by ANSI
October 29, 2002

Reaffirmed by ANSI
October 19, 2007

**ANSI S12.12-1992
(ASA 104-1992)**

Reaffirmed by ANSI
September 5, 2012

Reaffirmed by ANSI
September 29, 2017

**AMERICAN NATIONAL STANDARD
Engineering Method for the Determination of Sound
Power Levels of Noise Sources Using Sound Intensity**

Reaffirmed by ANSI
June 19, 2020

**ACCREDITED STANDARDS COMMITTEE S12,
NOISE**

ABSTRACT

This standard describes a method for *in situ* determination of the sound power level of noise sources in indoor or outdoor environments using sound intensity measurements. The standard contains information on instrumentation, installation and operation of the source, procedures for the selection of a measurement surface, methods for the sampling of sound intensity on the measurement surface, procedures for the calculation of sound power level, and techniques that can be used to qualify the measurement environment.

AMERICAN NATIONAL STANDARDS ON ACOUSTICS

The Acoustical Society of America provides the Secretariat for Accredited Standards Committees S1 on Acoustics, S2 on Mechanical Shock and Vibration, S3 on Bioacoustics, and S12 on Noise. These committees have wide representation from the technical community (manufacturers, consumers, and general-interest representatives). The standards are published by the Acoustical Society of America through the American Institute of Physics as American National Standards after approval by their respective standards committees and the American National Standards Institute.

These standards are developed and published as a public service to provide standards useful to the public, industry, and consumers, and to Federal, State, and local governments.


This standard was approved by the American National Standards Institute as ANSI S12.12-1992 on 27 January 1992.

An American National Standard implies a consensus of those substantially concerned with its scope and provisions. An American National Standard is intended as a guide to aid the manufacturer, the consumer, and the general public. The existence of an American National Standard does not in any respect preclude anyone, whether he has approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. American National Standards are subject to periodic review and users are cautioned to obtain the latest editions.

Caution Notice: An American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken to reaffirm, revise, or withdraw this standard no later than five years from the date of publication.

The American National Standards Institute, Inc. (ANSI) is the national coordinator of voluntary standards development and the clearing house in the U.S. for information on national and international standards.

The Acoustical Society of America (ASA) is an organization of scientists and engineers formed in 1929 to increase and diffuse the knowledge of acoustics and to promote its practical applications.



Published by the Acoustical Society of America through the American Institute of Physics

© 1992 by Acoustical Society of America. This standard may not be reproduced in whole or in part in any form for sale, promotion, or any commercial purpose, or any purpose not falling within the fair-use provisions of the Copyright Act of 1976, without prior written permission of the publisher. For permission, address the Standards Secretariat of the Acoustical Society of America.

FOREWORD

[This Foreword is not a part of the American National Standard Engineering Method for the Determination of Sound Power Levels of Noise Sources Using Sound Intensity, ANSI S12.12-1992 (ASA Catalog No. 104-1992).]

This American National Standard contains procedures for the determination of the sound power of noise sources from sound intensity measurements. Previous sound power determinations required special facilities such as anechoic, hemi-anechoic, or reverberation rooms and the estimation of the sound power of sources from sound pressure measurements which resulted in various errors. Using this new standard, the sound power of noise sources can be directly determined from sound intensity measurements made *in situ* even when intrusive background noise is present in the environment. The approaches in this standard are useful also for the relative ranking of machinery noise sources in terms of sound power.

This standard was developed using the American National Standards Institute (ANSI) Accredited Standards Committee Procedure under the Secretariat of the Acoustical Society of America.

Accredited Standards Committee S12, Noise, under whose jurisdiction this standard was developed, has the following scope:

Standards, specifications, and terminology in the field of acoustical noise pertaining to methods of measurement, evaluation, and control, including biological safety, tolerance, and comfort, and physical acoustics as related to environmental and occupational noise.

At the time this standard was submitted to Accredited Standards Committee S12, Noise, for approval, the membership was as follows:

W. Melnick, *Chairman*
R. Hillquist, *Vice-Chairman*
A. Brenig, *Secretary*

Acoustical Society of America ● W. Melnick, R. Hillquist (*Alt*)
Air-Conditioning and Refrigeration Institute ● R. Harold, H. C. Skarbek (*Alt*)
Aluminum Company of America (ALCOA) ● S. I. Roth
American Academy of Otolaryngology, Head and Neck Surgery, Inc. ● R. F. Naunton, L. A. Michael (*Alt*)
American Industrial Hygiene Association ● C. D. Bohl
American Occupational Medical Association ● T. R. Hatfield, J. Sataloff (*Alt*)
American Otological Society, Inc. ● J. Tonndorf
American Society of Heating, Refrigerating, and Air-Conditioning Engineers (ASHRAE) ● H. S. Pei, G. F. Hohn (*Alt*)
American Speech-Language-Hearing Association ● S. J. Barry, V. W. Byers (*Alt*)
American Welding Society ● vacant, M. E. Kennebeck (*Alt*)
Audio Engineering Society, Inc. ● M. R. Chial
Bruel & Kjaer Instruments, Inc. ● G. C. Michel (*Alt*)
Compressed Air and Gas Institute (CAGI) ● J. H. Addington
Computer and Business Equipment Manufacturers Association ● L. F. Luttrell
Electric Light and Power Group ● C. E. Hickman, M. C. Mingoia (*Alt*)
Fastener Industry Noise Control Research Program ● E. H. Toothman, J. C. McMurray (*Alt*)

Federal Aviation Administration ● R. Tedrick
Larson-Davis Laboratories ● D. Johnson, L. Davis (*Alt*)
Motor Vehicle Manufacturers Association ● D. B. Smith, T. J. Slavin (*Alt*)
National Institute of Standards and Technology (NIST) ● S. L. Yaniv, D. R. Flynn (*Alt*)
National Council of Acoustical Consultants ● G. C. Tocci, A. P. Nash (*Alt*)
National Electrical Manufacturers Association ● W. L. Billington
Naval Ship Research and Development ● vacant
Power Tool Institute, Inc. ● J. L. Bennett
Scantek, Inc. ● R. J. Peppin
U. S. Army Aeromedical Research Laboratory ● B. Mozo, J. H. Patterson (*Alt*)
U. S. Army Human Engineering Laboratory ● G. R. Price, J. Kalb (*Alt*)
U. S. Army Medical Corps ● R. M. Atack
U. S. Department of the Air Force ● H. E. von Gierke
U. S. Department of the Army, Environmental Office ● P. D. Schomer, R. Raspet (*Alt*)
U. S. Department of the Navy, Naval Medical Command ● J. Page

Individual experts of Accredited Standards Committee S12, Noise, were:

P. K. Baade	S. Feldman	W. W. Lang
R. G. Bartheld	R. S. Gales	G. C. Maling, Jr.
L. Batchelder	W. J. Galloway	A. H. Marsh
R. W. Benson	E. E. Gross, Jr.	W. Melnick
R. D. Bruce	R. Guernsey	J. Wesler
K. M. Eldred	R. K. Hillquist	G. Winzer
		R. W. Young

Working Group S12/WG21, Determination of Sound Power Using Sound Intensity Measurements, which assisted Accredited Standards Committee S12, Noise, in the development of this standard, had the following membership:

M. J. Crocker, *Chairman*

D. Bernhard	G. Krishnappa	J. Pope
J. Buffa	M. A. Lang	T. Reinhart
R. L. Cohen	G. P. Mathur	J. K. Thompson
P. R. Donavan	J. Nicolas	J. Tichy
B. Forssen	D. Yeager	L. W. Tweed
R. Hickling	A. Mielnicka Pate	Y. S. Wang
C. Holsclaw	R. J. Peppin	H. A. Wolf

Suggestions for improvement of this standard will be welcomed. They should be sent to **Accredited Standards Committee S12, Noise, in care of the ASA Standards Secretariat, 335 East 45th Street, New York, NY 10017-3483. Telephone (212) 661-9404.**

CONTENTS

1 INTRODUCTION	1
2 SCOPE	1
2.1 General.....	1
2.2 Field of Application	1
2.3 Measurement Uncertainty.....	1
3 DEFINITIONS	1
4 TEST ENVIRONMENT	4
4.1 General.....	4
4.2 Guidelines on the Test Environment.....	4
4.2.1 Physical site.....	4
4.2.2 Atmospheric conditions	4
4.2.3 Background noise	4
5 INSTRUMENTATION	4
5.1 General.....	4
5.2 Sound Pressure Microphone and Associated Cables.....	5
5.3 Sound Intensity Probe and Associated Cables	5
5.4 Frequency Response of Instrumentation System	5
5.5 Frequency Analyzer and Weighting Network	5
5.6 Transducer and Instrumentation System Calibration	6
5.7 Performance Verification	6
5.8 Field Check	6
6 INSTALLATION AND OPERATION OF SOURCE	6
6.1 General.....	6
6.2 Installation of Source	7
6.3 Auxiliary Equipment.....	7
6.4 Operation of Source During Test.....	7
7 DETERMINATION OF SOUND INTENSITY AND SOUND PRESSURE LEVELS ON MEASUREMENT SURFACE	7
7.1 General.....	7
7.1.1 Measurement Surface	7
7.1.2 Shape of Measurement Surface.....	7
7.1.3 Size of Measurement Surface.....	8
7.1.4 Placement and Orientation of Measurement Surface.....	8
7.2 Sampling of Sound on Measurement Surface	8
7.2.1 Measurement at Fixed Points	8
7.2.2 Scanning of Surface Segments.....	8
7.2.2.1 General.....	9
7.2.2.2 Scanning Procedures.....	9
7.3 Reduction of Number of Measurement Locations	9

7.4	Conditions of Measurements.....	10
7.4.1	General.....	10
7.4.2	Measurements with Exponential Averaging or Time-Period Averaging Systems.....	10
8	CALCULATION OF RADIATION QUANTITIES.....	10
8.1	Sound Intensity and Intensity Level Averaged Over Measurement Surface	11
8.2	Sound Pressure Level Averaged Over Measurement Surface	11
8.3	Data Quality Indicators.....	11
8.4	Sound Power Level.....	11
9	INFORMATION TO BE RECORDED.....	11
9.1	Sound Source Under Test	11
9.2	Test Environment.....	12
9.3	Instrumentation.....	12
9.4	Measurement Technique.....	12
9.5	Acoustical Data	12
10	INFORMATION TO BE REPORTED	12
APPENDIX A: COMPUTATIONAL PROCEDURES FOR CALCULATING A-WEIGHTED SOUND POWER LEVEL FROM OCTAVE OR ONE-THIRD OCTAVE BAND SOUND POWER LEVELS		13
APPENDIX B: DATA QUALITY INDICATORS		13
B.1	Source and Background Level Variability Indicators, <i>D1</i>	14
B.2	Parasitic Noise Level Indicators, <i>D2</i>	14
B.3	Surface Averaging Accuracy Indicators, <i>D3</i>	16
B.4	Sound Power Accuracy Indicators, <i>D4</i>	16
B.5	Phase Indicators, <i>D5</i>	16
B.6	Averaging Time Sufficiency Indicators, <i>D6</i>	17
B.7	Measurement Interference Indicators, <i>D7</i>	17
BIBLIOGRAPHY		18
FIGURE		
FIG. 1	2
TABLES		
TABLE 1	1
TABLE 2	5
TABLE 3	6
TABLE 4	6
TABLE 5	8
TABLE 6	9
TABLE 7	10

American National Standard Engineering Method for the Determination of Sound Power Levels of Noise Sources Using Sound Intensity

1 INTRODUCTION

In previous engineering standards on sound power determination (S12.31–S12.36), sound intensity (the primary quantity needed to determine sound power) is estimated by sound pressure measurement, and therefore is subject to several sources of errors and uncertainties. These uncertainties include errors due to measurements in the near field, test environment shortcomings, and errors caused by the lack of a vector measurement. Previous standards also required the use of special facilities such as anechoic, semi-anechoic, or reverberation rooms. By using this standard, the sound power can be determined in the usual operating or test environment of the sound source by measuring sound intensity directly.

2 SCOPE

2.1 General

This standard gives engineering methods for determining *in situ* sound power using sound intensity measurements on a measurement surface enclosing a source. It provides guidelines on the acoustical environment, including ambient noise, the measurement surface enclosing the source, and the number of measurements on the surface.

This standard does not include specification of instruments or calibration procedures. It is not limited to any one technique of intensity measurement.

2.2 Field of Application

This standard is intended for use primarily to determine the sound power radiated by sources *in situ* for the purpose of noise control, rating, and compliance with purchase specifications. There are no explicit restrictions on the size or shape of the source. The standard applies to any source emitting essentially steady noise during test, except those sources for which the measurement surface cannot be fixed in space. This, in most cases, should enable the sound power level of a noise source to be determined in its usual operating environment when the source of interest is a prominent source.

2.3 Measurement Uncertainty

Sound power determinations made in conformity with this standard should result in standard deviations which are equal to or less than those given in Table 1. This table includes uncertainty in the sound intensity measurement technique due to the test environment, background noise levels, and selection of the number of measurement points.

TABLE 1. Uncertainty in determining sound power levels for engineering measurements *in situ*.

Octave band center frequencies (Hz)	One-third octave band center frequencies (Hz)	Standard deviation (dB)
125	100 to 160	3.0
250 to 500	200 to 630	2.0
1000 to 4000	800 to 5000	1.5
8000	6300 to 10 000	2.5

NOTE: The standard deviations given in Table 1 reflect the cumulative effects of all causes of measurement uncertainty, excluding variations in the sound power level from machine to machine or from test to test which may be caused, for example, by changes in the mounting or operating conditions of the source. The reproducibility and repeatability of the test results may be considerably better (i.e., smaller standard deviations) than the uncertainties given in Table 1 would indicate.

3 DEFINITIONS

For the purpose of this standard, the following definitions apply:

measurement surface (see Fig. 1): an arbitrary boundary or hypothetical surface of area S chosen so that it encloses the source surface, and on which measurements are made. If present, reflective surfaces, e.g., a supporting or ground plane, usually terminate the measurement surface. Such reflective surfaces are not a part of the measurement surface.