

Reaffirmed by ANSI
July 10, 2001

Reaffirmed by ANSI
May 24, 2005

Reaffirmed by ANSI
April 22, 2010

Reaffirmed by ANSI
August 27, 2015

Reaffirmed by ANSI June
19, 2020

**ANSI S2.61-1989
(ASA 78-1989)**

AMERICAN NATIONAL STANDARD Guide to the Mechanical Mounting of Accelerometers

Accredited Standards Committee S2, Mechanical Shock and Vibration

ABSTRACT

This standard specifies methods for mounting contact accelerometers and delineates the limitations of the recommended methods such as frequency range of interest, amplitude, and phase measurement accuracy. Characteristics of the mounting arrangements used by specific accelerometers which should be specified by the manufacturer are established and guidance is provided to the user to optimize performance of a recommended mounting method.

AMERICAN NATIONAL STANDARDS ON ACOUSTICS

The Acoustical Society of America provides the Secretariat for Accredited Standards Committees S1 on Acoustics, S2 on Mechanical Shock and Vibration, S3 on Bioacoustics, and S12 on Noise. These committees have wide representation from the technical community (manufacturers, consumers, and general-interest representatives). The standards are published by the Acoustical Society of America through the American Institute of Physics as American National Standards after approval by their respective standards committees and the American National Standards Institute.

These standards are developed and published as a public service to provide standards useful to the public, industry, and consumers, and to Federal, State, and local governments.


This standard was approved by the American National Standards Institute as ANSI S2.61-1989 on 7 February 1989.

An American National Standard implies a consensus of those substantially concerned with its scope and provisions. An American National Standard is intended as a guide to aid the manufacturer, the consumer, and the general public. The existence of an American National Standard does not in any respect preclude anyone, whether he has approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. American National Standards are subject to periodic review and users are cautioned to obtain the latest editions.

Caution Notice: An American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken to reaffirm, revise, or withdraw this standard no later than five years from the date of publication.

The American National Standards Institute, Inc. (ANSI) is the national coordinator of voluntary standards development and the clearing house in the U.S. for information on national and international standards.

The Acoustical Society of America (ASA) is an organization of scientists and engineers formed in 1929 to increase and diffuse the knowledge of acoustics and to promote its practical applications.



Published by the Acoustical Society of America through the American Institute of Physics.

Copyright © 1989 by the Acoustical Society of America. No portion of this publication may be quoted or reproduced in any form without permission of the Acoustical Society of America.

Any request to reproduce this standard in whole or in part should be addressed to the Standards Secretariat, in care of the Acoustical Society of America, 335 East 45th Street, New York, New York 10017-3483

FOREWORD

[This Foreword is not a part of American National Standard Guide to the Mechanical Mounting of Accelerometers, ANSI S2.61-1989 (ASA Catalog No. 78-1989).]

This standard is the counterpart of ISO 5348-1987 Mechanical Vibration and Shock—Mechanical Mounting of Accelerometers, developed by ISO/TC 108/SC3 Use and Calibration of Vibration and Shock Measuring Instruments, under the International Organization for Standardization (ISO).

This standard was developed under the jurisdiction of Accredited Standards Committee S2, Mechanical Shock and Vibration, using the American National Standards Institute (ANSI) Accredited Standards Committee Procedure. The Acoustical Society of America provides the Secretariat for Accredited Standards Committee S2, Mechanical Shock and Vibration.

Accredited Standards Committee S2, Mechanical Shock and Vibration, under whose jurisdiction this standard was developed, has the following scope:

Standards, specifications, methods of measurement and test, and terminology, in the fields of mechanical shock and vibration, but excluding those aspects which pertain to biological safety, tolerance, and comfort.

At the time this standard was submitted to Accredited Standards Committee S2, Mechanical Shock and Vibration, for final approval, the membership was as follows:

S. I. Hayek, *Chairman*
M. R. Serbyn, *Vice-Chairman*
A. Brenig, *Secretary*

Acoustical Society of America ● S. I. Hayek, M. R. Serbyn (*Alt*)
Institute of Environmental Sciences ● R. Brown, J. Ehmann (*Alt*)
National Bureau of Standards ● M. R. Serbyn, D. C. Robinson (*Alt*)
National Electrical Manufacturers Association ● D. V. Wright, E. LaBrush (*Alt*)
Naval Ship Research & Development Center ● A. Kilcullen
Schenck Trebel Corporation ● D. G. Stadelbauer
U.S. Department of the Air Force ● O. Maurer, J. P. Henderson (*Alt*)
U.S. Department of the Navy, Naval Sea Systems Command ● A. R. Paladino, D. D. Bernhard (*Alt*)

Individual experts of Accredited Standards Committee S2, Mechanical Shock and Vibration, were:

P. K. Baade
R. G. Bartheld
L. Batchelder
G. Booth
J. Catlin
K. M. Eldred

R. Eshleman
S. Feldman
D. Johnson
P. H. Maedel, Jr.
D. Muster
H. L. Rich

A. O. Sykes
H. E. von Gierke
D. Walker
S. P. Ying

FOREWORD

Working Group S2-81, Use and Calibration of Vibration and Shock Measuring Instruments, which assisted the Accredited Standards Committee, S2, Mechanical Shock and Vibration, in the development of this standard, had the following membership:

B. E. Douglas, *Chairman*

D. E. Bently
R. R. Bouche
W. F. Caplan
J. B. Catlin

E. Edwards
R. L. Eshleman
J. Heist
J. E. Judd

G. F. Lang
M. R. Serbyn

Suggestions for improvements in this standard will be welcomed. They should be sent to Accredited Standards Committee S2 at the Standards Secretariat, in care of the Acoustical Society of America, 335 East 45th Street, New York, NY 10017-3483.

CONTENTS

0 INTRODUCTION	1
1 SCOPE AND FIELD OF APPLICATION	1
2 REFERENCES	1
3 DEFINITIONS	1
4 SELECTION OF A MOUNTING METHOD	1
4.1 General Considerations.....	1
4.2 Areas of Application.....	1
5 MOUNTING METHODS	2
5.1 General.....	2
5.2 Machine Screw or Stud Mounting.....	2
5.3 Cementing.....	3
5.4 Mounting Fixtures.....	3
5.5 Miscellaneous Accelerometer Mounting Methods.....	4
6 CABLES	4
7 MOUNTED RESONANCE FREQUENCY	4
7.1 Frequency Range of Application.....	4
7.2 Vibration Excitation Method.....	6
7.3 Shock Excitation Method.....	7
8 CHARACTERISTICS TO BE SPECIFIED BY MANUFACTURERS OF ACCELEROMETERS	8
APPENDIX: THE DYNAMICS OF MOUNTING ACCELEROMETERS	8
FIGURES	
FIG. 1 Frequency response to a representative stud mounted accelerometer relative to the absolute acceleration of the structure at its attachment.....	2
FIG. 2 Frequency response of a representative cement (methyl cyanoacrylate) mounted accelerometer relative to the absolute acceleration of the structure at its attachment.....	3
FIG. 3 Frequency response of a magnetically mounted accelerometer relative to the absolute acceleration of the structure at its attachment.....	4
FIG. 4 Frequency response of an accelerometer mounted by double-sided adhesive tape relative to the absolute acceleration of the structure at its attachment.....	5
FIG. 5 Frequency response of an accelerometer mounted with beeswax relative to the absolute acceleration of the structure at its attachment.....	5
FIG. 6 Accelerometers with axial or radial connector.....	6
FIG. 7 Accelerometer test arrangement.....	7
FIG. 8 Accelerometer response to shock.....	8

American National Standard

Guide to the Mechanical Mounting of Accelerometers

0 INTRODUCTION

Vibration measurement transducers fall into two broad classes: contacting and noncontacting. Noncontacting structural response transducers are placed in close proximity to the structure and include such generic types as eddy current probes and optical proximity probes. Contacting transducers are placed in mechanical contact with the structural system and include such generic types as piezoelectric and piezoresistive accelerometers and seismic velocity transducers. This standard is concerned with contacting type accelerometers, which enjoy wide current popularity. A major concern with using such transducers is that the mechanical connection between the accelerometer and test structure will alter the response of the structure. This standard discusses methods of mounting accelerometers to a structure and parameters that control the accuracy of structural measurement using contact accelerometers.

1 SCOPE AND FIELD OF APPLICATION

This ANSI standard describes the mounting characteristics of accelerometers to be specified by the manufacturer and makes recommendations to the user for mounting accelerometers. The application of this standard is limited to the mounting of electromechanical transducers of the type that are attached on the surface of the structure in motion. It does not cover other types, such as relative motion pickups. This standard is in general accordance with ISO 5348-1987 Mechanical Vibration and Shock—Mechanical Mounting of Accelerometers.

2 REFERENCES

(1) International Organization for Standardization, Vibration and Shock Vocabulary, ISO 2041-1975.

(2) ISO 2954-1975 Mechanical Vibration of Rotating and Reciprocating Machinery—Requirements for Instruments for Measuring Vibration Severity.

(3) ISO 5347-1987 Methods of Calibration of Vibration and Shock Pickups.

(4) ISO/DP 7626/Part 1-1986 Methods for the Experimental Determination of Mechanical Mobility, Part 1, Basic Definitions and Transducers.

(5) B. E. Douglas, "Model Damping Determinations from Resonance Spectral Shape Measurements," J. Acoust. Soc. Am. Suppl. 1 58, S24 (1975).

(6) ISO/DP 7626/Part 2, Methods for the Experimental Determination of Mechanical Mobility, Part 2.

(7) ISO/DP 7626/Part 3, Methods for the Experimental Determination of Mechanical Mobility, Part 3.

3 DEFINITIONS

For the purpose of this ANSI standard, the terms used and definitions given in ISO 2041-1975 are applicable.

4 SELECTION OF A MOUNTING METHOD

4.1 General Considerations

A contact accelerometer will achieve optimal performance only under the following conditions:

- (a) The accelerometer follows as nearly as possible the motion of the structure under test at the accelerometer attachment.
- (b) The motion of the structure shall not be measurably affected by the addition of the accelerometer.
- (c) The effective (i.e., dynamic) mass of the accelerometer and mounting fixtures shall be small in comparison with the structure under test (see ISO 2954 and ISO/DP 7626/Part 1).

In order to achieve these conditions:

- (a) the accelerometer and its mounting shall be as rigid as practical (the mounting surfaces shall be as clean and flat as practical);
- (b) the mounting shall not introduce significant distorting motions of its own (note: This can best be achieved using simple symmetrical mountings); and
- (c) the mass of the accelerometer and mounting shall be small in comparison with the structure under test (see ISO 2954).

4.2 Areas of Application

The selection of a mounting method is strongly dependent on the requirements of the measurement (i.e., frequency range of interest, phase accuracy, measurement, environment, etc.). Screw or stud mounting and cementing methods are the preferred methods in that the mounted resonance frequency is high and associat-