

October 15, 2003

Reaffirmed by ANSI
October 28, 2008

Reaffirmed by ANSI
November 26, 2013

ANSI S3.1-1999
(Revision of ANSI S3.1-1991)

AMERICAN NATIONAL STANDARD

Maximum Permissible Ambient Noise Levels for Audiometric Test Rooms


ANSI S3.1-1999

Accredited Standards Committee S3, Bioacoustics

Standards Secretariat
Acoustical Society of America
35 Pinelawn Road, Suite 114 E
Melville, NY 11747-3177

The American National Standards Institute, Inc. (ANSI) is the national coordinator of voluntary standards development and the clearinghouse in the U.S.A. for information on national and international standards.

The Acoustical Society of America (ASA) is an organization of scientists and engineers formed in 1929 to increase and diffuse the knowledge of acoustics and to promote its practical applications.



ANSI S3.1-1999

(Revision of
ANSI S3.1-1991)

American National Standard

**Maximum Permissible Ambient
Noise Levels for Audiometric Test Rooms**

Secretariat

Acoustical Society of America

Approved 3 August 1999

American National Standards Institute, Inc.

Abstract

This Standard specifies maximum permissible ambient noise levels (MPANLs) allowed in an audiometric test room that produce negligible masking (≤ 2 dB) of test signals presented at reference equivalent threshold levels specified in ANSI S3.6-1996 *American National Standard Specification of Audiometers*. The MPANLs are specified from 125 to 8000 Hz in octave and one-third octave band intervals for two audiometric testing conditions (ears covered and ears not covered) and for three test frequency ranges (125 to 8000 Hz, 250 to 8000 Hz, and 500 to 8000 Hz). The Standard is intended for use by all persons testing hearing and for distributors, installers, designers, and manufacturers of audiometric test rooms. This standard is a revision of ANSI S3.1-1991 *American National Standard Maximum Permissible Ambient Noise Levels for Audiometric Test Rooms*.

AMERICAN NATIONAL STANDARDS ON ACOUSTICS

The Acoustical Society of America (ASA) provides the Secretariat for Accredited Standards Committees S1 on Acoustics, S2 on Mechanical Vibration and Shock, S3 on Bioacoustics, and S12 on Noise. These committees have wide representation from the technical community (manufacturers, consumers, and general-interest representatives). The standards are published by the Acoustical Society of America through the American Institute of Physics as American National Standards after approval by their respective standards committees and the American National Standards Institute.


These standards are developed and published as a public service to provide standards useful to the public, industry, and consumers, and to Federal, State, and local governments.

Each of the Accredited Standards Committees (operating in accordance with procedures approved by ANSI) is responsible for developing, voting upon, and maintaining or revising its own standards. The ASA Standards Secretariat administers committee organization and activity, and provides liaison between the Accredited Standards Committees and ANSI. After the Standards have been produced and adopted by the Accredited Standards Committees, and approved as American National Standards by ANSI, the ASA Standards Secretariat arranges for their publication and distribution.

An American National Standard implies a consensus of those substantially concerned with its scope and provisions. Consensus is established when, in the judgment of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered and that a concerted effort be made toward their resolution.

The use of American National Standards is completely voluntary. Their existence does not in any respect preclude anyone, whether he or she has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this Standard.



Standards Secretariat
Acoustical Society of America
120 Wall Street, 32nd Floor
New York, New York 10005-3993
USA

Telephone: +1 212 248 0373
Telefax: +1 212 248 0146
E-mail: asastds@aip.org
Internet: <http://asa.aip.org>

© 1999 by the Acoustical Society of America. This Standard may not be reproduced in whole or in part in any form for sale, promotion, or any commercial purpose, or any purpose not falling within the provisions of the Copyright Act of 1976, without prior written permission of the publisher. For permission, address a written request to the Standards Secretariat of the Acoustical Society of America.

Contents

	Page
Foreword.	ii
1 Scope, purpose, and applications	1
2 References	2
3 Definitions of terms	2
4 Recommended ambient noise levels and measurements	4
5 Compliance with Standard	7
Annexes	
A Derivation of ambient noise levels	8
B Present compared with previous (1991) ambient noise levels	11
C Interim low frequency ambient noise levels	12
D High frequency ambient noise levels	13
E Techniques for physical measurement	14
F Passive noise-reducing enclosures	15
G General considerations in the planning of an audiometric test room	15
Tables	
1 Ears Covered: Octave band MPANLs for supra-aural earphones and for insert earphones for three test frequency ranges; in dB re: 20 μ Pa to nearest 0.5 dB	4
2 Ears Covered: One-third octave band MPANLs for supra-aural earphones and for insert earphones for three test frequency ranges; in dB re: 20 μ Pa to the nearest 0.5 dB	5
3 Ears Not Covered: Octave and one-third octave band MPANLs for three test frequency ranges; in dB re: 20 μ Pa to the nearest 0.5 dB	5
A.1 Derivation of the MPANLs for ears not covered and ears covered using a supra-aural earphone and insert earphone for the test frequency range 125 to 8000 Hz as reported in tables 1–3. Values in parentheses are interpolated or extrapolated from existing data	9
A.2 Average standard deviations of the attenuation provided by a supra-aural and insert earphone. Values in parentheses are interpolated or extrapolated from existing data	10
B.1 Differences between the present and the 1991 octave band MPANLs for ears covered and ears not covered for each test frequency range; in dB re: 20 μ Pa to the nearest 0.5 dB	11
B.2 Differences between the present and the 1991 one-third octave band MPANLs for ears covered and ears not covered for each test frequency range; in dB re: 20 μ Pa to the nearest 0.5 dB	12
C.1 Ears Covered: Interim octave band MPANLs using a supra-aural and insert earphone for each test frequency range; in dB re: 20 μ Pa to the nearest 0.5 dB	13
C.2 Ears Covered: Interim one-third octave band MPANLs using a supra-aural and insert earphone for each test frequency range; in dB re: 20 μ Pa to the nearest 0.5 dB	13
C.3 Ears Not Covered: Interim octave and one-third octave band MPANLs for each test frequency range; in dB re: 20 μ Pa to the nearest 0.5 dB	13

Foreword

[This Foreword is for information only and is not an integral part of ANSI S3.1-1999 *American National Standard Maximum Permissible Ambient Noise Levels for Audiometric Test Rooms*.]

This Standard is a revision of ANSI S3.1-1991 *American National Standard Maximum Permissible Ambient Noise Levels for Audiometric Test Rooms*. The 1991 Standard specified maximum permissible ambient noise levels (MPANLs) from 125 to 8000 Hz in octave and one-third octave bands allowed in an audiometric test room that would produce negligible masking of pure tones presented at reference equivalent threshold sound pressure levels specified in ANSI S3.6-1989 for two audiometric conditions (ears covered using a supra-aural earphone and ears not covered) and for three test frequency ranges (125–8000 Hz, 250–8000 Hz, and 500–8000 Hz). The MPANLs were derived using reference equivalent threshold sound pressure levels measured in a NBS 9-A coupler.

This revision also specifies MPANLs from 125 to 8000 Hz in octave and one-third octave bands allowed in an audiometric test room that would produce negligible masking of pure tones for two audiometric conditions and three test frequency ranges. However, the MPANLs specified in this revision were derived using a different computational method based on sound field thresholds, principles of masking for a given threshold shift, power-law summation of masking, and earphone attenuation values. Since the new computational method does not use NBS 9-A coupler based measurements as done in 1991, MPANLs have been specified for an insert as well as for a supra-aural earphone. The present Standard allows slightly more low frequency and slightly less very high frequency ambient noise in an audiometric test room than specified in 1991.

The annexes detail the derivation of the MPANLs and provide information for interim low frequency MPANLs, high frequency ambient noise levels, techniques for physical measurement of ambient noise, supra-aural earphones encased in passive noise-reducing enclosures, and general considerations in planning for an audiometric test room.

Further experimental work concerning the influence of masking on hearing thresholds is encouraged, especially masking of low frequency noise on higher frequency hearing thresholds, very low and high frequency earphone attenuation, and the attenuation provided by passive noise-reducing earphone enclosures. As a result of this research, a more precise specification of permissible ambient noise levels may be developed.

This Standard, a revision of ANSI S3.1-1991 *American National Standard Maximum Permissible Ambient Noise Levels for Audiometric Test Rooms*, was developed under the jurisdiction of Accredited Standards Committee S3, Bioacoustics, which has the following scope:

Standards, specifications, methods of measurement and test, and terminology, in the fields of psychological and physiological acoustics, including aspects of general acoustics, shock and vibration which pertain to biological safety, tolerance, and comfort.

At the time this Standard was submitted to Accredited Standards Committee S3, Bioacoustics, for final approval, the membership was as follows:

R. F. Burkard, *Chair*
 J. Franks, *Vice Chair*
 A. Brenig, *Secretary*

Acoustical Society of America	R. F. Burkard J. Franks (<i>Alt.</i>)
American Academy of Otolaryngology, Head and Neck Surgery, Inc.	G. A. Gates L. A. Michael (<i>Alt.</i>)
American Industrial Hygiene Association.	L. H. Royster J. F. Meagher (<i>Alt.</i>)
American College of Occupational Medicine.	P. J. Brownson J. Sataloff (<i>Alt.</i>)
American Otological Society, Inc.	R. F. Naunton
American Speech-Hearing-Language Association	J. D. Royster M. E. Thompson (<i>Alt.</i>)
AT&T	R. M. Sachs M. S. Mayer (<i>Alt.</i>)
Audio Engineering Society	R. H. Campbell M. R. Chial (<i>Alt.</i>)
Hearing Industries Association (HIA)	S. Armstrong C. Rogin (<i>Alt.</i>)
Industrial Safety Equipment Association (ISEA)	D. K. Shipp
Industrial Hearing Society	P. Mercola
National Institute of Standards and Technology	E. D. Burnett V. Nedzelnitsky (<i>Alt.</i>)
Power Tool Institute, Inc.	G. Resigno M. Brown (<i>Alt.</i>)
U. S. Air Force	C. Nixon
U. S. Army Aeromedical Research Laboratory	B. Mozo
U. S. Army Human Engineering Laboratory	G. Garinther G. R. Price (<i>Alt.</i>)
U. S. Army Walter Reed Army Medical Center Audiology and Speech Center	R. W. Danielson
U. S. Navy Environmental Health Center	J. Page L. Marshall (<i>Alt.</i>)

Individual experts of Accredited Standards Committee S3, Bioacoustics, were:

J. R. Bareham	J. L. Fletcher	E. S. Thalheimer
S. J. Barry	R. S. Gales	H. E. von Gierke
R. W. Benson	W. J. Galloway	D. E. Wasserman
A. J. Brammer	D. L. Johnson	L. A. Wilber
K. M. Eldred	K. D. Kryter	W. A. Yost
L. S. Finegold	R. L. McKinley	R. W. Young
	J. D. Royster	

Working Group S3/WG56, Criteria for Background Noise for Audiometric Testing, which assisted Accredited Standards Committee S3, Bioacoustics, in the development of this Standard, had the following membership:

T. Frank, *Chair*

I. M. Blood
E. H. Berger
J. D. Durrant

J. M. Lovrinic
E. Singer

G. A. Studebaker
L. A. Wilber

Suggestions for improvements will be welcomed. Send suggestions for improvement in writing to Accredited Standards Committee S3, Bioacoustics, in care of the ASA Standards Secretariat, 120 Wall Street, 32nd floor, New York, New York 10005-3993, USA.

Telephone: +1 212 248 0373
Fax: +1 212 248 0146
E-mail: asastds@aip.org

American National Standard

Maximum Permissible Ambient Noise Levels for Audiometric Test Rooms

1 Scope, purpose, and applications

1.1 Scope

Hearing measurements in an audiometric test room may be conducted for different audiometric purposes over different test frequency ranges. However, if the ambient noise level is excessively high, some hearing threshold levels measured in that environment will be inaccurate. This occurs because excessively high ambient noise will create an elevation of hearing threshold levels. This is a psychoacoustic phenomenon called masking.

Exclusion of all ambient noise from an audiometric test room is not feasible nor practical from structural and cost considerations. However, to ensure that hearing tests are not influenced by excessive ambient noise masking, there is a need to specify maximum permissible ambient noise levels (MPANLs) that can be allowed in an audiometric test room. This is true particularly when testing is done at sound pressure levels for pure tones corresponding to 0 dB hearing level. Further, there is a need to specify MPANLs for different audiometric conditions, test frequency ranges, and earphone types that will permit testing at reference equivalent threshold levels specified in ANSI S3.6-1996 *American National Standard Specification for Audiometers*.

This Standard includes the specification and measurement of MPANLs in an audiometric test room. Within this Standard, MPANLs are specified for octave and one-third octave band intervals from 125 to 8000 Hz for the audiometric conditions of testing with ears covered using a supra-aural and insert earphone and ears not covered for the test frequency ranges 125 to 8000 Hz, 250 to 8000 Hz, and 500 to 8000 Hz.

1.2 Purpose

The purpose of this Standard is to specify MPANLs in an audiometric test room and measurement procedures that will produce negligible masking (≤ 2 dB) of pure tones presented at reference equivalent threshold levels as specified in ANSI S3.6-1996 *American National Standard Specifications for Audiometers* for different audiometric conditions and test frequency ranges.

This Standard is intended for use by all persons conducting hearing tests to ensure that ambient noise in an audiometric test room will not have a masking effect or influence on hearing measurements. This Standard is also intended for distributors, installers, designers, and manufacturers so that they can plan and construct appropriate acoustical treatment for the reduction of ambient noise in audiometric test rooms.

1.3 Applications

This Standard is a revision of ANSI S3.1-1991 *American National Standard Maximum Permissible Ambient Noise Level for Audiometric Test Rooms*. This Standard pertains to all audiometric testing conditions, particularly to those conditions that require measurement of hearing thresholds. Specifically, the ears covered MPANLs shall apply when both ears are covered by an earphone in a supra-aural cushion or when an insert earphone is placed in each ear canal, as would be the usual case for pure tone air conduction audiometry. The ears not covered MPANLs shall apply when audiometric testing is conducted with one ear not covered by an earphone, as would be typical for pure tone bone conduction audiometry. The ears not covered MPANLs also shall apply when audiometric testing is conducted with both ears not covered, as would be typical for sound field audiometry.

The Standard assumes that ambient noise masking of signals below reference equivalent threshold levels is tolerable. This assumption is not appropriate in applications where hearing thresholds lower than reference equivalent threshold levels must be measured. However, the audibility of all signals presented at levels equal to or greater than reference equivalent threshold levels must be such that the threshold of an average normally hearing listener will be free of any significant effects from ambient noise masking.

This Standard specifies that ambient noise in an audiometric test room shall be measured at octave