

ANSI/ASA S3.22-2014
Reaffirmed by ANSI June 5, 2020

AMERICAN NATIONAL STANDARD

Specification of Hearing Aid Characteristics

ANSI/ASA S3.22-2014

Accredited Standards Committee S3, Bioacoustics

Standards Secretariat
Acoustical Society of America
1305 Walt Whitman Road, Suite 300
Melville, NY 11747

The American National Standards Institute, Inc. (ANSI) is the national coordinator of voluntary standards development and the clearinghouse in the U.S.A. for information on national and international standards.

The Acoustical Society of America (ASA) is an organization of scientists and engineers formed in 1929 to increase and diffuse the knowledge of acoustics and to promote its practical applications.



ANSI/ASA S3.22-2014
(Revision of ANSI/ASA S3.22-2009)

AMERICAN NATIONAL STANDARD

Specification of Hearing Aid Characteristics

Secretariat:

Acoustical Society of America

Approved on November 10, 2014, by:

American National Standards Institute, Inc.

Abstract

The standard describes air-conduction hearing aid measurement methods that are particularly suitable for specification and tolerance purposes. Among the test methods described are output sound pressure level (SPL) with a 90-dB input SPL, full-on gain, frequency response, harmonic distortion, equivalent input noise, current drain, and induction-coil sensitivity. Specific configurations are given for measuring the input SPL to a hearing aid. Allowable tolerances in relation to values specified by the manufacturer are given for certain parameters. Annexes are provided to describe an equivalent substitution method, characteristics of battery simulators, static and dynamic characteristics of automatic gain control (AGC) hearing aids, and additional tests to characterize more completely the electroacoustic performance of hearing aids.

AMERICAN NATIONAL STANDARDS ON ACOUSTICS

The Acoustical Society of America (ASA) provides the Secretariat for Accredited Standards Committees S1 on Acoustics, S2 on Mechanical Vibration and Shock, S3 on Bioacoustics, S3/SC 1 on Animal Bioacoustics, and S12 on Noise. These committees have wide representation from the technical community (manufacturers, consumers, trade associations, organizations with a general interest, and government representatives). The standards are published by the Acoustical Society of America as American National Standards after approval by their respective Standards Committees and the American National Standards Institute (ANSI).


These standards are developed and published as a public service to provide standards useful to the public, industry, and consumers, and to Federal, State, and local governments.

Each of the Accredited Standards Committees (operating in accordance with procedures approved by ANSI) is responsible for developing, voting upon, and maintaining or revising its own Standards. The ASA Standards Secretariat administers Committee organization and activity and provides liaison between the Accredited Standards Committees and ANSI. After the Standards have been produced and adopted by the Accredited Standards Committees, and approved as American National Standards by ANSI, the ASA Standards Secretariat arranges for their publication and distribution.

An American National Standard implies a consensus of those substantially concerned with its scope and provisions. Consensus is established when, in the judgment of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered and that a concerted effort be made towards their resolution.

The use of an American National Standard is completely voluntary. Their existence does not in any respect preclude anyone, whether he or she has approved the Standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the Standards.

NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this Standard.



Acoustical Society of America
ASA Secretariat
1305 Walt Whitman Road, Suite 300
Melville, New York 11747
Telephone: 1 (631) 390-0215
Fax: 1 (631) 923-2875
E-mail: asastds@acousticalsociety.org

© 2014 by Acoustical Society of America. This standard may not be reproduced in whole or in part in any form for sale, promotion, or any commercial purpose, or any purpose not falling within the provisions of the U.S. Copyright Act of 1976, without prior written permission of the publisher. For permission, address a request to the Standards Secretariat of the Acoustical Society of America.

Contents

1	Scope, applications and purpose	1
1.1	Scope	1
1.2	Applications	1
1.3	Purpose	1
2	Normative references	1
3	Definitions	2
3.1	Terms relating to hearing aids	2
3.2	Terms relating to test equipment and test conditions	3
3.3	Terms relating to measurements with acoustic inputs	3
3.4	Terms relating to measurements with inductive inputs	4
4	Test equipment	4
4.1	Test space	4
4.2	Measurement configuration for nondirectional hearing aids	5
4.3	Measurement configuration for directional hearing aids	5
4.4	Sound source	6
4.5	Test signal	6
4.6	Frequency accuracy	7
4.7	Earphone coupler	7
4.8	The rms response	7
4.9	Averaging time constant for noise measurement	7
4.10	Current measurement	7
5	Standard conditions	8
5.1	Ambient conditions	8
5.2	Operating conditions	8
6	Recommended measurements, specifications and tolerances	13
6.1	Curves	13
6.2	OSPL90 curve	13
6.3	HFA-OSPL90	13
6.4	Full-on gain	14
6.5	HFA full-on gain (HFA-FOG)	14
6.6	Adjustment of the gain control to the reference test setting (RTS)	14
6.7	Reference test gain (RTG)	14
6.8	Frequency response curve	15
6.9	Frequency range	15
6.10	Tolerance method for frequency response curve	15
6.11	Harmonic distortion	16
6.12	Equivalent input noise level (EIN)	17
6.13	Battery current	18
6.14	Induction coil response	18
6.15	Interpretation of tolerances	21

Annex A (Informative) Equivalent substitution method	22
A.1 Introduction	22
A.2 Purpose	22
A.3 Application	22
A.4 Special equipment	22
A.5 Procedure	22
Annex B (Informative) Recommended voltage and resistance values to be used in battery simulators	25
B.1 Introduction	25
B.2 Battery simulator characteristics	25
Annex C (Informative) Guidelines for additional optional tests to characterize hearing aids	26
C.1 Characteristics of the gain control	26
C.2 Characterization of battery current as a function of quiescent current and maximum current.....	26
C.3 Effect of tone-control setting on frequency response	27
C.4 Effect of output limiting control setting on OSPL90 and full-on frequency response	27
C.5 Effect of gain control setting on frequency response	28
C.6 Effect of power supply voltage variation on acoustic gain and OSPL90	28
C.7 Effect of power supply impedance variation on acoustic gain and OSPL90	28
C.8 Hearing aid output noise spectrum.....	29
C.9 AGC hearing aids	32
C.10 Total harmonic distortion in acoustic mode as a function of input SPL.....	34
C.11 Total harmonic distortion in acoustic mode as a function of frequency	35
C.12 Maximum induction coil sensitivity measurement	36
C.13 Total harmonic distortion for induction coil mode as a function of input magnetic field strength.....	36
C.14 Difference frequency distortion.....	37
Bibliography.....	40

Figures

Figure 1 — Measurement configuration for nondirectional hearing aids	5
Figure 2 — Measurement configuration for directional hearing aids	6
Figure 3 — HA-1 configuration	9
Figure 4 — HA-2 configuration	10
Figure 5 — HA-2B configuration	11
Figure 6 — HA-3 configuration	11
Figure 7 — HA-4 configuration	12
Figure 8 — Example of OSPL90 and frequency response curves	14

Revision of ANSI/ASA S3.22-2009

Figure 9 — Example of construction of tolerance template for frequency response curve 16

Figure 10 — Hearing aids on TMFS for SPLITS test; BTE is shown for left ear test..... 18

Figure 11 — Telephone magnetic field simulator..... 19

Figure 12 — Example orientation of an ITE aid for the SPLIV test..... 20

Figure 13 — Example orientation of a BTE aid for the SPLIV test..... 20

Figure A.1 — Setup for measurement of sound level corrections..... 23

Figure A.2 — Setup for equivalent substitution method of hearing aid measurement..... 24

Figure C.1 — Example of a test equipment noise spectrum in one-third-octave bands 30

Figure C.2 — Example of a hearing aid output noise spectrum in one-third-octave bands..... 30

Figure C.3 — Example of acoustic gain for sinusoidal input signals at the frequencies shown,
for the hearing aid used for Figure C.2 31

Figure C.4 — Equivalent input noise in one-third-octave bands for the hearing aid used for
Figures C.2 and C.3 31

Figure C.5 — Example of a steady-state input-output function illustrating compression ratio
and expansion ratio 33

Figure C.6 — Schematic illustration of input signal envelope (top) and output signal envelope
(bottom) with attack and release time 34

Figure C.7 — Two examples of total harmonic distortion as a function of the input sound
pressure level 35

Figure C.8 — Example plot of THD vs. frequency 35

Figure C.9 — Example of total harmonic distortion as a function of magnetic input field
strength..... 36

Figure C.10 — Definition of difference frequency distortion products..... 38

Figure C.11 — Fundamental and difference frequency distortion..... 38

Figure C.12 — Total difference frequency distortion..... 39

Tables

Table B.1 — Battery simulator characteristics 25

Foreword

[This Foreword is for information only and is not an integral part of ANSI/ASA S3.22-2014 American National Standard Specification of Hearing Aid Characteristics. As such, this Foreword may contain material that has not been subjected to public review or a consensus process. In addition, it does not contain requirements necessary for conformance to the standard.]

This standard comprises a part of a group of definitions, standards, and specifications for use in bioacoustics. It was developed and approved by Accredited Standards Committee S3, Bioacoustics, under its approved operating procedures. Those procedures have been accredited by the American National Standards Institute (ANSI). The Scope of Accredited Standards Committee S3 is as follows:

Standards, specifications, methods of measurement and test, and terminology in the fields of psychological and physiological acoustics, including aspects of general acoustics which pertain to biological safety, tolerance and comfort.

The following is a list of the major non-editorial revisions to this edition of the standard:

- added an annex for an optional test of difference frequency distortion;
- moved Annex C.11 Induction Coil Performance With Loops into the main body of the standard;
- deleted the vertical reference angle;
- added a definition of a hearing aid;
- added graphs of attack and release time measurements to annex for Dynamic AGC characteristics;
- added to Annex C optional total harmonic distortion tests at more frequencies;
- the HA-1, HA-2, HA-3 and HA-4 coupler configurations previously defined in ANSI/ASA S3.7 have been updated and included in clause 5.2.3. New figures of these configurations have also been included.

This standard is a revision of ANSI/ASA S3.22-2009, developed originally to establish measurement and specification methods for several definitive hearing aid characteristics and to provide tolerances for some of them. The original purpose of the standard was to provide a means of determining whether a production hearing aid as shipped was as stated by a manufacturer for a particular model, within the tolerances specified in the standard. In the 2009 revision of the standard, considerable effort was made to achieve harmonization with IEC 60118-7.

In the 1996 revision of the standard, the gain control was set to reference test position for automatic gain control (AGC) hearing aids as has been done for all other types of hearing aids. To reduce ambiguity in specifying this procedure, and to reflect common practices in the hearing aid industry at this time, in the 2003 revision of the standard, AGC hearing aids are tested in AGC mode only for tests associated with AGC functions and are operated in non-AGC mode for all other tests. That is, for all hearing aids, for measurements to determine OSPL90, full-on gain, the Reference Test Setting of the gain control (RTS), total harmonic distortion, equivalent input noise, battery current drain, and induction coil sensitivity the hearing aid is set to operate in non-AGC mode. For AGC hearing aids, tests for input-output characteristic and attack and release times are made with the hearing aid operating in AGC mode. In the 2009 revision of the standard, the AGC tests were moved to Annex C and provision was made to use a stationary noise signal as well a sinusoidal signal for some tests.

Although intended for the characterization of air-conduction hearing aids, the methods in this standard can be applied to other systems which have acoustic inputs and outputs, such as personal sound amplification products.

Although the purpose of this standard is quality measurements using the 2 cm³ coupler, the methods described herein may also be used for design, research, or product development using the 2 cm³ coupler, occluded ear simulator, or other specialized couplers or ear simulators. Additional advanced methods are found in the annexes.

This standard contains several informative annexes which are not considered to be part of this standard.

Since 1976, earlier versions of this standard have been incorporated into regulations of the United States Food and Drug Administration and have given guidance to manufacturers and consumers of hearing aids and to those who serve the hearing-impaired population.

At the time this Standard was submitted to Accredited Standards Committee S3, Bioacoustics for approval, the membership was as follows:

C.J. Struck, *Chair*
P. Nelson, *Vice-Chair*

S.B. Blaeser, *Secretary*

Acoustical Society of America	C.J. Struck P. Nelson (Alt.)
American Academy of Audiology	C. Schweitzer T. Ricketts (Alt.)
American Academy of Otolaryngology, Head and Neck Surgery, Inc.	R.A. Dobie A. Kim (Alt.)
American Industrial Hygiene Association	T.K. Madison D. Driscoll (Alt.)
American Speech-Language-Hearing Association (ASHA)	L.A. Wilber N. DiSarno (Alt.)
Beltone/GN Resound	S. Petrovic
Council for Accreditation in Occupational Hearing Conservation (CAOHC)	L.D. Hager
Diagnostic Group	T. McColley P. Dobrowski (Alt.)
ETS – Lindgren Acoustic Systems	S. Dunlap D. Winker (Alt.)
Etymotic Research, Inc.	M.C. Killion J.K. Stewart (Alt.)
Food and Drug Administration	S-C Peng
G.R.A.S. Sound & Vibration	B. Schustrich J. Soendergaard (Alt.)
Hearing Industries Association	A. Bopp M. Jones (Alt.)

National Electrical Manufacturers Association, Signaling Protection & Communication Section (NEMA – 3SB)	J. McNamara R. Reiswig (Alt.)
National Hearing Conservation Association	G.L. Poling R. Danielson (Alt.)
National Institute for Occupational Safety and Health (NIOSH)	M.R. Stephenson W.J. Murphy (Alt.)
National Institute of Standards and Technology	R.P. Wagner S. Fick (Alt.)
National Park Service	M. McKenna K. Fristrup (Alt.)
Natus Medical, Inc.	Y. Hekimoglu
Ocean Conservation Research	M. Stocker
Starkey Hearing Technologies	D.A. Preves T.H. Burns (Alt.)
U.S. Air Force	R.L. McKinley B.D. Simpson (Alt.)
U.S. Army Aeromedical Research Lab	W.A. Ahroon
U.S. Army CERL	D.K. Delaney M.J. White (Alt.)
U.S. Army Research Laboratory, Human Research and Engineering Directorate	A. Scharine M.A. Ericson (Alt.)
University of Cincinnati Animal Audiology Clinic/Bioacoustics Lab	P.M. Scheifele D.K. Brown (Alt.)

Individual Experts of the Accredited Standards Committee S3, Bioacoustics, were:

A.J. Brammer
R.F. Burkard
M.D. Burkhard
A.J. Campanella

R.L. McKinley
V. Nedzelnitsky
P.D. Schomer

C.J. Struck
H. Teder
L.A. Wilber
W.A. Yost

Working group S3/WG48, Hearing Aids, which assisted Accredited Standards Committee, S3, Bioacoustics, in the preparation of this standard, had the following membership who worked actively on this revision:

David Preves, Chair
William Cole, S3.22 Sub-group Chair
Steve Armstrong, Secretary

R. Brennan
T. Burns
J. DaBell
V. Dasika
B. Fedor
G. Frye
A. Gebert
G. Haas

D. Hotvet
J. Kane
M. Killion
B. Kruger
F. Kruger
V. Nedzelnitsky
R. Olson
S-C Peng

S. Petrovic
G. Ravn
L. Revit
O. Saltykov
C. Struck
T. Victorian
R. Wagner
D. Warren

Suggestions for improvements to this standard will be welcomed. They should be sent to Accredited Standards Committee S3, Bioacoustics, in care of the Standards Secretariat of the Acoustical Society of America, 1305 Walt Whitman Road, Suite 300, Melville, New York 11747. Telephone: 631-390-0215; Fax: (631) 923-2875; E-mail: asastds@acousticalsociety.org.

Introduction

This standard describes hearing aid measurements that are particularly suitable for specification and tolerance purposes.

The procedures of this standard employ a reference test setting of the gain control (see 6.6) to which the hearing aid is adjusted for certain measurements such as frequency response, harmonic distortion, and equivalent input noise level. The rationale underlying the use of the reference test setting is that the gain control setting for certain tests should be related to the saturation output capability of the hearing aid. Among the advantages of the gain control setting specified herein are (1) the gain control is set fairly close to a typical "use" setting, and (2) harmonic distortion measurements are made with a setting appropriately related to the maximum output capability of the hearing aid.

NOTE The concept of using a gain control setting related to the output capability of a hearing aid was originally put into practice by the National Bureau of Standards in their work for the Veterans Administration. This standard uses similar principles.

An important feature of the procedures described in this standard is the method of determining the input SPL at the microphone opening(s) of the hearing aid (see 3.2.2 and 4.2).

American National Standard

Specification of Hearing Aid Characteristics

1 Scope, applications and purpose

1.1 Scope

This standard describes certain hearing aid measurements and parameters that are deemed useful in determining the electroacoustic performance of an air-conduction hearing aid. Some of these lend themselves to setting of tolerances for the purpose of maintaining product uniformity and for compliance with the performance specified for a model.

It is not the intent of this document to restrict the variety of hearing aid performance available nor to inhibit in any way advances in the state of the art.

This standard is limited to the specification of certain electroacoustic characteristics based on sinusoidal and other steady-state test signals. Measurements described in this standard are not intended to reflect *in situ* performance of hearing aids, directional performance of directional hearing aids, or certain performance properties of digital hearing aids, such as processing delay.

Notes in the standard are not considered to be part of the standard.

1.2 Applications

Tolerances are given relative to specified characteristics supplied by a manufacturer. In the case of "custom" or "made-to-order" hearing aids, individual test data are to be supplied. The tolerances specified in the standard will apply to the individual test data supplied.

1.3 Purpose

This standard is intended to meet the need for specifications of air-conduction hearing aid performance parameters and their tolerances. The quantities suggested for specifications and tolerances are considered to be useful for comparing performance characteristics of different hearing aids or for comparing performance characteristics of a hearing aid with published specifications.

2 Normative references

The following referenced documents are indispensable for the application of this standard. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

[1] ANSI/ASA S1.1 *American National Standard Acoustical Terminology*

[2] ANSI/ASA S1.6 *American National Standard Preferred Frequencies, Frequency Levels, and Band Numbers for Acoustical Measurements*

[3] ANSI/ASA S3.42-1992/Part 1 (R 2012) *American National Standard Testing Hearing Aids with a Broad-Band Noise Signal*