

**ANSI S3.42-1992
(ASA 103-1992)**

Reaffirmed by ANSI
April 22, 2002

Reaffirmed by ANSI
May 18, 2007

Reaffirmed by ANSI
June 15, 2012

Reaffirmed by ANSI
July 12, 2017

**AMERICAN NATIONAL STANDARD
Testing Hearing Aids with a
Broad-Band Noise Signal**

**ACCREDITED STANDARDS COMMITTEE S3,
BIOACOUSTICS**

ABSTRACT

This standard describes techniques for characterizing the steady-state performance of hearing aids with a broad-band noise signal. The need for such a standard arises from the importance of assessing the performance of hearing aids in environments more nearly representing their real-world use. The noise test signal specified herein has been employed by the National Bureau of Standards for over 20 years in testing hearing aids. Among the tests described are noise saturation sound pressure level, noise gain, frequency response, family of frequency response curves and output versus input characteristic. Additionally, the appendix recommends use of the coherence function to indicate the validity of frequency response measures and distinguishes between use of random and pseudo-random noise and asynchronous versus synchronous analysis.

AMERICAN NATIONAL STANDARDS ON ACOUSTICS

The Acoustical Society of America provides the Secretariat for Accredited Standards Committees S1 on Acoustics, S2 on Mechanical Shock and Vibration, S3 on Bioacoustics, and S12 on Noise. These committees have wide representation from the technical community (manufacturers, consumers, and general-interest representatives). The standards are published by the Acoustical Society of America through the American Institute of Physics as American National Standards after approval by their respective standards committees and the American National Standards Institute.

These standards are developed and published as a public service to provide standards useful to the public, industry, and consumers, and to Federal, State, and local governments.


This standard was approved by the American National Standards Institute as ANSI S3.42-1992 on 2 June 1992.

An American National Standard implies a consensus of those substantially concerned with its scope and provisions. An American National Standard is intended as a guide to aid the manufacturer, the consumer, and the general public. The existence of an American National Standard does not in any respect preclude anyone, whether he has approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. American National Standards are subject to periodic review and users are cautioned to obtain the latest editions.

Caution Notice: An American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken to reaffirm, revise, or withdraw this standard no later than five years from the date of publication.

The American National Standards Institute, Inc. (ANSI) is the national coordinator of voluntary standards development and the clearing house in the U.S. for information on national and international standards.

The Acoustical Society of America (ASA) is an organization of scientists and engineers formed in 1929 to increase and diffuse the knowledge of acoustics and to promote its practical applications.



Published by the Acoustical Society of America through the American Institute of Physics

© 1992 by the Acoustical Society of America. This Standard may not be reproduced in whole or in part in any form for sale, promotion, or any commercial purpose, or any purpose not falling within the fair-use provisions of the Copyright Act of 1976, without prior written permission of the publisher. For permission, address the Standards Secretariat of the Acoustical Society of America.

FOREWORD

[This Foreword is not a part of American National Standard ANSI S3.42-1992—Testing Hearing Aids with a Broad-Band Noise Signal, ASA Catalog No. 103-1992.]

American National Standards have traditionally utilized pure tone signals to test hearing aids. However, new types of hearing aids have been developed over the last few years with increasingly complex and non-linear signal processing algorithms. The salient features of these devices are often not well characterized by pure tone measurements. Thus, the use of pure tones to assess the performance of these newer hearing aids is limited to quality control purposes. This document represents an initial effort to develop a national standard for the measurement of hearing aid performance with a steady-state complex input signal. The signal recommended herein is a random noise that has been spectrally shaped to represent the short-term average speech spectrum.

In addition to addressing methods of expressing gain, saturation sound pressure level and frequency response using the noise input signal, the use of the coherence function is recommended to validate the frequency response, and indirectly as an indicator of the amount of noise and distortion produced by a hearing aid.

This standard has been developed under the jurisdiction of Accredited Standards Committee S3, Bioacoustics, using the American National Standards Institute (ANSI) Accredited Standards Committee Procedure. The Acoustical Society of America provides the Secretariat for Accredited Standards Committee S3, Bioacoustics.

Accredited Standards Committee S3, Bioacoustics, under whose jurisdiction this standard was developed, has the following scope:

Standards, specifications, methods of measurement and test, and terminology, in the fields of psychological and physiological acoustics, including aspects of general acoustics, shock and vibration which pertain to biological safety, tolerance, and comfort.

At the time this Standard was submitted to Accredited Standards Committee S3, Bioacoustics, for approval, the membership was as follows:

J. D. Royster, *Chairman*
J. L. Fletcher, *Vice-Chairman*
A. Brenig, *Secretary*

Acoustical Society of America ● J. D. Royster, J. L. Fletcher (*Alt*)
American Academy of Otolaryngology, Head and Neck Surgery, Inc. ● R. F. Naunton, L.A. Michael (*Alt*)
American Industrial Hygiene Association ● C. D. Bohl
American Institute of Ultrasound in Medicine ● J. Zagzebski
American College of Occupational Medicine ● P. J. Brownson, J. Sataloff (*Alt*)
American Otological Society, Inc. ● R. F. Naunton
American Speech-Language-Hearing Association ● R. Kasten
AT&T Bell Laboratories ● R. M. Sachs, M. S. Mayer (*Alt*)
Audio Engineering Society, Inc. ● R. H. Campbell, M. R. Chial (*Alt*)
Bruel and Kjaer Instruments Inc. ● J. R. Bareham, G. C. Michel (*Alt*)
Compressed Air and Gas Institute ● J. H. Addington
Devco Corporation ● K. T. Chandy
Exchange Carriers Standards Association ● B. Kushler
Fastener Industry Noise Control Research Program ● E. H. Toothman, J. C. McMurray (*Alt*)
Hearing Industries Association (HIA) ● H. Teder, C. Rogin (*Alt*)
Industrial Safety Equipment Association (ISEA) ● A. M. Bovi, R. Campbell (*1st Alt*), F. E. Wilcher, Jr. (*2nd Alt*)
National Hearing Aid Society ● W. F. S. Hopmeier
National Institute of Standards and Technology ● E. D. Burnett, V. Nedzelnitsky (*Alt*)

Power Tool Institute, Inc. ● J. L. Bennett, M. Brown (*Alt*)
U.S. Air Force ● C. Nixon
U.S. Army Aeromedical Research Laboratory ● J. H. Patterson, B. Mozo (*Alt*)
U.S. Army Human Engineering Laboratory ● B. Garinther, G. R. Price (*Alt*)
U.S. Army Medical Corps ● R. M. Attack
U.S. Department of the Navy, Bureau of Medicine and Surgery ● J. Page, L. Marshall (*Alt*)

Individual experts of the Accredited Standards Committee S3, Bioacoustics were:

S. J. Barry	J. C. Guignard	W. Melnick
R. W. Benson	D. L. Johnson	H. E. von Gierke
K. M. Eldred	K. D. Kryter	D. E. Wasserman
R. S. Gales	H. Levitt	L. A. Wilber
W. J. Galloway	S. F. Lybarger	W. Yost
R. Guernsey	R. McKinley	R. W. Young

Working Group S3/WG48 Hearing Aids, which assisted Accredited Standards Committee, S3, Bioacoustics, in the development of this standard, has the following membership:

D. A. Preves, *Chairman*
W. O. Olsen, *Secretary*

J. R. Bareham	W. A. Cole	B. Kruger
L. B. Beck	G. J. Frye	I. Leonard
A. J. Becker	H. Goldberg	S. Lybarger
R. Brander	W. F. S. Hopmeier	R. F. Sullivan
E. D. Burnett	R. Kasten	H. Teder
E. Carlson	J. Kates	G. P. Widin

Suggestions for improvements in this Standard will be welcomed. They should be sent to **Accredited Standards Committee S3 at the Standards Secretariat, in care of the Acoustical Society of America, 335 East 45th Street, New York, NY 10017-3483. Telephone (212) 661-9404.**

CONTENTS

0 INTRODUCTION	1
REFERENCES	2
1 SCOPE	2
2 DEFINITIONS	2
2.1 NSPL90	2
2.2 Full-on noise gain.....	2
2.3 Spectrum	2
2.4 Spectrum level.....	2
2.5 Auto-spectrum (power spectrum)	2
2.6 Cross-spectrum	2
2.7 Coherence	2
2.8 Crest factor.....	2
2.9 Spectrum analyzer	2
2.10 RMS	3
2.11 Butterworth filter	3
2.12 Random noise.....	3
2.13 Pseudo-random noise	3
2.14 <i>In-situ</i>	3
2.15 Simulated <i>in-situ</i>	3
2.16 dB SPL.....	3
2.17 Synchronous analysis	3
3 NOISE INPUT SIGNAL	3
3.1 Noise type and crest factor	3
3.2 Spectrum	4
4 SPECTRUM ANALYSIS EQUIPMENT	5
5 TEST ENVIRONMENT	5
6 TEST PROCEDURES	6
6.1 Equalization	6
6.2 Hearing Aid Settings	6
6.3 NSPL 90	6
6.4 Full-on Noise Gain.....	6
6.5 Frequency Response	6
6.6 Family of Frequency Response Curves	7
6.7 Output Versus Input Characteristic.....	7
APPENDIX A: FAST FOURIER SPECTRUM ANALYSIS FOR DEVELOPING HEARING AID FREQUENCY RESPONSES	8
A.1 Introduction.....	8
A.2 Synchronous analysis with pseudo-random noise versus analysis with random noise.....	9

A.3	Auto-Spectrum analysis.....	9
A.3.1	Synchronous auto-spectrum analysis with pseudo-random noise....	9
A.3.2	Auto-spectrum analysis with random noise	9
A.4	Cross-spectrum analysis	10
A.4.1	Cross-spectrum analysis with pseudo-random noise	10
A.4.2	Cross-spectrum analysis with random noise	10
A.5	Coherence function	10
A.5.1	Coherence with cross-spectrum analysis	10
A.5.2	Coherence using synchronous analysis with a pseudo-random noise input signal	12
A.5.3	Bias and variance in coherence measurements	12
A.6	Coherent and noncoherent power	12
APPENDIX B:	SMOOTHED DATA PRESENTATION	12
B.1	3-point or 5-point running average.....	13
B.2	Third-octave running average method.....	13
B.3	Tenth-octave running method	13
APPENDIX C:	SUMMARY OF TESTING CHARACTERISTICS TO BE SPECIFIED	14

American National Standard Testing Hearing Aids with a Broad-Band Noise Signal

0 INTRODUCTION

The frequency response of electroacoustic systems has traditionally been obtained with a swept pure tone input signal whose level is held constant while the system output is monitored over the frequency range of interest. Heretofore, this has also been the method of testing hearing aid frequency response. However, another procedure has evolved for obtaining frequency responses of electronic systems as a result of the recent proliferation of digital spectrum analyzers which utilize steady-state broad-band noise as one of their test signals. A time-stationary, steady-state broad-band noise, which is more typical of the complex input signals that hearing aids are required to process in non-laboratory real-world environments, may be a more suitable test signal for depicting performance, particularly for those hearing aids with level dependent gain circuitry. However, to date, no standardized document exists that denotes procedures for testing hearing aids with a broad-band input signal.

For those hearing aids which do not have automatic gain control (AGC) or other forms of adaptive signal processing circuitry, or for hearing aids having such circuitry but tested with input levels below their activation point, the same frequency response should result whether a swept pure tone is used or a broad-band noise is used, as long as the hearing aid is operating linearly and the signal-to-noise ratio is adequate. The intent of the present ANSI hearing aid testing standard S3.22-1987 is to describe the frequency response of hearing aids in their linear mode regardless of whether automatic signal processing such as AGC is incorporated or not. However, a method is needed to demonstrate the change in steady-state frequency response of AGC or other adaptive circuit action as a function of input signal level because of the effect it may have on the speech recognition abilities of hearing aid wearers. Knowledge of such effects may be used to provide a better selection of frequency-gain characteristics for hearing aid fitting purposes. Care in selecting the most appropriate methods for characterizing AGC systems is but one manifestation of a growing awareness in hearing aid measurements: the more sophisticated signal processing techniques are employed such as, for example, frequency-selective input or output compression, the more the selection of appropriate measurement signals and measurement techniques be-

comes crucial to realizing the goal of obtaining meaningful performance measures.

Frequency response curves developed for hearing aids with level dependent frequency response, or for AGC hearing aids with frequency dependent compression threshold using swept pure tones at varying input levels may not be representative of the response using complex signals. This occurs with the swept-tone method because only one frequency is presented at a time and the control system responds to each frequency individually. In the case of AGC hearing aids with frequency dependent compression threshold tested at high input levels, compression may vary with frequency, producing a flattened frequency response curve not representative of the response obtained with a complex input signal.

For an input signal more representative of real use conditions, such as speech, many frequency components are present simultaneously. There is considerable precedent for testing hearing aids either with linear circuitry or with automatic signal processing circuitry using a shaped, steady-state, broad-band noise for an input signal. The National Institute of Standards and Technology (formerly the National Bureau of Standards) and the Veterans Administration have historically used a broad-band, speech spectrum-like noise input signal to determine performance characteristics of hearing aids. The test signal specified in this standard has spectral characteristics similar to those of the short-term spectrum of speech but is not representative of other important characteristics such as the temporal nature and the amplitude probability distribution of speech. With this signal, an AGC detector will respond to a single level, from contributions at many frequencies, not to the individual frequency components. Thus, with this signal, the individual frequency components affected by an AGC loop will retain their relative amplitude relationships.

Care should be exercised in interpreting measurements made with a steady-state noise signal because hearing aids whose frequency response is changed by the dynamic characteristics of the input signal cannot be fully characterized by this time-invariant signal, e.g., hearing aids that have adaptive AGC time constants based on the temporal pattern of the input signal.

References

ANSI S3.22-1987 Specification of Hearing Aid Characteristics.

ANSI S3.35-1985(R 1990) Method of Measurement of Performance Characteristics of Hearing Aids Under Simulated *in-situ* Working Conditions.

Bendat, J. S. and Piersol, A. G. (1980): *Engineering Applications of Correlation and Spectral Analysis*, Wiley & Sons, New York.

Prokakis, J. and Manolakis, D. (1988): *Introduction to Digital Signal Processing*, Macmillan Publishing Company, New York.

Rife, D. and Vanderkooy, J. (1989): Transfer function measurement with maximum-length sequences, *J. Audio Eng. Soc.*, 37, 6: 419-443.

1 SCOPE

The purpose of this document is to define a test method with which to characterize the steady-state frequency response and input/output characteristics of hearing aids as the input level varies. This method is particularly useful for those hearing aids that have automatic gain control or other types of adaptive circuitry.

2 DEFINITIONS

2.1 NSPL90

The NSPL90 is the root-mean-square (RMS) output sound pressure level (SPL) produced by a hearing aid with its gain set at maximum and with a 90 dB RMS speech-weighted noise input SPL. The standard measurement bandwidth is limited to 200 to 5000 Hz because of diffraction problems and equipment accuracy tolerances at high frequencies. If a measurement bandwidth other than 200 to 5000 Hz is used, it shall be stated.

2.2 Full-on noise gain

With the gain control on the hearing aid in the full-on position, the noise gain is a single figure expressed in decibels and is derived by subtracting the overall RMS input SPL to the hearing aid from the RMS output SPL produced by the hearing aid. The standard measurement bandwidth is 200 to 5000 Hz. E.g., with a 50 dB RMS noise input sound pressure level, full-on noise gain = output SPL - 50 dB.

2.3 Spectrum

The distribution of amplitude and phase of the components of a signal as a function of frequency.

2.4 Spectrum level

The level of that part of the signal contained within a band 1 Hz wide, centered at the particular frequency.

2.5 Auto-spectrum (power spectrum)

The auto-spectrum represents the power level of either the input signal (G_{AA}) to or the output signal (G_{BB}) from a hearing aid in the frequency domain. It is computed by multiplying the Fourier transform of the signal by the complex conjugate of the Fourier transform of the signal.

2.6 Cross-spectrum

The cross-spectrum (G_{AB}) indicates the degree to which the same signal frequencies are mutually present in the input and output of a hearing aid. It is computed by multiplying the complex conjugate of the Fourier transform of the input signal to the hearing aid by the Fourier transform of the output signal from the hearing aid.

2.7 Coherence

A number ranging from 0 to 1 showing to what degree the output from a hearing aid is correlated to the input. Coherence for a random noise test signal is reduced by nonlinearity and by system noise. For a more thorough definition, see Appendix A.

2.8 Crest factor

The ratio of the peak value of a signal to its root-mean-square (RMS) value as follows:

Crest factor (dB) = $20\log_{10}$

$$\left(\frac{\text{peak voltage or pressure}}{\text{RMS voltage or pressure}} \right).$$

2.9 Spectrum analyzer

An instrument generally used to display the power distribution of a signal as a function of frequency. A dual channel analyzer permits calculation of the fre-