

ANSI/ASA S3.6-2010
(Revision of ANSI S3.6-2004)

AMERICAN NATIONAL STANDARD

Specification for Audiometers

ANSI/ASA S3.6-2010

Accredited Standards Committee S3, Bioacoustics

Standards Secretariat
Acoustical Society of America
35 Pinelawn Road, Suite 114 E
Melville, NY 11747-3177

The American National Standards Institute, Inc. (ANSI) is the national coordinator of voluntary standards development and the clearinghouse in the U.S.A. for information on national and international standards.

The Acoustical Society of America (ASA) is an organization of scientists and engineers formed in 1929 to increase and diffuse the knowledge of acoustics and to promote its practical applications.



ANSI/ASA S3.6-2010
(Revision of ANSI S3.6-2004)

AMERICAN NATIONAL STANDARD
Specification for Audiometers

Secretariat:

Acoustical Society of America

Approved November 2, 2010 by:

American National Standards Institute, Inc.

Abstract

The audiometers covered in this specification are devices designed for use in determining the hearing threshold level of an individual in comparison with a chosen standard reference threshold level. This standard provides specifications and tolerances for pure tone, speech, and masking signals and describes the minimum test capabilities of different types of audiometers.

AMERICAN NATIONAL STANDARDS ON ACOUSTICS

The Acoustical Society of America (ASA) provides the Secretariat for Accredited Standards Committees S1 on Acoustics, S2 on Mechanical Vibration and Shock, S3 on Bioacoustics, S3/SC 1 on Animal Bioacoustics, and S12 on Noise. These committees have wide representation from the technical community (manufacturers, consumers, trade associations, organizations with a general interest, and government representatives). The standards are published by the Acoustical Society of America as American National Standards after approval by their respective Standards Committees and the American National Standards Institute (ANSI).


These standards are developed and published as a public service to provide standards useful to the public, industry, and consumers, and to Federal, State, and local governments.

Each of the Accredited Standards Committees (operating in accordance with procedures approved by ANSI) is responsible for developing, voting upon, and maintaining or revising its own Standards. The ASA Standards Secretariat administers Committee organization and activity and provides liaison between the Accredited Standards Committees and ANSI. After the Standards have been produced and adopted by the Accredited Standards Committees, and approved as American National Standards by ANSI, the ASA Standards Secretariat arranges for their publication and distribution.

An American National Standard implies a consensus of those substantially concerned with its scope and provisions. Consensus is established when, in the judgment of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered and that a concerted effort be made towards their resolution.

The use of an American National Standard is completely voluntary. Their existence does not in any respect preclude anyone, whether he or she has approved the Standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the Standards.

NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this Standard.



Acoustical Society of America
ASA Secretariat
35 Pinelawn Road, Suite 114E
Melville, New York 11747-3177
Telephone: 1 (631) 390-0215
Fax: 1 (631) 390-0217
E-mail: asastds@aip.org

© 2010 by Acoustical Society of America. This standard may not be reproduced in whole or in part in any form for sale, promotion, or any commercial purpose, or any purpose not falling within the provisions of the U.S. Copyright Act of 1976, without prior written permission of the publisher. For permission, address a request to the Standards Secretariat of the Acoustical Society of America.

Contents

1	Scope, purpose, and applications	1
1.1	Scope	1
1.2	Purpose	1
1.3	Applications	1
2	Normative references	1
3	Terms and definitions	3
4	Requirements for specific types of audiometers	8
4.1	Type designation	8
5	General requirements	8
5.1	Electrical safety requirements	8
5.2	Acoustic safety requirements	9
5.3	Power supply variation and environment conditions	9
5.4	Unwanted acoustic signals	11
5.5	Subject response system	13
5.6	Monitor earphone or loudspeaker	13
5.7	Talk-back system	13
6	Signal sources	13
6.1	Pure-tone source	13
6.2	Speech source	16
6.3	Masking sounds	18
7	Signal level controls	20
7.1	Marking	20
7.2	Accuracy of sound pressure and vibratory force level for pure-tone and speech	20
7.3	Hearing level control	20
7.4	Masking sound level control	21
7.5	Tone switching	22
8	Reference-signal facilities	23
8.1	Frequencies	24
8.2	Reference-signal level control	24
9	Transducers (specifications, RETSPLs and RETFLs)	24
9.1	Air conduction, supra-aural earphones	24
9.2	Air conduction, circumaural earphones	26
9.3	Air conduction, insert earphones	27
9.4	Bone conduction	31
9.5	Sound field produced via loudspeakers	32
10	Device marking and instruction manual	34
10.1	Marking	34
10.2	Instruction manual	34
Annex A	(normative) Correction figures for free-field equivalent output for certain types of commonly used earphones	36

Annex B (normative) Recordings of speech test material	38
Annex C (normative) Calibration of circumaural earphones for extended high frequency testing	39
C.1 General	39
C.2 Calibration	39
C.3 Microphone	39
C.4 Adapters for making extended high frequency measurements on an IEC 60318-1 ear simulator	39
C.5 Coupler configurations using a Type 1 adapter	39
C.6 Coupler configuration using a Type 2 adapter	42
C.7 Environmental conditions	42
C.8 References	42
Annex D (informative) Procedures for transfer of reference equivalent threshold values	43
D.1 General	43
D.2 Probe-tube method	43
D.3 Loudness-balance method	44
D.4 Threshold-determination method	44
D.5 Standard reference earphone with other-than-specified coupler	44
D.6 Standard earphones with other-than-specified cushions	45
D.7 Earphone and coupler both different from standard types	45
Annex E (informative) Audiogram format	46
E.1 General	46
E.2 Reference levels	46
E.3 Vertical scale	46
E.4 Horizontal scale	46
E.5 Scale proportions	46
Annex F (informative) Supra-aural cushions for earphones	47
Annex G (informative) Requirements for periodic calibration	49
G.1 Minimum periodic calibration requirements	49
G.2 Routine checking and subjective tests	49
G.3 Equipment needed to perform audiometer calibrations	53
G.4 Calibration of test equipment	54
Bibliography	55

Tables

Table 1 — Minimum required facilities for specific types of audiometers	9
Table 2a — Required frequencies and hearing levels for various audiometers	14
Table 2b — Required frequencies and hearing levels for EHF audiometers	14
Table 3 — Maximum permissible harmonic distortion, expressed in percent	15
Table 4 — Narrow-band masking noise: upper and lower cut-off frequencies and correction factors for determining RETSPLs	19

Table 5 — Reference equivalent threshold sound pressure levels (RETSPLs) (dB re 20 μ Pa) for supra-aural earphones in common use.....	26
Table 6 — Reference equivalent threshold sound pressure levels (RETSPLs) (dB re 20 μ Pa) for circumaural earphones.....	27
Table 7a — Reference equivalent threshold sound pressure levels (RETSPLs) (dB re 20 μ Pa) for ER-3A and EARTone 3A insert earphones.....	29
Table 7b — Reference equivalent threshold sound pressure levels (RETSPLs) (dB re 20 μ Pa) for ER-2 insert earphones for use at Extended High Frequency	30
Table 8 — Reference equivalent threshold force levels (RETFLs) for bone vibrators.....	32
Table 9 — Reference equivalent threshold sound pressure levels (RETSPLs) (dB re 20 μ Pa) for sound-field testing	33
Table A.1 — Difference between the free-field equivalent sensitivity level GF and the coupler or acoustic coupler or ear simulator level GC for two types of earphones.....	37

Figures

Figure 1 — Rise/fall envelope of test tones.....	23
Figure 2 — Coupling of the insert earphone to the calibration coupler – Figure 2(a) shows an occluded ear simulator (as described in ANSI/ASA S3.25 and IEC 60318-4 (IEC 60711)), see subclause 9.3.1; Figure 2(b) an HA-2 acoustic coupler with rigid tube attachment (as described in ANSI/ASA S3.7), see subclause 9.3.2; and Figure 2(c) an HA-1 acoustic coupler (as described in ANSI/ASA S3.7), see subclause 9.3.3.....	28
Figure C.1 — Type 1 adapter for use with IEC 60318-1 ear simulator	40
Figure C.2 — Position of Type 1 adapter on the IEC 60318-1 ear simulator.....	40
Figure C.3 — Type 2 adapter for use with IEC 60318-1 ear simulator	41
Figure C.4 — Position of Type 2 adapter on the IEC 60318-1 ear simulator.....	41
Figure F.1 — Earphone supra-aural cushion MX 41/AR.....	47
Figure F.2 — Earphone supra-aural cushion Type 51	48

Foreword

[This Foreword is for information only and is not a part of ANSI/ASA S3.6-2010 American National Standard Specification for Audiometers.]

This standard comprises a part of a group of definitions, standards, and specifications for use in bioacoustics. It was developed and approved by Accredited Standards Committee S3 Bioacoustics, under its approved operating procedures. Those procedures have been accredited by the American National Standards Institute (ANSI). The Scope of Accredited Standards Committee S3 is as follows:

Standards, specifications, methods of measurement and test, and terminology in the fields of psychological and physiological acoustics, including aspects of general acoustics which pertain to biological safety, tolerance and comfort.

This standard is a revision of ANSI S3.6-2004, which has been technically revised to incorporate changes to keep this standard compatible with other equivalent international (IEC and ISO) standards. This standard contains comparable information to current ISO Standards for pure tone thresholds, bone conduction and sound field.

At the time this Standard was submitted to Accredited Standards Committee S3, Bioacoustics for approval, the membership was as follows:

C.A. Champlin, *Chair*
G.J. Frye, *Vice-Chair*

S.B. Blaeser, *Secretary*

Acoustical Society of America	C.A. Champlin
.....	M.D. Burkhard (Alt.)
American Academy of Audiology	D. Ostergren
.....	S. Gordon-Salant (Alt.)
American Academy of Otolaryngology - Head and Neck Surgery	R.A. Dobie
.....	L.A. Michael (Alt.)
American Industrial Hygiene Association	T.K. Madison
.....	D. Driscoll (Alt.)
American Speech-Language-Hearing Association (ASHA)	L.A. Wilber
.....	V. Gladstone (Alt.)
Beltone/GN Resound	S. Petrovic
Council for Accreditation in Occupational Hearing Conservation (CAOHC)	L.D. Hager
ETS-Lindgren Acoustic Systems	S. Dunlap
.....	D. Winker (Alt.)
Etymotic Research, Inc.	M.C. Killion
Food and Drug Administration	J.K. Kane
.....	S-C Peng (Alt.)

Frye Electronics, Inc.	G.J. Frye K.E. Frye (Alt.)
G.R.A.S. Sound & Vibration	B. Schustrich
Hearing Industries Association	T.A. Victorian C.M. Rogin (Alt.)
National Electrical Manufacturers Association, Signaling Protection & Communication Section (NEMA – 3SB)	J. McNamara R. Reiswig (Alt.)
National Hearing Conservation Association	T. Schulz
National Institute for Occupational Safety and Health (NIOSH)	M. Stephenson W.J. Murphy (Alt.)
National Institute of Standards and Technology	V. Nedzelitsky R. Wagner (Alt.)
National Park Service	G.R. Stanley K. Fristrup (Alt.)
Natus Medical, Inc.	Y. Hekimoglu P. Becke (Alt.)
Ocean Conservation Research	M. Stocker
Quest Technologies, Inc.	M. Wurm P.J. Battenberg (Alt.)
Starkey Laboratories	D.A. Preves T.H. Burns (Alt.)
U.S. Air Force	R.L. McKinley B.D. Simpson (Alt.)
U.S. Army Aeromedical Research Laboratory	W.A. Ahroon
U.S. Army CERL	D.K. Delaney M.J. White (Alt.)
U.S. Army Research Laboratory, Human Research and Engineering Directorate	T.R. Letowski P. Henry (Alt.)
University of Cincinnati Animal Audiology Clinic/Bioacoustics Lab	P.M. Scheifele D.K. Brown (Alt.)

Individual Experts of the Accredited Standards Committee S3, Bioacoustics, were:

J.R. Bareham
A.J. Brammer
R.F. Burkard

A.J. Campanella
K.D. Kryter
R.L. McKinley
P.D. Schomer

H. Teder
L.A. Wilber
W.A. Yost

Working Group S3/WG 35, Audiometric Equipment, which assisted Accredited Standards Committee S3, Mechanical Vibration and Shock, in the development of this standard, had the following membership:

R.L. Grason, Chair

D. Brewer
C. Clemen
S. Fournier
G. Haas
M. Killion

J.E. Kimmel
T.R. Letowski
D. Mayou
S. Savre
R.S. Schlauch

D.P. Speidel
J. Stewart
R. Veillette
L.A. Wilber

Suggestions for improvements to this standard will be welcomed. They should be sent to Accredited Standards Committee S3, Bioacoustics, in care of the Standards Secretariat of the Acoustical Society of America, 35 Pinelawn Road, Suite 114E, Melville, New York 11747-3177. Telephone: 631-390-0215; Fax: 631-390-0217; E-mail: asastds@aip.org.

American National Standard

Specification for Audiometers

1 Scope, purpose, and applications

1.1 Scope

Audiometers are electronic instruments intended for the testing of human hearing. This standard includes specifications and tolerances for audiometers and standard reference threshold levels for audiometric transducers such as supra-aural, circumaural, and insert earphones; bone vibrators; and loudspeakers.

This standard classifies audiometers according to the type of signal generated (pure tones, speech, or both), mode of operation, and complexity or range of auditory functions tested. Audiometers are also classified by type. Audiometers used for diagnostic assessment are classified as Types 1, 2, or 3, while audiometers having air-conduction facilities only are classified as Type 4.

1.2 Purpose

The purpose of this standard is to ensure that tests of hearing, and particularly tests of hearing threshold, performed on the same individual with different audiometers complying with this standard shall give equivalent results when used under comparable test conditions. Further, the test results shall accurately and validly represent the level difference, if any, that may exist between the threshold of the individual tested and the reference threshold of hearing. This standard covers the general specifications for audiometers that require behavioral responses from a listener, as well as specific requirements governing the functional units of audiometric equipment such as type or types of signal, signal level control mechanisms, and types of transducers. In addition, instruments which are not audiometers but are used primarily for audiometric testing purposes, which incorporate calibrated sound sources (e.g., acoustic immittance systems, otoacoustic emission systems, or narrow-band noise generators), shall, whenever possible, conform to clauses 4 through 10 of this standard.

1.3 Applications

This standard specifies requirements for audiometers used primarily to determine hearing threshold levels with respect to standard reference threshold levels.

2 Normative references

The following referenced documents are indispensable for the application of this standard. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ANSI S1.1-1994 (R 2004) *American National Standard Acoustical Terminology*

ANSI S1.4-1983 (R 2006) *American National Standard Specification for Sound Level Meters*

ANSI/ASA S3.1-1999 (R 2008) *American National Standard Maximum Permissible Ambient Noise Levels for Audiometric Test Rooms*