

ANSI/ASA S3.7-2016
Reaffirmed by ANSI June 5, 2020

AMERICAN NATIONAL STANDARD

Method for Measurement and Calibration of Earphones

ANSI/ASA S3.7-2016

Accredited Standards Committee S3, Bioacoustics

Standards Secretariat
Acoustical Society of America
1305 Walt Whitman Road, Suite 300
Melville, NY 11747

The American National Standards Institute, Inc. (ANSI) is the national coordinator of voluntary standards development and the clearinghouse in the U.S.A. for information on national and international standards.

The Acoustical Society of America (ASA) is an organization of scientists and engineers formed in 1929 to increase and diffuse the knowledge of acoustics and to promote its practical applications.



ANSI/ASA S3.7-2016
(Revision of ANSI/ASA S3.7-1995 (R2008))

AMERICAN NATIONAL STANDARD

Method for Measurement and Calibration of Earphones

Secretariat:

Acoustical Society of America

Approved on August 22, 2016, by:

American National Standards Institute, Inc.

ABSTRACT

This standard describes measurement methods for earphones and earphone transducers using couplers or ear simulators. Guidance is provided for the selection of the appropriate coupler or ear simulator for a given earphone and application. Methods for measurement of calibrated frequency response, input-output linearity, electrical impedance, and non-linear distortion, are described.

AMERICAN NATIONAL STANDARDS ON BIOACOUSTICS

The Acoustical Society of America (ASA) provides the Secretariat for Accredited Standards Committees S1 on Acoustics, S2 on Mechanical Vibration and Shock, S3 on Bioacoustics, S3/SC 1 on Animal Bioacoustics, and S12 on Noise. These committees have wide representation from the technical community (manufacturers, consumers, trade associations, organizations with a general interest, and government representatives). The standards are published by the Acoustical Society of America as American National Standards after approval by their respective Standards Committees and the American National Standards Institute (ANSI).


These standards are developed and published as a public service to provide standards useful to the public, industry, and consumers, and to Federal, State, and local governments.

Each of the Accredited Standards Committees (operating in accordance with procedures approved by ANSI) is responsible for developing, voting upon, and maintaining or revising its own Standards. The ASA Standards Secretariat administers Committee organization and activity and provides liaison between the Accredited Standards Committees and ANSI. After the Standards have been produced and adopted by the Accredited Standards Committees, and approved as American National Standards by ANSI, the ASA Standards Secretariat arranges for their publication and distribution.

An American National Standard implies a consensus of those substantially concerned with its scope and provisions. Consensus is established when, in the judgment of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered and that a concerted effort be made towards their resolution.

The use of an American National Standard is completely voluntary. Their existence does not in any respect preclude anyone, whether he or she has approved the Standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the Standards.

NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this Standard.



Acoustical Society of America
ASA Secretariat
1305 Walt Whitman Road, Suite 300
Melville, New York 11747
Telephone: 1 (631) 390-0215
Fax: 1 (631) 923-2875
E-mail: asastds@acousticalsociety.org

© 2016 by Acoustical Society of America. This standard may not be reproduced in whole or in part in any form for sale, promotion, or any commercial purpose, or any purpose not falling within the provisions of the U.S. Copyright Act of 1976, without prior written permission of the publisher. For permission, address a request to the Standards Secretariat of the Acoustical Society of America.

Contents

1	Scope.....	1
2	Normative references	1
3	Terms and definitions	2
4	Coupler / ear simulator selection	3
	4.1 General	3
	4.2 Earphone types.....	4
5	Measurements of earphone response.....	5
	5.1 Overview	5
	5.2 Standard ambient reference conditions.....	5
	5.3 Measurement system and test environment	5
	5.4 Mounting the earphone or receiver on the coupler or ear simulator	6
	5.5 Frequency response	7
	5.6 Input/output characteristic	9
	5.7 Electrical impedance.....	10
	5.8 Harmonic distortion	12
	Annex A (informative) Test configurations for various insert earphone types	16
	A.1 Acoustic connection to the coupler or ear simulator.....	16
	A.2 The HA-1 configuration.....	16
	A.3 The HA-2 configuration.....	17
	A.4 The HA-2B configuration	18
	A.5 The HA-3 configuration.....	18
	A.6 The HA-4 configuration.....	19
	Annex B (informative) Two-tone distortion measurements.....	21
	B.1 Difference frequency distortion.....	21
	B.2 Procedure for measuring difference frequency distortion.....	22
	B.3 Intermodulation distortion	24
	B.4 Procedure for measuring intermodulation distortion.....	24
	Annex C (informative) Left-right tracking response for stereo earphones	27
	Annex D (informative) Tolerances and measurement uncertainty.....	28
	Bibliography	30

Figures

Figure 1 – High- and low-acoustic-impedance earphones.	4
Figure 2 – Earphone types with respect to coupling to the outer ear.	4
Figure 3 – Test system configuration for earphone or receiver measurements.	5
Figure 4 – Example of an earphone frequency response.	8
Figure 5 – Example of earphone input/output characteristics at several frequencies, showing fundamentals and corresponding total distortion.	10
Figure 6 – Electrical impedance test circuit using a series detection resistor.	11
Figure 7 – Example of the impedance magnitude of an earphone.	12
Figure 8 – Example of Total Harmonic Distortion of an earphone.	14
Figure 9 – Example of Fundamental and Harmonics of an earphone.	15
Figure A.1 – HA-1 configuration.	16
Figure A.2 – HA-2 configuration.	17
Figure A.3 – HA-2B configuration.	18
Figure A.4 – HA-3 configuration.	19
Figure A.5 – HA-4 configuration.	20
Figure B.1 – Definition of difference frequency distortion products and orders.	21
Figure B.2 – Example of Difference Frequency Distortion of an earphone.	22
Figure B.3 – Example of Total Difference Frequency Distortion of an earphone.	23
Figure B.4 – Definition of intermodulation distortion products and orders.	24
Figure B.5 – Example of Intermodulation Distortion of an earphone.	25
Figure B.6 – Example of Total Intermodulation Distortion of an earphone.	26
Figure C.1 – Example of a Left-Right Tracking Response for a stereo earphone.	27
Figure D.1 – Relationship between tolerance limits, corresponding acceptance intervals and the maximum permitted uncertainty of measurement, U_{MAX}	28

Tables

Table 1 – Couplers and ear simulators, associated standards, and earphone applications.3

Table 2 – Decibel conversion to and from MKS (SI) units.9

Table D.1 – Example uncertainty budget.....29

Foreword

[*This Foreword is for information only and is not a part of the American National Standard ANSI/ASA S3.7-2016 American National Standard Method for Measurement and Calibration of Earphones. As such, this Foreword may contain material that has not been subjected to public review or a consensus process. In addition, it does not contain requirements necessary for conformance to the standard.*]

This standard comprises a part of a group of definitions, standards, and specifications for use in bioacoustics. It was developed and approved by Accredited Standards Committee S3 Bioacoustics, under its approved operating procedures. Those procedures have been accredited by the American National Standards Institute (ANSI). The Scope of Accredited Standards Committee S3 is as follows:

Standards, specifications, methods of measurement and test, and terminology in the fields of psychological and physiological acoustics, including aspects of general acoustics which pertain to biological safety, tolerance and comfort.

This standard is a revision of ANSI/ASA S3.7-1995 (R2008) *American National Standard Method for Coupler Calibration of Earphones*. That version of the standard, in addition to coupler calibration methods, contained detailed information about the audiometric ear, the 6 cm³ coupler, and the 2 cm³ coupler. Since that time, IEC 60318 was reorganized into several parts, which now include the same information about the aforementioned couplers (IEC 60318, Parts 1, 3, and 5, respectively). Given that the manufacture of these couplers has changed little if at all in more than 25 years, the fact that the IEC and ANSI/ASA specifications for these couplers are essentially identical, and the fact that manufacturers of these couplers do not produce different versions of the devices to meet alternative versions of the standards, harmonization was deemed appropriate. The first step in this process was the move of the detailed coupler specifications to the nationally adopted international standard (NAIS) ANSI/ASA S3.55, Parts 1, 3, and 5, which correspond directly to their IEC 60318 counterparts. This revision of ANSI/ASA S3.7 focuses upon measurement methods for earphones, and now also points to the appropriate parts of NAIS ANSI/ASA S3.55 where the detailed coupler information now resides. References in other standards that previously pointed to ANSI/ASA S3.7 for a particular coupler will now point to the appropriate NAIS ANSI/ASA S3.55 part instead, as these documents are revised and updated.

Major changes in this document since the last revision include:

- Revised title and scope.
- Updated references, including ANSI/ASA S3.55/IEC 60318, Parts 1, 3, and 5.
- Terms and definitions are updated.
- Reference is made to IEC 60268-7 for earphones for consumer applications.
- Clauses describing couplers and artificial ears are replaced with information about which coupler or ear simulator to use for a particular application, referencing the aforementioned documents.
- Only one measurement method is given. This is fully revised and updated.
- Measurements at constant available power are removed.
- All figures are updated and redrawn.
- Information about adaptors and configurations for insert earphones is updated and moved to an informative annex. This same information also appears in ANSI/ASA S3.22-2014.
- An informative annex has been added describing difference frequency and intermodulation distortion measurements.
- An informative annex has been added regarding the uncertainty of measurement including an example uncertainty budget calculation for a typical earphone test.
- The Bibliography is revised with updated references.

ANSI/ASA S3.7-2016

At the time this Standard was submitted to Accredited Standards Committee S3, Bioacoustics for approval, the membership was as follows:

C.J. Struck, *Chair*
P.B. Nelson, *Vice-Chair*

S.B. Blaeser, *Secretary*

Acoustical Society of America	C.J. Struck
.....	P.B. Nelson (Alt.)
American Academy of Audiology	C. Schweitzer
.....	T. Ricketts (Alt.)
American Academy of Otolaryngology, Head and Neck Surgery, Inc.	R.A. Dobie
.....	A. Kim (Alt.)
American Industrial Hygiene Association	T.K. Madison
.....	D. Driscoll (Alt.)
American Speech-Language-Hearing Association (ASHA)	L.A. Wilber
.....	N. DiSarno (Alt.)
Beltone/GN Resound	S. Petrovic
Council for Accreditation in Occupational Hearing Conservation (CAOHC)	B. Charlton
.....	B. Kirchner (Alt.)
Diagnostic Group	T. McColley
.....	P. Dobrowski (Alt.)
ETS – Lindgren Acoustic Systems	S. Dunlap
.....	D. Winker (Alt.)
Etymotic Research, Inc.	M.C. Killion
.....	J.K. Stewart (Alt.)
Food and Drug Administration	S-C Peng
.....	V. Dasika (Alt.)
G.R.A.S. Sound & Vibration	B. Schustrich
Hearing Industries Association	A. Bopp
.....	M. Jones (Alt.)
International Hearing Society	R. Giles
National Electrical Manufacturers Association, Signaling Protection & Communication Section (NEMA – 3SB)	R. Reiswig
National Hearing Conservation Association	E. Fallon
.....	R. Danielson (Alt.)
National Institute for Occupational Safety and Health (NIOSH)	W.J. Murphy
.....	C.L. Themann (Alt.)

National Institute of Standards and Technology	R.P. Wagner
.....	S. Fick (Alt.)
National Park Service	M. McKenna
.....	K. Frstrup (Alt.)
Ocean Conservation Research	M. Stocker
Starkey Hearing Technologies	D.A. Preves
U.S. Air Force	R.L. McKinley
.....	B.D. Simpson (Alt.)
U.S. Army Aeromedical Research Laboratory	W.A. Ahroon
U.S. Army CERL	D.K. Delaney
.....	M.J. White (Alt.)
U.S. Army Research Laboratory, Human Research and Engineering Directorate	A.A. Scharine
.....	A. Foots (Alt.)
University of Cincinnati Animal Audiology Clinic/Bioacoustics Lab	P.M. Scheifele
.....	D.K. Brown (Alt.)

Individual Experts of the Accredited Standards Committee S3, Bioacoustics, were:

R. F. Burkard	A. J. Campanella	C. J. Struck
M. D. Burkhard	R. L. McKinley	L. A. Wilber
	P. D. Schomer	

Working Group S3/WG 37, Couplers, Ear Simulators, and Earphones, which assisted Accredited Standards Committee S3, Bioacoustics, in the development of this standard, had the following membership.

C. J. Struck, Chair

D. Barrentine	G. Hess	S. Petrovic
D. A. Preves	J. K. Stewart	R. Wagner
D. Wiggins	M. Wille	K. A. Woo

Suggestions for improvements to this standard will be welcomed. They should be sent to Accredited Standards Committee S3, Bioacoustics, in care of the Standards Secretariat of the Acoustical Society of America, 1305 Walt Whitman Road, Suite 300, Melville, New York 11747. Telephone: 631-390-0215; FAX: 631-923-2875; E-mail: asastds@acousticalsociety.org.

Introduction

This standard describes test methods and procedures for the measurement and calibration of earphones and receivers using couplers or ear simulators. Only the electroacoustic response is considered. Digital, wireless, and systems with signal processing are beyond the scope of this standard. Although there are no performance tolerances in this standard, the results of the tests in this standard may be used as the basis for establishing performance tolerances.

The earphone response, as perceived by the listener, is largely dependent upon the insertion gain of the system, i.e., the ratio of the pressure response as measured *in-situ* to the open-ear response in a given sound field (e.g., free field, diffuse field). This correction to an equivalent sound field pressure is beyond the scope of this standard.

Noise isolation afforded by earphones (analogous to hearing protectors) as well as measurements or limit recommendations for the maximum output of earphones in order to protect the hearing of the listener are also beyond the scope of this standard.

Couplers used for earphone measurements are not designed to represent the real ear, but rather only to provide a convenient, stable, and reproducible acoustical load to enable a simple and ready means for the exchange of specifications and physical data. Often, as in production testing, exact correlation between the real-ear and coupler response is not required. Couplers are specified in ANSI/ASA S3.55-2014/Part 1/IEC 60318-1:2009, ANSI/ASA S3.55-2015/Part 3/IEC 60318-3:2014, and ANSI/ASA S3.55-2014/Part 5/IEC 60318-5:2006.

Alternatively, ear simulators (as described in ANSI/ASA S3.25-2009) may be used to provide an acoustic impedance approximating the median adult human ear and equivalent sound pressure at the eardrum. The occluded ear simulator is used in test systems such as manikins. For telephony measurements, the ear simulators described in ITU-T Rec. P.57 are used.

ANSI/ASA S3.36-2012 describes a manikin (a.k.a. head and torso simulator) which is intended to simulate the acoustics of a median human adult, including diffractions affecting the in-situ performance of electroacoustic devices used on or near a person, where the effects of the pinna and realistic positioning of an earphone are also considered.

This is a preview of "ANSI ASA S3.7-2016 (...". [Click here to purchase the full version from the ANSI store.](#)

American National Standard

Method for Measurement and Calibration of Earphones

1 Scope

The electroacoustic test methods described in this standard are primarily for use with circumaural (over-the-ear), supra-aural (against-the-ear) and insert (ear-canal) type earphones, but may be applied to other types as well. Although these methods are generally applicable to earphones for all uses, the most common application of these methods are for earphones intended for hearing aids and audiometric testing and for receivers or other electroacoustic transducers intended for use in an earphone. In all cases, connection to the device under test is analog electrical; Digital, USB, or wireless systems are not considered. For additional requirements and tests for earphones for consumer audio applications, reference is made to IEC 60268-7.

2 Normative references

The following referenced documents are indispensable for the application of this standard. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ANSI/ASA S1.6-2016 American National Standard Preferred Frequencies and Filter Band Center Frequencies for Acoustical Measurements

ANSI/ASA S3.6 American National Standard Specifications for Audiometers

ANSI/ASA S3.20 American National Standard Bioacoustical Terminology

ANSI/ASA S3.22 American National Standard Specification of Hearing Aid Characteristics

ANSI/ASA S3.25-2009 American National Standard for an Occluded Ear Simulator

ANSI/ASA S3.36-2012 American National Standard Specification for a Manikin for *in-situ* Airborne Acoustic Measurements

ANSI/ASA S3.55/Part 1/IEC 60318-1 American National Standard Electroacoustics – Simulators of Human Head and Ear – Part 1: Ear Simulator for the Measurement of Supra-aural and Circumaural Earphones (a nationally adopted international standard)

ANSI/ASA S3.55/Part 3/IEC 60318-3 American National Standard Electroacoustics – Simulators of Human Head and Ear – Part 3: Acoustic Coupler for the Calibration of Supra-aural Earphones Used in Audiometry (a nationally adopted international standard)

ANSI/ASA S3.55/Part 5/IEC 60318-5 MOD American National Standard Electroacoustics – Simulators of Human Head and Ear – Part 5: 2 cm³ Coupler for the Measurement of Hearing Aids and Earphones Coupled to the Ear by Means of Ear Inserts (a modified nationally adopted international standard)

IEC 60263:1982 Scales and sizes for plotting frequency characteristics and polar diagrams

IEC 60268-5:2003 Sound system equipment – Part 5: Loudspeakers