

ANSI S1.40-2006
(Revision of ANSI S1.40-1984)

AMERICAN NATIONAL STANDARD

**Specifications and Verification Procedures for
Sound Calibrators**

ANSI S1.40-2006

Accredited Standards Committee S1, Acoustics

Standards Secretariat
Acoustical Society of America
35 Pinelawn Road, Suite 114 E
Melville, NY 11747-3177

The American National Standards Institute, Inc. (ANSI) is the national coordinator of voluntary standards development and the clearinghouse in the U.S.A. for information on national and international standards.

The Acoustical Society of America (ASA) is an organization of scientists and engineers formed in 1929 to increase and diffuse the knowledge of acoustics and to promote its practical applications.



ANSI S1.40-2006
(Revision of ANSI S1.40-1984)

AMERICAN NATIONAL STANDARD

Specifications and Verification Procedures for Sound Calibrators

Secretariat

Acoustical Society of America

Approved 20 September 2006:

American National Standards Institute, Inc.

Abstract

This Standard specifies performance requirements for coupler-type sound calibrators. It replaces ANSI S1.40-1984 American National Standard Specifications for Acoustical Calibrators, and is technically equivalent (except for the absence of requirements for radio-frequency emissions) to International Standard IEC 60942:2003, *Electroacoustics—Sound calibrators*. The standard specifies performance requirements for the sound pressure level, frequency, and total distortion generated by a sound calibrator. It also provides requirements for the influence of environmental conditions, for electromagnetic compatibility, and for instrument marking and documentation. The standard gives details of the tests necessary to verify that a model of sound calibrator conforms to all the requirements, as well as details of the method for periodic testing of a sound calibrator. The tests require the determination of the actual uncertainties of measurement which are to not exceed the maximum uncertainties allowed for laboratory measurements.

AMERICAN NATIONAL STANDARDS ON ACOUSTICS

The Acoustical Society of America (ASA) provides the Secretariat for Accredited Standards Committees S1 on Acoustics, S2 on Mechanical Vibration and Shock, S3 on Bioacoustics, and S12 on Noise. These committees have wide representation from the technical community (manufacturers, consumers, trade associations, organizations with a general interest, and government representatives). The standards are published by the Acoustical Society of America as American National Standards after approval by their respective Standards Committees and the American National Standards Institute.


These standards are developed and published as a public service to provide standards useful to the public, industry, and consumers, and to Federal, State, and local governments.

Each of the accredited Standards Committees [operating in accordance with procedures approved by American National Standards Institute (ANSI)] is responsible for developing, voting upon, and maintaining or revising its own Standards. The ASA Standards Secretariat administers Committee organization and activity and provides liaison between the Accredited Standards Committees and ANSI. After the Standards have been produced and adopted by the Accredited Standards Committees, and approved as American National Standards by ANSI, the ASA Standards Secretariat arranges for their publication and distribution.

An American National Standard implies a consensus of those substantially concerned with its scope and provisions. Consensus is established when, in the judgment of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered and that a concerted effort be made towards their resolution.

The use of an American National Standard is completely voluntary. Their existence does not in any respect preclude anyone, whether he or she has approved the Standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the Standards.

NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this Standard.



Acoustical Society of America
ASA Secretariat
35 Pinelawn Road, Suite 114E
Melville, New York 11747-3177
Telephone: 1 (631) 390-0215
Fax: 1 (631) 390-0217
E-mail: asastds@aip.org

© 2006 by Acoustical Society of America. This Standard may not be reproduced in whole or in part in any form for sale, promotion, or any commercial purpose, or any purpose not falling within the provisions of the U.S. Copyright Act of 1976, without prior written permission of the publisher. For permission, address a request to the Standards Secretariat of the Acoustical Society of America.

Contents

1	Scope	1
2	Normative references	2
3	Terms and definitions	2
4	Reference environmental conditions	4
5	Requirements	4
5.1	General	4
5.2	Sound pressure level	6
5.3	Frequency	8
5.4	Influence of static pressure, air temperature, and humidity	9
5.5	Total distortion	10
5.6	Power supply	11
5.7	Specification and calibration of microphones	11
5.8	Electromagnetic compatibility	12
6	Instrument marking and documentation	13
6.1	Marking	13
6.2	Individual calibration chart for a class LS sound calibrator	13
6.3	Instruction Manual	13
Annex A	(normative) Verification	16
A.1	Introduction	16
A.2	Verification sample size	16
A.3	Principal values	17
A.4	Marking of the sound calibrator and supplied documentation	17
A.5	Performance tests at and around reference environmental conditions	17
A.6	Environmental tests	23
A.7	Electromagnetic compatibility	32
Annex B	(normative) Periodic tests	35
B.1	Introduction	35
B.2	Submission for test	35
B.3	Preliminary inspection	35
B.4	Performance tests	36
B.5	Calibration of the sound calibrator with other models of microphone	38
B.6	Documentation	39
B.7	Bibliography for Annex B	40

Tables

Table 1 — Tolerance limits for sound pressure level and short-term level fluctuation, at and around reference environmental conditions	7
Table 2 — Tolerance limits for the influence of supply voltage on sound pressure level, at and around reference environmental conditions	7
Table 3 — Tolerance limits for frequency, at and around reference environmental conditions.....	8
Table 4 — Tolerance limits for sound pressure level, over the specified range of environmental conditions	9
Table 5 — Tolerance limits for frequency, over the specified range of environmental conditions.....	10
Table 6 — Maximum total distortion.....	11
Table A.1 — Maximum permitted expanded uncertainty of measurement for sound pressure level and short-term level fluctuation, at and around reference environmental conditions	20
Table A.2 — Maximum permitted expanded uncertainty of measurement for frequency, at and around reference environmental conditions	21
Table A.3 — Maximum permitted expanded uncertainty of measurement for total distortion, over the appropriate range of environmental conditions.....	22
Table A.4 — Maximum permitted expanded uncertainty of measurement for sound pressure level, over the specified range of environmental conditions.....	24
Table A.5 — Maximum permitted expanded uncertainty of measurement for frequency, over the specified range of environmental conditions.....	28

Foreword

[This Foreword is for information only, and is not an integral part of the American National Standard ANSI S1.40—2006 American National Standard Specifications and Verification Procedures for Sound Calibrators.]

This Standard comprises a part of a group of definitions, standards, and specifications for use in acoustics. It was developed and approved by Accredited Standards Committee S1 Acoustics, under its approved operating procedures. Those procedures have been accredited by the American National Standards Institute (ANSI). The Scope of Accredited Standards Committee S1 is as follows:

Standards, specifications, methods of measurement and test, and terminology in the field of physical acoustics, including architectural acoustics, electroacoustics, sonics and ultrasonics, and underwater sound, but excluding those aspects which pertain to biological safety, tolerances, and comfort.

This Standard replaces ANSI S1.40-1984 (R2001), American National Standard Specifications for Acoustical Calibrators.

This Standard is technically equivalent (except for the absence of requirements regarding radio-frequency emissions from calibrators) to International Standard IEC 60942:2003, Electroacoustics – Sound calibrators.

This Standard includes two normative annexes that are considered to be integral parts of this Standard.

At the time this Standard was submitted to Accredited Standards Committee S1, Acoustics for approval, the membership was as follows:

J.P. Seiler, *Chair*

G.S.K. Wong, *Vice-Chair*

S.B. Blaeser, *Secretary*

Acoustical Society of America	J.P. Seiler
.....	G.S.K. Wong (Alt.)
Air-Conditioning and Refrigeration Institute	S. Lind
.....	D. Brown (Alt.)
American Industrial Hygiene Association	D. Driscoll
.....	D. Sandfort (Alt.)
Bruel & Kjaer Instruments	M. Alexander
.....	J. Chou (Alt.)
Campanella Associates	A.J. Campanella
Information Technology Industry Council	R.D. Hellweg
.....	J. Rosenberg (Alt.)
International Association of Geophysical Contractors	D. Hedgeland
National Council of Acoustical Consultants	G.E. Winzer
.....	E. Logsdon (Alt.)

National Institute for Occupational Safety and Health	C.A. Kardous W.J. Murphy (Alt.)
National Institute of Standards & Technology (NIST)	V. Nedzelnitsky D.J. Evans (Alt.)
PCB Group	K. Cox L. Harbaugh (Alt.)
Quest Technologies, Inc.	P. Battenberg M. Wurm (Alt.)
Scantek, Inc.	R.J. Peppin M. Buzduga (Alt.)
U.S. Air Force	R. McKinley
U.S. Army Construction Engineering Research Laboratory	L. Pater M. White (Alt.)
U.S. Army Human Research and Engineering Directorate	T.R. Letowski A. Scharine (Alt.)

Individual Experts of Accredited Standards Committee S1, Acoustics,

S.L. Ehrlich
K.M. Eldred
W.J. Galloway

W.W. Lang
A.H. Marsh

P.D. Schomer
G.S.K. Wong
L. Wu

Working Group S1/WG 25, Specifications for Sound Calibrators, which assisted Accredited Standards Committee S1, Acoustics, in the development of this Standard, had the following membership.

P. Battenberg, Chair

P. Hanes
A.H. Marsh

V. Nedzelnitsky

M.P. Valoski
L. Wu

Suggestions for improvements of this Standard will be welcomed. They should be sent to Accredited Standards Committee S1, Acoustics, in care of the Standards Secretariat of the Acoustical Society of America, 35 Pinelawn Road, Suite 114E, Melville, New York 11747-3177. Telephone: 631-390-0215; FAX: 631-390-0217; E-mail: asastds@aip.org

American National Standard

Specifications and Verification Procedures for Sound Calibrators

1 Scope

Sound calibrators generate known sound pressure levels at one or more frequencies in a coupler into which a specified model of microphone is inserted in a specified configuration.

NOTE An example of a specified configuration may be with or without a grid.

This Standard specifies performance requirements and verification procedures for three classes of coupler-type sound calibrators:

- Class LS: Laboratory Standard calibrator with the smallest tolerance limits.
- Class 1: Sound calibrators that are generally intended for field use with class 1 sound level meters and similar instruments.
- Class 2: Sound calibrators that are generally intended for field use with class 2 sound level meters, dosimeters, and similar instruments.

For class LS calibrators, the Standard requires the use of a laboratory standard microphone as specified in ANSI S1.15-1997/Part 1. For class 1 and class 2 calibrators, the Standard requires the use of a working standard microphone as specified in IEC 61094-4:1995.

Tolerance limits in this Standard include maximum permitted expanded uncertainties of measurement as well as the tolerance limits allowed for design and manufacturing.

This Standard applies to sound calibrators that are used to check or adjust the sensitivity of sound measuring instruments and systems including conventional, integrating-averaging and integrating sound level meters, and personal noise dosimeters.

This Standard does not include requirements for the Instruction Manual to provide data for use in adjusting a sound level indicated by a sound level meter, or equivalent instrument, in response to application of a sound calibrator to the equivalent free-field or random-incidence sound level.

Sound calibrators may provide two or more sound pressure levels and operate at two or more frequencies. These multiple-level / multiple-frequency calibrators have a single class designation for all combinations of levels and frequencies for which the Instruction Manual states conformance to the requirements of this Standard.

Additional useful functions such as tonebursts may be provided. Requirements for these additional functions are not included in this Standard.