

This is a preview of "BS EN 60601-1-2:2007". [Click here to purchase the full version from the ANSI store.](#)

MARCH 2010

# Medical electrical equipment —

**Part 1-2: General requirements for basic  
safety and essential performance —  
Collateral standard: Electromagnetic  
compatibility — Requirements and tests**

ICS 11.040.01; 33.100.10; 33.100.20

This is a preview of "BS EN 60601-1-2:2007". [Click here to purchase the full version from the ANSI store.](#)

This British Standard is the UK implementation of EN 60601-1-2:2007, incorporating corrigendum March 2010. It was derived by CENELEC from IEC 60601-1-2:2007. It supersedes BS EN 60601-1-2:2002, which will be declared obsolescent and will be withdrawn on publication of the revised BS EN 60601-2 series.

The CENELEC common modifications have been implemented at the appropriate places in the text. The start and finish of each common modification is indicated in the text by tags **[C]** **[C]**.

The UK participation in its preparation was entrusted by Technical Committee CH/62, Electromedical equipment in medical practice, to Subcommittee CH/62/1, Common aspects of electrical equipment used in medical practice.

A list of organizations represented on this subcommittee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 28 September 2007

© BSI 2010

**Amendments/corrigenda issued since publication**

Date	Comments
31 July 2010	Implementation of CENELEC corrigendum March 2010. Date of withdrawal added to Foreword

**EUROPÄISCHE NORM**

July 2007

ICS 11.040.01; 33.100.10; 33.100.20

Supersedes EN 60601-1-2:2001 + A1:2006  
Incorporating corrigendum March 2010

English version

**Medical electrical equipment -  
Part 1-2: General requirements for basic safety  
and essential performance -  
Collateral standard: Electromagnetic compatibility -  
Requirements and tests  
(IEC 60601-1-2:2007, modified)**

Appareils électromédicaux -  
Partie 1-2: Exigences générales  
pour la sécurité de base  
et les performances essentielles -  
Norme collatérale:  
Compatibilité électromagnétique -  
Exigences et essais  
(CEI 60601-1-2:2007, modifiée)

Medizinische elektrische Geräte -  
Teil 1-2: Allgemeine Festlegungen  
für die Sicherheit einschließlich  
der wesentlichen Leistungsmerkmale -  
Ergänzungsnorm:  
Elektromagnetische Verträglichkeit -  
Anforderungen und Prüfungen  
(IEC 60601-1-2:2007, modifiziert)

This European Standard was approved by CENELEC on 2007-04-11. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

# CENELEC

European Committee for Electrotechnical Standardization  
Comité Européen de Normalisation Electrotechnique  
Europäisches Komitee für Elektrotechnische Normung

**Central Secretariat: rue de Stassart 35, B - 1050 Brussels**

This is a preview of "BS EN 60601-1-2:2007". [Click here to purchase the full version from the ANSI store.](#)

The text of document 62A/560/FDIS, future edition 3 of IEC 60601-1-2, prepared by SC 62A, Common aspects of electrical equipment used in medical practice, of IEC TC 62, Electrical equipment in medical practice, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 60601-1-2 on 2007-04-11.

The following date was fixed:

- latest date by which the EN has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2008-02-01
- latest date by which the national standards conflicting with the EN have to be withdrawn (dow) 2012-06-01

This European Standard supersedes EN 60601-1-2:2001 and its amendment A1:2006.

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and covers essential requirements of EC Directive 93/42/EEC. See Annex ZZ.

This European Standard constitutes a collateral standard to EN 60601-1:2006, hereafter referred to as the general standard.

This EN 60601-1-2 was revised to structurally align it with EN 60601-1:2006 and to implement the decision of IEC SC 62A that the clause numbering structure of collateral standards written to EN 60601-1:2006 would adhere to the form specified in ISO/IEC Directives, Part 2:2004. The principle technical changes are in Clause 4, which now recognizes that there is a general requirement for a risk management process in EN 60601-1:2006.

In the 60601 series of publications, collateral standards specify general requirements for safety applicable to

- a subgroup of MEDICAL ELECTRICAL EQUIPMENT (e.g. radiological equipment); or
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the general standard (e.g. alarm systems).

In this collateral standard the following print types are used:

- requirements and definitions: in roman type;
- *test specifications: in italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS COLLATERAL STANDARD OR AS NOTES: IN SMALL CAPITALS.

NOTE Defined terms are not printed in SMALL CAPITALS in Table 1 through Table 8, in the tables in Annex C and in statements required to appear in the technical description or instructions for use because they are intended for the OPERATOR or RESPONSIBLE ORGANIZATION, who may not be familiar with the defined terms of EN 60601 standards.

In referring to the structure of this standard, the term

- “clause” means one of the six numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 6 includes Subclauses 6.1, 6.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 6.1, 6.2 and 6.2.1 are all subclauses of Clause 6).

This is a preview of "BS EN 60601-1-2:2007". [Click here to purchase the full version from the ANSI store.](#)

number. References to subclauses within this standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Clauses, subclauses, items and definitions for which a rationale is provided in informative Annex A are marked with an asterisk (\*).

Annexes ZA and ZZ have been added by CENELEC.

---

### **Endorsement notice**

The text of the International Standard IEC 60601-1-2:2007 was approved by CENELEC as a European Standard with agreed common modifications.

---

This is a preview of "BS EN 60601-1-2:2007". [Click here to purchase the full version from the ANSI store.](#)

## CONTENTS

INTRODUCTION.....	7
1 Scope, object and related standards .....	9
1.1 * Scope .....	9
1.2 Object .....	9
1.3 Related standards .....	9
2 Normative references .....	9
3 Terms and definitions .....	11
4 General requirements .....	14
4.1 General requirements for ELECTROMAGNETIC COMPATIBILITY of ME EQUIPMENT and ME SYSTEMS.....	14
4.2 * SINGLE FAULT CONDITION for ME EQUIPMENT.....	15
5 Identification, marking and documents .....	15
5.1 Marking on the outside of ME EQUIPMENT or ME EQUIPMENT parts.....	15
5.2 ACCOMPANYING DOCUMENTS .....	16
6 ELECTROMAGNETIC COMPATIBILITY .....	38
6.1 EMISSIONS.....	38
6.2 IMMUNITY .....	41
Annex A (informative) General guidance and rationale.....	57
Annex B (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS.....	87
Annex C (informative) Example completion of Table 1 through Table 8 .....	90
Annex D (informative) Guidance in classification according to CISPR 11.....	102
Annex E (informative) Guidance in the application of IEC 60601-1-2 to particular standards .....	105
Annex F (informative) ELECTROMAGNETIC ENVIRONMENTS .....	108
Annex G (informative) Guidance for determining if electrical equipment that is not ME EQUIPMENT and that is used in an ME SYSTEM is exempt from the EMC testing requirements of this collateral standard .....	109
Annex H (informative) Mapping between the elements of the second edition of IEC 60601-1-2 as amended and IEC 60601-1-2:2007 .....	111
Annex ZA (normative) Normative references to international publications with their corresponding European publications .....	123
Annex ZZ (informative) Coverage of Essential Requirements of EC Directives .....	125
Bibliography .....	119
Index of defined terms used in this collateral standard.....	121
Figure 1 – Instructions for completing Table 1 for CISPR 11 ME EQUIPMENT and ME SYSTEMS .....	20
Figure 2 – Instructions for completing Table 1 for CISPR 14 and CISPR 15 ME EQUIPMENT .....	21
Figure 3 – Instructions for completing Table 2 .....	24

This is a preview of "BS EN 60601-1-2:2007". [Click here to purchase the full version from the ANSI store.](#)

Figure 4 – Instructions for completing Table 3 and Table 5 for LIFE-SUPPORTING ME EQUIPMENT and ME SYSTEMS .....	30
Figure 5 – Instructions for completing Table 4 and Table 6 for ME EQUIPMENT and ME SYSTEMS that are not LIFE-SUPPORTING .....	31
Figure A.1 – Example of cable arrangement for radiated IMMUNITY test.....	85
Figure A.2 – Examples showing maximum dimension for ME EQUIPMENT with one and with two cables .....	86
Figure G.1 – Procedure for determining if electrical equipment that is not ME EQUIPMENT and that is used in an ME SYSTEM is exempt from the EMC testing requirements of this collateral standard .....	110
Table 1 – Guidance and MANUFACTURER’S declaration – ELECTROMAGNETIC EMISSIONS – for all ME EQUIPMENT and ME SYSTEMS .....	19
Table 2 – Guidance and MANUFACTURER’S declaration []– electromagnetic IMMUNITY – for all ME EQUIPMENT and ME SYSTEMS .....	23
Table 3 – Guidance and MANUFACTURER’S declaration – electromagnetic IMMUNITY – for LIFE-SUPPORTING ME EQUIPMENT and ME SYSTEMS .....	26
Table 4 – Guidance and MANUFACTURER’S declaration – electromagnetic IMMUNITY – for ME EQUIPMENT and ME SYSTEMS that are not LIFE-SUPPORTING .....	27
Table 5 – Recommended separation distances between portable and mobile RF communications equipment and the ME EQUIPMENT or ME SYSTEM – for LIFE-SUPPORTING ME EQUIPMENT and ME SYSTEMS .....	28
Table 6 – Recommended separation distances between portable and mobile RF communications equipment and the ME EQUIPMENT or ME SYSTEM – for ME EQUIPMENT and ME SYSTEMS that are not LIFE-SUPPORTING .....	29
Table 7 – Guidance and MANUFACTURER’S declaration – electromagnetic IMMUNITY – for LIFE-SUPPORTING ME EQUIPMENT and ME SYSTEMS that are specified for use only in a shielded location.....	35
Table 8 – Guidance and MANUFACTURER’S declaration – electromagnetic IMMUNITY – for ME EQUIPMENT and ME SYSTEMS that are not LIFE-SUPPORTING and are specified for use only in a shielded location.....	36
Table 9 – Modulation frequency, PHYSIOLOGICAL SIMULATION FREQUENCY, and OPERATING FREQUENCY.....	45
Table 10 – IMMUNITY TEST LEVELS for voltage dips .....	54
Table 11 – IMMUNITY TEST LEVEL for voltage interruption.....	54
Table B.1 – Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts.....	87
Table B.2 – ACCOMPANYING DOCUMENTS, instructions for use.....	88
Table B.3 – ACCOMPANYING DOCUMENTS, technical description.....	89
Table C.1 – Example (1) of completed Table 1 .....	90
Table C.2 – Example (2) of completed Table 1 .....	91
Table C.3 – Example (3) of completed Table 1 .....	92
Table C.4 – Example of completed Table 2 .....	93
Table C.5 – Example (1) test, IMMUNITY and COMPLIANCE LEVELS .....	94
Table C.6 – Example of completed Table 3 .....	95
Table C.7 – Example of completed Table 5 .....	96
Table C.8 – Example of completed Table 4 .....	97
Table C.9 – Example of completed Table 6 .....	98
Table C.10 – Example (2) test, IMMUNITY and COMPLIANCE LEVELS .....	98

This is a preview of "BS EN 60601-1-2:2007". [Click here to purchase the full version from the ANSI store.](#)

Table C.11 – Example of completed Table 7 .....	99
Table C.12 – Example (3) test, IMMUNITY and COMPLIANCE LEVELS .....	100
Table C.13 – Example of completed Table 8 .....	101
Table F.1 – ELECTROMAGNETIC ENVIRONMENTS.....	108
Table H.1 – Mapping between the elements of the second edition of IEC 60601-1-2 as amended and IEC 60601-1-2:2007 .....	111



This is a preview of "BS EN 60601-1-2:2007". [Click here to purchase the full version from the ANSI store.](#)

## INTRODUCTION

The need for establishing specific ELECTROMAGNETIC COMPATIBILITY standards for MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS is well recognized.

In particular, the existence of ELECTROMAGNETIC EMISSION standards is essential for the protection of:

- safety services;
- other MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS;
- electrical equipment that is not ME EQUIPMENT (e.g. computers);
- telecommunications (e.g. radio/TV, telephone, radio-navigation).

Of even more importance, the existence of ELECTROMAGNETIC IMMUNITY standards is essential to assure safety of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS. ELECTROMAGNETIC COMPATIBILITY (see Definition 3.4) differs from other aspects of safety covered by IEC 60601-1 because the electromagnetic phenomena exist, with varying degrees of severity, in the normal use environment of all MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS and by definition the equipment must “perform satisfactorily” within its intended environment in order to establish ELECTROMAGNETIC COMPATIBILITY. This means that the conventional single fault approach to safety is not appropriate for application to ELECTROMAGNETIC COMPATIBILITY standards. The ELECTROMAGNETIC DISTURBANCE environment can be compared to ambient temperature, humidity and atmospheric pressure. MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS may experience environmental conditions within the expected range at any time, and for extended periods of time. As with atmospheric pressure and humidity, the OPERATOR of the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM may not be aware of ambient levels on a continuous basis. The IMMUNITY TEST LEVELS specified in this collateral standard (IEC 60601 TEST LEVELS) represent the range found in the general medical use environment. Therefore, under these conditions, the performance of the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM would also be expected to be normal.

MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS are used in the practice of medicine because they provide needed FUNCTIONS. If MEDICAL ELECTRICAL EQUIPMENT or a MEDICAL ELECTRICAL SYSTEM does not provide its needed FUNCTION, because of a lack of IMMUNITY to events expected in the normal use environment, this interferes with the practice of medicine and cannot be considered an acceptable situation.

This edition recognizes that there is a shared responsibility between MANUFACTURERS, RESPONSIBLE ORGANIZATIONS and OPERATORS to ensure that MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS are designed and operated as intended. The MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM MANUFACTURER’S responsibility is to design and manufacture to meet the requirements of this collateral standard and to disclose information to the RESPONSIBLE ORGANIZATION or OPERATOR so that a compatible ELECTROMAGNETIC ENVIRONMENT can be maintained in order that the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM will perform as intended.

This is a preview of "BS EN 60601-1-2:2007". [Click here to purchase the full version from the ANSI store.](#)

Because the practice of medicine involves many specialities, there will by necessity be MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS that are designed to perform a variety of FUNCTIONS. Some FUNCTIONS involve, for example, measurement of signals from a PATIENT that are of very low levels when compared to ELECTROMAGNETIC NOISE levels that can be coupled into MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS during the ELECTROMAGNETIC IMMUNITY testing specified in this collateral standard. Because of the proven benefits of many such MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, this collateral standard allows the IMMUNITY TEST LEVELS to be lowered, provided there is sufficient justification based on physical, technological or physiological limitations. In this case, the MANUFACTURER is required to disclose the levels at which the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM meets the performance requirements of this collateral standard and to specify the characteristics of the ELECTROMAGNETIC use environment and how this environment is established, in which the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM will perform as intended.

This collateral standard also recognizes that for certain environments, higher IMMUNITY LEVELS may be required. Research necessary to determine how to identify the environments that may require higher IMMUNITY LEVELS, as well as what the levels should be, is in progress.

Finally, this collateral standard recognizes that for LIFE-SUPPORTING MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, higher levels of IMMUNITY are necessary in order to establish a broader safety margin, even for use in the general medical use environment. Therefore, this collateral standard specifies additional requirements for LIFE-SUPPORTING MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS.

This collateral standard is based on existing IEC standards prepared by subcommittee 62A, technical committee 77 (electromagnetic compatibility between electrical equipment including networks) and CISPR (International special committee on radio interference).

The ELECTROMAGNETIC COMPATIBILITY requirements specified by this collateral standard are generally applicable to MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS as defined in 3.63 and 3.64 in the general standard. For certain types of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, these requirements may need to be modified by the special requirements of a particular standard. Writers of particular standards are encouraged to refer to Annex E for guidance in the application of this collateral standard.

This is a preview of "BS EN 60601-1-2:2007". [Click here to purchase the full version from the ANSI store.](#)

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility – Requirements and tests

## 1 Scope, object and related standards

### 1.1 \* Scope

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, hereafter referred to as ME EQUIPMENT and ME SYSTEMS.

This collateral standard applies to ELECTROMAGNETIC COMPATIBILITY of ME EQUIPMENT and ME SYSTEMS.

### 1.2 Object

The object of this collateral standard is to specify general requirements and tests for ELECTROMAGNETIC COMPATIBILITY of ME EQUIPMENT and ME SYSTEMS. They are in addition to the requirements of the general standard and serve as the basis for particular standards.

### 1.3 Related standards

#### 1.3.1 IEC 60601-1

For ME EQUIPMENT and ME SYSTEMS, this collateral standard complements IEC 60601-1.

When referring to IEC 60601-1 or to this collateral standard, either individually or in combination, the following conventions are used:

- "the general standard" designates IEC 60601-1 alone;
- "this collateral standard" designates IEC 60601-1-2 alone;
- "this standard" designates the combination of the general standard and this collateral standard.

#### 1.3.2 Particular standards

A requirement in a particular standard takes priority over the corresponding requirement in this collateral standard.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60417, *Graphical symbols for use on equipment*