

This is a preview of "BS EN 45502-1:2015". [Click here to purchase the full version from the ANSI store.](#)

BS EN 45502-1:2015



BSI Standards Publication

Implants for surgery — Active implantable medical devices

Part 1: General requirements for safety, marking and for information to be provided by the manufacturer

bsi.

...making excellence a habit.™

This is a preview of "BS EN 45502-1:2015". [Click here to purchase the full version from the ANSI store.](#)

This British Standard is the UK implementation of EN 45502-1:2015. It supersedes BS EN 45502-1:1998 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/150/2, Cardiovascular implants.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2015.
Published by BSI Standards Limited 2015

ISBN 978 0 580 72639 2

ICS 11.040.40

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 June 2015.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

This is a preview of "BS EN 45502-1:2015". [Click here to purchase the full version from the ANSI store.](#)

EUROPÄISCHE NORM

May 2015

ICS 11.040.01

Supersedes EN 45502-1:1997

English Version

Implants for surgery - Active implantable medical devices - Part 1: General requirements for safety, marking and for information to be provided by the manufacturer

Dispositifs médicaux implantables actifs - Partie 1: Règles générales de sécurité, marquage et informations fournies par le fabricant


Aktive implantierbare medizinische Geräte - Teil 1: Allgemeine Festlegungen für die Sicherheit, Aufschriften und vom Hersteller zur Verfügung zu stellende Informationen

This European Standard was approved by CENELEC on 20 April 2015. CEN and CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN and CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN and CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN and CENELEC members are the national standards bodies and national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

This is a preview of "BS EN 45502-1:2015". [Click here to purchase the full version from the ANSI store.](#)

Contents

Page

Foreword.....	4
Introduction	5
1 Scope	6
2 Normative references	6
3 Terms and definitions	7
4 Symbols and abbreviations (optional).....	12
5 General requirements for ACTIVE IMPLANTABLE MEDICAL DEVICES	12
6 Requirements for particular ACTIVE IMPLANTABLE MEDICAL DEVICES	14
7 General arrangement of the packaging	14
8 General MARKINGS for ACTIVE IMPLANTABLE MEDICAL DEVICES	14
9 MARKINGS on the SALES PACKAGING.....	15
10 Construction of the SALES PACKAGING.....	16
11 MARKINGS on the STERILE PACK.....	17
12 Construction of the NON-REUSABLE PACK	18
13 MARKINGS on the ACTIVE IMPLANTABLE MEDICAL DEVICE	18
14 Protection from unintentional biological effects being caused by the ACTIVE IMPLANTABLE MEDICAL DEVICE	19
15 Protection from HARM to the patient or user caused by external physical features of the ACTIVE IMPLANTABLE MEDICAL DEVICE	21
16 Protection from HARM to the patient caused by electricity	21
17 Protection from HARM to the patient caused by heat.....	22
18 Protection from ionizing radiation released or emitted from the ACTIVE IMPLANTABLE MEDICAL DEVICE	22
19 Protection from unintended effects caused by the ACTIVE IMPLANTABLE MEDICAL DEVICE	23
20 Protection of the ACTIVE IMPLANTABLE MEDICAL DEVICE from damage caused by external defibrillators	24
21 Protection of the ACTIVE IMPLANTABLE MEDICAL DEVICE from changes caused by electrical fields applied directly to the patient	27
22 Protection of the ACTIVE IMPLANTABLE MEDICAL DEVICE from changes caused by miscellaneous medical treatments	28
23 Protection of the ACTIVE IMPLANTABLE MEDICAL DEVICE from mechanical forces	29
24 Protection of the ACTIVE IMPLANTABLE MEDICAL DEVICE from damage caused by electrostatic discharge	30
25 Protection of the ACTIVE IMPLANTABLE MEDICAL DEVICE from damage caused by atmospheric pressure changes	31
26 Protection of the ACTIVE IMPLANTABLE MEDICAL DEVICE from damage caused by temperature changes	31
27 Protection of the ACTIVE IMPLANTABLE MEDICAL DEVICE from electromagnetic non-ionizing radiation.....	31
28 Accompanying documentation	32

This is a preview of "BS EN 45502-1:2015". [Click here to purchase the full version from the ANSI store.](#)

Annex A (informative) General guidance and rationale	37
Annex ZA (normative).....	47
Annex ZZ (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 90/385/EEC on Active Implantable Medical Devices	48
Bibliography	59

Figures

Figure 1 – Damped sinus defibrillation waveform	25
Figure 2 – Defibrillation test voltage generator.....	25
Figure 3 – Timing sequence used for Test 1 and Test 2.....	26
Figure 4 – Test setup for truncated exponential DEFIBRILLATION waveform	26
Figure 5 – Biphasic DEFIBRILLATION waveform for Test 2	27
Figure A.1 – RLC implementation for generating a damped sinus defibrillation waveform	42
Figure A.2 – Positioning and scanning the ultrasound field exposure upon the implantable part	44

Tables

Table 1 – Timing parameters of test signal for Test 2.....	26
--	----

This is a preview of "BS EN 45502-1:2015". [Click here to purchase the full version from the ANSI store.](#)

Foreword

This document (EN 45502-1:2015) has been prepared by CEN/CLC/JWG AIMD "CEN/CENELEC Joint Working Group on Active Implantable Medical Devices".

The following dates are fixed:

- latest date by which this document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2016-04-20
- latest date by which the national standards conflicting with this document have to be withdrawn (dow) 2018-04-20

This document supersedes EN 45502-1:1997.

EN 45502-1:2015 includes the following significant technical changes with respect to EN 45502-1:1997:

- a) update according to the modified AIMD;
- b) update of normative references to the "state of the art";
- c) implementation of usability issues;
- d) implementation of links to information security;
- e) implementation of elements according to EN 14971:2012;
- f) improvement of Clause 14 "Protection from unintentional biological effects being caused by the active implantable medical device";
- g) improvement of Clause 20 "Protection of the active implantable medical device from damage caused by external defibrillators";
- h) improvement of Clause 22 "Protection of the active implantable medical device from changes caused by miscellaneous medical treatments" especially for ultrasonic diagnostic devices.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For the relationship with EU Directive(s) see informative Annex ZZ, which is an integral part of this document.

This is a preview of "BS EN 45502-1:2015". [Click here to purchase the full version from the ANSI store.](#)

Introduction

This European Standard specifies general requirements for ACTIVE IMPLANTABLE MEDICAL DEVICES to provide basic assurance of safety for both patients and users.

To minimize the likelihood of a device being misused, this European Standard also details comprehensive requirements for MARKINGS and for other information to be supplied as part of the documentation with any ACTIVE IMPLANTABLE MEDICAL DEVICE.

For particular types of ACTIVE IMPLANTABLE MEDICAL DEVICE, the general requirements can be supplemented or modified by the requirements of other parts of EN 45502. A requirement of a particular part of EN 45502 takes priority over the corresponding requirement of this general part of EN 45502. Where particular parts of EN 45502 exist, this general standard of EN 45502 is not intended to be used alone. Special care is required when applying this general standard part of EN 45502 alone to ACTIVE IMPLANTABLE MEDICAL DEVICES for which no particular part of EN 45502 has yet been published.

This is a preview of "BS EN 45502-1:2015". [Click here to purchase the full version from the ANSI store.](#)

1 Scope

This part of EN 45502 specifies requirements that are generally applicable to ACTIVE IMPLANTABLE MEDICAL DEVICES.

NOTE 1 For particular types of ACTIVE IMPLANTABLE MEDICAL DEVICES, these general requirements are supplemented or modified by the requirements of particular standards which form additional parts of this European Standard.

The tests that are specified in EN 45502 are type tests and are to be carried out on samples of an ACTIVE IMPLANTABLE MEDICAL DEVICE to show compliance.

This part of EN 45502 is applicable not only to ACTIVE IMPLANTABLE MEDICAL DEVICES that are electrically powered but also to those powered by other energy sources (for example by gas pressure or by springs).

This part of EN 45502 is also applicable to some non-implantable parts and accessories of the ACTIVE IMPLANTABLE MEDICAL DEVICES.

NOTE 2 The device that is commonly referred to as an ACTIVE IMPLANTABLE MEDICAL DEVICE can be a single device, a combination of devices, or a combination of a device or devices and one or more accessories. Not all of these parts are required to be either partially or totally implantable, but there is a need to specify some requirements of non-implantable parts and accessories if they could affect the safety or performance of the implantable device.

NOTE 3 In this part of EN 45502, terms printed in small capital letters are used as defined in Clause 3. Where a defined term is used as a qualifier in another term, it is not printed in small capital letters unless the concept thus qualified is also defined.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 60068-2-14:2009, *Environmental testing – Part 2 14: Tests – Test N: Change of temperature (IEC 60068-2-14:2009)*

EN 60068-2-27:2009, *Environmental testing – Part 2 27: Tests – Test Ea and guidance: Shock (IEC 60068-2-27:2008)*

EN 60068-2-47:2005, *Environmental testing – Part 2 47: Tests – Mounting of specimens for vibration, impact and similar dynamic tests (IEC 60068-2-47:2005)*

EN 60068-2-64:2008, *Environmental testing – Part 2 64: Tests – Test Fh: Vibration, broadband random and guidance (IEC 60068-2-64:2008)*

EN 60601-1:2006, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance (IEC 60601-1:2005)*

EN 60601-1:2006/A1:2013, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance (IEC 60601-1:2005/A1:2012)*

EN 62304:2006, *Medical devices software – Software life-cycle processes (IEC 62304:2006)*

EN 62366:2008, *Medical devices – Application of usability engineering to medical devices (IEC 62366:2007)*

EN ISO 10993-1:2009, *Biological testing of medical devices – Part 1: Evaluation and testing within a risk management process (ISO 10993-1:2003)*

EN ISO 11607-1:2006, *Packaging for terminally sterilized medical devices – Part 1: Requirements for materials, sterile barrier systems and packaging systems (ISO 11607-1:2006)*