

This is a preview of "BS EN ISO 8362-4:2011...". Click here to purchase the full version from the ANSI store.

BS EN ISO 8362-4:2011



BSI Standards Publication

Injection containers and accessories

Part 4: Injection vials made of moulded glass (ISO 8362-4:2011)

NO COPYING WITHOUT BSI PERMISSION EXCEPT AS PERMITTED BY COPYRIGHT LAW

raising standards worldwide™



This is a preview of "BS EN ISO 8362-4:201...". [Click here to purchase the full version from the ANSI store.](#)

This British Standard is the UK implementation of EN ISO 8362-4:2011. It supersedes BS EN ISO 8362-4:2004 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/212, IVDs.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© BSI 2011

ISBN 978 0 580 73790 9

ICS 11.040.20

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 September 2011.

Amendments issued since publication

Date	Text affected
------	---------------

This is a preview of "BS EN ISO 8362-4:201...". Click here to purchase the full version from the ANSI store.

EUROPÄISCHE NORM

September 2011

ICS 11.040.20

Supersedes EN ISO 8362-4:2004

English Version

Injection containers and accessories - Part 4: Injection vials made of moulded glass (ISO 8362-4:2011)

Réipients et accessoires pour produits injectables - Partie
4: Flacons en verre moulé (ISO 8362-4:2011)

Injektionsbehältnisse und Zubehör - Teil 4:
Injektionsflaschen aus Hüttenglas (ISO 8362-4:2011)

This European Standard was approved by CEN on 31 August 2011.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

This is a preview of "BS EN ISO 8362-4:201...". [Click here to purchase the full version from the ANSI store.](#)

Foreword

This document (EN ISO 8362-4:2011) has been prepared by Technical Committee ISO/TC 76 "Transfusion, infusion and injection, and blood processing equipment for medical and pharmaceutical use".

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2012, and conflicting national standards shall be withdrawn at the latest by March 2012.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 8362-4:2004.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

Endorsement notice

The text of ISO 8362-4:2011 has been approved by CEN as a EN ISO 8362-4:2011 without any modification.

This is a preview of "BS EN ISO 8362-4:201...". Click here to purchase the full version from the ANSI store.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 8362-4 was prepared by Technical Committee ISO/TC 76, *Transfusion, infusion and injection, and blood processing equipment for medical and pharmaceutical use*.

This third edition cancels and replaces the second edition (ISO 8362-4:2003), of which it constitutes a minor revision.

The principle changes to the second edition are the updating of normative references to ISO 4802-1 and ISO 4802-2, and the revision of Figure 2 and Table 2.

ISO 8362 consists of the following parts, under the general title *Injection containers and accessories*:

- *Part 1: Injection vials made of glass tubing*
- *Part 2: Closures for injection vials*
- *Part 3: Aluminium caps for injection vials*
- *Part 4: Injection vials made of moulded glass*
- *Part 5: Freeze drying closures for injection vials*
- *Part 6: Caps made of aluminium-plastics combinations for injection vials*
- *Part 7: Injection caps made of aluminium-plastics combinations without overlapping plastics part*

Introduction

The purpose of this part of ISO 8362 is to specify the shape, dimensions and capacities of, and performance requirements for, glass vials intended for medical use. Containers made from moulded glass are considered to be suitable for the packaging and storage of injectable preparations until they are administered for medicinal purposes. Such containers can be made from different types of glass, which can affect the chemical resistance properties. For example, those made from borosilicate glass will have a very high level of chemical resistance, whereas those made from soda-lime-silica glass will have a lower chemical resistance but one that is adequate for the purpose for which the containers are intended. The chemical resistance of the internal surface of containers made from soda-lime-silica glass can be improved by a treatment during production to produce a chemical resistance equal to that of containers made from borosilicate glass for single use. This level of chemical resistance will be maintained as long as the interior surface is not destroyed by chemical attack, in which case it will be reduced to that of untreated soda-lime-silica glass.

Because containers can be made from different types of glass and because it is the chemical behaviour of the internal surface that is important when they are filled with injectable preparations, it is essential to specify test procedures by which this performance can be measured. The procedures specified in this part of ISO 8362 will allow this performance based on the hydrolytic resistance to be measured and, from the result of measurement, it is possible to classify containers into their correct category. The procedures also allow containers to be tested and to determine whether the hydrolytic resistance is due to the composition of the glass, or to a treatment of the internal surface.

This is a preview of "BS EN ISO 8362-4:201...". Click here to purchase the full version from the ANSI store.

Injection containers and accessories —

Part 4: Injection vials made of moulded glass

1 Scope

This part of ISO 8362 specifies the shape, dimensions and capacities of glass vials for injectable preparations. It also specifies the material from which such containers are made and the performance requirements for the containers.

It applies to colourless or amber glass containers moulded from borosilicate or soda-lime-silica glass, with or without an internal surface treatment, and intended to be used in the packaging, storage or transportation of products intended for injection.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 719:1985, *Glass — Hydrolytic resistance of glass grains at 98 °C — Method of test and classification*

ISO 720:1985, *Glass — Hydrolytic resistance of glass grains at 121 °C — Method of test and classification*

ISO 4802-1:2010, *Glassware — Hydrolytic resistance of the interior surfaces of glass containers — Part 1: Determination by titration method and classification*

ISO 4802-2:2010, *Glassware — Hydrolytic resistance of the interior surfaces of glass containers — Part 2: Determination by flame spectrometry and classification*

ISO 7458, *Glass containers — Internal pressure resistance — Test methods*

ISO 7459, *Glass containers — Thermal shock resistance and thermal shock endurance — Test methods*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 4802-1 and ISO 4802-2 apply.