

This is a preview of "BS ISO 11040-7:2015". [Click here to purchase the full version from the ANSI store.](#)

BS ISO 11040-7:2015



BSI Standards Publication

Prefilled syringes

Part 7: Packaging systems for sterilized subassembled syringes ready for filling

bsi.

...making excellence a habit.™

This is a preview of "BS ISO 11040-7:2015". [Click here to purchase the full version from the ANSI store.](#)

This British Standard is the UK implementation of ISO 11040-7:2015.

The UK participation in its preparation was entrusted to Technical Committee CH/212, IVDs.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2015. Published by BSI Standards Limited 2015

ISBN 978 0 580 76968 9

ICS 11.040.25

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 April 2015.

Amendments issued since publication

Date	Text affected
------	---------------

This is a preview of "BS ISO 11040-7:2015". [Click here to purchase the full version from the ANSI store.](#)

First edition
2015-04-01

Prefilled syringes —

Part 7:

Packaging systems for sterilized subassembled syringes ready for filling

Seringues préremplies —

*Partie 7: Systèmes d'emballage pour les seringues stérilisées prêtes à
l'emploi préremplissables*



Reference number
ISO 11040-7:2015(E)

© ISO 2015

This is a preview of "BS ISO 11040-7:2015". Click here to purchase the full version from the ANSI store.



COPYRIGHT PROTECTED DOCUMENT

© ISO 2015

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

This is a preview of "BS ISO 11040-7:2015". Click here to purchase the full version from the ANSI store.

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements for the packaging system	2
4.1 General.....	2
4.2 Nest.....	3
4.3 Tub.....	4
4.4 Insert liner.....	4
4.5 Sealing lid.....	4
4.6 Protective bag.....	5
5 Information to be provided by the manufacturer	5
6 Marking of the tub	5
7 Packaging of tubs in trading units/bundles	6
Annex A (informative) Design of nests	7
Annex B (informative) Determination of nest deflection	14
Annex C (informative) Design of tubs	17
Annex D (informative) Schematic illustrations of examples for the orientation of tubs within the protective bag	19
Annex E (informative) Design and dimensions of the protective bag	21
Annex F (informative) Test method to determine the distance between the edge of the protective bag to the rear end of the tub	24
Bibliography	27

This is a preview of "BS ISO 11040-7:2015". Click here to purchase the full version from the ANSI store.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#).

The committee responsible for this document is ISO/TC 76, *Transfusion, infusion and injection, and blood processing equipment for medical and pharmaceutical use*.

ISO 11040 consists of the following parts, under the general title *Prefilled syringes*:

- *Part 1: Glass cylinders for dental local anaesthetic cartridges*
- *Part 2: Plunger stoppers for dental local anaesthetic cartridges*
- *Part 3: Seals for dental local anaesthetic cartridges*
- *Part 4: Glass barrels for injectables and sterilized subassembled syringes ready for filling*
- *Part 5: Plunger stoppers for injectables*
- *Part 6: Plastics barrels for injectables*
- *Part 7: Packaging systems for sterilized subassembled syringes ready for filling*

The following part is under preparation:

- *Part 8: Requirements and test methods for finished prefilled syringes*

This is a preview of "BS ISO 11040-7:2015". [Click here to purchase the full version from the ANSI store.](#)

Introduction

At the start of prefilled syringe processing by the pharmaceutical industry, syringes made of tubing glass were delivered to the pharmaceutical companies in the form of a so called non-sterile "bulkware" only. The process steps like washing, drying, inner lubrication, sealing the syringe with a closure system, sterilization, as well as filling, and closing were then performed in the pharmaceutical companies. Processing of "bulkware" is still performed like this nowadays. Sterilized subassembled syringes have partially replaced the non-sterile "bulkware".

In the case of sterilized subassembled syringes ready for filling, responsibility for the aforementioned process steps relevant to the injectable product lies within the manufacturer of the primary packaging material. Following the assembly of the needle shield on syringes with a staked needle or tip caps for the luer cone version, the subassembled syringes are placed into so called nests. The nests, in turn, are placed into a plastic tub. The syringes in the nest are protected by means of an insert liner and the tub itself is sealed by a sealing lid (which is currently, and so far, primarily achieved using a porous material). Thus, the tub properly sealed with the sealing lid represents the "sterile barrier system". The sealed tub is then wrapped into a sealable bag and, thus, ready for sterilization which is currently, and so far, primarily performed using ethylene oxide.

In this form, the sterilized subassembled syringes ready for filling are delivered to the pharmaceutical companies in a sterile condition where they are processed on suitable machines.

The packaging design and material have to ensure sterility and should be compatible with the process of the customer. The packaging characteristics, material, thickness, shape, and resistance to deformations among others are such that they maintain, up to the point of use, the integrity of the product and a validated barrier against particulate and bacterial contamination. The packaging materials have to fulfil regional and national regulatory requirements.

This is a preview of "BS ISO 11040-7:2015". [Click here to purchase the full version from the ANSI store.](#)

This is a preview of "BS ISO 11040-7:2015". Click here to purchase the full version from the ANSI store.

Prefilled syringes —

Part 7:

Packaging systems for sterilized subassembled syringes ready for filling

1 Scope

This part of ISO 11040 specifies the packaging system that is used to deliver sterilized subassembled syringes ready for filling in tubs and nests.

Downstream processes (processes after filling such as in house/outside transport, reprocessing) can result in specific requirements on the packaging system used to deliver sterilized subassembled syringes ready for filling. However, these requirements are not within the scope of this part of ISO 11040.

NOTE 1 Glass barrels and sterilized subassembled syringes ready for filling, plungers, and plastic barrels for injectables are specified in ISO 11040-4, ISO 11040-5, and ISO 11040-6.

NOTE 2 ISO 11607-2 addresses validation requirements of sealing and packaging processes for medical devices.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11607-1, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

customer

business entity which purchases syringe barrels or sterilized subassembled syringes ready for filling and conducts further processing or filling as appropriate

3.2

insert liner

foil to cover the filled nest

3.3

manufacturer

business entity which performs or is otherwise responsible for the manufacturing of the syringe barrels (bulkware) or for the sterilized subassembled syringes ready for filling by the customer

3.4

nest

plastic plate with a defined hole pattern for the suspension of the syringe bodies