This is a preview of "BS EN ISO 18113-2:20...". Click here to purchase the full version from the ANSI store.

BS EN ISO 18113-2:2011

Incorporating corrigendum December 2011

BSI Standards Publication

In vitro diagnostic medical devices — Information supplied by the manufacturer (labelling)

Part 2: In vitro diagnostic reagents for professional use (ISO 18113-2:2009)

NO COPYING WITHOUT BSI PERMISSION EXCEPT AS PERMITTED BY COPYRIGHT LAW

raising standards worldwide™

This is a preview of "BS EN ISO 18113-2:20...". Click here to purchase the full version from the ANSI store.

This British Standard is the UK implementation of EN ISO 18113-2:2011. It is identical to ISO 18113-2:2009. It supersedes BS EN ISO 18113-2:2009, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/212, IVDs.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2012

ISBN 978 0 580 77328 0

ICS 11.100.10

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 November 2011.

Amendments/corrigenda issued since publication

Date	Text affected
31 January 2012	Implementation of CEN correction notice 9 November 2011: Corrected date of withdrawal in EN foreword

This is a preview of "BS EN ISO 18113-2:20...". Click here to purchase the full version from the ANSI store.

EUROPÄISCHE NORM

October 2011

ICS 11.100.10

Supersedes EN ISO 18113-2:2009

English Version

In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 2: In vitro diagnostic reagents for professional use (ISO 18113-2:2009)

Dispositifs médicaux de diagnostic in vitro - Informations fournies par le fabricant (étiquetage) - Partie 2: Réactifs de diagnostic in vitro à usage professionnel (ISO 18113-2:2009)

In-vitro-Diagnostika - Bereitstellung von Informationen durch den Hersteller - Teil 2: In-vitro-diagnostische Reagenzien für den Gebrauch durch Fachpersonal (ISO 18113-2:2009)

This European Standard was approved by CEN on 20 September 2011.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

© 2011 CEN All rights of exploitation in any form and by any means reserved worldwide for CEN national Members.

Ref. No. EN ISO 18113-2:2011: E

This is a preview of "BS EN ISO 18113-2:20...". Click here to purchase the full version from the ANSI store.

Foreword

This document (EN ISO 18113-2:2011) has been prepared by Technical Committee ISO/TC 212 "Clinical laboratory testing and in vitro diagnostic test systems" in collaboration with Technical Committee CEN/TC 140 “In vitro diagnostic medical devices” the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2012, and conflicting national standards shall be withdrawn at the latest by October 2014.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 18113-2:2009.

This new edition contains a revised Annex ZA.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive.

For relationship with EU Directive, see informative Annex ZA, which is an integral part of this document.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

Endorsement notice

The text of ISO 18113-2:2009 has been approved by CEN as EN ISO 18113-2:2011 without any modification.

This is a preview of "BS EN ISO 18113-2:20...". Click here to purchase the full version from the ANSI store.

Annex ZA
(informative)

Relationship between this European Standard and the Essential Requirements of the EU Directive 98/79/EC on “in vitro Diagnostic Medical Devices”

This European Standard has been prepared under a mandate given to CEN by the European Commission to provide a means of conforming to the Essential Requirements of the New Approach Directive 98/79/EC on “*in vitro* Diagnostic Medical Devices”.

Once this European Standard is cited in the Official Journal of the European Union under that Directive and has been implemented as national standard in at least one Member State, compliance with the clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this European Standard, a presumption of conformity with the corresponding Essential Requirements of that Directive and associated EFTA regulations.

Table ZA.1 — Correspondence between this European Standard and Directive 98/79/EC

Clauses of this European Standard	Essential requirements (ERs) of Directive 98/79/EC	Qualifying comments/Notes
7.7	B.3.1	This standard only covers the second sentence of ER B.3.1 namely the labelling requirements.
5, 6, 7	B.8.1	Presumption of conformity with ER B.8.1 also requires compliance with clauses 4.1, 4.2.1 and 4.6 of EN ISO 18113-1.
5.8, 6.8, 7.10	B.8.3	**NOTE 2**
5.1, 6.2	B.8.4 (a)	**NOTE 1**
5.2.1, 5.3, 6.3.1, 6.4	B.8.4 (b)	
5.2.2, 6.3.2	B.8.4 (d)	Full compliance with ER B.8.4 (d) requires the use of EN 980, clause 5.4: symbol (LOT).
5.7, 6.7	B.8.4 (e)	
5.5, 6.5	B.8.4 (g)	
5.6, 6.6	B.8.4 (h)	
5.8, 6.8	B.8.4 (j)	**NOTE 2**
5.4, 7.3	B.8.5	
5.2.2, 6.3.2	B.8.6	
7.1, 7.2, 7.9, 7.10	B.8.7 (a)	Presumption of conformity with ER B.8.7 (a) requires also compliance with EN ISO 18113-1, clause 4.5, as well as an indication of the in vitro use of the device. **NOTE 1, NOTE 3**
7.6	B.8.7 (b)	
7.9	B.8.7 (c)	**NOTE 3**
7.16	B.8.7 (d)	
7.7	B.8.7 (e)	
7.11	B.8.7 (f)	

This is a preview of "BS EN ISO 18113-2:20...". Click here to purchase the full version from the ANSI store.

Clauses of this European Standard	Essential requirements (ERs) of Directive 98/79/EC	Qualifying comments/Notes
7.12	B.8.7 (g)	
7.4, 7.8, 7.16, 7.18	B.8.7 (h)	Full compliance with ER B.8.7 (h) requires, where applicable, an indication of any particular training needed to operate the device.
7.14, 7.15	B.8.7 (i)	
7.18, 7.13	B.8.7 (j)	
7.5, 7.13	B.8.7 (k)	
7.17	B.8.7 (l)	
7.7	B.8.7 (m)	
7.8	B.8.7 (o)	
7.10	B.8.7 (r)	**NOTE 3**
7.10	B.8.7 (s)	**NOTE 3**

NOTE 1 In the European Union, the name and address of the manufacturer's "EC Authorized representative" is required on the outer container label or in the instructions for use, if the legal manufacturer is not located within the European Union.

NOTE 2 Essential requirement B.8.3 of Directive 98/79/EC should be consulted for a comprehensive list of the information required.

NOTE 3 Essential requirement B.8.7 of Directive 98/79/ EC should be consulted for a comprehensive list of the information required.

This is a preview of "BS EN ISO 18113-2:20...". Click here to purchase the full version from the ANSI store.

Contents

Page

Foreword iv

Introduction v

1 Scope 1

2 Normative references 1

3 Terms and definitions 1

4 General 2
4.1 Essential requirements 2
4.2 Identification of kit components 2

5 Content of the outer container label 2
5.1 Manufacturer 2
5.2 Identification of the IVD reagent 2
5.3 Contents 2
5.4 Intended use 2
5.5 *In vitro* diagnostic use 3
5.6 Storage and handling conditions 3
5.7 Expiry date 3
5.8 Warnings and precautions 3

6 Content of the immediate container label 3
6.1 General provisions 3
6.2 Manufacturer 4
6.3 Identification of the IVD reagent 4
6.4 Contents 4
6.5 *In vitro* diagnostic use 4
6.6 Storage and handling conditions 4
6.7 Expiry date 4
6.8 Warnings and precautions 4

7 Content of the instructions for use 5
7.1 Manufacturer 5
7.2 Identification of the IVD reagent 5
7.3 Intended use 5
7.4 Principles of the examination method 5
7.5 Traceability of values assigned to calibrators and trueness-control materials 5
7.6 Components 6
7.7 Additional required equipment 6
7.8 Reagent preparation 6
7.9 Storage and shelf life after first opening 6
7.10 Warnings and precautions 6
7.11 Primary sample collection, handling and storage 7
7.12 Examination procedure 7
7.13 Control procedure 7
7.14 Calculation of examination results 7
7.15 Interpretation of results 7
7.16 Performance characteristics 7
7.17 Biological reference intervals 8
7.18 Limitations of the examination procedure 8
7.19 Literature references 8

Bibliography 9

© ISO 2011

This is a preview of "BS EN ISO 18113-2:20...". Click here to purchase the full version from the ANSI store.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 18113-2 was prepared by Technical Committee ISO/TC 212, *Clinical laboratory testing and* in vitro *diagnostic test systems*.

ISO 18113 consists of the following parts, under the general title In vitro *diagnostic medical devices — Information supplied by the manufacturer (labelling)*:

— *Part 1: Terms, definitions and general requirements*

— *Part 2:* In vitro *diagnostic reagents for professional use*

— *Part 3*: In vitro *diagnostic instruments for professional use*

— *Part 4:* In vitro *diagnostic reagents for self-testing*

— *Part 5:* In vitro *diagnostic instruments for self-testing*

© ISO 2011

This is a preview of "BS EN ISO 18113-2:20...". Click here to purchase the full version from the ANSI store.

Introduction

Manufacturers of *in vitro* diagnostic (IVD) reagents for professional use supply users with information to enable their safe use and the expected performance of their devices. The type and level of detail varies according to the intended uses and country-specific regulations.

The Global Harmonization Task Force (GHTF) encourages convergence of the evolution of regulatory systems for medical devices at the global level. Eliminating differences among regulatory jurisdictions could allow patients earlier access to new technologies and treatments. See Reference [9]. This part of ISO 18113 provides a basis for harmonization of labelling requirements for IVD reagents for professional use.

This part of ISO 18113 is concerned solely with information supplied with IVD reagents, calibrators and control materials intended for professional use. It is intended to be used in conjunction with ISO 18113-1, which contains the general requirements for information supplied by the manufacturer and definitions of general labelling concepts.

This part of ISO 18113 is based on EN 375:2001[5]. The text has been modified to conform to Part 2 of the ISO/IEC Directives[4], but the requirements, including those in ISO 18113-1, are substantially equivalent to the original European harmonized standard. This part of ISO 18113 is intended to support the essential labelling requirements of all the GHTF partners, as well as other countries that have or plan to enact labelling regulations for IVD medical devices.

For IVD reagents, calibrators and/or control materials that are intended to be used as a system with an instrument provided by the same manufacturer, this part of ISO 18113 is also intended to be used together with ISO 18113-1 and ISO 18113-3[2].

This is a preview of "BS EN ISO 18113-2:20...". Click here to purchase the full version from the ANSI store.

This is a preview of "BS EN ISO 18113-2:20...". Click here to purchase the full version from the ANSI store.

In vitro diagnostic medical devices — Information supplied by the manufacturer (labelling) —

Part 2: *In vitro* diagnostic reagents for professional use

1 Scope

This part of ISO 18113 specifies requirements for information supplied by the manufacturer of IVD reagents for professional use.

This part of ISO 18113 also applies to information supplied by the manufacturer with calibrators and control materials intended for use with IVD medical devices for professional use.

This part of ISO 18113 can also be applied to accessories.

This part of ISO 18113 applies to the labels for outer and immediate containers and to the instructions for use.

This part of ISO 18113 does not apply to

a) IVD instruments or equipment,

b) IVD reagents for self-testing.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 8601, *Data elements and interchange formats — Information interchange — Representation of dates and times*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 15223-1, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

ISO 18113-1:—, In vitro *diagnostic medical devices — Information supplied by the manufacturer (labelling) — Part 1: Terms, definitions and general requirements*

EN 980, *Symbols for use in the labelling of medical devices*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 18113-1 apply.

© ISO 2011