

This is a preview of "BS EN 13205-4:2014". [Click here to purchase the full version from the ANSI store.](#)

BS EN 13205-4:2014



BSI Standards Publication

Workplace exposure — Assessment of sampler performance for measurement of airborne particle concentrations

Part 4: Laboratory performance test based
on comparison of concentrations

bsi.

...making excellence a habit.™

This is a preview of "BS EN 13205-4:2014". [Click here to purchase the full version from the ANSI store.](#)

This British Standard is the UK implementation of EN 13205-4:2014. Together with BS EN 13205-1:2014, BS EN 13205-2:2014, PD CEN/TR 13205-3, BS EN 13205-5:2014 and BS EN 13205-6:2014 it supersedes BS EN 13205:2002 which will be withdrawn upon publication of all parts of the series.

The UK participation in its preparation was entrusted to Technical Committee EH/2/2, Work place atmospheres.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2014.
Published by BSI Standards Limited 2014

ISBN 978 0 580 78061 5
ICS 13.040.30

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 June 2014.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

This is a preview of "BS EN 13205-4:2014". [Click here to purchase the full version from the ANSI store.](#)

EUROPÄISCHE NORM

June 2014

ICS 13.040.30

Supersedes EN 13205:2001

English Version

**Workplace exposure - Assessment of sampler performance for
measurement of airborne particle concentrations - Part 4:
Laboratory performance test based on comparison of
concentrations**

Exposition sur les lieux de travail - Évaluation des performances des dispositifs de prélèvement pour la mesure des concentrations de particules en suspension dans l'air - Partie 4: Essai de performances en laboratoire par comparaison des concentrations

Exposition am Arbeitsplatz - Beurteilung der Leistungsfähigkeit von Sammlern für die Messung der Konzentration luftgetragener Partikel - Teil 4: Laborprüfung der Leistungsfähigkeit basierend auf dem Vergleich der Konzentrationen

This European Standard was approved by CEN on 7 May 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

This is a preview of "BS EN 13205-4:2014". [Click here to purchase the full version from the ANSI store.](#)

Contents	Page
Foreword.....	4
Introduction	5
1 Scope	6
2 Normative references	6
3 Terms and definitions	6
4 Symbols and abbreviations	6
4.1 Symbols	6
4.1.1 Latin	6
4.1.2 Greek.....	8
4.2 Enumerating subscripts.....	8
4.3 Abbreviations	8
5 Principle.....	9
6 Test method.....	9
6.1 General.....	9
6.2 Test conditions	9
6.3 Test variables	10
6.3.1 General.....	10
6.3.2 Particle size	11
6.3.3 Wind speed.....	12
6.3.4 Wind direction.....	12
6.3.5 Aerosol composition	12
6.3.6 Collected mass or internally separated mass	12
6.3.7 Specimen variability	12
6.3.8 Excursion from the nominal flow rate	12
7 Experimental requirements	13
8 Calculation of sampler bias and expanded uncertainty	14
8.1 Sampler bias.....	14
8.2 Correction factor.....	15
8.3 Sources of uncertainty (of measurement)	15
8.3.1 Principle.....	15
8.3.2 Test aerosol concentration, as determined using the validated sampler(s)	15
8.3.3 Validated sampler	16
8.3.4 Candidate sampler bias	16
8.3.5 Individual candidate sampler variability	17
8.3.6 Excursion from the nominal flow rate	17
8.4 Combined standard uncertainty.....	19
8.4.1 General.....	19
8.4.2 Candidate sampler without any coupling between the flow rate and internal penetration	19
8.4.3 Candidate sampler with a coupling between the flow rate and internal penetration	20
8.4.4 Combined uncertainty per influence variable value	20
8.4.5 Distinction between different values of the influence variables.....	21
8.4.6 Non-distinction between different values of the influence variables	21
8.5 Expanded uncertainty	22
9 Test report	22
9.1 General.....	22

This is a preview of "BS EN 13205-4:2014". [Click here to purchase the full version from the ANSI store.](#)

9.2	Testing laboratory details and sponsoring organisation	23
9.3	Description of the candidate sampler and validated sampler	23
9.4	Critical review of sampling process	23
9.5	Test facilities	23
9.6	Details of experimental design	24
9.7	Presentation of experimental results	24
9.8	Data analysis	24
9.9	Candidate sampler performance	24
9.10	Summary and information for the user of the sampler	25
	Bibliography	26

This is a preview of "BS EN 13205-4:2014". [Click here to purchase the full version from the ANSI store.](#)

Foreword

This document (EN 13205-4:2014) has been prepared by Technical Committee CEN/TC 137 "Assessment of workplace exposure to chemical and biological agents", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by December 2014 and conflicting national standards shall be withdrawn at the latest by December 2014.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document together with EN 13205-1, EN 13205-2, CEN/TR 13205-3, EN 13205-5 and EN 13205-6 supersedes EN 13205:2001.

EN 13205, *Workplace exposure – Assessment of sampler performance for measurement of airborne particle concentrations*, consists of the following parts:

- *Part 1: General requirements;*
- *Part 2: Laboratory performance test based on determination of sampling efficiency;*
- *Part 3: Analysis of sampling efficiency data [Technical Report];*
- *Part 4: Laboratory performance test based on comparison of concentrations (the present document);*
- *Part 5: Aerosol sampler performance test and sampler comparison carried out at workplaces;*
- *Part 6: Transport and handling tests.*

Significant technical changes from the previous edition, EN 13205:2001:

- This part of EN 13205 is based on Annex B of the previous edition, EN 13205:2001.
- The scope has been limited to aerosol samplers, and the current version of the standard is not (directly) applicable to other types of aerosol instruments.
- As this is now a standard in its own right, a clause on the used symbols has been added. All definitions are now given either in EN 1540, *Workplace exposure — Terminology* or in Part 1 or Part 2 of this standard.
- The method of calculating the uncertainty of a sampler or a measuring procedure has been revised in order to comply with ENV 13005. The concept of "accuracy" is no longer used, instead the concept of "expanded uncertainty" is used.
- The five major sources of uncertainty due to aspects of the sampling performance of an aerosol sampler (calibration of sampler test system, estimation of sampled concentration, bias relative to the sampling convention, individual sampler variability and excursion from nominal flow rate) are described with equations on how to incorporate these uncertainties into the expanded uncertainty of a sampler.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

This is a preview of "BS EN 13205-4:2014". [Click here to purchase the full version from the ANSI store.](#)

Introduction

EN 481 defines sampling conventions for the particle size fractions to be collected from workplace atmospheres in order to assess their impact on human health. Conventions are defined for the inhalable, thoracic and respirable aerosol fractions. These conventions represent target specifications for aerosol samplers, giving the ideal sampling efficiency as a function of particle aerodynamic diameter.

In general, the sampling efficiency of real aerosol samplers will deviate from the target specification, and the aerosol mass collected will therefore differ from that which an ideal sampler would collect. In addition, the behaviour of real samplers is influenced by many factors such as external wind speed. In many cases there is an interaction between the influence factors and fraction of the airborne particle size distribution of the environment in which the sampler is used.

The laboratory performance test for samplers for the inhalable, thoracic or respirable aerosol fractions described in this document is based on a comparison of concentrations sampled from three laboratory test atmospheres by a candidate sampler and a (previously) validated sampler.

EN 13205 (all parts) enables manufacturers and users of aerosol samplers to adopt a consistent approach to sampler validation, and provide a framework for the assessment of sampler performance with respect to EN 481 and EN 482.

It is the responsibility of the manufacturer of aerosol samplers to inform the user of the sampler performance under the laboratory conditions¹⁾ specified in this part of EN 13205. It is the responsibility of the user to ensure that the actual conditions of intended use are within what the manufacturer specifies as acceptable conditions according to the performance test.

1) The inhalable convention is undefined for particle sizes in excess of 100 µm or for wind speeds greater than 4 m/s. The tests required to assess performance are therefore limited to these conditions. Should such large particle sizes or wind speeds actually exist at the time of sampling, it is possible that different samplers meeting this part of EN 13205 give different results.

This is a preview of "BS EN 13205-4:2014". [Click here to purchase the full version from the ANSI store.](#)

1 Scope

This European Standard specifies a method for testing aerosol samplers based on comparison of concentrations under prescribed laboratory conditions in order to verify whether the performance of a candidate sampler fulfils the requirements of EN 13205-1:2014

This part of EN 13205 is applicable to all samplers used for the health-related sampling of particles in workplace air.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 481, *Workplace atmospheres - Size fraction definitions for measurement of airborne particles*

EN 1540, *Workplace exposure - Terminology*

EN 13205-1:2014, *Workplace exposure — Assessment of sampler performance for measurement of airborne particle concentrations — Part 1: General requirements*

EN 13205-2:2014, *Workplace exposure — Assessment of sampler performance for measurement of airborne particle concentrations — Part 2: Laboratory performance test based on determination of sampling efficiency*

EN ISO 13137, *Workplace atmospheres - Pumps for personal sampling of chemical and biological agents - Requirements and test methods (ISO 13137)*

3 Terms and definitions

For the purpose of this document, the terms and definitions given in EN 1540, EN 13205-1:2014 and EN 13205-2:2014 apply.

NOTE With regard to EN 1540, in particular, the following terms are used in this document: total airborne particles, respirable fraction, sampling efficiency, static sampler, thoracic fraction, inhalable fraction, measuring procedure, non-random uncertainty, random uncertainty, expanded uncertainty, standard uncertainty, combined standard uncertainty, uncertainty (of measurement), coverage factor, precision and analysis.

4 Symbols and abbreviations

4.1 Symbols

4.1.1 Latin

c	candidate sampler correction factor for bias correction, either prescribed by sampler manufacturer or measuring procedure, or assigned the value $c = 1.00$, [-]
N_{A_i}	number of test aerosols for influence variable value ζ_i
N_{IV}	number of values for the other influence variables at which tests were performed