BS EN 14683:2014



BSI Standards Publication

Medical face masks — Requirements and test methods



...making excellence a habit."

This British Standard is the UK implementation of EN 14683:2014. It supersedes BS EN 14683:2005 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/205/1, Medical textiles.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2014. Published by BSI Standards Limited 2014

ISBN 978 0 580 78520 7

ICS 11.140

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 April 2014.

Amendments issued since publication

Date Text affected

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

© 2014 CEN All rights of exploitation in any form and by any means reserved

2014 CEN All rights of exploitation in any form and by any means reserve worldwide for CEN national Members.



00

This is a preview of "BS EN 14683:2014". Click here to purchase the full version from the ANSI store.

EUROPÄISCHE NORM

ICS 11.140

English Version

Medical face masks - Requirements and test methods

Masques à usage médical - Exigences et méthodes d'essai

This European Standard was approved by CEN on 4 February 2014.

Medizinische Gesichtsmasken - Anforderungen und Prüfverfahren

Supersedes EN 14683:2005

BS EN 14683:2014

ENI 1/693

March 2014

Contents

Page

Foreword3			
Introdu	Introduction		
1	Scope	5	
2	Normative references	5	
3	Terms and definitions	5	
4	Classification	6	
5 5.1 5.2	Requirements General Performance requirements	.7 7	
6	Labelling and information to be supplied	8	
Annex	A (informative) Information for users	9	
Annex	B (normative) Method for <i>in-vitro</i> determination of bacterial filtration efficiency (BFE)	10	
B.1	Principle	10	
B.2	Reagents and materials	10	
B.3	Apparatus	11	
B.4	Test specimens	11	
B.5	Preparation of bacterial challenge	11	
B.6	Procedure	12	
B.7	Calculation of bacterial filtration efficiency	12	
B.8	Test report	12	
Annex C (normative) Method for determination of breathability (differential pressure)		15	
C.1	Principle	15	
C.2	Apparatus	16	
C.3	Test specimens	16	
C.4	Procedure	16	
C.5	Calculation of differential pressure	16	
C.6	Test report	16	
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on medical devices			
Bibliog	3ibliography		

Foreword

This document (EN 14683:2014) has been prepared by Technical Committee CEN/TC 205 "Non-active medical devices", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2014 and conflicting national standards shall be withdrawn at the latest by September 2014.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 14683:2005.

In respect to EN 14683:2005, the following changes have been made:

- a) change/extension of title and scope to the more general and broader use for medical face masks;
- b) adjustment to ISO 22609 concerning the request for resistance to liquid splashes;
- c) addition of requirements for microbiological purity and general biocompatibility;
- d) adjustment of Table 1 on performance requirements for medical face masks;
- e) update of Annex A on user information;
- f) complete revision of Annex B on method for in-vitro determination of the bacterial filter performance in particular with regard to the testing conditions and the structure of the test apparatus;
- g) complete editorial revision, including update of all normative references, the Bibliography and Annex ZA on the relationships to the EU Directive 93/42/EEC.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

The transmission of infective agents during surgical procedures in operating theatres and other medical settings can occur in several ways. Sources are, for example, the noses and mouths of members of the surgical team. The main intended use of medical face masks is to protect the patient from infective agents and, additionally, in certain circumstances to protect the wearer against splashes of potentially contaminated liquids and viable particles. Medical face masks may also be intended to be worn by patients and other persons to reduce the risk of spread of infections, particularly in epidemic or pandemic situations.

1 Scope

This European Standard specifies construction, design, performance requirements and test methods for medical face masks intended to limit the transmission of infective agents from staff to patients during surgical procedures and other medical settings with similar requirements. A medical face mask with an appropriate microbial barrier can also be effective in reducing the emission of infective agents from the nose and mouth of an asymptomatic carrier or a patient with clinical symptoms.

This European Standard is not applicable to masks intended exclusively for the personal protection of staff.

NOTE 1 Standards for masks for use as respiratory personal protective equipment are available.

NOTE 2 Annex A provides information for the users of medical face masks.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 10993-1, Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process (ISO 10993-1)

EN ISO 10993-5, *Biological evaluation of medical devices — Part 5: Tests for in vitro cytotoxicity (ISO 10993-5)*

EN ISO 10993-10, Biological evaluation of medical devices — Part 10: Tests for irritation and skin sensitization (ISO 10993-10)

EN ISO 11737-1, Sterilization of medical devices — Microbiological methods — Part 1: Determination of a population of microorganisms on products (ISO 11737-1)

ISO 22609, Clothing for protection against infectious agents — Medical face masks — Test method for resistance against penetration by synthetic blood (fixed volume, horizontally projected)

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

medical face mask

medical device covering the mouth and nose providing a barrier to minimise the direct transmission of infective agents between staff and patient

Note 1 to entry: Transmission of blood-borne agents from patients to staff may occur via splashes.

3.2

bacterial filtration efficiency (BFE)

efficiency of the medical face mask material(s) as a barrier to bacterial penetration

Note 1 to entry: The BFE test method is used to measure the bacterial filtration efficiency (BFE) of medical face mask materials.

3.3

differential pressure

air permeability of the mask, measured by determining the difference of pressure across the mask under specific conditions of air flow, temperature and humidity