

This is a preview of "PD IEC/TS 62129-3:20...". Click here to purchase the full version from the ANSI store.

PD IEC/TS 62129-3:2014



BSI Standards Publication

Calibration of wavelength/optical frequency measurement instruments

Part 3: Optical frequency meters using
optical frequency combs

bsi.

...making excellence a habit.™

This is a preview of "PD IEC/TS 62129-3:20...". [Click here to purchase the full version from the ANSI store.](#)

This Published Document is the UK implementation of IEC/TS 62129-3:2014.

The UK participation in its preparation was entrusted to Technical Committee GEL/86, Fibre optics.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2014.

Published by BSI Standards Limited 2014

ISBN 978 0 580 84250 4

ICS 33.180.30

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 31 August 2014.

Amendments/corrigenda issued since publication

Date	Text affected
-------------	----------------------

This is a preview of "PD IEC/TS 62129-3:20...". [Click here to purchase the full version from the ANSI store.](#)



Edition 1.0 2014-02

TECHNICAL SPECIFICATION



Calibration of wavelength/optical frequency measurement instruments – Part 3: Optical frequency meters using optical frequency combs

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

PRICE CODE

R

ICS 33.180.30

ISBN 978-2-8322-1421-3

Warning! Make sure that you obtained this publication from an authorized distributor.

This is a preview of "PD IEC/TS 62129-3:20...". Click here to purchase the full version from the ANSI store.

CONTENTS

FOREWORD.....	3
INTRODUCTION.....	5
1 Scope.....	6
2 Normative references	6
3 Terms and definitions	6
4 Calibration test requirements	7
4.1 Preparation.....	7
4.2 Reference test conditions	8
4.3 Traceability.....	8
4.3.1 General.....	8
4.3.2 National standard.....	8
4.3.3 Transfer standard.....	9
4.3.4 Working standard.....	9
5 Optical frequency calibration	9
5.1 General.....	9
5.2 Establishing the calibration conditions	11
5.3 Calibration procedure.....	11
5.3.1 General.....	11
5.3.2 Measurement configuration	11
5.3.3 Detailed procedure.....	13
5.4 Calibration uncertainty.....	13
5.5 Reporting the results.....	13
Annex A (normative) Mathematical basis	14
A.1 General.....	14
A.2 Type A evaluation of uncertainty	14
A.3 Type B evaluation of uncertainty	15
A.4 Determining the combined standard uncertainty.....	15
A.5 Reporting.....	16
Annex B (informative) References of optical frequency comb source	17
B.1 Method A (mode-locked fibre laser + carrier-envelope phase lock).....	17
B.2 Method B (stabilized laser + electro-optical modulator)	17
B.3 Method C (stabilized laser + supercontinuum source)	18
Annex C (informative) Frequency-dependence of uncertainty.....	19
Bibliography.....	20
Figure 1 – Traceability chain using optical frequency measurement scheme	9
Figure 2 – Schematic configuration of optical frequency measurement technique that uses optical comb	10
Figure 3 – Optical spectra of lasers and optical frequency combs	11
Figure 4 – Optical frequency meter measurement using a reference source.....	12
Figure 5 – Optical frequency meter measurement using a reference optical frequency meter.....	12
Figure B.1 – Mode-locked laser + nonlinear optical effect	17
Figure B.2 – Electro-optical modulator type comb source.....	18
Figure B.3 – Supercontinuum source	18

INTERNATIONAL ELECTROTECHNICAL COMMISSION

CALIBRATION OF WAVELENGTH/OPTICAL FREQUENCY MEASUREMENT INSTRUMENTS –

Part 3: Optical frequency meters using optical frequency combs

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. In exceptional circumstances, a technical committee may propose the publication of a technical specification when

- the required support cannot be obtained for the publication of an International Standard, despite repeated efforts, or
- the subject is still under technical development or where, for any other reason, there is the future but no immediate possibility of an agreement on an International Standard.

Technical specifications are subject to review within three years of publication to decide whether they can be transformed into International Standards.

IEC/TS 62129-3, which is a technical specification, has been prepared by IEC technical committee 86: Fibre optics.

This is a preview of "PD IEC/TS 62129-3:20...". [Click here to purchase the full version from the ANSI store.](#)

The text of this technical specification is based on the following documents:

Enquiry draft	Report on voting
86/461/DTS	86/465/RVC

Full information on the voting for the approval of this technical specification can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts in the IEC 62129 series, published under the general title *Calibration of wavelength/optical frequency measurement instruments*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- transformed into an International standard,
- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

This is a preview of "PD IEC/TS 62129-3:20...". Click here to purchase the full version from the ANSI store.

INTRODUCTION

It is essential for realizing fibre optic systems that optical channels are defined in the optical frequency domain, not the wavelength domain. One example, the anchor frequency of the ITU-T grid is 193,1 THz, and the channel spacings of the ITU-T grid are 12,5 GHz, 25 GHz, 50 GHz, and 100 GHz [2]¹.

ITU-T has also discussed λ -interface systems such as “black link” [3]. “Black link” includes WDM MUX/DEMUX and a transmission fibre, and provides λ -interfaces. Especially in DWDM systems (channel spacing <100 GHz), the uncertainty in specifying optical frequency needs to be minimized.

To implement future telecom systems, it is expected that optical frequency measurements will need to be extremely precise. For example, to achieve the channel spacing of 25 GHz, signal optical frequency uncertainty ($U_{f_{\text{sig}}}$) and required measurement uncertainty ($U_{f_{\text{meas}}}$) need to be 2 GHz to 200 MHz ($U_{f_{\text{sig}}} / f = 10^{-5}$ to 10^{-6}) and 200 MHz to 2 MHz ($U_{f_{\text{meas}}} / f = 10^{-6}$ to 10^{-8}), respectively. Unfortunately, conventional wavelength meters have measurement uncertainties of 10^{-6} to 10^{-7} . The solution is to use optical frequency measurements since measurement uncertainties can be as small as 10^{-15} to 10^{-16} , which satisfies the above telecom requirement ($U_{f_{\text{meas}}} / f = 10^{-6}$ to 10^{-8}). Therefore, an optical frequency measurement scheme is necessary for the calibration of future telecom systems.

Optical frequency measurement technology is progressing rapidly. Many fundamental papers have examined the use of equally-spaced “optical frequency comb” lines (spacing of up to 50 GHz) from an optical frequency comb as a “ruler” for optical frequency measurement [4-15]. For example, mode-locked lasers with carrier-envelope phase locked enable ultra-low measurement uncertainties of 10^{-15} to 10^{-16} . Some examples of practical optical frequency combs are shown in Annex B (mode-locked fibre laser + carrier-envelope phase lock, stabilized laser + electro-optical modulator, and stabilized laser + supercontinuum source). Frequency measurements provide more accurate values than interferometric wavelength measurements in air by eliminating the effects of refractive indices. Furthermore, they allow the measurement devices to be significantly smaller than wavelength meters.

¹ Numbers between square brackets refer to the Bibliography.

This is a preview of "PD IEC/TS 62129-3:20...". [Click here to purchase the full version from the ANSI store.](#)

CALIBRATION OF WAVELENGTH/OPTICAL FREQUENCY MEASUREMENT INSTRUMENTS –

Part 3: Optical frequency meters using optical frequency combs

1 Scope

This part of IEC 62129, which is a technical specification, describes the calibration of optical frequency meters. It is applicable to instruments measuring the optical frequency emitted from sources that are typical for the fibre-optic communications industry. It is assumed that the optical radiation will be coupled to the optical frequency meter by a single-mode optical fibre.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60793-2-50, *Optical fibres – Part 2-50: Product specifications – Sectional specification for class B single-mode fibres*

IEC 60825-1, *Safety of laser products – Part 1: Equipment classification and requirements*

IEC 60825-2, *Safety of laser products – Part 2: Safety of optical fibre communication systems (OFCS)*

IEC/TR 61931, *Fibre optic – Terminology*

ISO/IEC 98-3, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO/IEC 17025:2005, *General requirements for the competence of testing and calibration laboratories*

3 Terms and definitions

For the purposes of this document, the terms and definitions contained in IEC/TR 61931, as well as the following terms and definitions, apply.

3.1

accredited calibration laboratory

calibration laboratory authorized by the appropriate national organization to issue calibration certificates with a minimum specified uncertainty, which demonstrate traceability to national measurement standards

3.2

calibration

set of operations that establish, under specified conditions, the relationship between the values of quantities indicated by a measuring instrument and the corresponding values realized by measurement standards