

This is a preview of "PD ISO/TS 22077-2:20...". Click here to purchase the full version from the ANSI store.

PD ISO/TS 22077-2:2015



BSI Standards Publication

Health informatics — Medical waveform format

Part 2: Electrocardiography

bsi.

...making excellence a habit.™

This is a preview of "PD ISO/TS 22077-2:20...". [Click here to purchase the full version from the ANSI store.](#)

This Published Document is the UK implementation of ISO/TS 22077-2:2015.

The UK participation in its preparation was entrusted to Technical Committee IST/35, Health informatics.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2015. Published by BSI Standards Limited 2015

ISBN 978 0 580 86356 1

ICS 35.240.80

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 31 August 2015.

Amendments issued since publication

Date	Text affected
------	---------------

This is a preview of "PD ISO/TS 22077-2:20...". [Click here to purchase the full version from the ANSI store.](#)

First edition
2015-08-01

Health informatics — Medical waveform format —

Part 2: Electrocardiography

*Informatique de santé — Forme d'onde médicale —
Partie 2: Electrocardiographie*



Reference number
ISO/TS 22077-2:2015(E)

© ISO 2015

This is a preview of "PD ISO/TS 22077-2:20...". Click here to purchase the full version from the ANSI store.



COPYRIGHT PROTECTED DOCUMENT

© ISO 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

This is a preview of "PD ISO/TS 22077-2:20...". Click here to purchase the full version from the ANSI store.

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Symbols and abbreviated terms	2
5 Encoding format	2
5.1 Primary description.....	2
5.1.1 Sampling attributes.....	3
5.1.2 Frame attributes.....	3
5.1.3 Waveform.....	4
5.1.4 Channel.....	6
5.2 Data alignment.....	7
5.3 Abstract waveform.....	7
5.4 Lead calculation.....	8
5.5 Filter information.....	9
5.5.1 Description of filter-processed data.....	9
5.5.2 Description of filter use information.....	9
5.6 Unique identifier.....	9
6 Measurement information	10
6.1 Measurement date/time.....	10
6.2 Measurement time (classification point).....	10
6.3 Measurement value.....	11
6.4 Measurement information classification.....	11
6.4.1 Observation event.....	11
6.4.2 Waveform ancillary information.....	11
6.4.3 Recording/display condition.....	11
6.5 Power supply frequency.....	12
6.6 Electrode condition.....	12
6.7 Calibration waveform.....	12
6.8 Artefact contamination.....	12
6.9 Automatic interpretation code, etc.....	13
6.9.1 MFER interpretation code and heart beat code encoding rules.....	13
6.10 Patient information.....	14
6.10.1 Patient name.....	14
6.10.2 Patient ID.....	14
6.10.3 Age and date of birth.....	14
6.10.4 Gender.....	14
6.11 Comment.....	15
Annex A (informative) MFER Conformance statement	16
Annex B (informative) Waveform alignment	17
Annex C (informative) Encoding of waveform recognition point and measurement values	26
Annex D (informative) Reference table of coding scheme	33
Annex E (informative) Waveform verification rule between ECG provider and user	37
Bibliography	38

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 215, *Health informatics*.

ISO/TS 22077 consists of the following parts, under the general title *Health informatics — Medical waveform format*:

- *Part 1: Encoding rules*
- *Part 2: Electrocardiography*
- *Part 3: Long term electrocardiography*

This is a preview of "PD ISO/TS 22077-2:20...". [Click here to purchase the full version from the ANSI store.](#)

Introduction

The standard 12-lead electrocardiogram (ECG) is one of the most widely used medical waveforms in clinical sites. In particular, the increased usage of electronic medical records provides the environment in which these ECGs can be accurately utilized; however, it is essential that to address the therapeutic requirements, ECG use is not constrained to specific machine types and manufacturers. Furthermore, there is great interest in the various kinds of patient information contained in ECGs that are extensively studied and shared between health care providers.

This Technical Specification defines the detailed rules for electrocardiogram waveform format that is encoded according to the medical waveform format encoding rules (MFER). In addition to electrocardiogram waveform format encoding, there are rules for other waveforms such as long-term ECG (Holter ECG), stress ECG, etc. that are contained in other MFER technical specifications. Please refer to those specifications for additional information.

About MFER

Medical waveforms such as electrocardiogram, electroencephalogram, and blood pressure waveforms are widely utilized in clinical areas such as physiological examinations, electronic medical records, medical investigations, research, education, etc. Medical waveforms are used in various combinations and document types according to the intended diagnostic purpose. For example, ECG waveforms are utilized extensively in the clinical arena, with resting 12-lead ECG being used the most. A cardiologist makes diagnoses using 10 s to 15 s ECG waveform measurements; however, longer periods are sometimes required to recognize patient heart conditions such as arrhythmia. Also, there are many other methods using ECG such as Holter ECG, physiologic monitoring ECG, stress ECG, intracardiac ECG, VCG, EEG with ECG, blood pressure with ECG, PSG, etc. MFER can describe not only ECG for physiological examinations conducted in ICU and operating room acute care contexts, but also EEG, respiration waveform, and pulse.

Simple and easy

MFER is a specialized representation for medical waveforms that removes unnecessary coded elements ("tags") for waveform description. For example, a standard 12-lead ECG can be described simply only using a common sampling condition and the lead condition, making waveform synchronization and correct lead calculation much easier.

Using with other appropriate standards

It is recommended that MFER only describes medical waveforms. Other information can be described using appropriate standards such as HL7, DICOM, IEEE, etc. For example, clinical reports that include patient demographics, order information, medication, etc. are supported in other standards such as HL7 Clinical Document Architecture (CDA); by including references to MFER information in these documents, implementation for message exchange, networking, database management that includes waveform information becomes simple and easy.

Separation between supplier and consumer of medical waveforms

The MFER specification concentrates on data format instead of paper-based recording. For example, recorded ECG is processed by filter, data alignment, and other parameters, so that the ECG waveform can be easily displayed using an application viewer. However, it is not as useful for other purposes such as data processing for research investigations. A design goal of MFER is that a waveform is described in raw format with as complete as possible recording detail. When the waveform is used, appropriate processing of the data are supported like filtering, view alignment and so on. In this way, the medical waveform described in MFER can be used for multiple purposes.

Product capabilities are not limited

Standards often support only a minimum set of requirements, so the expansion of product features can be greatly limited. MFER can describe medical waveform information without constraining the potential features of a product. Also, medical waveform display must be very flexible, and thus MFER

This is a preview of "PD ISO/TS 22077-2:20...". [Click here to purchase the full version from the ANSI store.](#)

has mechanisms supporting not only a machine-readable coded system for abstract data, but also human-readable representation.

The MFER specification can support both present and future product implementations. MFER supports the translation of stored waveform data that was encoded using other standards, enabling harmonization and interoperability. This capability supports not only existing waveform format standards, but can be extended to support future formats as well.

This is a preview of "PD ISO/TS 22077-2:20...". Click here to purchase the full version from the ANSI store.

Health informatics — Medical waveform format —

Part 2: Electrocardiography

1 Scope

This Technical Specification defines the application of medical waveform format encoding rules (MFER) to describe standard electrocardiogram waveforms measured in physiological laboratories, hospital wards, clinics, and primary care medical checkups. It covers electrocardiograms such as 12-lead, 15-lead, 18-lead, Cabrera lead, Nehb lead, Frank lead, XYZ lead, and exercise tests that are measured by inspection equipment such as electrocardiographs and patient monitors that are compatible with MFER.

Medical waveforms that are not in the scope of this Technical Specification include Holter ECG, exercise stress ECG, and real-time ECG waveform encoding used for physiological monitors.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 22077-1, *Health informatics — Medical waveform format — Part 1: Encoding rules*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

dominant beat

typical heart beat used for measurement and analysis in standard 12-lead ECG

Note 1 to entry: In general, it is the primary heart beat excepting extrasystole or drifts of baseline.

3.2

average beat

typical heart beat used for measurement and analysis in standard 12-lead ECG

Note 1 to entry: This is averaged for waveforms excluding abnormal beats for each lead.

3.3

median beat

typical heart beat used for measurement and analysis in standard 12-lead ECG

Note 1 to entry: This is a waveform with the median value of waveforms excluding the abnormal beats for each lead.

3.4

tag

identifier code for a semantic concept