

This is a preview of "BS ISO 16000-39:2019". Click here to purchase the full version from the ANSI store.



BSI Standards Publication

Indoor air

Part 39: Determination of amines — Analysis of amines by (ultra-) high-performance liquid chromatography coupled to high resolution or tandem mass spectrometry

This is a preview of "BS ISO 16000-39:2019". Click here to purchase the full version from the ANSI store.

National foreword

This British Standard is the UK implementation of ISO 16000-39:2019.

The UK participation in its preparation was entrusted to Technical Committee EH/2/5, Emissions to internal environments.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2019
Published by BSI Standards Limited 2019

ISBN 978 0 580 89882 2

ICS 13.040.20

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 June 2019.

Amendments/corrigenda issued since publication

Date	Text affected

This is a preview of "BS ISO 16000-39:2019". Click [here](#) to purchase the full version from the ANSI store.

First edition
2019-06-13

Indoor air —

Part 39:

**Determination of amines —
Analysis of amines by (ultra-) high-
performance liquid chromatography
coupled to high resolution or tandem
mass spectrometry**



Reference number
ISO 16000-39:2019(E)

This is a preview of "BS ISO 16000-39:2019". Click here to purchase the full version from the ANSI store.



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

This is a preview of "BS ISO 16000-39:2019". Click here to purchase the full version from the ANSI store.

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Amines in indoor air	2
4.1 Properties of amines	2
4.2 Origin and occurrence of amines	2
5 Analytical procedure	2
5.1 Preparation of sample solutions	2
5.2 High-performance liquid chromatography	2
5.3 Tandem mass spectrometry	3
5.4 High resolution mass spectrometry	3
5.5 Sample sequence and external calibration	3
6 Testing of the suitability of the equipment and the instrumental analysis	3
6.1 General	3
6.2 High-performance or ultra-performance liquid chromatography (HPLC or UHPLC)	3
6.3 Tandem mass spectrometry (MS-MS)	4
6.4 High resolution mass spectrometry (HRMS)	4
7 Reporting requirements	4
Annex A (informative) List of included amines	6
Annex B (informative) Performance data of the analytical method	9
Bibliography	12

This is a preview of "BS ISO 16000-39:2019". Click here to purchase the full version from the ANSI store.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 146, *Air quality*, Subcommittee SC 6, *Indoor air*.

A list of all parts in the ISO 16000 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This is a preview of "BS ISO 16000-39:2019". Click [here](#) to purchase the full version from the ANSI store.

Introduction

ISO 16000 (all parts) describe general requirements relating to the measurement of indoor air pollutants and the important conditions to be observed before or during the sampling of individual pollutants or groups of pollutants, as well as the measurements procedures themselves.

The definition of indoor environment is given by ISO 16000-1. Dwellings [living rooms, bedrooms, do-it-yourself (DIY) rooms, sport rooms and cellars, kitchens and bathrooms], workrooms or workplaces in buildings which are not subject to health and safety inspections with respect to air pollutants (e.g. offices, salesrooms), public buildings (e.g. restaurants, theatres, cinemas and other meeting rooms) and passenger cabins of motor vehicles and public transportation are among the most important types of indoor environment.

This is a preview of "BS ISO 16000-39:2019". Click here to purchase the full version from the ANSI store.

This is a preview of "BS ISO 16000-39:2019". Click here to purchase the full version from the ANSI store.

Indoor air —

Part 39:

Determination of amines — Analysis of amines by (ultra-) high-performance liquid chromatography coupled to high resolution or tandem mass spectrometry

1 Scope

This document, along with ISO 16000-38, specifies the measurement method for determining the mass concentration of primary, secondary and tertiary aliphatic and aromatic amines in indoor air using accumulated sampling and high-performance liquid chromatography (HPLC) coupled with tandem mass spectrometry (MS-MS) or high-resolution mass spectrometry (HRMS). The analytical procedure is covered by this document. The sampling procedure and the manufacturing of the samplers are covered by ISO 16000-38.

This document describes specifications for the chromatography and the mass spectroscopy for the amines. Measurement results are expressed in $\mu\text{g}/\text{m}^3$.

Although primarily intended for the measurement of amines listed in [Tables A.1](#) and [A.2](#), it can also be used for the measurement of other amines in indoor air. This document gives instructions and describes procedures for the inclusion of other amines.

The range of application of this document concerning the concentrations of amines in indoor air depends on the linear range of the calibration line and hence on the gas sample volume (here: from 5 l up to 100 l), the eluate volume (from 1 ml up to 5 ml), the injection volume (from 1 μl up to 10 μl) and the sensitivity of the analytical equipment (e.g. linear range from 2 pg up to 2 ng amine). The range of application can be expected to be from approximately 0,002 $\mu\text{g}/\text{m}^3$ (100 l sample) up to 2 000 $\mu\text{g}/\text{m}^3$ (5 l sample) for a common analytical equipment (e.g. Waters „TQD“) for the majority of the amines listed in [Tables A.1](#) and [A.2](#). The analysis of derivatives of ethanolamine is usually about 10 times more sensitive and the analysis of short-chained aliphatic amines is usually about 10 times less sensitive than the analysis of an average amine.

The performance data of the analytical method is given in Annex B, particularly in [Tables B.1](#) and [B.2](#).

This document can be used also for the determination of amines in water if the detection limit is sufficient.

This document does not cover the determination of isocyanates in indoor air (nor in water samples) as corresponding amines (covered by ISO 17734-1 and ISO 17734-2).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 16000-38, *Indoor air — Part 38: Determination of amines in indoor and test chamber air — Active sampling on samplers containing phosphoric acid impregnated filters*