

This is a preview of "BS EN ISO 10079-1:20...". Click here to purchase the full version from the ANSI store.

Incorporating corrigendum February 2016



BSI Standards Publication

Medical suction equipment

Part 1: Electrically powered suction equipment

bsi.

...making excellence a habit.™

This is a preview of "BS EN ISO 10079-1:20...". [Click here to purchase the full version from the ANSI store.](#)

This British Standard is the UK implementation of EN ISO 10079-1:2015. It supersedes BS EN ISO 10079-1:2009 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/121, Anaesthetic and respiratory equipment.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2016.
Published by BSI Standards Limited 2016

ISBN 978 0 580 93408 7

ICS 11.040.10

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 November 2015.

Amendments/corrigenda issued since publication

Date	Text affected
29 February 2016	Implementation of CEN Correction Notice 2 December 2015: EN Foreword and Annex ZA updated

This is a preview of "BS EN ISO 10079-1:20...". Click here to purchase the full version from the ANSI store.

EUROPÄISCHE NORM

November 2015

ICS 11.040.10

Supersedes EN ISO 10079-1:2009

English Version

Medical suction equipment - Part 1: Electrically powered suction equipment (ISO 10079-1:2015)

Appareils d'aspiration médicale - Partie 1: Appareils électriques d'aspiration (ISO 10079-1:2015)

Medizinische Absauggeräte - Teil 1: Elektrisch betriebene Absauggeräte (ISO 10079-1:2015)

This European Standard was approved by CEN on 13 May 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

This is a preview of "BS EN ISO 10079-1:20...". [Click here to purchase the full version from the ANSI store.](#)

European foreword

This document (EN ISO 10079-1:2015) has been prepared by Technical Committee ISO/TC 121 "Anaesthetic and respiratory equipment" in collaboration with Technical Committee CEN/TC 215 "Respiratory and anaesthetic equipment" the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2016, and conflicting national standards shall be withdrawn at the latest by November 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 10079-1:2009.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 10079-1:2015 has been approved by CEN as EN ISO 10079-1:2015 without any modification.

This is a preview of "BS EN ISO 10079-1:20...". Click here to purchase the full version from the ANSI store.

Annex ZA
(informative)

Relationship between this European Standard and the essential requirements of Directive 93/42/EEC [OJ L 169] aimed to be covered

This European Standard has been prepared under a Commission's standardization request M/023 concerning the development of European Standards related to medical devices to provide one voluntary means of conforming to essential requirements of Council Directive 93/42/EEC of 14 June 1993 concerning medical devices [OJ L 169].

Once this standard is cited in the Official Journal of the European Union under that Directive, compliance with the normative clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding essential requirements of that Directive and associated EFTA regulations.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Directive 93/42/EEC as amended by 2007/47/EC. This means that risks have to be reduced 'as far as possible', 'to a minimum', 'to the lowest possible level', 'minimized' or 'removed', according to the wording of the corresponding essential requirement.

NOTE 2 The manufacturer's policy for determining acceptable risk must be in compliance with Essential Requirements 1, 2, 5, 6, 7, 8, 9, 11 and 12 of the Directive.

NOTE 3 When an Essential Requirement does not appear in Table ZA.1, it means that it is not addressed by this European Standard.

Table ZA.1 — Correspondence between this European Standard and Annex I of Directive 93/42/EEC [OJ L 169]

Essential Requirements of Directive 93/42/EEC	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
7.1 Third indent only	4.4	
7.2	5; 7.5	Partly covered There are no requirements for packaging.
7.3 First part only	6.1.3	
7.6	6.2.3; 6.5; 7.5.1; 7.5.2	
8.1	4.2; 5; 7.5.1	
8.7	11.3 c)	
9.1 First sentence only	6.2; 6.3	
9.2	4; 6.1.3	Partly covered Electrical safety is by ref to IEC 60601-1 and risk management by ref to ISO 14971.

This is a preview of "BS EN ISO 10079-1:20...". Click here to purchase the full version from the ANSI store.

Essential Requirements of Directive 93/42/EEC	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
10.1	6.4.6	Partly covered. There are no requirements for the manufacturer to disclose the accuracy of the vacuum level indicator.
10.2	6.4	
10.3	11.3 i)	Covered for volume measurements only
12.1	4	Covered by ref to IEC 60601-1
12.1a)	4	Covered by ref to IEC 60601-1
12.2	4	Covered by ref to IEC 60601-1 although suction equipment is not considered life-support equipment.
12.5	4	Covered by ref to IEC 60601-1 and thereby to IEC 60601-1-2
12.6	4; 6.5	Covered by ref to IEC 60601-1
12.7.1	6.1.3; 7.4	
12.7.2	4	Covered by ref to IEC 60601-1
12.7.3	7.6	
12.7.4	4	Covered by ref to IEC 60601-1
12.7.5	4	Covered by ref to IEC 60601-1
12.8.2 Second sentence only	7.5.3.2	
12.9	11.3 i); j); k); l); m); n); o); p); q); r)	
13.1	11	
13.2	11.2	
13.3a)	11.3 a)	
13.3b)	11.3 b)	
13.3c)	11.3 c)	
13.3d)	11.3 d)	
13.3e)	11.3 e)	
13.3 f)	11.3 f)	
13.3 k)	11.4 c); q); y)	
13.3 l)	11.3 d)	
13.3 m)	11.4 i)	
13.4	11.4 b)	
13.6 a)	11.4	Not covered for the requirement of ER 13.3b)
13.6 b)	11.4 d); e)	

This is a preview of "BS EN ISO 10079-1:20...". Click here to purchase the full version from the ANSI store.

Essential Requirements of Directive 93/42/EEC	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
13.6 c)	11.4 d) ;k)	
13.6 d)	11.4 d); j); v)	Calibration is not covered
13.6 f)	11.4 x)	
13.6 h) First two paragraphs only	11.4 i)	
13.6 i)	11.4 j)	
13.6 q)	11.4 z)	

WARNING 1 — Presumption of conformity stays valid only as long as a reference to this European Standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

WARNING 2 — Other Union legislation may be applicable to the products falling within the scope of this standard.

This is a preview of "BS EN ISO 10079-1:20...". Click here to purchase the full version from the ANSI store.

Contents

Page

Foreword	viii
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 General requirements	4
4.1 Risk management.....	4
4.2 Usability.....	5
4.3 Clinical investigation.....	5
4.4 Biophysical or modelling research.....	5
4.5 Test methods.....	5
5 Cleaning, disinfection and sterilization	5
6 Design requirements	6
6.1 Collection container.....	6
6.1.1 General.....	6
6.1.2 Container capacity.....	6
6.1.3 Container strength.....	6
6.2 Connections.....	6
6.2.1 Tubing connectors for collection containers.....	6
6.2.2 Inlet port.....	7
6.2.3 Exhaust port.....	7
6.3 Suction tubing.....	7
6.4 Vacuum level indicators.....	7
6.5 Spillage on electrical suction equipment.....	8
7 Operational requirements	8
7.1 Ease of operation.....	8
7.2 Dismantling and reassembly.....	8
7.3 Mechanical shock.....	8
7.4 Stability.....	8
7.5 Protective devices.....	9
7.5.1 Contamination protection.....	9
7.5.2 Overfill protection devices.....	9
7.5.3 Pressure protection.....	9
7.6 Noise.....	9
7.6.1 Low vacuum/low flowrate equipment.....	9
7.6.2 Suction equipment other than that specified in 7.6.1.....	9
7.7 Air leakage.....	10
7.7.1 Collection containers for general use.....	10
7.7.2 Collection containers for thoracic drainage.....	10
8 Physical requirements for suction equipment for field use	10
8.1 (*) Dimensions.....	10
8.2 Mass.....	10
9 Performance requirements for vacuum level and flowrate	10
9.1 High vacuum/high flowrate equipment.....	10
9.2 Medium vacuum equipment.....	11
9.3 Low vacuum/low flowrate equipment.....	11
9.4 Low vacuum/high flowrate equipment.....	11
9.5 Thoracic drainage equipment for adults.....	11
9.6 Intermittent vacuum equipment.....	11
9.7 Vacuum regulators with fixed setting.....	11
9.8 Vacuum regulators with variable setting.....	12
9.9 Equipment intended for pharyngeal suction.....	12

This is a preview of "BS EN ISO 10079-1:20...". [Click here to purchase the full version from the ANSI store.](#)

9.10	Battery powered transportable suction equipment.....	12
9.11	Interruption of the power supply	12
10	(*) Resistance to environment of suction equipment for field and/or transport use.....	12
10.1	Operating conditions	12
10.2	Storage	12
11	Information to be supplied by the manufacturer (labelling and instructions for use).....	13
11.1	Information supplied by the manufacturer shall comply with EN 1041.....	13
11.3	Labelling of equipment.....	13
11.4	Instructions for use.....	14
Annex A (normative) Test methods.....		16
Annex B (informative) Rationale statement.....		27
Annex C (informative) Lumen size and its effect on flowrate.....		28
Annex D (informative) Schematic of suction equipment.....		29
Bibliography.....		30

This is a preview of "BS EN ISO 10079-1:20...". Click here to purchase the full version from the ANSI store.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 8, *Suction devices for hospital and emergency care use*.

This third edition cancels and replaces the second edition (ISO 10079-1:1999), which has been technically revised.

ISO 10079 consists of the following parts, under the general title *Medical suction equipment*:

- *Part 1: Electrically powered suction equipment*
- *Part 2: Manually powered suction equipment*
- *Part 3: Suction equipment powered from a vacuum or positive pressure gas source*

[Annex A](#) forms a normative part of this part of ISO 10079 while [Annex B](#), [Annex C](#), and [Annex D](#) are for information only.

[Annex B](#) contains rationale statements for some of the requirements of this part of ISO 10079. The clauses and subclauses marked with an asterisk (*) at the beginning of the paragraph have corresponding rationale contained in [Annex B](#) included to provide additional insight into the reasoning that led to the requirements and recommendations that have been incorporated into this part of ISO 10079. It is considered that knowledge of the reasons for the requirements will not only facilitate the proper application of this part of ISO 10079, but will expedite any subsequent revisions.

[Annex D](#) illustrates the three parts of ISO 10079 by providing a schematic for typical systems.

This is a preview of "BS EN ISO 10079-1:20...". Click here to purchase the full version from the ANSI store.

Medical suction equipment —

Part 1:

Electrically powered suction equipment

1 Scope

This part of ISO 10079 specifies safety and performance requirements for electrically powered medical and surgical suction equipment. It applies to equipment used in health care facilities such as hospitals, for domiciliary care of patients and for field and transport use.

This part of ISO 10079 does not apply to the following:

- a) central power supply (by vacuum/compressed air generation), piping systems of vehicles and buildings, and wall connectors;
- b) end-pieces such as suction catheters, drains, curettes, Yankauer suckers and suction tips;
- c) syringes;
- d) dental suction equipment;
- e) anaesthetic gas scavenging systems;
- f) laboratory suction;
- g) autotransfusion systems;
- h) mucus extractors including neonatal mucus extractors;
- i) suction equipment where the collection container is downstream of the vacuum pump;
- j) ventouse (obstetric) equipment;
- k) suction equipment marked for endoscopic use only;
- l) plume evacuation systems.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3744, *Acoustics — Determination of sound power levels and sound energy levels of noise sources using sound pressure — Engineering methods for an essentially free field over a reflecting plane*

ISO 5356-1, *Anaesthetic and respiratory equipment — Conical connectors — Part 1: Cones and sockets*

ISO 7000, *Graphical symbols for use on equipment — Registered symbols*¹⁾

ISO 14155, *Clinical investigation of medical devices for human subjects — Good clinical practice*

ISO 14971, *Medical devices — Application of risk management to medical devices*

1) The graphical symbol collections of ISO 7000, ISO 7001, and ISO 7010 are also available on the Online Browsing Platform <http://www.iso.org/obp>.