

This is a preview of "BS ISO 16000-6:2021". [Click here to purchase the full version from the ANSI store.](#)



BSI Standards Publication

Indoor air

Part 6: Determination of organic compounds (VVOC, VOC, SVOC) in indoor and test chamber air by active sampling on sorbent tubes, thermal desorption and gas chromatography using MS or MS FID

This is a preview of "BS ISO 16000-6:2021". [Click here to purchase the full version from the ANSI store.](#)

National foreword

This British Standard is the UK implementation of ISO 16000-6:2021. It supersedes BS ISO 16000-6:2011, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee EH/2/5, Emissions to internal environments.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2021
Published by BSI Standards Limited 2021

ISBN 978 0 580 99090 8

ICS 13.040.20

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 August 2021.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

This is a preview of "BS ISO 16000-6:2021". [Click here to purchase the full version from the ANSI store.](#)

Third edition
2021-08-20

Indoor air —

Part 6:

Determination of organic compounds (VVOC, VOC, SVOC) in indoor and test chamber air by active sampling on sorbent tubes, thermal desorption and gas chromatography using MS or MS FID

Air intérieur —

Partie 6: Dosage des composés organiques (COTV, COV, COSV) dans l'air intérieur et l'air de chambre d'essai par prélèvement actif sur tubes à sorbant, désorption thermique et chromatographie en phase gazeuse avec détection MS ou MS-FID



Reference number
ISO 16000-6:2021(E)

© ISO 2021

This is a preview of "BS ISO 16000-6:2021". [Click here to purchase the full version from the ANSI store.](#)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2021, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

This is a preview of "BS ISO 16000-6:2021". Click here to purchase the full version from the ANSI store.

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Abbreviated terms	3
5 Principle	3
6 Reagents and materials	3
6.3 Sorbents.....	3
6.3.1 General.....	3
6.4 Preparing calibration standards on sorbent tubes.....	4
6.4.1 Gas-phase standards.....	4
6.4.2 Loading sorbent tubes with gas-phase standards.....	4
6.4.3 Calibration blend solution for liquid spiking.....	4
6.4.4 Loading sorbent tubes with liquid standards.....	5
6.4.5 Commercial, pre-loaded standard tubes.....	5
7 Apparatus	5
7.1 Sorbent tubes of stainless steel or glass,.....	6
7.1.1 General.....	6
7.1.2 Sorbent tubes — Combinations and options.....	6
7.2 Sorbent tube unions.....	8
7.2.1 Types of thermal desorber.....	8
7.2.2 General requirements.....	8
7.9 Injection facility.....	9
8 Conditioning and storage of sorbent tubes	9
8.1 Conditioning.....	9
8.2 Storage of conditioned sorbent tubes before sampling.....	10
9 Sampling	10
9.1 Air sampling.....	10
9.2 Sampling volumes.....	10
9.3 Storage of loaded samples.....	11
9.4 Field blanks.....	11
10 Analysis	11
10.1 Determination of VOC, VVOC and SVOC.....	11
10.1.1 Analytical System.....	11
10.1.2 Identification of target compounds.....	13
10.1.3 Quantification of target compounds and compounds according to task list.....	14
10.1.4 Determining the lower limit of quantification.....	15
10.2 Identified non-target compounds and unidentified compounds.....	16
11 Concentration of analytes in the sampled air	16
12 Performance characteristics	17
13 Test report	18
14 Quality control	19
Annex A (informative) Total volatile organic compounds (TVOC) and total semi-volatile organic compounds (TSVOC)	20
Annex B (informative) Examples of compounds detected in indoor air and from building products in test chambers	22

This is a preview of "BS ISO 16000-6:2021". [Click here to purchase the full version from the ANSI store.](#)

Annex C (informative) Description of sorbents	27
Annex D (informative) Guide on sorbent usage	28
Annex E (informative) Safe sampling volumes for selected organic vapours	30
Annex F (informative) Storage recovery of solvents on sorbent tubes	33
Bibliography	35

This is a preview of "BS ISO 16000-6:2021". [Click here to purchase the full version from the ANSI store.](#)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 146, *Air quality*, Subcommittee SC 6, *Indoor air*.

This third edition cancels and replaces the second edition (ISO 16000-6:2011), which has been technically revised. The main changes compared to the previous edition are as follows:

- other sorbents than Tenax TA® are allowed to be used;
- descriptions of VVOC and SVOC measurements are included in the mandatory part of the document.

A list of all parts in the ISO 16000 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This is a preview of "BS ISO 16000-6:2021". Click here to purchase the full version from the ANSI store.

Introduction

ISO 16000-1 establishes general requirements relating to the measurement of indoor air pollutants and the important conditions to be observed before or during the sampling of individual pollutants or groups of pollutants. Aspects of the determination (sampling and analysis) and the sampling strategy of specific pollutants or groups of pollutants are specified in the subsequent parts of ISO 16000 (see Foreword).

ISO 16000-5 (dealing with VOC sampling strategy) is a link between ISO 16000-1 (a generic standard establishing the principles) and this part of ISO 16000, which deals with sampling and analytical measurements.

ISO 16017 (see [Clause 2](#) and Reference [8]) and ISO 12219 [3]-[7] also focus on measuring vapour-phase organic chemicals in air.

This document can be applied to measure vapour phase organic compounds in indoor environments that include buildings with varying designs and purposes and cabins for different modes of transport, as well as measurement in product emission test chambers. These measurements can be for a range of purposes as described in ISO 16000-1 and ISO 16000-5, therefore the requirement for the measurement may be well defined by the task descriptor or may be quite open. For example, the task may be to determine a specific list of target chemicals with a defined sampling time and sensitivity of measurement or it may be to investigate the cause of a reported and poorly understood indoor air quality problem. Depending upon the task of measurement the user of this document should select the most appropriate sampling and analytical instrumentation and conditions. This document provides that information in the normative part combined with informative guidance. [Figure 1](#) refers to the most critical parts of the standard with regard to selection of the most appropriate methodology for the task to be undertaken. Tenax TA^{® 1)} only or multisorbents can be used to capture ranges of vapour phase organic compounds. Multisorbents are used for wider ranges and may improve recovery of compounds.




Figure 1 — Measurement scheme showing different ways of analysing air samples depending on the respective task including target compounds

1) Tenax TA[®] is a trade name of a product supplied by Buchem. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of the product named. Equivalent products may be used, if they can be shown to lead to the same results.

This is a preview of "BS ISO 16000-6:2021". Click here to purchase the full version from the ANSI store.

Indoor air —

Part 6:

Determination of organic compounds (VVOC, VOC, SVOC) in indoor and test chamber air by active sampling on sorbent tubes, thermal desorption and gas chromatography using MS or MS FID

1 Scope

This document specifies a method for determination of volatile organic compounds (VOC) in indoor air and in air sampled for the determination of the emission from products or materials used in indoor environments (according to ISO 16000-1) using test chambers and test cells. The method uses sorbent sampling tubes with subsequent thermal desorption (TD) and gas chromatographic (GC) analysis employing a capillary column and a mass spectrometric (MS) detector with or without an additional flame ionisation detector (FID) [[13](#)].

The method is applicable to the measurement of most GC-compatible vapour-phase organic compounds at concentrations ranging from micrograms per cubic metre to several milligrams per cubic metre. Many very volatile organic compounds (VVOC) and semi-volatile organic compounds (SVOC) can be analysed depending on the sorbents used.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 16000-1, *Indoor air — Part 1: General aspects of sampling strategy*

EN 13137, *Workplace atmospheres – Pumps for personal sampling of chemical and biological agents – Requirements and test methods*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

semi-volatile organic compound SVOC

organic compound eluting after n-hexadecane on a gas chromatographic column specified as a 5 % phenyl 95 % methyl polysiloxane phase capillary gas chromatographic column

Note 1 to entry: The vapour-fraction of SVOC ranging in volatility to n-C₄₄ can also be analysed by thermal desorption GC-MS but requires specific sampling and analytical conditions for optimum performance [[22](#) , [25](#)].