



BSI Standards Publication

Sterilization of health care products — Microbiological methods

Part 1: Determination of a population of microorganisms on products

This is a preview of "BS EN ISO 11737-1:20...". Click here to purchase the full version from the ANSI store.

National foreword

This British Standard is the UK implementation of EN ISO 11737-1:2018+A1:2021. It is identical to ISO 11737-1:2018, incorporating amendment 1:2021. It supersedes BS EN ISO 11737-1:2018, which is withdrawn.

The start and finish of text introduced or altered by amendment is indicated in the text by tags. Tags indicating changes to ISO text carry the number of the ISO amendment. For example, text altered by ISO amendment 1 is indicated by **A1** **A1**.

The UK participation in its preparation was entrusted to Technical Committee CH/198, Sterilization and Associated Equipment and Processes.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

This publication has been prepared under a mandate given to the European Standards Organizations by the European Commission and the European Free Trade Association. It is intended to support essential requirements of the EU legislation detailed in the European Foreword. A European Annex, usually Annex ZA or ZZ, describes how this publication relates to that EU legislation.

For the Great Britain market (England, Scotland and Wales), if UK Government has designated this publication for conformity with UKCA marking (or similar) legislation, it may contain an additional National Annex. Where such a National Annex exists, it shows the correlation between this publication and the relevant UK legislation. If there is no National Annex of this kind, the relevant Annex ZA or ZZ in the body of the European text will indicate the relationship to UK regulation applicable in Great Britain. References to EU legislation may need to be read in accordance with the UK designation and the applicable UK law. Further information on designated standards can be found at www.bsigroup.com/standardsandregulation.

For the Northern Ireland market, UK law will continue to implement relevant EU law subject to periodic confirmation. Therefore Annex ZA/ZZ in the European text, and references to EU legislation, are still valid for this market.

This is a preview of "BS EN ISO 11737-1:20...". Click here to purchase the full version from the ANSI store.

UK Government is responsible for legislation. For information on legislation and policies relating to that legislation, consult the relevant pages of www.gov.uk.

© The British Standards Institution 2021
Published by BSI Standards Limited 2021

ISBN 978 0 539 02423 4

ICS 07.100.10; 11.080.01

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 May 2018.

Amendments/corrigenda issued since publication

Date	Text affected
31 May 2018	Implementation of CEN correction notice 14 March 2018: European foreword has been updated and Annexes ZA, ZB and ZC have been added
30 June 2018	Implementation of CEN correction notice 18 April 2018: Z Annexes updated
30 June 2021	Implementation of ISO amendment 1:2021 with CEN endorsement A1:2021 and updating of Annexes ZA and ZB

This is a preview of "BS EN ISO 11737-1:20...". [Click here to purchase the full version from the ANSI store.](#)

EUROPÄISCHE NORM

June 2021

ICS 07.100.10; 11.080.01

English Version

Sterilization of health care products - Microbiological methods - Part 1: Determination of a population of microorganisms on products (ISO 11737-1:2018)

Stérilisation des produits de santé - Méthodes microbiologiques - Partie 1: Détermination d'une population de microorganismes sur des produits (ISO 11737-1:2018)

Sterilisation von Produkten für die Gesundheitsfürsorge - Mikrobiologische Verfahren - Teil 1: Bestimmung der Population von Mikroorganismen auf Produkten (ISO 11737-1:2018)

This European Standard was approved by CEN on 6 December 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

This is a preview of "BS EN ISO 11737-1:20...". Click here to purchase the full version from the ANSI store.

European foreword

This document (EN ISO 11737-1:2018) has been prepared by Technical Committee ISO/TC 198 "Sterilization of health care products" in collaboration with Technical Committee CEN/TC 204 "Sterilization of medical devices" the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2018, and conflicting national standards shall be withdrawn at the latest by July 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 11737-1:2006.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive.

For relationship with EU Directive(s), see informative Annex ZA, ZB, and ZC, which are integral parts of this document.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

The following referenced documents are indispensable for the application of this document. For undated references, the edition of the referenced document (including any amendments) listed below applies. For dated references, only the edition cited applies. However, for any use of this standard within the meaning of Annex ZA, ZB or ZC, the user should always check that any referenced document has not been superseded and that its relevant contents can still be considered the generally acknowledged state-of-art.

When an IEC or ISO standard is referred to in the ISO standard text, this should be understood as a normative reference to the corresponding EN standard, if available, and otherwise to the dated version of the ISO or IEC standard as listed below.

NOTE 1 The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

Table — Correlation between normative references and dated EN and ISO standards

Normative references as listed in Clause 2 of the ISO standard	Equivalent dated standard	
	EN	ISO
ISO 10012	EN ISO 10012:2003	ISO 10012:2003
ISO 15189	EN ISO 15189:2012	ISO 15189:2012
ISO/IEC 17025	EN ISO/IEC 17025:2017	ISO/IEC 17025:2017
ISO 13485	EN ISO 13485:2016	ISO 13485:2016

NOTE 2 Many of the standards normatively referred to by ISO 11737-1 are undated. These referred standards also include normative references themselves to other dated and undated standards. For undated normative references, it should always be assumed that the latest edition applies.

This is a preview of "BS EN ISO 11737-1:20...". Click here to purchase the full version from the ANSI store.

The text of ISO 11737-1:2018 has been approved by CEN as EN ISO 11737-1:2018 without any modification.

Foreword to amendment A1

This document (EN ISO 11737-1:2018/A1:2021) has been prepared by Technical Committee ISO/TC 198 "Sterilization of health care products" in collaboration with Technical Committee CEN/TC 204 "Sterilization of medical devices" the secretariat of which is held by BSI.

This Amendment to the European Standard EN ISO 11737-1:2018 shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by December 2021, and conflicting national standards shall be withdrawn at the latest by December 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s) and Regulations, see informative [Annex ZA](#) and [ZB](#), which are an integral part of this document.

The following referenced documents are indispensable for the application of this document. For undated references, the edition of the referenced document (including any amendments) listed below applies. For dated references, only the edition cited applies. However, for any use of this standard within the meaning of [Annex ZA](#) or [ZB](#), the user should always check that any referenced document has not been superseded and that its relevant contents can still be considered the generally acknowledged state-of-art.

When an IEC or ISO standard is referred to in the ISO standard text, this should be understood as a normative reference to the corresponding EN standard, if available, and otherwise to the dated version of the ISO or IEC standard as listed below.

NOTE The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

Table — Correlation between normative references and dated EN and ISO standards

Normative references as listed in Clause 2 of the ISO standard	Equivalent dated standard	
	EN	ISO
ISO 10012	EN ISO 10012:2003	ISO 10012:2003
ISO 13485	EN ISO 13485:2016	ISO 13485:2016
ISO 15189	EN ISO 15189:2012	ISO 15189:2012
ISO/IEC 17025	EN ISO/IEC 17025:2017	ISO/IEC 17025:2017

NOTE The standards normatively referred to by ISO 11737-1:2018/Amd 1:2021 are undated. These referred standards also include normative references to other dated and undated standards. For undated normative references, it should always be assumed that the latest edition applies.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

This is a preview of "BS EN ISO 11737-1:20...". [Click here to purchase the full version from the ANSI store.](#)

Endorsement notice

The text of ISO 11737-1:2018/Amd1:2021 has been approved by CEN as EN ISO 11737-1:2018/A1:2021 without any modification.

This is a preview of "BS EN ISO 11737-1:20...". Click here to purchase the full version from the ANSI store.

Annex ZA (informative)

Relationship between this European standard and the General Safety and Performance Requirements of Regulation (EU) 2017/745 aimed to be covered

This European standard has been prepared under a Commission's standardisation request to provide one voluntary means of conforming to the General Safety and Performance Requirements of Regulation (EU) 2017/745 of 5 April 2017 concerning medical devices [O] L 117].

This document is an adoption of an International Standard. As the scope of the applicable regulatory requirements differ from nation to nation and region to region, the scope of this document can differ from the scope of the European Regulations that it supports. This document supports European regulatory requirements only to the extent of the scope of the European regulations for medical devices ((EU) 2017/745).

Once this standard is cited in the Official Journal of the European Union under that Regulation, compliance with the normative clauses of this standard given in [Table ZA.1](#) confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding General Safety and Performance Requirements of that Regulation, and associated EFTA regulations.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Regulation (EU) 2017/745. This means that risks have to be 'reduced as far as possible', 'reduced to the lowest possible level', 'reduced as far as possible and appropriate', 'removed or reduced as far as possible', 'eliminated or reduced as far as possible', 'removed or minimized as far as possible', or 'minimized', according to the wording of the corresponding General Safety and Performance Requirement.

NOTE 2 The manufacturer's policy for determining acceptable risk must be in compliance with General Safety and Performance Requirements 1, 2, 3, 4, 5, 8, 9, 10, 11, 14, 16, 17, 18, 19, 20, 21 and 22 of the Regulation.

NOTE 3 This [Annex ZA](#) is based on normative references according to the table of references in the European Foreword, replacing the references in the core text.

NOTE 4 When a General Safety and Performance Requirement does not appear in [Table ZA.1](#), it means that it is not addressed by this European Standard.

This is a preview of "BS EN ISO 11737-1:20...". [Click here to purchase the full version from the ANSI store.](#)

Table ZA.1 — Correspondence between this European standard and Annex I of Regulation (EU) 2017/745 [OJ L 117]

General Safety and Performance Requirements of Regulation (EU) 2017/745	Clause(s) / sub-clause(s) of this EN	Remarks / Notes
11.3	4,5,6,7,8,9	<p>This standard addresses the determination of bioburden in the validation and maintenance of a sterilization process for medical devices. It could also be applied in the development, validation and routine control of a process for attainment of a specific microbial state other than sterility.</p> <p>This relevant General Safety and Performance Requirement is only partly addressed in this European Standard. Design and packaging for maintenance of a specific microbial state during transportation and storage are not covered. Aspects of manufacture other than those related to determination of bioburden in attainment of a specific microbial state are not covered.</p>
11.4 first sentence only	4,5,6,7,8,9	<p>This standard addresses the determination of bioburden in the validation and maintenance of a sterilization process for medical devices.</p> <p>This relevant General Safety and Performance Requirement is only partly addressed in this European Standard. Design and packaging for maintenance of sterility during transportation and storage are not covered. Aspects of manufacture other than those related to determination of bioburden in definition, validation and maintenance of a sterilization process are not covered. Evidence that the integrity of the packaging is maintained to the point of use is not covered.</p>
11.5	4,5,6,7,8,9	<p>This standard addresses the determination of bioburden in the validation and maintenance of a sterilization process for medical devices.</p> <p>This relevant General Safety and Performance Requirement is only partly addressed in this European Standard. Packaging for maintenance of sterility are not covered. Aspects of manufacture other than those related to determination of bioburden in definition, validation and maintenance of a sterilization process are not covered.</p>

WARNING 1 Presumption of conformity stays valid only as long as a reference to this European standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

WARNING 2 Other Union legislation may be applicable to the product(s) falling within the scope of this standard.

This is a preview of "BS EN ISO 11737-1:20...". Click here to purchase the full version from the ANSI store.

Annex ZB (informative)

Relationship between this European standard and the General Safety and Performance Requirements of Regulation (EU) 2017/746 aimed to be covered

This European standard has been prepared under a Commission's standardisation request to provide one voluntary means of conforming to the General Safety and Performance Requirements of Regulation (EU) 2017/746 of 5 April 2017 concerning in vitro diagnostic medical devices [O] L 117].

This document is an adoption of an International Standard. As the scope of the applicable regulatory requirements differ from nation to nation and region to region, the scope of this document can differ from the scope of the European Regulations that it supports. This document supports European regulatory requirements only to the extent of the scope of the European regulations for in vitro diagnostic medical devices ((EU) 2017/746).

Once this standard is cited in the Official Journal of the European Union under that Regulation, compliance with the normative clauses of this standard given in [Table ZB.1](#) confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding General Safety and Performance Requirements of that Regulation, and associated EFTA regulations.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Regulation (EU) 2017/746. This means that risks have to be 'reduced as far as possible', 'reduced to a level as low as reasonably practicable', 'reduced to the lowest possible level', 'reduced as far as possible and appropriate', 'removed or reduced as far as possible', 'eliminated or reduced as far as possible', 'prevented' or 'minimized', according to the wording of the corresponding General Safety and Performance Requirement.

NOTE 2 The manufacturer's policy for determining acceptable risk must be in compliance with General Safety and Performance Requirements 1, 2, 3, 4, 5, 8, 10, 11, 13, 15, 16, 17, 18 and 19 of the Regulation.

NOTE 3 This [Annex ZB](#) is based on normative references according to the table of references in the European Foreword, replacing the references in the core text.

NOTE 4 When a General Safety and Performance Requirement does not appear in [Table ZB.1](#), it means that it is not addressed by this European Standard.

This is a preview of "BS EN ISO 11737-1:20...". [Click here to purchase the full version from the ANSI store.](#)

Table ZB.1 — Correspondence between this European standard and Annex I of Regulation (EU) 2017/746 [OJ L 117]

General Safety and Performance Requirements of Regulation (EU) 2017/746	Clause(s) / sub-clause(s) of this EN	Remarks / Notes
11.2	4,5,6,7,8,9	<p>This standard addresses the determination of bioburden in the validation and maintenance of a sterilization process for medical devices. It could also be applied to the development or validation of a process for attainment of a specific microbial state other than sterility.</p> <p>This relevant General Safety and Performance Requirement is only partly addressed in this European Standard. Design and packaging for maintenance of a sterility or another specific microbial state during transportation and storage are not covered. Aspects of manufacture other than those related to determination of bioburden in attainment of a specific microbial state are not covered</p>
11.3	4,5,6,7,8,9	<p>This standard addresses the determination of bioburden in the, validation and maintenance of a sterilization process for medical devices.</p> <p>This relevant General Safety and Performance Requirement is only partly addressed in this European Standard. Packaging for maintenance of sterility is not covered. Aspects of manufacture other than those related to determination of bioburden in definition, validation and maintenance of a sterilization process are not covered.</p>

WARNING 1 Presumption of conformity stays valid only as long as a reference to this European standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

WARNING 2 Other Union legislation may be applicable to the product(s) falling within the scope of this standard.

This is a preview of "BS EN ISO 11737-1:20...". Click here to purchase the full version from the ANSI store.

Annex ZC (informative)

Relationship between this European Standard and the Essential Requirements of EU Directive 98/79/EC on in vitro diagnostic medical devices [OJ L 331] aimed to be covered

This European standard has been prepared under a Commission's standardisation request, M/252, concerning the development of European standards relating to in vitro diagnostic medical devices, to provide one voluntary means of conforming to essential requirements of Directive 98/79/EC of the European Parliament and of the Council of 27 October 1998 on in vitro diagnostic medical devices [OJ L 331].

Once this standard is cited in the Official Journal of the European Union under that Directive, compliance with the normative clauses of this standard given in [Table ZC.1](#) confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding Essential Requirements of that Directive and associated EFTA regulations.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with 98/79/EC. This means that risks have to be reduced 'as far as possible', 'to a minimum', 'to the lowest possible level', 'minimized' or 'removed', according to the wording of the corresponding essential requirement.

NOTE 2 The manufacturer's policy for determining acceptable risk must be in compliance with essential requirements Part A: 1, 2 and 5; Part B: 1.2, 2, 3, 5, 6, and 7 of the Directive.

NOTE 3 This [Annex ZC](#) is based on normative references according to the table of references in the European foreword, replacing the references in the core text.

NOTE 4 When an Essential Requirement does not appear in [Table ZC.1](#), it means that it is not addressed by this European Standard.

Table ZC.1 — Correspondence between this European Standard and Annex I of Directive 98/79/EC [OJ L 331]

Essential Requirements (ERs) of Directive 98/79/EC	Clauses of this EN	Qualifying remarks/Notes
B.2.3	4,5,6,7,8,9	<p>This standard addresses the determination of the population of microorganisms on or in a medical device as part of the validation and routine control of a sterilization process.</p> <p>This relevant Essential Requirement is partly addressed in this European Standard and only in conjunction with the applicable standard for validation and routine control of the sterilization process being employed. Design and packaging for maintenance of sterility during transportation and storage are not covered. Aspects of manufacture other than those related to sterilization are not covered.</p>

This is a preview of "BS EN ISO 11737-1:20...". [Click here to purchase the full version from the ANSI store.](#)

Essential Requirements (ERs) of Directive 98/79/EC	Clauses of this EN	Qualifying remarks/Notes
B.2.4	4,5,6,7,8,9	This relevant Essential requirement is addressed only with regard to determination of the population of microorganisms for the validation and routine control of sterilization.

WARNING 1 — Presumption of conformity stays valid only as long as a reference to this European Standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

WARNING 2 — Other Union legislation may be applicable to the products falling within the scope of this standard.

This is a preview of "BS EN ISO 11737-1:20...". Click here to purchase the full version from the ANSI store.

Contents

	Page
Foreword	xii
Introduction	xiv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
A1 4 General	4
5 Selection of products	5
5.1 General.....	5
5.2 Sample item portion (SIP).....	5
6 Methods of determination and microbial characterization of bioburden	6
6.1 Determination of bioburden.....	6
6.1.1 Selection of an appropriate method.....	6
6.1.2 Neutralization of inhibitory substances.....	6
6.1.3 Removal of microorganisms.....	6
6.1.4 Culturing of microorganisms.....	7
6.1.5 Enumeration of microorganisms.....	7
6.2 Microbial characterization of bioburden.....	7
7 Validation of the method for determining bioburden	7
7.1 General.....	7
7.2 Validation.....	7
8 Routine determination of bioburden and interpretation of data	8
8.1 General.....	8
8.2 Limits of detection and plate counting.....	8
8.3 Microbial characterization.....	8
8.4 Bioburden data for extent of treatment.....	8
8.5 Bioburden spikes.....	8
8.6 Bioburden levels.....	8
8.7 Data analysis.....	8
8.8 Statistical methods.....	8
9 Maintenance of the method for determining bioburden	9
9.1 Changes to the product and/or manufacturing process.....	9
9.2 Changes to the method for determining bioburden.....	9
9.3 Requalification of the method for determining bioburden.....	9
Annex A (informative) Guidance on the determination of a population of microorganisms on products	10
Annex B (informative) Guidance on methods to determine bioburden	27
Annex C (informative) Validation of bioburden recovery efficiency	37
Annex D (informative) Typical assignment of responsibilities	45
Bibliography	47

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

This third edition cancels and replaces the second edition (ISO 11737-1:2006), which has been technically revised. It also incorporates the Technical Corrigendum ISO 11737-1:2006/Cor.1:2007.

The main changes compared to the previous edition are as follows:

- the term “bioburden spikes” has been introduced as a normal and consistent part of the bioburden, and examples of data have been provided;
- clarification has been added that package testing is not typically done except when it is an integral part of the product;
- more information has been provided on the most probable number (MPN) technique and its applications;
- details have been provided on ways to improve limit of detection (LOD) and correct use of the data;
- some discussion has been deleted of statistical methods for the evaluation of bioburden data where information was not typical or not required;
- a table has been added with criteria for selection of a bioburden recovery efficiency approach, the use of the correction factor (CF) has been explained, and the bioburden recovery efficiency value of < 50 % mentioned for technique modifications has been eliminated;
- more information has been provided on the application and performance of a bioburden method suitability test;
- a section has been added to detail rules for direct plate counts, estimated counts and counts beyond the ideal range;
- a table has been added to clarify where typical responsibilities reside for the manufacturer or the laboratory;