



BSI Standards Publication

Cardiovascular implants — Endovascular devices

Part 3: Vena cava filters

This is a preview of BS EN ISO 25539-3:2024. [Click here to purchase the full version from the ANSI store.](#)

National foreword

This British Standard is the UK implementation of EN ISO 25539-3:2024. It is identical to ISO 25539-3:2024. It supersedes BS EN ISO 25539-3:2011, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/150/2, Cardiovascular implants.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2024
Published by BSI Standards Limited 2024

ISBN 978 0 539 17110 5

ICS 11.040.40

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 October 2024.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

This is a preview of BS EN ISO 25539-3:2024. [Click here to purchase the full version from the ANSI store.](#)

EUROPÄISCHE NORM

October 2024

ICS 11.040.40

Supersedes EN ISO 25539-3:2011

English Version

Cardiovascular implants - Endovascular devices - Part 3: Vena cava filters (ISO 25539-3:2024)

Implants cardiovasculaires - Dispositifs
endovasculaires - Partie 3 : Filtres de veine cave (ISO
25539-3:2024)

Kardiovaskuläre Implantate - Endovaskuläre
Implantate - Teil 3: Hohlvenenfilter (ISO 25539-
3:2024)

This European Standard was approved by CEN on 23 August 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

This is a preview of BS EN ISO 25539-3:2024. [Click here to purchase the full version from the ANSI store.](#)

European foreword

This document (EN ISO 25539-3:2024) has been prepared by Technical Committee ISO/TC 150 "Implants for surgery" in collaboration with Technical Committee CEN/TC 285 "Non-active surgical implants" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2025, and conflicting national standards shall be withdrawn at the latest by April 2025.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 25539-3:2011.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 25539-3:2024 has been approved by CEN as EN ISO 25539-3:2024 without any modification.

This is a preview of BS EN ISO 25539-3:2024. [Click here to purchase the full version from the ANSI store.](#)

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 General requirements	5
4.1 Classification.....	5
4.2 Materials of construction for filter system.....	5
4.3 Configuration and size designation for filters.....	6
4.4 Intended clinical use designation.....	6
5 Intended performance	6
6 Design attributes	6
6.1 General.....	6
6.2 Filter system.....	6
6.3 Vena cava filter.....	6
6.4 Optional filter.....	7
6.5 Convertible filter.....	7
6.6 Retrieval system.....	7
6.7 Conversion system.....	7
6.8 Filter system, retrieval system and conversion system.....	7
6.9 Coating on delivery system or filter.....	8
6.10 Absorbable filter or coating.....	8
7 Materials	8
8 Design evaluation	8
8.1 General.....	8
8.2 Sampling.....	9
8.3 Conditioning of test samples.....	10
8.4 Reporting.....	10
8.5 Bench and analytical tests.....	11
8.5.1 General.....	11
8.5.2 Filter system.....	11
8.5.3 Vena cava filter.....	13
8.5.4 Optional filter.....	15
8.5.5 Convertible filter.....	16
8.5.6 Retrieval system.....	16
8.5.7 Conversion system.....	17
8.5.8 Filter system, retrieval system and conversion system.....	18
8.5.9 Coating on the delivery system, the retrieval system, the conversion system or the filter.....	19
8.5.10 Absorbable filter or coating.....	19
8.6 Preclinical in vivo evaluation.....	20
8.6.1 Purpose.....	20
8.6.2 Specific aims.....	20
8.6.3 Protocol considerations.....	21
8.6.4 Data acquisition.....	21
8.6.5 Test report and additional information.....	23
8.7 Clinical evaluation.....	24
8.7.1 Purpose.....	24
8.7.2 Specific aims.....	25
8.7.3 Protocol considerations.....	25
8.7.4 Data acquisition.....	26
8.7.5 Final report.....	29

This is a preview of BS EN ISO 25539-3:2024. [Click here to purchase the full version from the ANSI store.](#)

10	Manufacturing	30
11	Sterilization	30
	11.1 Products supplied sterile.....	30
	11.2 Sterilization residuals.....	30
12	Packaging	31
	12.1 General.....	31
	12.1.1 General.....	31
	12.1.2 Unit container.....	31
	12.1.3 Outer container.....	31
	12.1.4 Shipping container.....	31
	12.1.5 Maintenance of sterility in transit.....	31
	12.2 Labelling.....	31
	12.2.1 Container label.....	31
	12.2.2 Filter systems.....	31
	12.2.3 Retrieval systems.....	32
	12.2.4 Conversion systems.....	32
	12.2.5 Record label.....	32
	12.3 Information supplied by the manufacturer.....	33
	12.3.1 General.....	33
	12.3.2 Information and instructions for use.....	33
	Annex A (informative) Relationship between testing requirements, device attributes and potential failure modes, and guidance for the creation of a device evaluation strategy	35
	Annex B (informative) Description of clinical effects of failure	53
	Annex C (informative) Description of device effects of failure	55
	Annex D (informative) Test methods	56
	Annex E (informative) Examples of general issues related to the clinical use of vena cava filters	92
	Bibliography	94

This is a preview of BS EN ISO 25539-3:2024. [Click here to purchase the full version from the ANSI store.](#)

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC2, *Cardiovascular implants and extracorporeal systems*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 285, *Non-active surgical implants*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

The second edition cancels and replaces the first edition (ISO 25539-3:2011), which has been technically revised.

The main changes are as follows:

- the testing and clinical use related to vena cava filters has been updated;
- the consistency in nomenclature and reporting requirements has been improved.

A list of all parts in the ISO 25539 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This is a preview of BS EN ISO 25539-3:2024. [Click here to purchase the full version from the ANSI store.](#)

This document was prepared to provide guidance on the minimum requirements for vena cava filter systems. The rationale for the requirements for bench tests and analyses to assess device performance and safety, guidance on the identification of appropriate testing to evaluate a specific device design, and guidance for developing test methods are provided in informative annexes. Further clarification of terminology is provided in [Annexes B, C and E](#).

This document has been updated to reflect current knowledge regarding the testing and clinical use related to vena cava filters, reflected in modifications to the requirements in the main body and in the guidance for developing test methods in [Annex D](#). In addition, revisions have been made to improve consistency in nomenclature and reporting and to enhance the utility of this document.

This revised document introduces methodology to identify appropriate testing and analyses applicable to intended clinical use, design and potential failure modes for a specific vena cava filter system, designated as the device evaluation strategy. The requirement regarding the device evaluation strategy is in the main body. [Annex A](#) provides guidance for developing a focused device evaluation strategy table that is specific to the unique characteristics of a device, device design modifications or changes in intended use. [Annex A](#) also provides guidance for the development of a comprehensive device evaluation strategy table that may be used when it is not sufficient to focus only on the unique characteristics or changes.

The guidance on the development of methods to address the requirement for evaluating fatigue and durability through computational analyses has been modified to include recommendations regarding verification of the solution and validation of the computational model, as well as reporting. The guidance on the model development for simulated use has also been significantly revised to improve the clinical relevance of this testing.

In addition to modifications to specific design evaluation requirements, guidance has been provided regarding the assessment of the acceptability of test results. When the requirement is to quantitatively appraise or analyse a parameter, test results generally may be compared to a quantitative value (i.e. acceptance criteria). For characterization tests, it is appropriate to provide an explanation of the relevance of the results. Additionally, some testing can include a comparison to test data or existing data from a previously evaluated device.

For design evaluation, requirements regarding sampling, conditioning of test samples and reporting have been incorporated in the main body. Guidance on these elements of testing and documentation were previously only included in [Annex D](#).

The revisions to the annexes to this document are given in [Table 1](#).

Table 1 — Revisions to the annexes in this document

Annex of ISO 25539-3:2011	Changes in ISO 25539-3:2024
Annex A – Attributes of endovascular devices – Vena cava filters – Technical and clinical considerations	Annex A now includes the relationship between testing requirements, device attributes, and potential failure modes and guidance for the creation of a device evaluation strategy.
Annex B – Descriptions of potential device effects of failure and failure modes and descriptions of detrimental clinical effects	Annex B now includes a description of potential clinical effects of failure.
Annex C – Bench and analytical tests	The list of tests is included in Table D.1 . Annex C now includes a description of potential device effects of failure.
Annex D – Test methods	Annex D – Test methods

Many filter systems have been shown to be safe and effective in clinical use – this update is not intended to require additional evaluations of these devices to remain in compliance with this document as the testing would not provide useful information regarding the expected clinical performance of the device. Manufacturers may rely on historical data gathered under the guidance of the previous edition of this document (i.e. ISO 25539-3:2011). Similarly, for device modifications or changes in intended clinical use, this

This is a preview of BS EN ISO 25539-3:2024. [Click here to purchase the full version from the ANSI store.](#)

NOTE The relationship between testing requirements, device attributes and potential failure modes is provided in [Clause A.1](#). [Clause A.1](#) also provides general information regarding device evaluation strategies. [Tables A.3](#) to [A.9](#) provide the rationale for the requirements specified in this document for bench tests and analyses to assess device performance and safety. An explanation of the table headings for [Tables A.3](#) to [A.9](#) is given in [Table A.1](#).

Guidance for the creation of a device-specific evaluation strategy is provided in [Clause A.2](#). Two approaches to create a device-specific evaluation strategy are provided:

- a) focused device evaluation strategy in [A.2.1](#);
- b) comprehensive device evaluation strategy in [A.2.2](#).

[Annex B](#) provides a description of the potential clinical effects of failure identified in [Annex A](#).

[Annex C](#) provides a description of the potential device effects of failure identified in [Annex A](#).

Additional descriptions of clinical and device effects of failure are included in [Annexes B](#) and [C](#), respectively.

[Annex D](#) provides information to consider in developing appropriate bench test and analytical methods.

[Annex E](#) provides examples of terms for clinical use related to vena cava filters.

This is a preview of BS EN ISO 25539-3:2024. [Click here to purchase the full version from the ANSI store.](#)

This is a preview of BS EN ISO 25539-3:2024. [Click here to purchase the full version from the ANSI store.](#)

Cardiovascular implants — Endovascular devices —

Part 3: Vena cava filters

1 Scope

This document specifies the requirements for the evaluation of vena cava filter systems (filters and delivery systems) and the requirements with respect to nomenclature, design attributes and information supplied by the manufacturer. Guidance for the development of in vitro test methods is included in [Annex D](#). This document is intended to be used in conjunction with ISO 14630, which specifies general requirements for the performance of non-active surgical implants.

NOTE 1 Due to the variations in the design of implants covered by this document, and in some cases due to the emergence of novel types of such implants, acceptable standardized in vitro tests and clinical results are not always available. As further scientific and clinical data become available, a revision of this document will be necessary.

This document is applicable to vena cava filters intended to prevent symptomatic pulmonary embolism by capturing blood clots in the inferior vena cava (IVC). While this document can be useful with respect to filters implanted in other venous locations (e.g. superior vena cava, iliac veins), it does not specifically address the use of filters in other implantation sites.

This document is also applicable to permanent filters together with their associated delivery systems, optional filters that can be retrieved and their associated retrieval systems, and convertible filters and their associated conversion systems. While this document can be useful with respect to the evaluation of repositioning filters after chronic implantation, it does not specifically address filter repositioning.

This document is not applicable to

- temporary filters (e.g. tethered) that need to be removed after a defined period of time,
- issues associated with viable tissues and non-viable biological materials, and
- procedures and devices (e.g. venous entry needle) used prior to the vena cava filter procedure.

Although absorbable filters and filters with absorbable coatings are within the scope of this document, this document is not comprehensive with respect to the absorbable properties of these devices.

NOTE 2 Absorbable implants are covered in ISO/TS 17137.

Although coated filters and coated filter systems are within the scope of this document, this document is not comprehensive with respect to coatings.

NOTE 3 Vascular device-drug combination products are covered in ISO 12417-1 and some coating properties are covered in ISO 25539-4.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10993 (all parts), *Biological evaluation of medical devices*