



BSI Standards Publication

Intravascular catheters — Sterile and single-use catheters

Part 1: General requirements

This is a preview of BS EN ISO 10555-1:2023. [Click here to purchase the full version from the ANSI store.](#)

National foreword

This British Standard is the UK implementation of EN ISO 10555-1:2023. It is identical to ISO 10555-1:2023. It supersedes BS EN ISO 10555-1:2013+A1:2017, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/84, Catheters and syringes.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2023
Published by BSI Standards Limited 2023

ISBN 978 0 539 20327 1

ICS 11.040.25

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 December 2023.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

This is a preview of BS EN ISO 10555-1:2023. [Click here to purchase the full version from the ANSI store.](#)

EUROPÄISCHE NORM

November 2023

ICS 11.040.25

Supersedes EN ISO 10555-1:2013

English Version

Intravascular catheters - Sterile and single-use catheters - Part 1: General requirements (ISO 10555-1:2023)

Cathéters intravasculaires - Cathéters stériles et non
réutilisables - Partie 1: Exigences générales (ISO
10555-1:2023)

Intravaskuläre Katheter - Sterile Katheter zur
einmaligen Verwendung - Teil 1: Allgemeine
Anforderungen (ISO 10555-1:2023)

This European Standard was approved by CEN on 24 November 2023.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

This is a preview of BS EN ISO 10555-1:2023. [Click here to purchase the full version from the ANSI store.](#)

European foreword

This document (EN ISO 10555-1:2023) has been prepared by Technical Committee ISO/TC 84 "Devices for administration of medicinal products and catheters" in collaboration with Technical Committee CEN/TC 205 "Non-active medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2024, and conflicting national standards shall be withdrawn at the latest by May 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 10555-1:2013.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 10555-1:2023 has been approved by CEN as EN ISO 10555-1:2023 without any modification.

This is a preview of BS EN ISO 10555-1:2023. [Click here to purchase the full version from the ANSI store.](#)

Contents

	Page
Foreword	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements	5
4.1 Risk approach.....	5
4.2 Usability engineering.....	5
4.3 Sterilization.....	5
4.4 Shelf life.....	5
4.5 Detectability.....	5
4.6 Biocompatibility.....	5
4.7 Surface.....	5
4.8 Corrosion resistance.....	6
4.9 Peak tensile force.....	6
4.10 Freedom from leakage during pressurization.....	7
4.11 Freedom from leakage during aspiration.....	7
4.12 Hubs.....	7
4.13 Flowrate.....	8
4.14 Power injection burst pressure.....	8
4.15 Packaging system.....	8
4.16 Simulated use, kink and/or torque testing to consider depending on device design, intended use, and risk analysis.....	8
4.17 Coating integrity and/or particulate testing to consider depending on device design, intended use, and risk analysis.....	9
4.18 Distal tip stiffness testing to consider for neurovascular applications.....	9
5 Designation of nominal size	9
5.1 Nominal outside diameter.....	9
5.2 Nominal inside diameter.....	9
5.3 Nominal effective length.....	9
6 Information to be supplied with the catheter	10
6.1 General.....	10
6.2 Marking on the device and/or primary packaging.....	10
6.3 Instructions for use.....	11
6.4 Marking on the secondary packaging.....	11
Annex A (normative) Test method for corrosion resistance	12
Annex B (normative) Method for determining peak tensile force	13
Annex C (normative) Test method for liquid leakage under pressure	16
Annex D (normative) Test method for air leakage into hub assembly during aspiration	18
Annex E (normative) Determination of flowrate through catheter	20
Annex F (normative) Test for burst pressure under static conditions	22
Annex G (normative) Power injection tests for flowrate and device pressure (only for products indicated for power injection)	25
Annex H (informative) Units of measurement systems other than those specified in this document	30
Annex I (normative) Test method for air leakage under water	32
Annex J (informative) Rationale and guidance	34
Annex K (informative) Test methods for distal tip stiffness for neurovascular applications	41

This is a preview of BS EN ISO 10555-1:2023. [Click here to purchase the full version from the ANSI store.](#)

Bibliography	43
---------------------------	-----------

This is a preview of BS EN ISO 10555-1:2023. [Click here to purchase the full version from the ANSI store.](#)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 84, *Devices for administration of medicinal products and catheters*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 205, *Non-active medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 10555-1:2013), which has been technically revised. It also incorporates the amendment ISO 10555-1:2013/Amd 1:2017.

The main changes are as follows:

- added definitions for “inside diameter”, “gauge length”, and “coating” in [Clause 3](#);
- added clarification on requirements ([Clause 4](#)) related to:
 - peak tensile force (revised the NOTE in [Table 1](#));
 - leakage during pressurization: option for air pressure test ([Annex I](#));
 - power injection burst pressure.
- added new requirements ([Clause 4](#)) related to:
 - risk approach;
 - usability engineering;
 - shelf life;
 - packaging system;
 - simulated use, kink and torque;

This is a preview of BS EN ISO 10555-1:2023. [Click here to purchase the full version from the ANSI store.](#)

- coating integrity, particulate;
- distal tip stiffness.
- removed the requirements on side holes and distal tip;
- added text on “Nominal inside diameter for some applications” ([Clause 5](#));
- added test details in the instructions for use for power injection ([Clause 6](#));
- added reporting of maximum, minimum, standard deviation for variable data analysis in test reports;
- clarified “conditioning time” and “gauge length” ([Annex B](#));
- clarified “minimum outside pressure requirement” ([Annex D](#));
- introduced alternative test method using constant flowrate source ([Annex G](#));
- replaced Figure H.1 in previous version with the new [Table H.1](#);
- added new [Annex I](#) for alternative leakage under pressurization using air pressure;
- added new [Annex J](#) for rationale.

A list of all parts in the ISO 10555 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user’s national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This is a preview of BS EN ISO 10555-1:2023. Click here to purchase the full version from the ANSI store.

Intravascular catheters — Sterile and single-use catheters —

Part 1: General requirements

1 Scope

This document specifies general requirements for intravascular catheters, supplied sterile and intended for single use, for any application.

This document does not apply to intravascular catheter accessories, e.g. those covered by ISO 11070.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 7886-1, *Sterile hypodermic syringes for single use — Part 1: Syringes for manual use*

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 11607-1, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 15223-1, *Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements*

ISO 80369-7, *Small-bore connectors for liquids and gases in healthcare applications — Part 7: Connectors for intravascular or hypodermic applications*

IEC 62366-1, *Medical devices — Part 1: Application of usability engineering to medical devices*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

intravascular catheter

tubular device, single or multilumen, designed to be partially or totally inserted or implanted into the vascular system for diagnostic and/or therapeutic purposes