



BSI Standards Publication

Pressure regulators for use with medical gases

Part 1: Pressure regulators and pressure regulators with flow-metering devices

This is a preview of BS EN ISO 10524-1:2019+A1:2023. [Click here to purchase the full version from the ANSI store](#)

National foreword

This British Standard is the UK implementation of EN ISO 10524-1:2019+A1:2023. It is identical to ISO 10524-1:2018, incorporating amendment 1:2023. It supersedes BS EN ISO 10524-1:2019, which is withdrawn.

The start and finish of text introduced or altered by amendment is indicated in the text by tags. Tags indicating changes to ISO text carry the number of the ISO amendment. For example, text altered by ISO amendment 1 is indicated by A1 A1.

The UK participation in its preparation was entrusted to Technical Committee CH/121/6, Medical gas supply systems.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2023
Published by BSI Standards Limited 2023

ISBN 978 0 539 25069 5

ICS 11.040.10

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 28 February 2019.

Amendments/corrigenda issued since publication

Date	Text affected
30 November 2023	Implementation of ISO amendment 1:2023 with CEN endorsement A1:2023

This is a preview of BS EN ISO 10524-1:2019+A1:2023. [Click here to purchase the full version from the ANSI store](#)

EUROPÄISCHE NORM

November 2023

ICS 11.040.10

Supersedes EN ISO 10524-1:2006

English Version

Pressure regulators for use with medical gases - Part 1: Pressure regulators and pressure regulators with flow- metering devices (ISO 10524-1:2018)

Détendeurs pour l'utilisation avec les gaz
médicaux- Partie 1: Détendeurs et détendeurs-
débitmètres (ISO 10524-1:2018)

Druckminderer zur Verwendung mit medizinischen
Gasen - Teil 1: Druckminderer und Druckminderer
mit Durchflussmessgeräten (ISO 10524-1:2018)

This European Standard was approved by CEN on 13 December 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

This is a preview of BS EN ISO 10524-1:2019+A1:2023. [Click here to purchase the full version from the ANSI store](#)

This document (EN ISO 10524-1:2019) has been prepared by Technical Committee ISO/TC 121 "Anaesthetic and respiratory equipment" in collaboration with Technical Committee CEN/TC 215 "Respiratory and anaesthetic equipment" the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2019, and conflicting national standards shall be withdrawn at the latest by July 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 10524-1:2006.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 10524-1:2018 has been approved by CEN as EN ISO 10524-1:2019 without any modification.

European foreword to amendment 1

This document (EN ISO 10524-1:2019/A1:2023) has been prepared by Technical Committee ISO/TC 121 "Anaesthetic and respiratory equipment" in collaboration with Technical Committee CEN/TC 215 "Respiratory and anaesthetic equipment" the secretariat of which is held by BSI.

This Amendment to the European Standard EN ISO 10524-1:2019 shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2024, and conflicting national standards shall be withdrawn at the latest by May 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 10524-1:2018/Amd 1:2023 has been approved by CEN as EN ISO 10524-1:2019/A1:2023 without any modification.

This is a preview of BS EN ISO 10524-1:2019+A1:2023. [Click here to purchase the full version from the ANSI store](#)

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Nomenclature	4
5 General requirements	4
5.1 Safety.....	4
5.2 Usability.....	4
5.3 Alternative construction.....	4
5.4 Materials.....	4
6 Design requirements	5
6.1 General.....	5
6.2 Indicator for cylinder pressure or cylinder content.....	6
6.2.1 General.....	6
6.2.2 PRESSURE GAUGES, pressure indicators, and FLOWGAUGES.....	6
6.3 Integrated electronic device.....	6
6.4 Connections.....	6
6.4.1 Inlet connector.....	6
6.4.2 Outlet connectors.....	7
6.5 * Requirements for outlet pressure.....	7
6.5.1 PRESSURE OUTLET.....	7
6.5.2 FLOW OUTLET.....	8
6.6 Flow-metering device.....	8
6.7 Flow control and indication.....	8
6.8 Pressure-adjusting device.....	8
6.9 * Filtration.....	9
6.10 * PRESSURE-RELIEF DEVICE.....	9
6.11 Leakage.....	9
6.12 Mechanical strength.....	10
6.12.1 Resistance of the high-pressure side.....	10
6.12.2 Resistance of the low-pressure side to pneumatic pressure.....	10
6.12.3 Resistance of the low-pressure side to inlet pressure, P_1	10
6.13 * Resistance to ignition.....	10
6.14 Requirements for PRESSURE REGULATORS with FLOWMETERS.....	10
6.14.1 Calibration.....	10
6.14.2 ACCURACY OF FLOW.....	10
6.14.3 Stability of flow.....	11
6.14.4 Legibility.....	11
6.14.5 Flow control device.....	11
6.15 Requirements for PRESSURE REGULATORS fitted with FLOWGAUGES.....	11
6.15.1 Calibration.....	11
6.15.2 ACCURACY OF FLOW.....	11
6.15.3 Stability of flow.....	11
6.16 Requirements for PRESSURE REGULATORS fitted with fixed ORIFICES.....	12
6.16.1 Stability and ACCURACY OF FLOW.....	12
6.16.2 * Flow setting torque.....	12
6.16.3 Removal of a fixed ORIFICE.....	12
6.16.4 Legibility.....	12
6.17 Endurance.....	12
6.17.1 FLOW SELECTOR.....	12
6.17.2 PRESSURE REGULATOR.....	13

This is a preview of BS EN ISO 10524-1:2019+A1:2023. [Click here to purchase the full version from the ANSI store](#)

7	Construction requirements	13
7.1	* Cleanliness.....	13
7.2	Lubricants.....	13
7.3	Loosening torques.....	13
8	Test methods for type tests	14
8.1	General conditions.....	14
8.1.1	General.....	14
8.1.2	Ambient conditions.....	14
8.1.3	Test gas.....	14
8.1.4	Reference conditions.....	14
8.2	Test schedule.....	14
8.3	Test methods for outlet pressure.....	16
8.3.1	Test equipment.....	16
8.3.2	Test methods for determining outlet pressure limits for a PRESSURE REGULATOR fitted with a PRESSURE OUTLET.....	17
8.3.3	Test method for determining the outlet pressure limit for a PRESSURE REGULATOR fitted with a FLOW OUTLET.....	18
8.4	Test method for a PRESSURE-RELIEF DEVICE.....	18
8.5	Test methods for leakage.....	18
8.5.1	External leakage.....	18
8.5.2	Internal leakage.....	19
8.6	Test method for mechanical strength.....	19
8.6.1	Test method for the high-pressure side.....	19
8.6.2	Test method for resistance of the low-pressure side to pneumatic pressure.....	19
8.6.3	Test method for the resistance of the low pressure side to P_1	19
8.7	Test method for resistance to ignition.....	20
8.7.1	General.....	20
8.7.2	Test procedure for ADJUSTABLE and PRE-SET PRESSURE REGULATORS.....	20
8.8	Test method for ACCURACY OF FLOW of PRESSURE REGULATORS fitted with FLOWMETERS or FLOWGAUGES.....	20
8.9	Test method for the stability of flow of PRESSURE REGULATOR fitted with FLOWMETERS or FLOWGAUGES.....	20
8.10	Test method for stability and ACCURACY OF FLOW of PRESSURE REGULATORS fitted with fixed ORIFICES.....	21
8.11	Test method for flow setting and loosening torques.....	21
8.11.1	General.....	21
8.11.2	Test method for verifying no stable position between two settings.....	21
8.12	Test method for durability of markings and colour coding.....	21
8.13	*FLOW SELECTOR endurance test.....	21
8.14	PRESSURE REGULATOR endurance test.....	22
9	Marking, colour coding, and packaging	22
9.1	Marking.....	22
9.2	Colour coding.....	24
9.3	Packaging.....	24
10	Information to be supplied by the manufacturer	24
Annex A (informative) Rationale		26
Annex B (informative) Reported regional and national deviations of colour coding and nomenclature for medical gases		31
Bibliography		33

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 6, *Medical gas systems*.

This second edition cancels and replaces the first edition (ISO 10524-1:2006), which has been technically revised.

The main changes compared to the previous edition are as follows:

- the common requirements have been aligned with ISO 10524-2 and ISO 10524-3;
- this document has been restructured according to the new ISO template and associated renumbering;
- a complete schedule has been introduced;
- all type tests have been reviewed.

A list of all parts in the ISO 10524 series can be found on the ISO website.

Introduction

PRESSURE REGULATORS are used to reduce high cylinder pressure to a lower pressure suitable for use with medical equipment or for delivery of gas directly to a patient.

These functions cover a wide range of inlet and outlet pressures and flows which require specific design characteristics. It is important that the operating characteristics of PRESSURE REGULATORS are specified and tested in a defined manner.

A PRESSURE REGULATOR normally has coupled to it a device which controls the flow, such as a flow control device or a fixed ORIFICE. The flow can be indicated by a FLOWMETER or by a FLOWGAUGE.

It is essential that regular inspection and maintenance be undertaken to ensure that the PRESSURE REGULATOR continues to meet the requirements of this document.

This document pays particular attention to

- use of suitable materials,
- safety (mechanical strength, leakage, safe relief of excess pressure and resistance to ignition),
- GAS SPECIFICity,
- cleanliness,
- type testing,
- marking, and
- information supplied by the manufacturer.

[Annex A](#) contains rationale statements for some of the requirements of this document. The clauses and subclauses marked with an asterisk (*) after their number have corresponding rationale included to provide additional insight into the reasoning that led to the requirements and recommendations that have been incorporated into this document. It is considered that knowledge of the reasons for the requirements will not only facilitate the proper application of this document, but will expedite any subsequent revisions.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in [Annex A](#).

In this document, the following print types are used:

- requirements and definitions: roman type;
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- *test specifications: italic type;*
- TERMS DEFINED IN CLAUSE 3 OF THIS DOCUMENT OR AS NOTED: SMALL CAPITALS TYPE.

Pressure regulators for use with medical gases —

Part 1:

Pressure regulators and pressure regulators with flow-metering devices

1 Scope

This document specifies the design, construction, type testing, and marking requirements for PRESSURE REGULATORS (as defined in [3.18](#)) intended for the administration of medical gases and their mixtures in the treatment, management, diagnostic evaluation and care of patients or for gases used for driving surgical tools.

Examples of gases include oxygen, medical air and oxygen/nitrous oxide mixtures.

This document applies to PRESSURE REGULATORS:

- a) intended to be connected to cylinders by the operator;
- b) with integral flow-metering devices intended to be connected to cylinders by the operator;
- c) that are an integral part of medical equipment (e.g. anaesthetic workstations, lung ventilators, resuscitators).

A PRESSURE REGULATOR can be provided with PRESSURE OUTLET or FLOW OUTLET, and can be adjustable or pre-set.

PRESSURE REGULATORS are intended to be fitted to refillable cylinders with a WORKING PRESSURE up to 30 000 kPa (300 bar) and can be provided with devices which control and measure the flow of the medical gas delivered.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 32, *Gas cylinders for medical use — Marking for identification of content*

ISO 407, *Small medical gas cylinders — Pin-index yoke-type valve connections*

ISO 5145, *Cylinder valve outlets for gases and gas mixtures — Selection and dimensioning*

ISO 7000, *Graphical symbols for use on equipment — Registered symbols*

ISO 9170-1, *Terminal units for medical gas pipeline systems — Part 1: Terminal units for use with compressed medical gases and vacuum*

ISO 10297:2014, *Gas cylinders — Cylinder valves — Specification and type testing*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 15001:2010, *Anaesthetic and respiratory equipment — Compatibility with oxygen*