



BSI Standards Publication

Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices

Part 2: Non-critical medical devices

This is a preview of BS EN ISO 17664-2:2023. [Click here to purchase the full version from the ANSI store.](#)

National foreword

This British Standard is the UK implementation of EN ISO 17664-2:2023. It is identical to ISO 17664-2:2021. It supersedes BS ISO 17664-2:2021, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/198, Sterilization and Associated Equipment and Processes.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

This publication has been prepared under a mandate given to the European Standards Organizations by the European Commission and the European Free Trade Association. It is intended to support requirements of the EU legislation detailed in the European Foreword. A European Annex, usually Annex ZA or ZZ, describes how this publication relates to that EU legislation.

For the Great Britain market (England, Scotland and Wales), if UK Government has designated this publication for conformity with UKCA marking (or similar) legislation, it may contain an additional National Annex. Where such a National Annex exists, it shows the correlation between this publication and the relevant UK legislation. If there is no National Annex of this kind, the relevant Annex ZA or ZZ in the body of the European text will indicate the relationship to UK regulation applicable in Great Britain. References to EU legislation may need to be read in accordance with the UK designation and the applicable UK law. Further information on designated standards can be found at www.bsigroup.com/standardsandregulation.

For the Northern Ireland market, UK law will continue to implement relevant EU law subject to periodic confirmation. Therefore Annex ZA/ZZ in the European text, and references to EU legislation, are still valid for this market.

UK Government is responsible for legislation. For information on legislation and policies relating to that legislation, consult the relevant pages of www.gov.uk.

This is a preview of BS EN ISO 17664-2:2023. [Click here to purchase the full version from the ANSI store.](#)

ISBN 978 0 539 26399 2

ICS 11.080.01

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 28 February 2021.

Amendments/corrigenda issued since publication

Date	Text affected
31 January 2024	This corrigendum renumbers BS ISO 17664-2:2021 as BS EN ISO 17664-2:2023

This is a preview of BS EN ISO 17664-2:2023. [Click here to purchase the full version from the ANSI store.](#)

This is a preview of BS EN ISO 17664-2:2023. [Click here to purchase the full version from the ANSI store.](#)

EUROPÄISCHE NORM

December 2023

ICS 11.080.01

English Version

Processing of health care products - Information to be
provided by the medical device manufacturer for the
processing of medical devices - Part 2: Non-critical medical
devices (ISO 17664-2:2021)

Traitement de produits de soins de santé -
Informations relatives au traitement des dispositifs
médicaux à fournir par le fabricant du dispositif -
Partie 2: Dispositifs médicaux non critiques (ISO
17664-2:2021)

Aufbereitung von Produkten für die
Gesundheitsfürsorge - Vom Medizinprodukt-Hersteller
bereitzustellende Informationen für die Aufbereitung
von Medizinprodukten - Teil 2: Nicht kritische
Medizinprodukte (ISO 17664-2:2021)

This European Standard was approved by CEN on 17 December 2023.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

The text of ISO 17664-2:2021 has been prepared by Technical Committee ISO/TC 198 "Sterilization of health care products" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 17664-2:2023 by Technical Committee CEN/TC 204 "Sterilization of medical devices" the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2024, and conflicting national standards shall be withdrawn at the latest by June 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a Standardization Request given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s) / Regulation(s).

For the relationship with EU Directive(s) / Regulation(s), see informative Annex ZA, which is an integral part of this document.

Any feedback and questions on this document should be directed to the users' national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 17664-2:2021 has been approved by CEN as EN ISO 17664-2:2023 without any modification.

This is a preview of BS EN ISO 17664-2:2023. [Click here to purchase the full version from the ANSI store.](#)

Annex ZA (informative)

Relationship between this European Standard and the General Safety and Performance requirements of Regulation (EU) 2017/745 aimed to be covered

This European standard has been prepared under M/575 to provide one voluntary means of conforming to the General Safety and Performance Requirements of Regulation (EU) 2017/745 of 5 April 2017 concerning medical devices [O] L 117] and to system or process requirements including those relating to quality management systems, risk management, post-market surveillance systems, clinical investigations, clinical evaluation or post-market clinical follow-up.

Once this standard is cited in the Official Journal of the European Union under that Regulation, compliance with the normative clauses of this standard given in Table ZA.1 and application of the edition of the normatively referenced standards as given in Table ZA.2 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding General Safety and Performance Requirements of that Regulation, and associated EFTA Regulations.

Where a definition in this standard differs from a definition of the same term set out in Regulation (EU) 2017/745, the differences shall be indicated in this Annex Z. For the purpose of using this standard in support of the requirements set out in Regulation (EU) 2017/745, the definitions set out in this Regulation prevail. In this context, the definition of 'medical device' in EN ISO 17664-2 is a modified version of the definition prepared by the Global Harmonization Task Force with modification to the Note in the definition.

Where the European standard is an adoption of an International Standard, the scope of this standard can differ from the scope of the European Regulation that it supports. As the scope of the applicable regulatory requirements differ from nation to nation and region to region, the standard can only support European regulatory requirements to the extent of the scope of the European regulation for medical devices (EU) 2017/745).

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Regulation (EU) 2017/745. This means that risks have to be 'reduced as far as possible', 'reduced to the lowest possible level', 'reduced as far as possible and appropriate', 'removed or reduced as far as possible', 'eliminated or reduced as far as possible', 'removed or minimized as far as possible', or 'minimized', according to the wording of the corresponding General Safety and Performance Requirement.

NOTE 2 The manufacturer's policy for determining **acceptable risk** must be in compliance with General Safety and Performance Requirements 1, 2, 3, 4, 5, 8, 9, 10, 11, 14, 16, 17, 18, 19, 20, 21 and 22 of the Regulation.

NOTE 3 When a General Safety and Performance Requirement does not appear in Table ZA.1, it means that it is not addressed by this European Standard.

This is a preview of BS EN ISO 17664-2:2023. [Click here](#) to purchase the full version from the ANSI store.

Table ZA.1 — Correspondence between this European Standard and Annex I of Regulation (EU) 2017/745 [OJ L 117] and to system or process requirements including those relating to quality management systems, risk management, post-market surveillance systems, clinical investigations, clinical evaluation or post-market clinical follow-up

General Safety and Performance Requirements of Regulation (EU) 2017/745	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
23.4.i	4,5,6,7	23.4.i is covered only for the disinfection of devices to make ready for first use.
23.4.n	4,5,6,7	

Table ZA.2 — Applicable Standards to confer presumption of conformity as described in this Annex ZA

Column 1 Reference in Clause 2	Column 2 International Standard Edition	Column 3 Title	Column 4 Corresponding European Standard Edition
ISO 14971	ISO 14971:2019	Medical devices — Application of risk management to medical devices	EN ISO 14971:2019 EN ISO 14971:2019+A11:2021

The documents listed in the Column 1 of table ZA.2, in whole or in part, are normatively referenced in this document and are indispensable for its application. The achievement of the presumption of conformity is subject to the application of the edition of Standards as listed in Column 4 or, if no European Standard Edition exists, the International Standard Edition given in Column 2 of table ZA.2.

WARNING 1 — Presumption of conformity stays valid only as long as a reference to this European Standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

WARNING 2 — Other Union legislation may be applicable to the product(s) falling within the scope of this standard.

This is a preview of BS EN ISO 17664-2:2023. [Click here to purchase the full version from the ANSI store.](#)

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Risk analysis	4
5 Validation of the processes identified in the information provided by the medical device manufacturer	5
6 Information to be provided by the medical device manufacturer	5
6.1 General.....	5
6.2 Processing instructions.....	6
6.3 Limitations and restrictions on processing.....	7
6.4 Preparation before processing.....	7
6.5 Cleaning.....	7
6.5.1 General.....	7
6.5.2 Manual cleaning.....	7
6.5.3 Automated cleaning.....	8
6.6 Disinfection.....	8
6.6.1 General.....	8
6.6.2 Manual disinfection.....	8
6.6.3 Automated disinfection.....	9
6.7 Drying.....	9
6.8 Inspection and maintenance.....	9
6.9 Packaging.....	10
6.10 Storage.....	10
6.11 Transportation.....	10
7 Presentation of the information	10
Annex A (informative) Commonly utilized processing methods	11
Annex B (informative) Example processing instructions for non-critical reusable medical devices	15
Annex C (informative) Processing classification and grouping of medical devices	17
Annex D (informative) Additional guidance on information to be provided by the medical device manufacturer	19
Annex E (informative) Examples of medical devices and their relationship to this document	20
Bibliography	24

This is a preview of BS EN ISO 17664-2:2023. [Click here to purchase the full version from the ANSI store.](#)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

A list of all parts in the ISO 17664 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This is a preview of BS EN ISO 17664-2:2023. [Click here to purchase the full version from the ANSI store.](#)

Introduction

This document applies to manufacturers of non-critical medical devices that are intended to be cleaned and/or disinfected by the processor to be made ready for use or reuse. This includes:

- medical devices that are intended for reuse and require processing to take them from their state after clinical use to the state of being ready for their next use;
- single-use medical devices that require processing before use and are intended to be used in a clean and/or disinfected state.

Significant advances in technology and knowledge have resulted in the development of complex medical devices to support the delivery of healthcare to patients. These advances have led to medical devices being designed that are potentially more difficult to clean and/or disinfect.

Cleaning and disinfecting technologies have also undergone significant change in the past decade, resulting in new systems and approaches that can be applied in the processing of medical devices. This has led to a greater appreciation of the need for validation of processing including cleaning and/or disinfection in order to ensure that medical devices are effectively processed. These developments have led to the need to ensure that manufacturers of medical devices provide adequate instructions that support the end users to undertake safe and effective processing of medical devices, utilizing the available equipment and processes.

A medical device requiring processing is supplied with detailed processing instructions in order to ensure that, when followed correctly, the risks of transmission of infectious agents are minimized. In addition, effective processing minimizes the risk of other adverse effects on medical devices.

Cleaning is an important step in rendering a used medical device safe for subsequent use. Failure to remove contaminants (e.g. blood, tissues, microorganisms, cleaning agents and lubricants) from surfaces of medical devices could compromise the correct functioning of the medical device, its safe use and (if required) any subsequent disinfection process. Single-use medical devices provided by the medical device manufacturer for processing prior to use can also require cleaning prior to further processing.

After cleaning, other factors can affect the safe and effective use of a medical device. For example, procedures for inspection and functional testing might be necessary to ensure that a medical device does not pose a risk to safety when used. Manufacturers of medical devices can assist users by providing instructions on how this inspection and testing should be performed.

Manufacturers of medical devices that are to be processed have a responsibility to ensure that the design of the medical devices facilitates achievement of effective processing. This includes consideration of commonly available validated processes; examples are shown in [Annex A](#). This annex can be used as a guide to validate procedures.

This is a preview of BS EN ISO 17664-2:2023. [Click here to purchase the full version from the ANSI store.](#)

This is a preview of BS EN ISO 17664-2:2023. [Click here to purchase the full version from the ANSI store.](#)

Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices —

Part 2: Non-critical medical devices

1 Scope

This document specifies requirements for the information to be provided by the medical device manufacturer for the processing of non-critical medical devices not intended to be sterilized (i.e. a medical device that is intended to come into contact with intact skin only or a medical device not intended for direct patient contact).

This includes information for processing prior to use or reuse of the medical device.

Processing instructions are not defined in this document. Rather, this document specifies requirements to assist manufacturers of medical devices in providing detailed processing instructions that consist of the following activities, where applicable:

- a) preparation before processing;
- b) cleaning;
- c) disinfection;
- d) drying;
- e) inspection and maintenance;
- f) packaging;
- g) storage;
- h) transportation.

This document excludes processing of:

- 1) critical and semi-critical medical devices;
- 2) medical devices intended to be sterilized;
- 3) textile medical devices used in patient draping systems or surgical clothing;
- 4) medical devices specified by the manufacturer for single use only and supplied ready for use.

NOTE See [Annex E](#) for further guidance on the application of the ISO 17664 series to a medical device.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 14971, *Medical devices — Application of risk management to medical devices*