



BSI Standards Publication

Packaging for terminally sterilized medical devices

Part 2: Sterilization wrap — Requirements and test methods

This is a preview of BS EN 868-2:2025. [Click here to purchase the full version from the ANSI store.](#)

National foreword

This British Standard is the UK implementation of EN 868-2:2025. It supersedes BS EN 868-2:2017, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/198, Sterilization and Associated Equipment and Processes.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2025
Published by BSI Standards Limited 2025

ISBN 978 0 539 28557 4

ICS 11.080.30; 55.040

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 April 2025.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

This is a preview of BS EN 868-2:2025. [Click here to purchase the full version from the ANSI store.](#)

EUROPÄISCHE NORM

April 2025

ICS 11.080.30

Supersedes EN 868-2:2017

English Version

Packaging for terminally sterilized medical devices - Part 2: Sterilization wrap - Requirements and test methods

Emballages des dispositifs médicaux stérilisés au stade terminal - Partie 2 : Enveloppe de stérilisation - Exigences et méthodes d'essai

Verpackungen für in der Endverpackung zu sterilisierende Medizinprodukte - Teil 2: Sterilisierverpackung - Anforderungen und Prüfverfahren

This European Standard was approved by CEN on 14 March 2025.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents		Page
European foreword		3
Introduction		5
1	Scope	6
2	Normative references	6
3	Terms and definitions	7
4	General requirements	7
5	Performance requirements and test methods	8
5.1	General	8
5.2	Specific requirements for wrap made of creped paper	8
5.3	Specific requirements for wrap made of nonwoven material	9
5.4	Specific requirements for wrap made of woven textile material	9
6	Sterilization compatibility	10
7	Labelling	10
7.1	General	10
7.2	Sales packaging	10
7.3	Inner packaging	10
8	Information to be provided	11
8.1	Application of the sterilization wrap	11
8.2	Environmental declarations	11
Annex A (informative) Method for the determination of drape		12
Annex B (normative) Method for the determination of water repellency		14
Annex C (normative) Method for the determination of pore size		17
Annex D (informative) Repeatability and reproducibility of test methods		22
Annex E (informative) Environmental aspects		24
Bibliography		27

This is a preview of BS EN 868-2:2025. [Click here to purchase the full version from the ANSI store.](#)

European foreword

This document (EN 868-2:2025) has been prepared by Technical Committee CEN/TC 102 “Sterilizers and associated equipment for processing of medical devices”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2025, and conflicting national standards shall be withdrawn at the latest by October 2025.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 868-2:2017.

EN 868-2:2025 includes the following significant technical changes with respect to EN 868-2:2017:

- a) The scope has been revised for clarity; plain paper has been removed as material for sterilization wrap.
- b) The document was renumbered to limit the list numbering to 3 levels only for better readability.
- c) Clause 4 “General requirements” was slightly revised for clarity and aligned with the other parts of the EN 868 series, and a statement was added clarifying when requirements apply.
- d) Clause 5 “Performance requirements and test methods”: materials used solely for irradiation sterilization are excluded from the requirements regarding wet tensile strength of the wrap.
- e) Clause 6 “Sterilization compatibility” was added and aligned with the other parts of the EN 868 series.
- f) Subclause 8.2 “Environmental declarations” was added and aligned with the other parts of EN 868 series.
- g) The list of major changes was moved to the foreword, thus the Annex with “Details of significant technical changes between this European Standard and the previous edition” (former Annex A) was deleted.
- h) New Clause “Environmental aspects for testing” was added to each test method in Annexes A to C.
- i) New Annex E regarding environmental aspects was added.

EN 868 consists of the following parts, under the general title *Packaging for terminally sterilized medical devices*:

- *Part 2: Sterilization wrap — Requirements and test methods;*
- *Part 3: Paper for use in the manufacture of paper bags (specified in EN 868-4) and in the manufacture of pouches and reels (specified in EN 868-5) — Requirements and test methods;*
- *Part 4: Paper bags — Requirements and test methods;*
- *Part 5: Sealable pouches and reels of porous materials and plastic film construction — Requirements and test methods;*
- *Part 6: Paper for low temperature sterilization processes — Requirements and test methods;*

This is a preview of BS EN 868-2:2025. [Click here to purchase the full version from the ANSI store.](#)

- *Part 7: Adhesive coated paper for low temperature sterilization processes — Requirements and test methods;*
- *Part 8: Re-usable sterilization containers for steam sterilizers conforming to EN 285 — Requirements and test methods;*
- *Part 9: Uncoated nonwoven materials of polyolefines — Requirements and test methods;*
- *Part 10: Adhesive coated nonwoven materials of polyolefines — Requirements and test methods.*

Any feedback and questions on this document should be directed to the users' national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

This is a preview of BS EN 868-2:2025. [Click here to purchase the full version from the ANSI store.](#)

Introduction

The EN ISO 11607 series of standards consists of two parts under the general title “Packaging for terminally sterilized medical devices”. EN ISO 11607-1 specifies general requirements and test methods for materials, preformed sterile barrier systems, sterile barrier systems and packaging systems that are intended to maintain sterility of terminally sterilized medical devices to the point of use. EN ISO 11607-2 specifies validation requirements for forming, sealing and assembly processes.

General requirements for all types of sterile barrier systems are provided by EN ISO 11607-1 and EN ISO 11607-2 [1].

The EN 868 series of standards have been developed mainly for materials and sterile barrier systems used in health care facilities sterilization processes. The EN 868 series of standards can be used to demonstrate compliance with one or more of the requirements specified in EN ISO 11607-1.

Considering CEN guide 4 [2] and the CEN environmental checklists, this revision has been complemented with a new annex with guidance to encourage users to also include environmental aspects when applying the EN 868 series of standards with the objective to minimize the environmental impact. Environmental aspects have also been included into the description of test methods with the same objective.

1 Scope

This document specifies test methods and values for sterilization wrap made of

- single-use creped paper
- single-use nonwoven materials
- reusable woven textile materials

used as sterile barrier systems and/or packaging systems for terminally sterilized medical devices.

Other than the general requirements as specified in EN ISO 11607-1 and EN ISO 11607-2, this part of EN 868 specifies materials, test methods and values that are specific to the products covered by this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 187, *Paper, board and pulps — Standard atmosphere for conditioning and testing and procedure for monitoring the atmosphere and conditioning of samples (ISO 187)*

EN ISO 536, *Paper and board — Determination of grammage (ISO 536)*

EN ISO 811, *Textiles — Determination of resistance to water penetration — Hydrostatic pressure test (ISO 811)*

EN ISO 1924-2, *Paper and board — Determination of tensile properties — Part 2: Constant rate of elongation method (20 mm/min) (ISO 1924-2)*

EN ISO 1974, *Paper — Determination of tearing resistance — Elmendorf method (ISO 1974)*

EN ISO 2758, *Paper — Determination of bursting strength (ISO 2758:2014)*

EN ISO 9073-3, *Nonwovens — Test methods — Part 3: Determination of tensile strength and elongation at break using the strip method (ISO 9073-3)*

EN ISO 9237, *Textiles — Determination of permeability of fabrics to air (ISO 9237)*

EN ISO 11607-1:2020, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems (ISO 11607-1:2019)*

EN ISO 13937-1, *Textiles — Tear properties of fabrics — Part 1: Determination of tear force using ballistic pendulum method (Elmendorf) (ISO 13937-1)*

EN ISO 13938-1, *Textiles — Bursting properties of fabrics — Part 1: Hydraulic method for determination of bursting strength and bursting distension (ISO 13938-1)*

EN ISO 14021, *Environmental labels and declarations — Self-declared environmental claims (Type II environmental labelling) (ISO 14021)*

EN ISO 14025, *Environmental labels and declarations — Type III environmental declarations — Principles and procedures (ISO 14025)*