



May 2014

# EP10-A3-AMD

## Preliminary Evaluation of Quantitative Clinical Laboratory Measurement Procedures; Approved Guideline—Third Edition

This guideline provides experimental design and data analysis for preliminary evaluation of the performance of a measurement procedure or device.

A guideline for global application developed through the Clinical and Laboratory Standards Institute consensus process.

# Clinical and Laboratory Standards Institute

*Setting the standard for quality in clinical laboratory testing around the world.*

The Clinical and Laboratory Standards Institute (CLSI) is a not-for-profit membership organization that brings together the varied perspectives and expertise of the worldwide laboratory community for the advancement of a common cause: to foster excellence in laboratory medicine by developing and implementing clinical laboratory standards and guidelines that help laboratories fulfill their responsibilities with efficiency, effectiveness, and global applicability.

## Consensus Process

Consensus—the substantial agreement by materially affected, competent, and interested parties—is core to the development of all CLSI documents. It does not always connote unanimous agreement, but does mean that the participants in the development of a consensus document have considered and resolved all relevant objections and accept the resulting agreement.

## Commenting on Documents

CLSI documents undergo periodic evaluation and modification to keep pace with advancements in technologies, procedures, methods, and protocols affecting the laboratory or health care.

CLSI's consensus process depends on experts who volunteer to serve as contributing authors and/or as participants in the reviewing and commenting process. At the end of each comment period, the committee that developed the document is obligated to review all comments, respond in writing to all substantive comments, and revise the draft document as appropriate.

Comments on published CLSI documents are equally essential, and may be submitted by anyone, at any time, on any document. All comments are addressed according to the consensus process by a committee of experts.

## Appeals Process

If it is believed that an objection has not been adequately addressed, the process for appeals is documented in the CLSI Standards Development Policies and Process document.

All comments and responses submitted on draft and published documents are retained on file at CLSI and are available upon request.

## Get Involved—Volunteer!

Do you use CLSI documents in your workplace? Do you see room for improvement? Would you like to get involved in the revision process? Or maybe you see a need to develop a new document for an emerging technology? CLSI wants to hear from you. We are always looking for volunteers. By donating your time and talents to improve the standards that affect your own work, you will play an active role in improving public health across the globe.

For further information on committee participation or to submit comments, contact CLSI.

Clinical and Laboratory Standards Institute  
950 West Valley Road, Suite 2500  
Wayne, PA 19087 USA  
P: 610.688.0100  
F: 610.688.0700  
[www.clsi.org](http://www.clsi.org)  
[standard@clsi.org](mailto:standard@clsi.org)

ISBN 1-56238-622-0  
ISSN 0273-3099

EP10-A3-AMD  
Vol. 26 No. 34  
Replaces EP10-A2  
Vol. 22 No. 29

---

## Preliminary Evaluation of Quantitative Clinical Laboratory Measurement Procedures; Approved Guideline—Third Edition

Volume 26 Number 34

Jan S. Krouwer, PhD  
George S. Cembrowski, MD, PhD  
Daniel W. Tholen, MS

### Abstract

Clinical and Laboratory Standards Institute document EP10-A3-AMD—*Preliminary Evaluation of Quantitative Clinical Laboratory Measurement Procedures; Approved Guideline—Third Edition* is intended to facilitate a limited, preliminary evaluation of the performance of a measurement procedure or device. Using the experimental design and data analysis procedure described, determination of whether a device has problems that require further evaluation or referral to the manufacturer can be done with a minimum expenditure of time and material. Included in Appendixes A and B are sample data sheets that should facilitate the analysis of the data. Appendix C contains a more sophisticated, powerful, statistical method for determining the possible causes of imprecision.

Clinical and Laboratory Standards Institute (CLSI). *Preliminary Evaluation of Quantitative Clinical Laboratory Measurement Procedures; Approved Guideline—Third Edition*. CLSI document EP10-A3-AMD (ISBN 1-56238-622-0). Clinical and Laboratory Standards Institute, 950 West Valley Road, Suite 2500, Wayne, Pennsylvania 19087 USA, 2014.

The Clinical and Laboratory Standards Institute consensus process, which is the mechanism for moving a document through two or more levels of review by the health care community, is an ongoing process. Users should expect revised editions of any given document. Because rapid changes in technology may affect the procedures, methods, and protocols in a standard or guideline, users should replace outdated editions with the current editions of CLSI documents. Current editions are listed in the CLSI catalog and posted on our website at [www.clsi.org](http://www.clsi.org). If you or your organization is not a member and would like to become one, and to request a copy of the catalog, contact us at: Telephone: 610.688.0100; Fax: 610.688.0700; E-Mail: [customerservice@clsi.org](mailto:customerservice@clsi.org); Website: [www.clsi.org](http://www.clsi.org).



CLINICAL AND  
LABORATORY  
STANDARDS  
INSTITUTE®

Copyright ©2014 Clinical and Laboratory Standards Institute. Except as stated below, any reproduction of content from a CLSI copyrighted standard, guideline, companion product, or other material requires express written consent from CLSI. All rights reserved. Interested parties may send permission requests to [permissions@clsi.org](mailto:permissions@clsi.org).

CLSI hereby grants permission to each individual member or purchaser to make a single reproduction of this publication for use in its laboratory procedure manual at a single site. To request permission to use this publication in any other manner, e-mail [permissions@clsi.org](mailto:permissions@clsi.org).

### **Suggested Citation**

CLSI. *Preliminary Evaluation of Quantitative Clinical Laboratory Measurement Procedures; Approved Guideline—Third Edition*. CLSI document EP10-A3-AMD. Wayne, PA: Clinical and Laboratory Standards Institute; 2014.

#### **Proposed Guideline**

December 1985

#### **Tentative Guideline**

June 1989

#### **Tentative Guideline—Second Edition**

September 1993

#### **Approved Guideline**

May 1998

#### **Approved Guideline—Second Edition**

December 2002

#### **Approved Guideline—Third Edition**

November 2006

#### **Amended Third Edition**

May 2014

ISBN 1-56238-622-0

ISSN 0273-3099

## Committee Membership

The changes in this amendment were approved by the Consensus Committee on Evaluation Protocol as follows.

### Consensus Committee on Evaluation Protocols

**James F. Pierson-Perry**

**Chairholder**

**Siemens Healthcare Diagnostic**

**Newark, Delaware, USA**

**Mitchell G. Scott, PhD**

**Vice-Chairholder**

**Barnes-Jewish Hospital,**

**Washington University School of**

**Medicine**

**St. Louis, Missouri, USA**

Karl De Vore

Bio-Rad Laboratories, Inc.

Irvine, California, USA

Robert J. McEnroe, PhD

Roche Diagnostics Operations, Inc.

Indianapolis, Indiana, USA

James H. Nichols, PhD, DABCC,

FACB

Vanderbilt University School of

Medicine

Nashville, Tennessee, USA

Gene Pennello, PhD

FDA Center for Devices and

Radiological Health

Silver Spring, Maryland, USA

Megan E. Sawchuk, MT(ASCP)

Centers for Disease Control and

Prevention

Atlanta, Georgia, USA

### Acknowledgment

CLSI and the Consensus Committee on Evaluation Protocols gratefully acknowledge the following individuals for reviewing all data and providing all appropriate amendments to this document:

Karl De Vore

Bio-Rad Laboratories, Inc.

Irvine, California, USA

Nils B. Person, PhD, FACB

Siemens Healthcare Diagnostics

Flanders, New Jersey, USA

James F. Pierson-Perry

Siemens Healthcare Diagnostics

Newark, Delaware, USA

## Area Committee on Evaluation Protocols

**Luann Ochs, MS**  
**Chairholder**

**Roche Diagnostics Corporation**  
**Indianapolis, Indiana**

**Greg Cooper, CLS, MHA**

**Vice-Chairholder**

**Bio-Rad Laboratories, Inc., QSD**  
**Division**  
**Plano, Texas**

George S. Cembrowski, MD, PhD  
Provincial Laboratory for Public  
Health  
Edmonton, Alberta, Canada

David L. Duewer, PhD  
National Institute of Standards and  
Technology  
Gaithersburg, Maryland

Anders Kallner, MD, PhD  
Karolinska Hospital  
Stockholm, Sweden

Kristian Linnet, MD, PhD  
University of Copenhagen  
Copenhagen, Denmark

Donald R. Parker, PhD  
Bayer HealthCare, LLC  
Elkhart, Indiana

Daniel W. Tholen, MS  
American Association for  
Laboratory Accreditation  
Traverse City, Michigan

Lakshmi Vishnuvajjala, PhD  
FDA Ctr. for Devices/Rad. Health  
Rockville, Maryland

### Advisors

David A. Armbruster, PhD,  
DABCC, FACB  
Abbott Laboratories  
Abbott Park, Illinois

R. Neill Carey, PhD  
Peninsula Regional Medical Center  
Salisbury, Maryland

Carl C. Garber, PhD, FACB  
Quest Diagnostics, Incorporated  
Lyndhurst, New Jersey

Patricia E. Garrett, PhD  
SeraCare Life Sciences, Inc.  
Portland, Maine

Martin H. Kroll, MD  
Dallas VA Medical Center  
Dallas, Texas

Jan S. Krouwer, PhD  
Krouwer Consulting  
Sherborn, Massachusetts

Jacob (Jack) B. Levine, MBA  
Bayer Corporation  
Tarrytown, New York

Donald M. Powers, PhD  
Powers Consulting Services  
Pittsford, New York

Max Robinowitz, MD  
FDA Ctr. for Devices/Rad. Health  
Rockville, Maryland

Gian Alfredo Scassellati, PhD  
Ente Nazionale Italiano Di  
Unificazione  
Turin, Italy

Michele M. Schoonmaker, PhD  
Cepheid  
Sunnyvale, California

Jack Zakowski, PhD, FACB  
Beckman Coulter, Inc.  
Brea, California

## Working Group on Evaluation of Quantitative Clinical Laboratory Methods

**Jan S. Krouwer, PhD**

**Chairholder**

**Krouwer Consulting**  
**Sherborn, Massachusetts**

George S. Cembrowski, MD, PhD  
Provincial Laboratory for Public  
Health  
Edmonton, Alberta, Canada

Daniel W. Tholen, MS  
Dan Tholen Statistical Consulting  
Traverse City, Michigan

### Staff

Clinical and Laboratory Standards  
Institute  
Wayne, Pennsylvania

John J. Zlockie, MBA  
*Vice President, Standards*

Lois M. Schmidt, DA  
*Staff Liaison*

Patrice E. Polgar  
*Projects Coordinator*

Donna M. Wilhelm  
*Editor*

Melissa A. Lewis  
*Assistant Editor*

## Acknowledgment

CLSI, the Area Committee on Evaluation Protocols, and the Working Group on Evaluation of Quantitative Clinical Laboratory Methods gratefully acknowledge Stanley Bauer, MD, a longtime contributor to CLSI, who had the foresight to commission Cuthbert Daniel, an award-winning statistical consultant, to prepare efficient protocols to evaluate commercial analyzers, and John Kennedy, also a dedicated contributor to CLSI, who chaired the original subcommittee that developed the EP10 consensus guideline.

**Contents**

Abstract..... i

Committee Membership..... iii

Summary of Changes in EP10 Amendment ..... vii

Foreword..... ix

1 Scope..... 1

2 Introduction..... 1

3 Standard Precautions..... 1

4 Terminology..... 2

    4.1 A Note on Terminology ..... 2

    4.2 Definitions ..... 2

5 Materials ..... 5

    5.1 Reference Procedures ..... 6

6 Calibration and Sequence of Samples in a Run ..... 6

7 Number of Days and Runs..... 6

8 Preliminary Procedures ..... 6

9 Collection and Recording of Data..... 6

10 Initial Data Plotting and Inspection ..... 7

    10.1 Difference Plot of Data vs. Concentration..... 7

    10.2 Visual Inspection for Outliers..... 10

    10.3 Visual Inspection for Linearity..... 10

11 Analysis of the Data for Imprecision..... 10

    11.1 Interpretation..... 10

12 Preliminary Assessment of Bias ..... 11

    12.1 Assigned Values ..... 11

    12.2 Calculation of Bias..... 11

    12.3 Interpretation..... 11

13 Full Data Analysis Procedures..... 11

    13.1 A Comment on the Model..... 12

    13.2 Summarizing the Five Runs..... 12

14 An Alternative Procedure ..... 12

15 Use of EP10 by Manufacturers ..... 12

16 How to Perform Multiple Regression for EP10..... 12

**Contents (Continued)**

References.....	15
Symbols Used in Appendixes .....	16
Appendix A. Preliminary Performance Acceptability Check.....	17
Appendix B. Example Use of Data Sheets .....	23
Appendix C. Statistical Explanation .....	44
The Quality System Approach.....	48
Evaluation Protocols Documents, Descriptions, and Key Words.....	49



## Summary of Changes in EP10 Amendment

### Foreword

- Deleted existing reference to “Excel,” but added trade name statement regarding its use in Section 16, per CLSI policy.

### Laboratory Error Sources and CLSI Evaluation Protocols Documents

- For consistency with current CLSI publications, the flow chart on Laboratory Error Sources and CLSI Evaluation Protocols Documents was removed.

### Section 4.2

- For clarification, added definitions for “commutable” and “continuous flow analyzer.”
- For clarification, modified definition of linear drift to state “a change in measurement value that is directly proportional to duration.”

### Section 5

- For clarification, changed “useful interval” to “desired interval” in the first sentence of the third paragraph.

### Section 6

- For clarification, added the phrase, “in the exact sequence specified” in the fifth sentence of the first paragraph.
- For clarification, added the statement “It is important to understand the protocol used by the analyzer to determine which sample will be pipetted next, and, if necessary, configure or program the analyzer to process the samples in the strict order outlined above.”

### Section 8

- For clarification, added the statement “It is especially important to know what steps are necessary to ensure that the samples are tested in the strict order specified.”

### Section 11

- For clarification, the first paragraph was shortened to state the following:

The imprecision of this experiment can be done by estimating the components of variance due to within-run, between-run (if more than one run is done per day), and between-day factors, as given by the formulas and procedures described in Appendixes A and C. These components of variance should then be added and the square root calculated to yield the “corrected” imprecision of this experiment. The relative sizes of the components may then be examined to investigate the sources of imprecision. This provides a more robust estimate of imprecision compared to simple calculation of the standard deviation of the data collected.

### Section 11.1

- For clarification, reference to Data Summary Sheet #3 was changed to Appendix A and the following statement was added: “Note that the imprecision goals should take into account the confidence limits from the estimate of imprecision due to the small sample sizes used in this procedure. CLSI document EP15<sup>1</sup> covers this topic.”

### Section 12.1.1

- For clarification, the first sentence was revised to “Analysis of an aliquot should be done by a measurement procedure of known accuracy (a reference method is ideal).”

### Section 12.2.1

- For clarification, reference to page 30 was changed to Appendix A.

### Section 16

- Deleted reference to “Excel” in the section title.
- For clarification, the first paragraph was revised to state the following:

This section describes how to perform the multiple regression calculations using the software program Microsoft<sup>®</sup> Excel (or the equivalent). Users of other software should be able to follow the description and implement the steps as appropriate. **NOTE:** This section uses the example ethanol data from Appendix B.

- For clarification, deleted “and does not need to be purchased separately” from the NOTE in the second paragraph.

### Appendix A

- The first footnote of Data Summary Sheet #1 was revised to “\* See Appendix C for basis of calculations. Standard deviations use equation 1 of Appendix C. When using software such as a spreadsheet program or statistical software to facilitate the calculations, be sure to use the formula for sample standard deviation with  $n - 1$  as the denominator.”

### Appendixes A and B

- Realigned table formats in the data summary sheets.

### Summary of Consensus and Delegate Comments and Committee Responses

- The Summary of Consensus and Delegate Comments and Committee Responses was removed as part of this amendment. This summary is on file at the CLSI office and available upon request by contacting CLSI at 610.688.0100 or [standard@clsi.org](mailto:standard@clsi.org).

## Foreword

Before using a new measurement procedure or instrument for *in vitro* diagnostic use, the laboratory must make a preliminary decision about its acceptability. This initial performance check is neither a rigorous characterization of long-term performance nor an evaluation of the many factors that can affect results produced by the device. Rather, this experiment is a quick check to rule out major problems and a starting point for accumulating data and experience that will enable the user to make a final decision. The primary purpose of this document is to help detect performance problems that would warrant immediate correction, referral to the manufacturer, or expanded investigation before a new device is placed into service.

This document may also now be used by manufacturers to either establish the magnitude of factors that can affect performance or verify that such magnitude is acceptable.

Additional revisions since the last edition of EP10 (2002) include:

- a figure to illustrate which error sources the EP10 protocol can detect with respect to all error sources and other EP documents (see page viii);
- suggested sample sizes, so now the document is useful for manufacturers;
- instructions for the multiple regression calculations;
- revised references; and
- revised definitions.

## Key Words

Carry-over, comparison of methods, drift, evaluation protocol, experimental design, linearity, multiple regression, outlier, precision

**Note that the trade name Microsoft<sup>®</sup> Excel is included in Section 16 of this document. It is Clinical and Laboratory Standards Institute's policy to avoid using a trade name unless the product identified is the only one available, or it serves solely as an illustrative example of the procedure, practice, or material described. In this case, the working group and consensus committee believe the trade name is an important descriptive adjunct to the document. In such cases, it is acceptable to use the product's trade name, as long as the words, "or the equivalent" are added to the references. It should be understood that information on this product in this standard also applies to any equivalent products. Please include in your comments any information that relates to this aspect of EP10.**



## **Preliminary Evaluation of Quantitative Clinical Laboratory Measurement Procedures; Approved Guideline—Third Edition**

### **1 Scope**

Before starting a complete evaluation of a new measurement procedure, kit, or instrument for *in vitro* diagnostic use, it is often necessary to make a preliminary decision about its acceptability. This initial performance check is neither a rigorous investigation into the procedure's long-term performance, nor an evaluation of the many factors that can affect results produced by the device. The primary purpose of this document is to help detect problems that are severe enough to warrant immediate correction, referral to the manufacturer, or expanded investigation. Accreditation bodies may have requirements for verification or validation that exceed the procedures in this document (see CLSI document EP15<sup>1</sup>).

Manufacturers can also benefit by performing this protocol either as assays are developed or when they are validated. By performing more than five runs, manufacturers can detect trends in the effects estimated by EP10 or document their absence.

### **2 Introduction**

This document describes a procedure for the preliminary evaluation of linearity, proportional and constant bias, linear drift, sample carry-over, and precision of a clinical laboratory measurement procedure. Preliminary evaluations should be performed before new procedures are used to test patients' samples and when any modifications of procedures are made. This guideline is based on a protocol and procedure developed for continuous flow analyzers.<sup>2</sup> The rationale for recommending a protocol based on so old a system is explained in Section 13.1. The experiment is intended primarily for evaluating automated instruments but may be appropriate for kits, manual procedures, or other *in vitro* diagnostic devices. By repeating a sequence of only ten samples, performance characteristics may be evaluated by plotting the data and performing some simple calculations. Using a statistical technique called multiple linear regression analysis, further information about the factors influencing accuracy (such as sample carry-over linear drift, and nonlinearity) can be obtained. Instructions are given for simple data analysis, in case a computer is not available.

The experiment is intended to provide preliminary estimates of those performance characteristics that may be used to determine the ultimate acceptability of the device. The results should be used only to determine whether the device has grossly unacceptable performance.

The following sections outline the materials and procedures to be used. Many variations on this basic experiment are possible (such as extending the number of days or eliminating the priming samples when appropriate). Variations should be dictated by the complexities of the device, the particular characteristics of the measurement procedure, and the resources available to the user.

### **3 Standard Precautions**

Because it is often impossible to know what isolates or specimens might be infectious, all patient and laboratory specimens are treated as infectious and handled according to "standard precautions." Standard precautions are guidelines that combine the major features of "universal precautions and body substance isolation" practices. Standard precautions cover the transmission of all infectious agents and thus are more comprehensive than universal precautions which are intended to apply only to transmission of blood-borne pathogens. Standard and universal precaution guidelines are available from the U.S. Centers for Disease Control and Prevention (Garner JS, Hospital Infection Control Practices Advisory Committee. Guideline for isolation precautions in hospitals. *Infect Control Hosp Epidemiol.* 1996;17(1):53-80). For