

June 2014

M40-A2

Quality Control of Microbiological Transport Systems; Approved Standard—Second Edition

This document provides criteria to assist manufacturers and end users of transport devices in providing and selecting dependable products for the transport of microbiological clinical specimens.

A standard for global application developed through the Clinical and Laboratory Standards Institute consensus process.

Clinical and Laboratory Standards Institute

Setting the standard for quality in medical laboratory testing around the world.

The Clinical and Laboratory Standards Institute (CLSI) is a not-for-profit membership organization that brings together the varied perspectives and expertise of the worldwide laboratory community for the advancement of a common cause: to foster excellence in laboratory medicine by developing and implementing medical laboratory standards and guidelines that help laboratories fulfill their responsibilities with efficiency, effectiveness, and global applicability.

Consensus Process

Consensus—the substantial agreement by materially affected, competent, and interested parties—is core to the development of all CLSI documents. It does not always connote unanimous agreement but does mean that the participants in the development of a consensus document have considered and resolved all relevant objections and accept the resulting agreement.

Commenting on Documents

CLSI documents undergo periodic evaluation and modification to keep pace with advances in technologies, procedures, methods, and protocols affecting the laboratory or health care.

CLSI's consensus process depends on experts who volunteer to serve as contributing authors and/or as participants in the reviewing and commenting process. At the end of each comment period, the committee that developed the document is obligated to review all comments, respond in writing to all substantive comments, and revise the draft document as appropriate.

Comments on published CLSI documents are equally essential and may be submitted by anyone, at any time, on any document. All comments are managed according to the consensus process by a committee of experts.

Appeal Process

When it is believed that an objection has not been adequately considered and responded to, the process for appeal, documented in the CLSI *Standards Development Policies and Processes*, is followed.

All comments and responses submitted on draft and published documents are retained on file at CLSI and are available upon request.

Get Involved—Volunteer!

Do you use CLSI documents in your workplace? Do you see room for improvement? Would you like to get involved in the revision process? Or maybe you see a need to develop a new document for an emerging technology? CLSI wants to hear from you. We are always looking for volunteers. By donating your time and talents to improve the standards that affect your own work, you will play an active role in improving public health across the globe.

For additional information on committee participation or to submit comments, contact CLSI.

Clinical and Laboratory Standards Institute 950 West Valley Road, Suite 2500 Wayne, PA 19087 USA P: +1.610.688.0100 F: +1.610.688.0700 www.clsi.org standard@clsi.org ISBN 1-56238-963-7 (Print) ISBN 1-56238-964-5 (Electronic) ISSN 1558-6502 (Print) ISSN 2162-2914 (Electronic)

M40-A2 Vol. 34 No. 9

Quality Control of Microbiological Transport Systems; Approved Standard—Second Edition

Volume 34 Number 9

Paul Bourbeau, PhD, D(ABMM) Paul L. Cerwinka, MSc, RM(AAM) Jerry Abramson, PhD, F(AAM), SM(AAM) Susan Finn, MLT Musa Y. Hindiyeh, PhD, D(ABMM) Michael J. Loeffelholz, PhD, D(ABMM) Elaine Seavey Maliff Charles Thomas Nugent IV, PhD CDR Raquel Peat, PhD, MPH Norman Sharples Douglas J. Shedden Kenneth Van Horn, PhD, D(ABMM)

Abstract

Clinical and Laboratory Standards Institute document M40-A2—*Quality Control of Microbiological Transport Systems; Approved Standard*—*Second Edition* presents the criteria that shall be considered when choosing a microbiological transport device to facilitate sample preservation. QC considerations for the manufacturer and testing laboratory are presented, as well as techniques, control microorganisms, and acceptability criteria. This document provides a consistent protocol for initial testing of microbiological transport devices by manufacturers and a method by which laboratories can validate manufacturer claims and compare devices.

Clinical and Laboratory Standards Institute (CLSI). *Quality Control of Microbiological Transport Systems; Approved Standard— Second Edition*. CLSI document M40-A2 (ISBN 1-56238-963-7 [Print]; ISBN 1-56238-964-5 [Electronic]). Clinical and Laboratory Standards Institute, 950 West Valley Road, Suite 2500, Wayne, Pennsylvania 19087 USA, 2014.

The Clinical and Laboratory Standards Institute consensus process, which is the mechanism for moving a document through two or more levels of review by the health care community, is an ongoing process. Users should expect revised editions of any given document. Because rapid changes in technology may affect the procedures, methods, and protocols in a standard or guideline, users should replace outdated editions with the current editions of CLSI documents. Current editions are listed in the CLSI catalog and posted on our website at www.clsi.org. If you or your organization is not a member and would like to become one, and to request a copy of the catalog, contact us at: Telephone: 610.688.0100; Fax: 610.688.0700; E-Mail: customerservice@clsi.org; Website: www.clsi.org.



Number 9

M40-A2

Copyright [©]2014 Clinical and Laboratory Standards Institute. Except as stated below, any reproduction of content from a CLSI copyrighted standard, guideline, companion product, or other material requires express written consent from CLSI. All rights reserved. Interested parties may send permission requests to permissions@clsi.org.

CLSI hereby grants permission to each individual member or purchaser to make a single reproduction of this publication for use in its laboratory procedure manual at a single site. To request permission to use this publication in any other manner, e-mail permissions@clsi.org.

Suggested Citation

CLSI. *Quality Control of Microbiological Transport Systems; Approved Standard—Second Edition.* CLSI document M40-A2. Wayne, PA: Clinical and Laboratory Standards Institute; 2014.

Proposed Standard August 2002

Approved Standard December 2003

Approved Standard—Second Edition June 2014

ISBN 1-56238-963-7 (Print) ISBN 1-56238-964-5 (Electronic) ISSN 1558-6502 (Print) ISSN 2162-2194 (Electronic)

Volume 34

Committee Membership

Consensus Committee on Microbiology

Richard B. Thomson, Jr., PhD, D(ABMM), FAAM Chairholder Evanston Hospital, NorthShore University HealthSystem Evanston, Illinois, USA

John H. Rex, MD, FACP Vice-Chairholder AstraZeneca Pharmaceuticals Waltham, Massachusetts, USA

Thomas R. Fritsche, MD, PhD Marshfield Clinic Marshfield, Wisconsin, USA Patrick R. Murray, PhD BD Diagnostic Systems Sparks, Maryland, USA

Jean B. Patel, PhD, D(ABMM) Centers for Disease Control and Prevention Atlanta, Georgia, USA

Kerry Snow, MS, MT(ASCP) FDA Center for Drug Evaluation and Research Silver Spring, Maryland, USA

John D. Turnidge, MD SA Pathology at Women's and Children's Hospital North Adelaide, Australia Jeffrey L. Watts, PhD, RM(NRCM) Zoetis Kalamazoo, Michigan, USA

M40-A2

Nancy L. Wengenack, PhD, D(ABMM), FIDSA Mayo Clinic Rochester, Minnesota, USA

Barbara L. Zimmer, PhD Siemens Healthcare Diagnostics Inc. West Sacramento, California, USA

Document Development Committee on Quality Control of Microbiological Transport Systems

Paul Bourbeau, PhD, D(ABMM) Co-Chairholder BD Diagnostic Systems Sparks, Maryland, USA

Paul L. Cerwinka, MSc, RM(AAM) Co-Chairholder Collegeville, Pennsylvania, USA

Jerry Abramson, PhD, F(AAM), SM(AAM) Reisterstown, Maryland, USA

Susan Finn, MLT Starplex Scientific Etobicoke, Ontario, Canada

Elaine Seavey Maliff Puritan Medical Products Company LLC Guilford, Maine, USA CDR Raquel Peat, PhD, MPH FDA Center for Devices and Radiological Health Silver Spring, Maryland, USA

Andrew Rudsdale Newcastle General Hospital Newcastle Upon Tyne, United Kingdom

Norman Sharples Copan Diagnostics Inc. Murrieta, California, USA

Douglas J. Shedden Medical Wire and Equipment Corsham/Wiltshire, United Kingdom

Staff

Clinical and Laboratory Standards Institute Wayne, Pennsylvania, USA

Luann Ochs, MS Senior Vice President – Operations

Tracy A. Dooley, MLT(ASCP) *Staff Liaison*

Megan L. Tertel, MA Editorial Manager

Joanne P. Christopher, MA *Editor*

Patrice E. Polgar *Editor*

Acknowledgment

CLSI and the Consensus Committee on Microbiology gratefully acknowledge the following individuals for their help in preparing the second edition of this document:

Musa Y. Hindiyeh, PhD, D(ABMM) Caritas Baby Hospital Bethlehem, Palestine

Michael J. Loeffelholz, PhD, D(ABMM) The University of Texas Medical Branch Galveston, Texas, USA Charles Thomas Nugent IV, PhD Gen-Probe San Diego, California, USA

Kenneth Van Horn, PhD, D(ABMM) Focus Diagnostics Cypress, California, USA

Number 9

M40-A2

Volume 34		

Contents

Abstracti					
Committee Membershipiii					
Forewo	ord		. vii		
1	Scope1				
2	Introduction				
3	Standard Precautions				
4	Terminology		3		
	4.1 4.2 4.3	A Note on Terminology Definitions Abbreviations and Acronyms	3		
5	General Microbiological Transport System Considerations		5		
	5.1 5.2 5.3 5.4 5.5 5.6	Container Structural Requirements Assessment of Structural Requirements Biological Requirements of the Transport System Bio-Burden Product Information US Department of Transportation/International Air Transport Association Regulations Governing Shipping	6 7 7		
6	Quality	y Control	9		
	6.1	Sampling	9		
7	Quality	y Control of the Transport Container	9		
	7.1 7.2	Manufacturer Laboratory End User			
8	Bacteriological Swab Transport Devices		10		
	8.1 8.2 8.3 8.4 8.5 8.6 8.7 8.8 8.9 8.10 8.11 8.12	Storage Conditions Shelf-Life Studies Microorganism-Specific Standards Source of Control Strains Maintenance of Control Strains Preparation of Frozen Stock Cultures of Bacteria Working Control Cultures Turbidity Standard for Inoculum Preparation Direct Colony Suspension Inoculum Preparation Method Quality Control Protocols Swab Elution Method (Quantitative) Roll-Plate Method (Qualitative)	10 10 11 11 11 12 12 12 12 13 18		
9	Specia	lized Transport Devices	22		
	9.1 9.2 9.3	General Considerations Urine Culture Containers With Preservative Fecal Culture Containers With Preservative	22		

M40-A2

Number 9

Contents (Continued)

	9.4	Media That Are Used as Transport Devices	
10	Moleo	cular Transport and Specimen Types	
	10.1 10.2	General Considerations	
11	Viral	Specimen Collection and Transport Devices	
	11.1 11.2	Specimen Collection and Transport Viral Recovery	
Refe	rences		
Appe	endix A. S	Sources of Specimen Container Failure	
Appe	endix B. I	Regulatory Considerations	
The	Quality M	Aanagement System Approach	
Relat	ted CLSI	Reference Materials	

Volume 34

M40-A2

Foreword

In 1893, Councilman first described preparing transport swabs by wrapping cotton pledgets around the ends of wires, enclosing these wires in test tubes, and sterilizing them in a hot air sterilizer.¹ After sterilization, the test tubes were taken to wards where a wire was removed and used to rub the pharyngeal membranes of patients suspected of being infected with diphtheria. After collection, the test tubes were labeled and sent to a laboratory where specimens were inoculated on a culture medium.

The development of transport devices was a result of public health concerns.² Maintaining microorganism viability during transport to the public health laboratory was imperative for the isolation and identification of the agents responsible for relevant infectious diseases. During the 1930s, 1940s, and 1950s, such infections, particularly gonorrhea and bacterial diarrhea, were the driving forces behind the development of transport media and devices.³ Most studies focused on evaluation of performance rather than establishment of an acceptable standard of expected performance.⁴ It is difficult to determine when systematic QC began to be applied to transport systems. However, it was Rubbo, Benjamin, and Stuart⁵ who noted that certain batches of cotton wool used on swabs were associated with faster microbial death rates than others and that this phenomenon (toxicity) could be countered by the addition of serum onto the transport swabs.⁶ Additionally, the important contributions of Amies, Cary, and Blair for transport medium should be acknowledged.^{7,8}

Within the hospital setting, use of transport devices for various "routine cultures" began as investigators determined variability in recovery from specimens plated at the bedside compared to those routed to the laboratory via established mechanisms.⁵ Currently, a number of factors contribute to the increasing emphasis on the use of transport devices to maintain specimens for microbiological testing. These factors include an increased use of outpatient treatment that has accompanied shortened hospital stays, and the centralization of laboratory services due to both managed care and a shortage of individuals with expertise in clinical microbiology.

As new technologies provide the opportunity to redefine the method of recovery or detection of microorganisms of interest, standardizing the QC testing and acceptance criteria will become important in order to assure the highest level of care to patients. This document on QC of transport devices will assist in the standardization of the performance of these devices.

Since the publication of the first edition of this document, many studies have followed its recommendations.^{9,10-17} In revising M40, the committee updated the document where appropriate, using data generated from studies performed using the protocols that were published in the first edition. Testing protocols were updated to accommodate new types of swab collection devices that have been introduced since the first edition was published, and temperatures under which QC testing and specimen transport are conducted were better defined. Lastly, the committee added to, or expanded sections related to, QC of transport devices used for viruses, urine, and fecal specimens.

In the United States, basic manufacturing requirements for medical devices, including *in vitro* diagnostic devices, were established via the Medical Device Amendments of 1976. This legislation gave the US Food and Drug Administration (FDA) authority to regulate medical devices (premarket notification, [510(k)] and premarket approval), and develop consistent manufacturing requirements (good manufacturing practices [GMP]). GMP include the requirement to perform product QC testing before distribution. Each manufacturer is required to establish the type of testing to be performed, as well as acceptance criteria based on the product and its intended use. Additionally, the European Union has adopted the Medical Devices Directive 93/42/EEC¹⁸ and the *In Vitro* Diagnostic Device Directive 98/79/EC,¹⁹ which have requirements very similar to those in the United States. These directives include provisions to use harmonized standards as a method of demonstrating conformity to the directive requirements. Likewise, the FDA has formalized the use of these types of standards by manufacturers to

Number 9

demonstrate performance in premarket submissions. Further discussion of regulatory considerations for these markets can be found in Appendix B.

Key Words

Acceptable performance, acceptance criteria, biological properties, control strains, microbiological, microbiological testing, molecular transport, performance criteria, quality control, regulatory considerations, specimen transport, standards, storage conditions, transport devices, transport medium, transport temperature, viral transport

Volume 34

M40-A2

Quality Control of Microbiological Transport Systems; Approved Standard— Second Edition

1 Scope

The transport of clinical specimens is a critical component for accurate diagnosis. Preservation of inherent, interpretive attributes of microorganisms and/or nucleic acids can be quickly compromised when the transport conditions or transport devices are suboptimal. The advent of antigen detection methods, methods for amplification and detection of genetic elements, and the requirement for local or distant transport of these specimens to a testing facility has imposed further considerations on manufacturers to provide products that will not compromise reporting of clinically relevant laboratory data to physicians. Clinicians should be able to collect and submit specimens to the laboratory and laboratorians should be able to retrieve specimens from containers, devices, and transport media with a reasonable assurance that the viability of microorganisms and/or preservation of nucleic acids present in the specimen will be maintained.

This standard provides criteria to the manufacturers and end users of transport devices to assist in providing and choosing dependable products for the transport of microbiological clinical specimens. Manufacturers will be able to state whether or not the performance characteristics for specific groups of microorganisms and transport devices of a particular product satisfy the performance standards as specified in this document. Furthermore, manufacturers shall state whether or not any additional testing is required before the use of a particular product.

Secondary distributors and end users must assume the responsibility for storage and transport conditions of specimen collection devices before and after use, by adhering to the conditions specified by the manufacturer or those deemed optimum by the laboratory in order to ensure microorganism viability/stability.

In this document, except as specifically noted, QC consists of an assessment of the performance characteristics of a complete device, and not the individual components. There are multiple variables involved in the manufacture of a transport device, including, but not limited to, the container, transport medium, collection device, packaging, and environment. It is fundamental that the assessment of the device be based on measurable performance characteristics for the particular device.

This document is not intended to provide proprietary information on product development, but rather to provide assurance to the device's user that manufacturer claims are met following standardized testing and acceptance criteria. It provides guidance to the manufacturer in addressing critical issues related to specimen integrity specific to the type of testing to be performed, eg, bacterial and viral culture, or nucleic acid detection. This document does not address the technique of transport device manufacturing, but focuses on the methods for QC testing and acceptance criteria in order to provide a product suitable for the analysis of clinical specimens for agents of disease.

Transport devices are essential components of the preexamination process of microbiology laboratory testing. It is recognized that these early steps in the total testing process are critical to the production of clinically relevant information. Patients, physicians, health care providers, and laboratorians expect products that meet the highest standards of laboratory practice. This document will facilitate this goal.^a And while it is beyond the scope of this document to address the design of devices, it is imperative that

^a In the United States, the Clinical Laboratory Improvement Amendments guidelines place the responsibility for acceptance of quality specimens on the laboratorian.

[©]Clinical and Laboratory Standards Institute. All rights reserved.