

EP6-A  
Vol. 23 No. 16  
Replaces EP6-P2  
Vol. 21 No. 20

---

# Evaluation of the Linearity of Quantitative Measurement Procedures: A Statistical Approach; Approved Guideline

This document provides guidance for characterizing the linearity of a method during a method evaluation; for checking linearity as part of routine quality assurance; and for determining and stating a manufacturer's claim for linear range.

---

A guideline for global application developed through the NCCLS consensus process.



# NCCLS...

## Serving the World's Medical Science Community Through Voluntary Consensus

NCCLS is an international, interdisciplinary, nonprofit, standards-developing, and educational organization that promotes the development and use of voluntary consensus standards and guidelines within the healthcare community. It is recognized worldwide for the application of its unique consensus process in the development of standards and guidelines for patient testing and related healthcare issues. NCCLS is based on the principle that consensus is an effective and cost-effective way to improve patient testing and healthcare services.

In addition to developing and promoting the use of voluntary consensus standards and guidelines, NCCLS provides an open and unbiased forum to address critical issues affecting the quality of patient testing and health care.

### PUBLICATIONS

An NCCLS document is published as a standard, guideline, or committee report.

**Standard** A document developed through the consensus process that clearly identifies specific, essential requirements for materials, methods, or practices for use in an unmodified form. A standard may, in addition, contain discretionary elements, which are clearly identified.

**Guideline** A document developed through the consensus process describing criteria for a general operating practice, procedure, or material for voluntary use. A guideline may be used as written or modified by the user to fit specific needs.

**Report** A document that has not been subjected to consensus review and is released by the Board of Directors.

### CONSENSUS PROCESS

The NCCLS voluntary consensus process is a protocol establishing formal criteria for:

- the authorization of a project
- the development and open review of documents
- the revision of documents in response to comments by users
- the acceptance of a document as a consensus standard or guideline.

Most NCCLS documents are subject to two levels of consensus—"proposed" and "approved." Depending on

the need for field evaluation or data collection, documents may also be made available for review at an intermediate (i.e., "tentative") consensus level.

**Proposed** An NCCLS consensus document undergoes the first stage of review by the healthcare community as a proposed standard or guideline. The document should receive a wide and thorough technical review, including an overall review of its scope, approach, and utility, and a line-by-line review of its technical and editorial content.

**Tentative** A tentative standard or guideline is made available for review and comment only when a recommended method has a well-defined need for a field evaluation or when a recommended protocol requires that specific data be collected. It should be reviewed to ensure its utility.

**Approved** An approved standard or guideline has achieved consensus within the healthcare community. It should be reviewed to assess the utility of the final document, to ensure attainment of consensus (i.e., that comments on earlier versions have been satisfactorily addressed), and to identify the need for additional consensus documents.

NCCLS standards and guidelines represent a consensus opinion on good practices and reflect the substantial agreement by materially affected, competent, and interested parties obtained by following NCCLS's established consensus procedures. Provisions in NCCLS standards and guidelines may be more or less stringent than applicable regulations. Consequently, conformance to this voluntary consensus document does not relieve the user of responsibility for compliance with applicable regulations.

### COMMENTS

The comments of users are essential to the consensus process. Anyone may submit a comment, and all comments are addressed, according to the consensus process, by the NCCLS committee that wrote the document. All comments, including those that result in a change to the document when published at the next consensus level and those that do not result in a change, are responded to by the committee in an appendix to the document. Readers are strongly encouraged to comment in any form and at any time on any NCCLS document. Address comments to the NCCLS Executive Offices, 940 West Valley Road, Suite 1400, Wayne, PA 19087, USA.

### VOLUNTEER PARTICIPATION

Healthcare professionals in all specialties are urged to volunteer for participation in NCCLS projects. Please contact the NCCLS Executive Offices for additional information on committee participation.

EP6-A  
ISBN 1-56238-498-8  
ISSN 0273-3099

Volume 23 Number 16

---

## Evaluation of the Linearity of Quantitative Measurement Procedures: A Statistical Approach; Approved Guideline

Daniel W. Tholen, M.S., Chairholder  
Martin Kroll, M.D., FACB  
J. Rex Astles, Ph.D., FACB  
Albert L. Caffo, Ph.D.  
Thomas M. Happe, MT(ASCP)  
Jan Krouwer, Ph.D.  
Fred Lasky, Ph.D.

### Abstract

NCCLS document EP6-A— *Evaluation of the Linearity of Quantitative Measurement Procedures: A Statistical Approach; Approved Guideline* is intended to provide both manufacturers and users of quantitative analytical methods with an economical and user-friendly method of establishing or verifying the linear range. This guideline also can be used to demonstrate the extent to which a quantitative analytical method meets clinical requirements or manufacturer's linear range claims.

NCCLS. *Evaluation of the Linearity of Quantitative Measurement Procedures: A Statistical Approach; Approved Guideline*. NCCLS document EP6-A (ISBN 1-56238-498-8). NCCLS, 940 West Valley Road, Suite 1400, Wayne, Pennsylvania 19087-1898 USA, 2003.

THE NCCLS consensus process, which is the mechanism for moving a document through two or more levels of review by the healthcare community, is an ongoing process. Users should expect revised editions of any given document. Because rapid changes in technology may affect the procedures, methods, and protocols in a standard or guideline, users should replace outdated editions with the current editions of NCCLS documents. Current editions are listed in the *NCCLS Catalog*, which is distributed to member organizations, and to nonmembers on request. If your organization is not a member and would like to become one, and to request a copy of the *NCCLS Catalog*, contact the NCCLS Executive Offices. Telephone: 610.688.0100; Fax: 610.688.0700; E-Mail: [exoffice@nccls.org](mailto:exoffice@nccls.org); Website: [www.nccls.org](http://www.nccls.org)



This publication is protected by copyright. No part of it may be reproduced, stored in a retrieval system, transmitted, or made available in any form or by any means (electronic, mechanical, photocopying, recording, or otherwise) without prior written permission from NCCLS, except as stated below.

NCCLS hereby grants permission to reproduce limited portions of this publication for use in laboratory procedure manuals at a single site, for interlibrary loan, or for use in educational programs provided that multiple copies of such reproduction shall include the following notice, be distributed without charge, and, in no event, contain more than 20% of the document's text.

Reproduced with permission, from NCCLS publication EP6-A—*Evaluation of the Linearity of Quantitative Measurement Procedures: A Statistical Approach; Approved Guideline* (ISBN 1-56238-498-8). Copies of the current edition may be obtained from NCCLS, 940 West Valley Road, Suite 1400, Wayne, Pennsylvania 19087-1898, USA.

Permission to reproduce or otherwise use the text of this document to an extent that exceeds the exemptions granted here or under the Copyright Law must be obtained from NCCLS by written request. To request such permission, address inquiries to the Executive Director, NCCLS, 940 West Valley Road, Suite 1400, Wayne, Pennsylvania 19087-1898, USA.

Copyright ©2003. The National Committee for Clinical Laboratory Standards.

### **Suggested Citation**

(NCCLS. *Evaluation of the Linearity of Quantitative Measurement Procedures: A Statistical Approach; Approved Guideline*. NCCLS document EP6-A [ISBN 1-56238-498-8]. NCCLS, 940 West Valley Road, Suite 1400, Wayne, Pennsylvania 19087-1898 USA, 2003.)

### **Proposed Guideline**

October 1986

### **Proposed Guideline—Second Edition**

December 2001

### **Approved Guideline**

April 2003

ISBN 1-56238-498-8  
ISSN 0273-3099

## Committee Membership

### Area Committee on Evaluation Protocols

**Jan S. Krouwer, Ph.D.**  
Chairholder  
Krouwer Consulting  
Sherborn, Massachusetts

**Daniel W. Tholen, M.S.**  
Vice-Chairholder  
Dan Tholen Statistical Services  
Traverse City, Michigan

### Working Group on Linearity

**Daniel W. Tholen, M.S.**  
Chairholder  
Dan Tholen Statistical Services  
Traverse City, Michigan

Martin H. Kroll, M.D., FACB  
VA North Texas Healthcare System  
Dallas, Texas

J. Rex Astles, Ph.D., FACB  
Centers for Disease Control and  
Prevention  
Atlanta, Georgia

Albert L. Caffo, Ph.D.  
West Chester University  
West Chester, Pennsylvania

Thomas M. Happe, M.T. (ASCP)  
Maine Standards Company  
Portland, Maine

Jan S. Krouwer, Ph.D.  
Krouwer Consulting  
Sherborn, Massachusetts

Fred D. Lasky, Ph.D.  
Ortho Clinical Diagnostics  
Rochester, New York

#### Staff

Lois M. Schmidt, D.A.  
*Staff Liaison*  
NCCLS  
Wayne, Pennsylvania

Patrice E. Polgar  
*Project Manager, Editor*  
NCCLS  
Wayne, Pennsylvania

Donna M. Wilhelm  
*Assistant Editor*  
NCCLS  
Wayne, Pennsylvania



**Contents**

Abstract..... i

Committee Membership..... iii

Foreword..... vii

1 Scope..... 1

2 Introduction..... 1

    2.1 Purpose ..... 1

    2.2 Alternative Approaches ..... 2

    2.3 Definitions ..... 2

    2.4 Overview of the Experimental Approach ..... 4

3 Time Requirements..... 4

    3.1 Device Familiarization Period ..... 4

    3.2 Duration of the Experiment ..... 4

4 Evaluating Experimental Conditions ..... 4

    4.1 Specimen Requirement ..... 5

    4.2 Matrix Effects ..... 5

    4.3 Hierarchy of Acceptable Matrices ..... 5

    4.4 Selection of Materials Used to Supplement Samples ..... 7

    4.5 Analytical Sequence ..... 8

    4.6 Analyte Range ..... 8

    4.7 Sample Preparation and Value Assignment ..... 8

    4.8 Data Collection ..... 10

    4.9 Clinically Important Evaluation Concentrations ..... 10

5 Demonstration of Linear Range—Data Evaluation ..... 10

    5.1 Preliminary Data Examination ..... 10

    5.2 Outlier Test ..... 11

    5.3 Determination of the Linear Range ..... 11

    5.4 Considerations for Random Error ..... 13

6 Linear Range Claims..... 17

    6.1 Graphical Approaches ..... 17

    6.2 Verification of Claims for Linear Range ..... 17

    6.3 Allowable Error and Goal Setting ..... 17

    6.4 Effects of Other Errors ..... 17

    6.5 Assumptions ..... 18

    6.6 Guidelines for Stating a Claim for the Linear Range ..... 18

References..... 19

Appendix A. Dilution Schemes ..... 20

Appendix B. t table ..... 21

Appendix C. Examples ..... 22

**Contents (Continued)**

|  |    |
|--|----|
| Summary of Comments and Working Group Responses .....          | 30 |
| Summary of Delegate Comments and Working Group Responses ..... | 44 |
| The Quality System Approach.....                               | 46 |
| Related NCCLS Publications.....                                | 47 |



## Foreword

A quantitative analytical method is linear when there exists a mathematically verified straight-line relationship between the observed values and the true concentrations or activities of the analyte. The property of being linear is important for analytical and clinical laboratory methods. The linear relationship is valuable because it represents the simplest mathematical relationship and allows for simple and easy interpolation of results.

It has been argued that linearity is not necessary for such methods as competitive immunoassays, because the relationship between the response of the system and analyte concentration is inherently nonlinear. However, the mathematical relationship between the response and analyte concentrations should be sufficiently well defined to allow the selection of a suitable transformation of the dose-response curve to linear form. Furthermore, for an analytical method, one cannot interpolate between points unless one knows the results are linear.

The reason for ensuring linearity of analytical methods is important on clinical grounds. Clinicians know that the relationship between an analyte and the pathophysiologic process is usually nonlinear, but they expect that the results reported to them by the laboratory include a linear relationship between the result and the true concentration, count, or activity recovered. For example, if the true amount of an analyte in a sample were to double, clinicians expect that would be reflected in a doubling of the measured value.

### This Revision

The approach outlined within the approved version of EP6 represents a significant change in statistical approach from the previous version.

The first edition (EP6-P) used the statistical “Lack of Fit” test (LoF) as the sole basis for determining linearity. With this protocol, five equally spaced concentrations were analyzed with four replicates at each level. A regression line was fit to the points, and two variance estimates were produced – the pooled variance between replicates and the variance of the five means around the regression line. The ratio of these variances was the basis for the LoF test. One problem with this approach is that very precise systems have small variance between replicates, and in the ratio LoF test, very small deviations of the means were seen as statistically significant – or nonlinear. It was also possible – though less frequent – that extremely large variance between replicates could lead to a failure to detect clinically important nonlinearity.

This polynomial method (EP6-A) is conceptually very similar to the original LoF test. In both procedures, there are two alternative statistical models (linear and nonlinear); and they are assessed relative to which is most likely to be true. However, with the polynomial approach, there is a specific parametric model for the nonlinear alternative, to better identify specific nonlinear conditions, and pure repeatability precision is included in a positive way. Better precision leads to better decisions and poor precision is screened out before linearity decisions are made. This approved statistical approach:

- estimates the magnitude of nonlinearity at every level;
- controls for unacceptable repeatability in the assessment of nonlinearity;
- provides a testable statistical model;
- can be programmed easily with widely available software.

Another key concept from the previous version has been retained, i.e., the necessity to plot the data as a first step in the evaluation of linearity. The visual examination will help the user decide the extent to which the statistical assessment can be useful for determining whether there is significant nonlinearity, and the magnitude of the nonlinearity relative to predetermined goals.

### ***A Note on Terminology***

NCCLS, as a global leader in standardization, is committed to achieving global harmonization wherever possible. Harmonization is a process of recognizing, understanding, and explaining differences while taking steps to achieve worldwide uniformity. NCCLS recognizes that medical conventions in the global metrological community have evolved differently in the United States, Europe, and elsewhere; that these differences are reflected in NCCLS, ISO, and CEN documents; and that legally required use of terms, regional usage, and different consensus timelines are all obstacles to harmonization. In light of this, NCCLS recognizes that harmonization of terms facilitates the global application of standards and is an area of immediate attention. Implementation of this policy is an evolutionary and educational process that begins with new projects and revisions of existing documents.

In keeping with NCCLS's commitment to align terminology with that of ISO, the following terms are used in EP6: *Trueness* is used in this document when referring to the closeness of the agreement between the average value from a large series of measurements and to a true value of a measurand; *Repeatability* has replaced the term *Within-run precision* where appropriate, when describing the closeness of agreement between results of successive measurements of the same measurand carried out under the same conditions of measurement; *Measurement procedure* has replaced the term *Analytical method* for a set of operations, used in the performance of particular measurements according to a given method; *Measuring range* has replaced *Reportable range* when referring to a set of values of measurands for which the error of a measuring instrument is intended to lie within specified limits; *Measurement error/Error of measurement* is used instead of *Total error* to describe the result of a measurement minus a true value of the measurand.

At this time, the working group has chosen not to replace *Analyte* with *Measurand*, (i.e., particular quantity subject to measurement) due to user nonfamiliarity and for the sake of the practicability of the guideline.

Users of EP6-A should understand, however, that the fundamental meanings of the terms are very similar, and to facilitate understanding, the terms are defined along with explanatory notes in the guideline's Definitions section.

All terms and definitions will be reviewed again for consistency with international use, and revised appropriately during the next scheduled revision of this document.

### **Standard Precautions**

Because it is often impossible to know what might be infectious, all human blood specimens are to be treated as infectious and handled according to "standard precautions." Standard precautions are new guidelines that combine the major features of "universal precautions and body substance isolation" practices. Standard precautions cover the transmission of any pathogen and thus are more comprehensive than universal precautions which are intended to apply only to transmission of blood-borne pathogens. Standard precaution and universal precaution guidelines are available from the U.S. Centers for Disease Control and Prevention (*Guideline for Isolation Precautions in Hospitals*. Infection Control and Hospital Epidemiology. CDC. 1996;Vol 17;1:53-80), (MMWR 1987;36[*suppl* 2S]2S-18S), and (MMWR 1988;37:377-382, 387-388). For specific precautions for preventing the laboratory transmission of blood-borne infection from laboratory instruments and materials and for recommendations for the management of blood-borne exposure, refer to the most current edition of [NCCLS document M29—Protection of Laboratory Workers from Occupationally Acquired Infections](#).

### **Key Words**

Allowable difference, allowable error, linearity, matrix effects, measurement error, total error, uncertainty

## **Evaluation of the Linearity of Quantitative Measurement Procedures: A Statistical Approach; Approved Guideline**

### **1 Scope**

This document presents a method to establish, verify and/or demonstrate the linear range of a quantitative measurement procedure. These methods do not identify the causes of significant nonlinearity. The method employs increasing numbers of samples for more definitive examinations of linearity. Therefore, if a failed demonstration is evaluated and it is determined that the experiment needs to be repeated, it can be done with more replicates or with fewer levels to cover a smaller range.

- This protocol is to assess linearity, isolated as much as possible from conditions of precision and trueness. It is understood that poor precision will hinder an effective assessment of linearity, so a check for poor repeatability is included.
- These experiments should use samples with a matrix appropriate to the specimens being analyzed (serum, plasma, urine, etc.)
- This protocol requires laboratories to set goals for nonlinear error. It provides basic concepts for setting such goals, but does not recommend any specific protocol.

### **2 Introduction**

#### **2.1 Purpose**

The purpose of this guideline is to describe a statistical process for determining the linearity of a quantitative measurement procedure. This primary objective is to determine the concentration(s) where a method is not linear and the extent of the nonlinearity at that level. This guideline emphasizes the necessity that each user establishes his or her requirements for linearity, or the allowable error due to nonlinearity. It also places less importance on global tests for linearity across the tested range (such as the LoF test). Global tests merely indicate that statistically significant nonlinearity exists; they do not show where that nonlinearity is, nor do they show the magnitude of the error. Linearity tests can be helpful to assess bias, which is a component of measurement error, but nonlinearity is not the only component of bias.

Users should have an understanding of their needs for measurement error, bias, random error (or imprecision), and nonspecificity (or interferences). From these they can derive a goal for linearity. In this context, NCCLS document EP6-A is a part of a series of documents that guide users through the process of method evaluation. Also see the most recent version of [NCCLS document EP21—Total Analytical Error for Clinical Laboratory Methods](#).

This document is meant to cover a broad range of situations, such as establishing the linear range of a method, which requires testing across a wide range of concentrations, then progressively narrowing into a range of acceptable linearity. It is also intended to cover situations where the linear range has been determined elsewhere (e.g., by the manufacturer), but the user wishes to verify that range in their laboratory. The same procedure is used for all scenarios, but with different numbers of concentration levels and different numbers of replicates.