

AUTO3-A
Vol. 20 No. 30

Replaces AUTO3-P
Vol. 18 No. 18

**Laboratory Automation: Communications with Automated Clinical
Laboratory Systems, Instruments, Devices, and Information Systems;
Approved Standard**

This document provides standards to facilitate accurate and timely electronic exchange of data and information between the automated laboratory elements.

A standard for global application developed through the NCCLS consensus process.



This is a preview of "AUTO3-A". [Click here to purchase the full version from the ANSI store.](#)

Acknowledgements

NCCLS gratefully acknowledges the following sponsors of this project:

American Association for Clinical Chemistry (AACC)	Labotix
American Society for Clinical Laboratory Science (ASCLS)	Medical Laboratory Automation, Inc.
A & T Corporation	Methodist Hospitals of Memphis Lab
Abbott Diagnostics	Michigan Department of Community Health
ARUP Laboratories	Microscan
Auto Lab Systems	NCSU College of Veterinary Medicine
Bayer Diagnostics	NIST/Analytical Chemistry CSTL
Beckman Coulter Corporation	National Academy of Clinical Biochemistry
Becton Dickinson and Company	Olympus, Inc.
Children's Hospital Medical Center (Cincinnati, OH)	Ortho-Clinical Diagnostics A Johnson & Johnson company
Chiron Diagnostics Corporation	Pharmacia & Upjohn
Clinical Laboratory Management Association (CLMA)	Quest Diagnostics
College of American Pathologists (CAP)	Roche Diagnostics, Inc.
Compunet Clinical Labs	SMS
CRS Robotics	Sanofi Pasteur Diagnostics
Dade Behring Inc.	SmithKline Beecham Clinical Labs
Dean Medical Center	Systemex Corporation (TOA)
Duke University	Tecan
Enterprise Analysis Corporation	Trillium GmbH
Food and Drug Administration (FDA)	Triple G Corporation
HBO & Company	UMD of New Jersey University Hospital
Hartford Hospital	University Hospitals Lab Services Foundation
Hitachi Instruments, Inc.	University of Nebraska Medical Center
Kaiser Permanente	Wuesthoff Health Systems
Konelab Corporation	Wellmont Health Systems
LAB-InterLink	York Hospital

This is a preview of "AUTO3-A". [Click here to purchase the full version from the ANSI store.](#)

Laboratory Automation: Communications with Automated Clinical Laboratory Systems, Instruments, Devices, and Information Systems; Approved Standard

Abstract

NCCLS document AUTO3-A—*Laboratory Automation: Communications with Automated Clinical Laboratory Systems, Instruments, Devices, and Information Systems; Approved Standard* provides standards to facilitate accurate and timely electronic exchange of data and information between the automated laboratory elements. This will allow and encourage scalable, open systems, and extendibility and interoperability of the automated laboratory elements. Implementation of this standard will contribute to the development of a shared vision of future clinical laboratory automation communications.

NCCLS. *Laboratory Automation: Communications with Automated Clinical Laboratory Systems, Instruments, Devices, and Information Systems; Approved Standard*. NCCLS document AUTO3-A (ISBN 1-56238-428-7). NCCLS, 940 West Valley Road, Suite 1400, Wayne, Pennsylvania 19087-1898 USA, 2000

THE NCCLS consensus process, which is the mechanism for moving a document through two or more levels of review by the healthcare community, is an ongoing process. Users should expect revised editions of any given document. Because rapid changes in technology may affect the procedures, methods, and protocols in a standard or guideline, users should replace outdated editions with the current editions of NCCLS documents. Current editions are listed in the *NCCLS Catalog*, which is distributed to member organizations, and to nonmembers on request. If your organization is not a member and would like to become one, and to request a copy of the *NCCLS Catalog*, contact the NCCLS Executive Offices. Telephone: 610.688.0100; Fax: 610.688.0700; E-Mail: exoffice@nccls.org; Website: www.nccls.org

AUTO3-A
ISBN 1-56238-428-7
ISSN 0273-3099

Laboratory Automation: Communications with Automated Clinical
Laboratory Systems, Instruments, Devices, and Information Systems;
Approved Standard

Volume 20 Number 30

Charles D. Hawker, Ph.D., Chairholder
Andrzej J. Knafel, Ph.D., Vice-Chairholder
John F. Boje
John W. Elling, Ph.D.
Hendrik J. Keesom
Brad Kowalski
Mary Lock
Michael Newcomb
Lori Mauthe



This publication is protected by copyright. No part of it may be reproduced, stored in a retrieval system, or transmitted in any form or by any means (electronic, mechanical, photocopying, recording, or otherwise) without written permission from NCCLS, except as stated below.

NCCLS hereby grants permission to reproduce limited portions of this publication for use in laboratory procedure manuals at a single site, for interlibrary loan, or for use in educational programs provided that multiple copies of such reproduction shall include the following notice, be distributed without charge, and, in no event, contain more than 20% of the document's text.

Reproduced with permission, from NCCLS publication AUTO3-A—*Laboratory Automation: Communications with Automated Clinical Laboratory Systems, Instruments, Devices, and Information Systems; Approved Standard* (ISBN 1-56238-428-7). Copies of the current edition may be obtained from NCCLS, 940 West Valley Road, Suite 1400, Wayne, Pennsylvania 19087-1898, USA.

Permission to reproduce or otherwise use the text of this document to an extent that exceeds the exemptions granted here or under the Copyright Law must be obtained from NCCLS by written request. To request such permission, address inquiries to the Executive Director, NCCLS, 940 West Valley Road, Suite 1400, Wayne, Pennsylvania 19087-1898, USA.

HL7 Permission

Permission to use portions of the standard, Health Level Seven – *An Application Protocol for Electronic Exchange in Healthcare Environments, Version 2.4*, has been granted by Health Level Seven, Inc. The current standard may be obtained from Health Level Seven, Inc., 3300 Washtenaw Avenue, Suite 227, Ann Arbor, MI 48104-4261 or via www.HL7.org.

Copyright ©2000. The National Committee for Clinical Laboratory Standards.

Suggested Citation

(NCCLS. *Laboratory Automation: Communications with Automated Clinical Laboratory Systems, Instruments, Devices, and Information Systems; Approved Standard*. NCCLS document AUTO3-A [ISBN 1-56238-428-7]. NCCLS, 940 West Valley Road, Suite 1400, Wayne, PA 19087-1898 USA, 2000.)

Proposed Standard

December 1998

Approved Standard

December 2000

ISBN 1-56238-428-7

ISSN 0273-3099

Committee Membership

Area Committee on Automation

Rodney S. Markin, M.D., Ph.D.
Chairholder

University of Nebraska Medical Center
Omaha, Nebraska

Paul J. Mountain, M.Sc., M.T.(ASCP)
Vice-Chairholder

MDS Laboratories
Etobicoke, Ontario Canada

Subcommittee on Communications with Automated Systems

Charles D. Hawker, Ph.D.
Chairholder

ARUP Laboratories
Salt Lake City, Utah

Andrzej J. Knafel, Ph.D.
Vice-Chairholder

Roche Diagnostics Corporation
Rotkreuz, Switzerland

John F. Boje

LAB-InterLink, Inc.
Omaha, Nebraska

James Callaghan

FDA Ctr. for Devices/Rad. Health
Gaithersburg, Maryland

George S. Cembrowski, M.D., Ph.D.

University of Alberta Hospital Site
Edmonton, Alberta Canada

Robert L. Jones, Ph.D.

Centers for Disease Control and Prevention
Atlanta, Georgia

Hiromi Kataoka, M.T.(JMLT)

Kochi Medical School
Kochi, Japan

Hendrik J. Keesom

Ortho-Clinical Diagnostics
Rochester, New York

Brad Kowalski

Marshfield Laboratories
Marshfield, Wisconsin

Advisors

Hunter Bagwell

Roche Diagnostics, Inc.
Indianapolis, Indiana

William A. Berger

University of Iowa Hygienic Laboratory
Iowa City, Iowa

W. Gregory Cooper

Bio-Rad Laboratories
Irvine, California

Advisors (continued)

Marcio Carvalho Correa	Sociedade Brasileira de Analises Clinicas Rio De Janeiro-RJ Brazil
Randall R. Davis	Dade Behring Inc. Newark, Delaware
Louis J. Dunka, Jr., Ph.D.	Life Scan, Inc. Milpitas, California
John W. Elling, Ph.D.	Bio Reason Santa Fe, New Mexico
Greg Forbes	Triple G Corporation Markham, Ontario Canada
Arden W. Forrey, Jr., Ph.D.	University of Washington Seattle, Washington
Yasushi Isami	Sysmex Corporation Kobe, Japan
Carol L. Lemons	North Carolina College of Veterinary Medicine Raleigh, North Carolina
Mary Lock	BD Immunocytometry Systems San Jose, California
John D. Low	Instrumentation Laboratory Lexington, Massachusetts
Lori Mauthe	Scottsdale, Arizona
Kevin Ravenhill	MDS AutoLab Systems Whitby, Ontario Canada
Dr. Arturo M. Terres Speziale	Carpermor D.F. Del Cuauhtemoc, Mexico
Christoph F. Tschopp	A. I. Scientific Scarborough, Australia
Jacob Zikherman	Bayer Corporation Tarrytown, New York
Charlie F. Galanaugh <i>Special Liaison</i>	Becton Dickinson and Company (Retired) West Milford, NJ
Marc R. Schlank, M.T.(ASCP), M.S. <i>Staff Liaison</i>	NCCLS Wayne, Pennsylvania

Patrice E. Polgar
Editor

NCCLS
Wayne, Pennsylvania

Donna M. Wilhelm
Assistant Editor

NCCLS
Wayne, Pennsylvania

Active Membership (as of 1 October 2000)

Sustaining Members

Abbott Laboratories
 American Association for
 Clinical Chemistry
 Bayer Corporation
 Beckman Coulter, Inc.
 BD and Company
 bioMérieux, Inc.
 College of American Pathologists
 Dade Behring Inc.
 Nippon Becton Dickinson Co., Ltd.
 Ortho-Clinical Diagnostics, Inc.
 Pfizer Inc
 Roche Diagnostics, Inc.

Professional Members

American Academy of Family
 Physicians
 American Association of Blood
 Banks
 American Association for
 Clinical Chemistry
 American Association for
 Respiratory Care
 American Chemical Society
 American Medical Technologists
 American Public Health Association
 American Society for Clinical
 Laboratory Science
 American Society of Hematology
 American Society for Microbiology
 American Society of
 Parasitologists, Inc.
 American Type Culture
 Collection, Inc.
 Asociación Española Primera de
 Socorros (Uruguay)
 Asociacion Mexicana de
 Bioquímica Clínica A.C.
 Assn. of Public Health Laboratories
 Assoc. Micro. Clinici Italiani-
 A.M.C.L.I.
 Australasian Association of
 Clinical Biochemists
 British Society for Antimicrobial
 Chemotherapy
 Canadian Society for Medical
 Laboratory Science—Société
 Canadienne de Science de
 Laboratoire Médical
 Canadian Society of Clinical
 Chemists

Clinical Laboratory Management
 Association
 College of American Pathologists
 College of Medical Laboratory
 Technologists of Ontario
 College of Physicians and
 Surgeons of Saskatchewan
 Commission on Office Laboratory
 Accreditation
 Fundación Bioquímica Argentina
 International Association of Medical
 Laboratory Technologists
 International Council for
 Standardization in Haematology
 International Federation of
 Clinical Chemistry
 Italian Society of Clinical
 Biochemistry
 Japan Society of Clinical Chemistry
 Japanese Association of Medical
 Technologists (Tokyo)
 Japanese Committee for Clinical
 Laboratory Standards
 Joint Commission on Accreditation
 of Healthcare Organizations
 National Academy of Clinical
 Biochemistry
 National Society for
 Histotechnology, Inc.
 Ontario Medical Association
 Laboratory Proficiency Testing
 Program
 RCPA Quality Assurance Programs
 PTY Limited
 Sociedade Brasileira de Analises
 Clinicas
 Sociedade Brasileira de
 Patologia Clínica
 Sociedad Espanola de Quimica
 Clínica

Government Members

Armed Forces Institute of Pathology
 BC Centre for Disease Control
 Centers for Disease Control and
 Prevention
 Chinese Committee for Clinical
 Laboratory Standards
 Commonwealth of Pennsylvania
 Bureau of Laboratories
 Department of Veterans Affairs
 Deutsches Institut für Normung
 (DIN)

FDA Center for Devices and
 Radiological Health
 FDA Center for Veterinary
 Medicine
 FDA Division of Anti-Infective
 Drug Products
 Health Care Financing
 Administration/CLIA Program
 Health Care Financing
 Administration
 Iowa State Hygienic Laboratory
 Massachusetts Department of
 Public Health Laboratories
 National Association of Testing
 Authorities – Australia
 National Center of Infectious
 and Parasitic Diseases (Bulgaria)
 National Institute of Standards
 and Technology
 Ohio Department of Health
 Ontario Ministry of Health
 Saskatchewan Health-Provincial
 Laboratory
 Scientific Institute of Public Health;
 Belgium Ministry of Social
 Affairs, Public Health and the
 Environment
 South African Institute for Medical
 Research
 Swedish Institute for Infectious
 Disease Control
 Thailand Department of Medical
 Sciences

Industry Members

AB Biodisk
 Abbott Laboratories
 Abbott Laboratories, MediSense
 Products
 Accumetrics, Inc.
 Amersham Pharmacia Biotech
 Ammirati Regulatory Consulting
 Assessor
 AstraZeneca
 Aventis
 Avocet Medical, Inc.
 Bayer Corporation – Elkhart, IN
 Bayer Corporation – Middletown,
 VA
 Bayer Corporation – Tarrytown, NY
 Bayer Corporation – West Haven,
 CT
 Bayer Medical Ltd.
 BD

BD Biosciences – San Jose, CA
 BD Biosciences – Sparks, MD
 BD Consumer Products
 BD Italia S.P.A.
 BD VACUTAINER Systems
 Beckman Coulter, Inc.
 Beckman Coulter, Inc. Primary Care Diagnostics
 Beckman Coulter K.K. (Japan)
 Bio-Development SRL
 Bio-Inova Life Sciences International
 Biolog, Inc.
 bioMérieux, Inc.
 Biometrology Consultants
 Bio-Rad Laboratories, Inc.
 Biotest AG
 Bristol-Myers Squibb Company
 Canadian Reference Laboratory Ltd.
 Capital Management Consulting, Inc.
 CASCO•NERL Diagnostics
 Checkpoint Development Inc.
 Clinical Design Group Inc.
 COBE Laboratories, Inc.
 Combact Diagnostic Systems Ltd.
 Community Medical Center (NJ)
 Control Lab (Brazil)
 Copan Diagnostics Inc.
 Cosmetic Ingredient Review
 Cubist Pharmaceuticals
 Cytometrics, Inc.
 Dade Behring Inc. - Deerfield, IL
 Dade Behring Inc. - Glasgow, DE
 Dade Behring Inc. - Marburg, Germany
 Dade Behring Inc. - Sacramento, CA
 Dade Behring Inc. - San Jose, CA
 DAKO A/S
 Diagnostic Products Corporation
 Eiken Chemical Company, Ltd.
 Enterprise Analysis Corporation
 Fort Dodge Animal Health
 Gen-Probe
 Glaxo-Wellcome, Inc.
 Greiner Mediatech, Inc.
 Health Systems Concepts, Inc.
 Helena Laboratories
 Home Diagnostics, Inc.
 Hycor Biomedical Inc.
 I-STAT Corporation
 Instrumentation Laboratory
 International Technidyne Corporation
 Kendall Sherwood-Davis & Geck
 Labtest Diagnostica S.A.

LifeScan, Inc. (a Johnson & Johnson Company)
 Lilly Research Laboratories
 Medical Automation Systems
 Medical Device Consultants, Inc.
 Medical Laboratory Automation Inc.
 Medtronic, Inc.
 Merck & Company, Inc.
 mvi Sciences (MA)
 Nabi
 Neometrics is.
 Nichols Institute Diagnostics (Div. of Quest Diagnostics, Inc.)
 Nissui Pharmaceutical Co., Ltd.
 Nippon Becton Dickinson Co., Ltd.
 Norfolk Associates, Inc.
 Ortho-Clinical Diagnostics, Inc. (Raritan, NJ)
 Ortho-Clinical Diagnostics, Inc. (Rochester, NY)
 Oxoid Inc.
 Pfizer Global R & D
 Pfizer Inc
 Pharmacia & Upjohn
 Premier Inc.
 Procter & Gamble
 Pharmaceuticals, Inc.
 The Product Development Group
 Quest Diagnostics Incorporated
 Quintiles, Inc.
 Radiometer America, Inc.
 Radiometer Medical A/S
 David G. Rhoads Associates, Inc.
 Roche Diagnostics GmbH
 Roche Diagnostics, Inc.
 Roche Laboratories (Div. Hoffmann-La Roche Inc.)
 The R.W. Johnson
 Pharmaceutical Research Institute
 Sanofi Diagnostics Pasteur
 Sarstedt, Inc.
 SARL Laboratoire Carron (France)
 Schering Corporation
 Schleicher & Schuell, Inc.
 Second Opinion
 SenDx Medical, Inc.
 Showa Yakuhin Kako Company, Ltd.
 SmithKline Beecham, S.A.
 Streck Laboratories, Inc.
 Sysmex Corporation (Japan)
 Sysmex Corporation
 (Long Grove, IL)
 The Toledo Hospital (OH)
 Trek Diagnostic Systems, Inc.
 Vetoquinol S.A.
 Visible Genetics, Inc.

Vysis, Inc.
 Wallac Oy
 Wyeth-Ayerst
 Xyletech Systems, Inc.
 YD Consultant
 Yeongdong Pharmaceutical Corporation

Trade Associations

AdvaMed
 Association of Medical Diagnostic Manufacturers
 Japan Association Clinical Reagents Ind. (Tokyo, Japan)
 Medical Industry Association of Australia

Associate Active Members

67th CSH Wuerzburg, GE (NY)
 121st General Hospital (CA)
 Academisch Ziekenhuis-VUB (Belgium)
 Acadiana Medical Laboratories, LTD (LA)
 Advocate Laboratories (IL)
 The Aga Khan Hospital & Medical College, Karachi (Pakistan)
 Albany Medical Center Hospital (NY)
 Albemarle Hospital (NC)
 Allegheny General Hospital (PA)
 Allegheny University of the Health Sciences (PA)
 Allina Laboratories (MN)
 Alton Ochsner Medical Foundation (LA)
 American Medical Laboratories (VA)
 Anzac House (Australia)
 Arkansas Department of Health
 Armed Forces Research Institute of Medical Science (APO, AP)
 Asan Medical Center (Korea)
 Associated Regional & University Pathologists (UT)
 Aurora Consolidated Laboratories (WI)
 Bay Medical Center (MI)
 Baystate Medical Center (MA)
 Boulder Community Hospital (CO)
 Brantford General Hospital (Brantford, ON, Canada)
 Brasileiro De Promocao (Brazil)
 Brookdale Hospital Medical Center (NY)
 Brooke Army Medical Center (TX)

Brooks Air Force Base (TX)	Dynacare Memorial Hermann Laboratory Services (TX)	Jacobi Medical Center (NY)
Broward General Medical Center (FL)	E.A. Conway Medical Center (LA)	Jersey Shore Medical Center (NJ)
Calgary Laboratory Services	East Side Clinical Laboratory (RI)	John C. Lincoln Hospital (AZ)
Carilion Consolidated Laboratory (VA)	Elyria Memorial Hospital (OH)	John Peter Smith Hospital (TX)
CB Healthcare Complex (Sydney, NS, Canada)	Emory University Hospital (GA)	John Randolph Hospital (VA)
Central Kansas Medical Center	Esoterix Center for Infectious Disease (TX)	Johns Hopkins Medical Institutions (MD)
Centralized Laboratory Services (NY)	Fairfax Hospital (VA)	Kaiser Permanente (CA)
Centro Diagnostico Italiano (Milano, Italy)	Fairview-University Medical Center (MN)	Kaiser Permanente (MD)
Champlain Valley Physicians Hospital (NY)	Foothills Hospital (Calgary, AB, Canada)	Kaiser Permanente (NC)
Chang Gung Memorial Hospital (Taiwan)	Fort St. John General Hospital (Fort St. John, BC, Canada)	Kantousspital (Switzerland)
Children's Hospital (LA)	Fox Chase Cancer Center (PA)	Keller Army Community Hospital (NY)
Children's Hospital (NE)	Franklin Square Hospital Center (MD)	Kern Medical Center (CA)
Children's Hospital & Clinics (MN)	Fresenius Medical Care/Life Chem (NJ)	King Fahad National Guard Hospital (Saudi Arabia)
Children's Hospital King's Daughters (VA)	Fresno Community Hospital and Medical Center	Kings County Hospital Center (NY)
Children's Hospital Medical Center (Akron, OH)	Gambro Healthcare Laboratory (FL)	Klinicni Center (Slovenia)
Children's Hospital of Philadelphia (PA)	GDS Technology, Inc (IN)	LabCorp (NC)
Clarian Health-Methodist Hospital (IN)	Geisinger Medical Center (PA)	Laboratoire de Santé Publique du Quebec (Canada)
Clendo Lab (Puerto Rico)	Grady Memorial Hospital (GA)	Laboratório Fleury S/C Ltda. (Brazil)
CLSI Laboratories (PA)	Guthrie Clinic Laboratories (PA)	Laboratory Corporation of America (MO)
Columbus County Hospital (NC)	Harris Methodist Fort Worth (TX)	LAC and USC Healthcare Network (CA)
Commonwealth of Kentucky	Harris Methodist Northwest (TX)	Lakeland Regional Medical Center (FL)
Commonwealth of Virginia (DCLS)	Hartford Hospital (CT)	Lancaster General Hospital (PA)
CompuNet Clinical Laboratories (OH)	Health Alliance Laboratory (OH)	Langley Air Force Base (VA)
Consolidated Laboratory Services (CA)	Health Network Lab (PA)	LeBonheur Children's Medical Center (TN)
Covance Central Laboratory Services (IN)	Health Sciences Centre (Winnipeg, MB, Canada)	Lewis-Gale Medical Center (VA)
Danish Veterinary Laboratory (Copenhagen, Denmark)	Heartland Health System (MO)	Libero Instituto Univ. Campus BioMedico (Italy)
Danville Regional Medical Center (VA)	Hinsdale Hospital (IL)	Licking Memorial Hospital (OH)
Deaconess Hospital (MO)	Hoag Memorial Hospital Presbyterian (CA)	Louisiana State University Medical Center
Dean Medical Center (WI)	Holmes Regional Medical Center (FL)	Magee Womens Hospital (PA)
Delaware Public Health Laboratory	Holy Spirit Hospital (PA)	Magnolia Regional Health Center (MS)
Department of Health & Community Services (New Brunswick, Canada)	Holzer Medical Center (OH)	Martin Luther King/Drew Medical Center (CA)
Detroit Health Department (MI)	Hospital for Sick Children (Toronto, ON, Canada)	Massachusetts General Hospital (Microbiology Laboratory)
Diagnostic Laboratory Services, Inc. (HI)	Hospital Israelita Albert Einstein (Brazil)	Massachusetts General Hospital (Pathology Laboratory)
Duke University Medical Center (NC)	Hotel Dieu Hospital (Windsor, ON, Canada)	Mayo Clinic Scottsdale (AZ)
Durham Regional Hospital (NC)	Huddinge University Hospital (Sweden)	MDS Metro Laboratory Services (Burnaby, BC, Canada)
Duzen Laboratories (Turkey)	Hurley Medical Center (MI)	Medical Center of Southern Indiana
Dynacare Laboratories - Eastern Region (Ottawa, ON, Canada)	Indiana State Board of Health	Medical College of Virginia Hospital
	Indiana University	Medicare/Medicaid Certification, State of North Carolina
	Instituto Scientifico HS. Raffaele (Italy)	Memorial Hospital (CO)
	International Health Management Associates, Inc. (IL)	

Memorial Medical Center (Napoleon Ave., New Orleans, LA)	Our Lady of Lourdes Hospital (NJ)	St. Mary's Hospital & Medical Center (CO)
Memorial Medical Center (N. Jefferson Davis Pkwy., New Orleans, LA)	Our Lady of the Resurrection Medical Center (IL)	Ste. Justine Hospital (Montreal, PQ, Canada)
Memorial Medical Center (IL)	Pathology and Cytology Laboratories, Inc. (KY)	Salina Regional Health Center (KS)
Mercy Health System (PA)	Pathology Associates Laboratories (CA)	San Francisco General Hospital (CA)
Mercy Hospital (NC)	The Permanente Medical Group (CA)	Santa Cabrini Hospital (Montreal, PQ Canada)
Mercy Medical Center Des Moines (IA)	Pocono Hospital (PA)	Santa Clara Valley Medical Center (CA)
Mescalero Indian Hospital (NM)	Presbyterian Hospital (NC)	Seoul Nat'l University Hospital (Korea)
Methodist Hospital (TX)	Presbyterian Hospital of Dallas (TX)	Shanghai Center for the Clinical Laboratory (China)
Methodist Hospitals of Memphis (TN)	Providence Health System (OR)	Shands Healthcare (FL)
Michigan Department of Community Health	Providence Seattle Medical Center (WA)	SmithKline Beecham Clinical Laboratories (GA)
Mississippi Baptist Medical Center	Queen Elizabeth Hospital (Prince Edward Island, Canada)	South Bend Medical Foundation (IN)
Monte Tabor – Centro Italo - Brazileiro de Promocao (Brazil)	Queensland Health Pathology Services (Australia)	Southern California Permanente Medical Group
Montreal Children's Hospital (Canada)	Quest Diagnostics, Incorporated (AZ)	South Western Area Pathology Service (Australia)
Montreal General Hospital (Canada)	Quest Diagnostics Incorporated (CA)	Speare Memorial Hospital (NH)
Morton Plant Mease Health Care (FL)	Quintiles Laboratories, Ltd. (GA)	Speciality Laboratories, Inc. (CA)
Mount Sinai Hospital (NY)	Regions Hospital	Stanford Hospital and Clinics (CA)
Mount Sinai Medical Center (FL)	Research Medical Center (MO)	State of Washington Department of Health
MRL Reference Laboratory (CA)	Rex Healthcare (NC)	Stormont-Vail Regional Medical Center (KS)
National University Hospital (Singapore)	Rhode Island Department of Health Laboratories	Sun Health-Boswell Hospital (AZ)
Naval Surface Warfare Center (IN)	Riyadh Armed Forces Hospital (Saudi Arabia)	Sunrise Hospital and Medical Center (NV)
New Britain General Hospital (CT)	Royal Columbian Hospital (New Westminster, BC, Canada)	Touro Infirmary (LA)
New England Fertility Institute (CT)	Saint Mary's Regional Medical Center (NV)	Tri-City Medical Center (CA)
New England Medical Center Hospital (MA)	St. Alexius Medical Center (ND)	Trident Regional Medical Center (SC)
New York Hospital Medical Center of Queens	St. Anthony Hospital (CO)	Tripler Army Medical Center (HI)
New York State Department of Health	St. Barnabas Medical Center (NJ)	Truman Medical Center (MO)
NorDx (ME)	St. Boniface General Hospital (Winnipeg, Canada)	Tulane Medical Center Hospital & Clinic (LA)
North Carolina Laboratory of Public Health	St. Elizabeth Hospital (NJ)	UCSF Medical Center (CA)
North Mississippi Medical Center	St. John Hospital and Medical Center (MI)	UNC Hospitals (NC)
North Shore – Long Island Jewish Health System Laboratories (NY)	St. John Regional Hospital (St. John, NB, Canada)	Unilab Clinical Laboratories (CA)
Northridge Hospital Medical Center (CA)	St. Joseph Hospital (NE)	University Hospital (Gent) (Belgium)
Northwestern Memorial Hospital (IL)	St. Joseph Medical Center (MD)	University Hospital (TX)
Ohio Valley Medical Center (WV)	St. Joseph Mercy – Oakland (MI)	The University Hospitals (OK)
Olin E. Teague Medical Center (TX)	St. Joseph's Hospital – Marshfield Clinic (WI)	University of Alberta Hospitals (Canada)
O.L. Vrouwziekenhuis (Belgium)	St. Luke's Hospital (PA)	University of Chicago Hospitals (IL)
Ordre professionnel des technologistes médicaux du Québec	St. Luke's Regional Medical Center (IA)	University of Florida
Ospedali Riuniti (Italy)	St. Mary Medical Center (IN)	University of Medicine & Dentistry, NJ University Hospital
The Ottawa Hospital (Ottawa, ON, Canada)	St. Mary of the Plains Hospital (TX)	University of the Ryukyus (Japan)
		University of Texas M.D. Anderson Cancer Center

University of Virginia Medical Center
 University of Washington
 UPMC Bedford Memorial (PA)
 UZ-KUL Medical Center (Belgium)
 VA (Dayton) Medical Center (OH)
 VA (Denver) Medical Center (CO)
 VA (Martinez) Medical Center (CA)
 VA (San Diego) Medical Center (CA)
 VA (Tuskegee) Medical Center (AL)

VA Outpatient Clinic (OH)
 Vejle Hospital (Denmark)
 Virginia Department of Health
 Viridae Clinical Sciences, Inc. (Vancouver, BC, Canada)
 Warde Medical Laboratory (MI)
 Washoe Medical Center (NV)
 Watson Clinic (FL)
 Wilford Hall Medical Center (TX)
 William Beaumont Hospital (MI)
 Williamsburg Community Hospital (VA)

Winchester Hospital (MA)
 Winn Army Community Hospital (GA)
 Wishard Memorial Hospital (IN)
 Womack Army Medical Center (NC)
 Yan Chai Hospital (P.R. China)
 Yonsei University College of Medicine (Korea)
 York Hospital (PA)
 Zale Lipshy University Hospital (TX)

OFFICERS

F. Alan Andersen, Ph.D.,
 President
 Cosmetic Ingredient Review

Donna M. Meyer, Ph.D.,
 President Elect
 CHRISTUS Health

Robert F. Moran, Ph.D., FCCM,
 FAIC
 Secretary
 mvi Sciences

Gerald A. Hoeltge, M.D.
 Treasurer
 The Cleveland Clinic Foundation

William F. Koch, Ph.D.,
 Immediate Past President
 National Institute of Standards and Technology

John V. Bergen, Ph.D.,
 Executive Director

BOARD OF DIRECTORS

Susan Blonshine, RRT, RPFT,
 FAARC
 TechEd

Kurt H. Davis, FCSMLS, CAE
 Canadian Society for Medical Laboratory Science

Robert L. Habig, Ph.D.
 Cytometrics, Inc.

Thomas L. Hearn, Ph.D.
 Centers for Disease Control and Prevention

Elizabeth D. Jacobson, Ph.D.
 Food and Drug Administration

Carolyn D. Jones, J.D., M.P.H.
 Health Industry Manufacturers Association

Tadashi Kawai, M.D., Ph.D.
 International Clinical Pathology Center

J. Stephen Kroger, M.D., FACP
 COLA

Barbara G. Painter, Ph.D.
 Bayer Corporation

Emil Voelkert, Ph.D.
 Roche Diagnostics GmbH

Ann M. Willey, Ph.D., J.D.
 New York State Department of Health

Judith A. Yost, M.A., M.T.(ASCP)
 Health Care Financing Administration

Contents

Abstract.....i

Committee Membership.....v

Active Membership.....ix

Matrix of NCCLS Laboratory Automation Standards..... xviii

Preface to Laboratory Automation Standards.....xxi

Foreword.....xxv

1 Introduction.....1

 1.1 Scope.....1

 1.2 Limitations.....1

 1.3 Terminology.....1

2 Elements of an Automation System.....2

 2.1 Information Systems.....2

 2.2 Laboratory Automation Systems.....3

 2.3 Analytical Instruments.....3

 2.4 Process Instruments.....3

3 Definitions.....4

4 Laboratory Automation Architectures/Models.....14

 4.1 Overview of Architectures.....14

 4.2 Functional Control Model.....14

 4.3 Information Among Elements.....15

5 Communication Standard.....18

 5.1 Introduction and Overview.....18

 5.2 Relationship Among Existing Communications Standards and the
 NCCLS Standard.....18

6 HL7 Communication Standard for Laboratory Automation.....22

 6.1 Background and Introduction.....22

 6.2 Trigger Events and Message Definitions.....35

 6.3 Message Segments.....40

 6.4 Notes Regarding Usage.....72

 6.5 Outstanding Issues.....77

7 Low-Level Protocol Considerations.....77

 7.1 Requirements to Low-Level Protocol.....77

 7.2 Recommendations for the Low-Level Protocol.....78

8 Implementation Considerations.....80

 8.1 ADT, Patient, Episode-of-Care Functions.....80

 8.2 Order Entry Function.....81

 8.3 Results Function.....82

 8.4 Query Function.....83

 8.5 Network Management Function.....83

Contents (Continued)

8.6	Analyzer/Instrument Functions.....	83
8.7	Timing and Throughput Considerations	84
8.8	Recommended Character Set Support	84
8.9	Example Transactions.....	84
Appendix A. Laboratory Automation Architectures/Models.....		85
Appendix B. Survey of Plans for Automation in 1999 – 2000		88
References.....		89
Summary of Comments and Subcommittee Responses.....		90
Summary of Delegate Voting Comments and Subcommittee Responses		105
Related NCCLS Publications.....		106

Matrix of NCCLS Laboratory Automation Standards

The laboratory automation standards documents, AUTO1, AUTO2, AUTO3, AUTO4, and AUTO5 are interdependent with respect to their implementation in automated laboratory systems. The matrix describes the engineering relationships between the standards elements in each of the five documents. This matrix is provided so that designers and engineers, as well as users and customers, understand the relationships between the different standards' components in an automated system. The matrix format allows the users of one document to easily identify other standard elements, which relate to the standard elements in the document or documents from which they may be working, to design a system correctly.

How to Read the Matrix (See matrix on the next page.)

The numbers listed on the horizontal (X) and vertical (Y) axes contain multiple-digit numbers (e.g., (1)5.4, (5)5.4.1.3).

The 'first digit' (in parentheses) represents one of the five automation documents (e.g., (1)5.4 is from AUTO1; (5)5.4.1.3 is from AUTO5).

The 'remaining digits' represent the specific section of that document.

The symbol XX represents the direct 'engineering relationship' between two sections.

The symbol ## represents the section's 'self'; when it has been lined up with itself on the other axis.

Preface to Laboratory Automation Standards

Background

In late 1996, NCCLS agreed to undertake the complex and challenging task of managing an effort to develop standards for clinical laboratory automation, based upon the urgent request of many leading individuals and institutions in the field. Standardization was needed to overcome difficulties and unnecessary costs incurred by laboratories and manufacturers in their efforts to integrate and simplify laboratory functions using technology.

As a result of discussions at an annual meeting of the International Conference on Automation, Robotics, and Artificial Intelligence Applied to Analytical Chemistry and Laboratory Medicine (ICAR) in 1994, an interested group of individuals had formed the Clinical Testing Automation Standards Steering Committee (CTASSC). The CTASSC approached NCCLS's leadership seeking collaboration, and believing that the desired standards could best be developed utilizing the unique voluntary consensus process, resources, and expertise of NCCLS and its member organizations. It was expected that cooperation would also be necessary with other complementary standards-developing bodies, such as ASTM, IEEE, and HL7.

The original shared vision was to take advantage of market forces within the industry and of the benefits of implementing prospective standards in the context of market forces and industry support so that customers (laboratories) and vendors could enjoy products that function together, and buyers and suppliers could agree on a format for laboratory automation systems.

NCCLS accepted the challenge and committed to the following:

- NCCLS's **voluntary consensus process** would be utilized to ensure balance, fairness, and broad review of documents by all institutions affected by the effort.
- The project would be **global** in scope and participation.
- Sources and mechanisms for **funding** would be identified to ensure that the projects would be given high priority to achieve timely completion.

NCCLS surveyed the interest of all institutions likely to be affected by the proposed standards effort, and confirmed high interest in providing both expertise and financial support. NCCLS presented the proposal at several meetings in the United States, Japan, and Europe to increase awareness of the activity and to invite broad, global participation. Based upon favorable response to the proposals, the NCCLS Board of Directors authorized the creation of a new Area Committee on Automation, chaired by Dr. Rodney S. Markin, with Mr. Paul S. Mountain serving as its vice chairholder.

Mission Statement

The mission of the Area Committee on Automation is:

“...to identify the need for, set priorities for, and manage and coordinate the development of compatible standards and guidelines that address, in a prospective manner, the design and integration of automated clinical laboratory systems worldwide. In addition, the area committee will foster communication of the issues and developments worldwide.”

Preface to Laboratory Automation Standards (Continued)

Subcommittee Activities

Based upon the recommendations of the new area committee, the Board authorized establishment of five subcommittees to manage the development of the following documents:

- **AUTO1—*Specimen Container/Specimen Carrier*** contains standards for the design and manufacture of specimen containers and carriers used for collecting and processing samples, such as blood and urine, for testing on laboratory automation systems.
- **AUTO2—*Bar Codes for Specimen Container Identification*** provides specifications for linear bar codes on specimen containers for use on laboratory automation systems.
- **AUTO3—*Communications with Automated Systems*** facilitates accurate and timely electronic exchange of data and information among automated instruments, laboratory automation systems, and other information systems.
- **AUTO4—*Systems Operational Requirements, Characteristics, and Informational Elements*** provides standards of interest to operators for display of system status information such as specimen location, reagent supply, and warnings and alerts to support laboratory automation operations.
- **AUTO5—*Electromechanical Interfaces*** provides guidance for the standardization of electromechanical interfaces between instruments and/or specimen processing and handling devices and automation systems in the automated laboratory.

The five subcommittees began their efforts in the spring of 1997, with goals to develop proposed standards suitable for publication and review by the end of 1999 consistent with the formal NCCLS consensus process, and to advance them to the approved consensus stage in 2000.

Validation of Designs, Systems, and Software

The five laboratory automation standards are tools to help in the design, development, and implementation of Laboratory Automation Systems (LAS) for the clinical laboratory. Each standard may be used fully or in part, whether or not the intent is to design a completely automated or semiautomated system. These standards provide specifications that can be adhered to and verified during various phases of development for each LAS project. Adherence to standards alone does not ensure valid system design. Design validation confirms that the medical devices (LAS) meet user needs and intended use. Software validation is also a required component of the design validation of a medical device.^a Also refer to NCCLS document GP19—*Laboratory Instruments and Data Management Systems: Design of Software User Interfaces and End-User Software Systems Validation, Operation, and Monitoring*.

Attributes of Standards for Laboratory Automation Systems

It was agreed by the Area Committee on Automation that all of the laboratory automation system standards should share the following attributes:

^a A good source of information on these and related subjects, plus other medical device regulations can be found on FDA/CDRH web pages: <http://www.fda.gov/cdrh/>.

Preface to Laboratory Automation Standards (Continued)

- **Prescriptive** – Essential requirements should be prescriptive, and should define only those features essential for compatibility of instruments, devices, and laboratory automation systems.
- **Prospective** – Standards should describe the desired and necessary attributes which will enable and enhance the connectivity of laboratory automation system components in the future; the creation of a laboratory automation system from components should not be constrained by obsolete or inadequate technology which may be in current use.
- **Inclusive** – Current technology with widespread use should not be excluded unless it impedes connectivity; in some instances, a future date for discontinuation of a technology may be recommended to encourage upgrades, providing sufficient time for interested laboratories or suppliers to comply with new requirements.
- **Explanatory** – In cases where exclusions are recommended that are not obvious, or where consensus is not achieved, the documents should include a brief rationale and, possibly, a description of opposing viewpoints.
- **Differentiating** – In view of the complexity of the tasks, documents should differentiate between imperative prescriptions ("must" verbal forms) and discretionary recommendations ("should" verbal forms).
- **Enabling of Innovation** – The concept of "prescriptive, essential requirements" should be employed to ensure that performance requirements rather than design specifications are utilized to the extent possible.
- **Consistent** – Each document should be written to be "self-sufficient" with respect to the scope of its individual effort. The five documents are interrelated and interdependent, and presented in a consistent style using cross-references and a common glossary of terms (definitions) giving the appearance of a collection of documents.

The five interrelated automation standards are a system of related documents that are available separately or packaged in a manner similar to NCCLS "specialty collections."

The clinical laboratory automation standards effort has attempted to engage the broadest possible worldwide representation in committee deliberations. Consequently, it was reasonable to expect that controversies existed and issues remained unresolved at the time of publication of the initial proposed-level documents. A mechanism for resolving such controversies through the subcommittees and the Area Committee on Automation was employed during the review and comment process.

The NCCLS voluntary consensus process is dependent upon broad distribution of documents for review and comment and upon the expertise of reviewers worldwide whose comments add value to the effort. At the end of the comment period, each subcommittee was obligated to review all comments and to respond in writing to all which are substantive. Where appropriate, modifications were made to the respective document, and all comments, along with the subcommittee's responses, are included in the Summary of Comments and Committee Responses at the end of each document.

Preface to Laboratory Automation Standards (Continued)

Special Recognition of Global Participation

The NCCLS Board of Directors wishes to give special recognition and thanks to several organizations which have taken leadership roles in the development of these standards, including the Japanese Committee for Clinical Laboratory Standards (JCCLS), the Japanese Society for Clinical Chemistry (JSCC) and the International Federation of Clinical Chemistry (IFCC). These and other organizations have helped shape the global scope of these documents.

NCCLS can only succeed in fulfilling its responsibilities with the cooperation of other organizations and individuals. In view of the economic and quality benefits expected by laboratory practitioners and manufacturers upon implementation of standardization in automation, broad participation and cooperation was sought and obtained, and is gratefully acknowledged. NCCLS will continue to achieve a position of world leadership and influence in the development and harmonization of global standards for the healthcare community.

Recognition of the Efforts of Other Standards Organizations

NCCLS would like to acknowledge and thank the volunteers who are active participants in the related work of other standards organizations for their contributions to the laboratory automation program. Their effective leadership and outstanding volunteer service during the development and successful completion of the automation standards is greatly appreciated. This special recognition includes volunteers who are participants in the following standards organizations:

American National Standards Institute (ANSI) Health Informatics Standards Board (HISB)
ASTM Committee E31
Health Level 7 (HL7)
International Organization for Standardization Technical Committee 212 (ISO/TC 212)
Institute of Electrical and Electronics Engineers, Inc. (IEEE)
International Federation of Clinical Chemistry (IFCC)
Japanese Association of Healthcare Information Systems (JAHIS)
Japanese Committee for Clinical Laboratory Standards (JCCLS)
Japanese Society for Clinical Chemistry (JSCC)

Recognition of Laboratory Automation Fund Contributors

Many of the large instrument and automation system vendors and the users of the technology recognized the clear need to develop standards for clinical laboratory automation and information systems and actively supported NCCLS in meeting this need through the efforts of the Area Committee on Automation. To achieve standardization and ensure that automation projects do not compete with other NCCLS projects for resources, a Laboratory Automation Development Fund was created. We express our appreciation to all organizations that have supported this important program.

A list of Laboratory Automation Development Fund contributors is included on the inside front cover of this document.

Foreword

Clinical laboratory automation is defined as the integration of laboratory personnel and analytical, preanalytical, postanalytical, and computer systems to positively benefit the healthcare system, patient, and the laboratory's quality, economics, reliability, speed, and safety.

This standard addresses the communication of data between systems and provides information that is needed for developing automated laboratory systems. Recent standards developments have addressed messaging and data interchange, and test/result naming. Although some are used by a few instrument and Laboratory Information Systems (LIS) vendors, they have not been universally adopted, are only a small part of what will eventually be needed for true standard, compatible, device interconnectability, and lack a mechanism for validating vendor conformance.

These specifications are also intended to complement the following interrelated NCCLS standards developed by other automation subcommittees and support overall operational goals for future development in laboratory instrumentation and automation:

AUTO1—*Laboratory Automation: Specimen Container/Specimen Carrier;*

AUTO2—*Laboratory Automation: Bar Codes for Specimen Container Identification;*

AUTO4—*Laboratory Automation: Systems Operational Requirements, Characteristics, and Information Elements;* and

AUTO5—*Laboratory Automation: Electromechanical Interfaces.*

An important aspect of the development of this and all NCCLS documents should be emphasized, and that is the consensus process. Within the context and operation of NCCLS, the term "consensus" means more than agreement. In the context of standard development, "consensus" is a process by which NCCLS, its members, and interested parties (1) have the opportunity to review and comment on any NCCLS publication; and (2) are assured that their comments are given serious, competent consideration.

Any standard dealing with laboratory automation will be continually evolving. Open communication and exchange of ideas and information can be used to modify the standard through the consensus process.

Key Words

Analytical instruments, automation, clinical laboratory automation, Laboratory Automation Systems (LAS), Laboratory Information Systems (LIS), process instruments

Laboratory Automation: Communications with Automated Clinical Laboratory Systems, Instruments, Devices, and Information Systems; Approved Standard

1 Introduction

This standard on information transfer and control for automated laboratory systems was developed by the NCCLS Subcommittee on Communications with Automation Systems.

Clinical laboratory automation is defined as the integration of laboratory personnel and analytical, pre- and postanalytical, and computer systems to positively benefit the healthcare system, patient, and the laboratory's quality, economics, reliability, speed, and safety. The goal of this document is to facilitate accurate and timely electronic exchange of data and information between the automated laboratory elements. This will allow and encourage scalable, open systems, and extendibility and interoperability of the automated laboratory elements. Implementation of this standard will contribute to the development of a shared vision of future clinical laboratory automation communications.

1.1 Scope

This standard provides a protocol for communications between Laboratory Automation Systems (LAS), Laboratory Information Systems (LIS), automated instruments (analyzers), and pre- and postanalytical automated devices. Although the primary venue for this standard is the clinical laboratory, in the future, elements of this standard may be applicable in the anatomic pathology, cytology, and related laboratories, as well as to nonclinical (analytical) laboratories. Additionally, although the focus of this standard is clinical laboratory automation, elements of the standard may apply to related (non-automated) areas such as small analyzers or point-of-care devices.

This standard fits into the series of interrelated NCCLS automation standards (AUTO1—*Laboratory Automation: Specimen Container/Specimen Carrier*; AUTO2—*Laboratory Automation: Bar Codes for Specimen Container Identification*; AUTO4—*Laboratory Automation: Systems Operational Requirements, Characteristics, and Information Elements*; and AUTO5—*Laboratory Automation: Electromechanical Interfaces*).

1.2 Limitations

The focus of this standard is on the characteristics of the communications (low-level protocol) and the data to be transferred (high-level protocol). The low-level protocol was developed to meet the bandwidth and time characteristics required by automation. The high-level protocol defines specific messages and data to be transferred in automated communications.

It is recognized that there are old protocols in use in clinical laboratories that are not supported by this standard. The overall intent of this standard is to be prospective in nature and meet anticipated future needs for automation. Of necessity, therefore, the standard focuses on protocols that can meet the time and data characteristics for automation systems. Older (legacy) systems are not necessarily excluded, but are also not supported.

1.3 Terminology

Throughout this document and the other automation documents, terms such as "system," "automation system," "laboratory automation system," "total laboratory automation," or "clinical laboratory automation" may be used. These terms are used interchangeably. The authors have attempted to