

Archived Document


This archived document is no longer being reviewed through the CLSI Consensus Document Development Process. However, this document is technically valid as of January 2019. Because of its value to the laboratory community, it is being retained in CLSI's library.



November 1999

C37-A

Preparation and Validation of Commutable Frozen Human Serum Pools as Secondary Reference Materials for Cholesterol Measurement Procedures; Approved Guideline



This guideline details procedures for the manufacture and evaluation of human serum pools for cholesterol measurement.

A guideline for global application developed through the Clinical and Laboratory Standards Institute consensus process.

Clinical and Laboratory Standards Institute

Setting the standard for quality in medical laboratory testing around the world.

The Clinical and Laboratory Standards Institute (CLSI) is a not-for-profit membership organization that brings together the varied perspectives and expertise of the worldwide laboratory community for the advancement of a common cause: to foster excellence in laboratory medicine by developing and implementing medical laboratory standards and guidelines that help laboratories fulfill their responsibilities with efficiency, effectiveness, and global applicability.

Consensus Process

Consensus—the substantial agreement by materially affected, competent, and interested parties—is core to the development of all CLSI documents. It does not always connote unanimous agreement, but does mean that the participants in the development of a consensus document have considered and resolved all relevant objections and accept the resulting agreement.

Commenting on Documents

CLSI documents undergo periodic evaluation and modification to keep pace with advancements in technologies, procedures, methods, and protocols affecting the laboratory or health care.

CLSI's consensus process depends on experts who volunteer to serve as contributing authors and/or as participants in the reviewing and commenting process. At the end of each comment period, the committee that developed the document is obligated to review all comments, respond in writing to all substantive comments, and revise the draft document as appropriate.

Comments on published CLSI documents are equally essential, and may be submitted by anyone, at any time, on any document. All comments are managed according to the consensus process by a committee of experts.

Appeals Process

When it is believed that an objection has not been adequately considered and responded to, the process for appeals, documented in the CLSI Standards Development Policies and Processes, is followed.

All comments and responses submitted on draft and published documents are retained on file at CLSI and are available upon request.

Get Involved—Volunteer!

Do you use CLSI documents in your workplace? Do you see room for improvement? Would you like to get involved in the revision process? Or maybe you see a need to develop a new document for an emerging technology? CLSI wants to hear from you. We are always looking for volunteers. By donating your time and talents to improve the standards that affect your own work, you will play an active role in improving public health across the globe.

For additional information on committee participation or to submit comments, contact CLSI.

Clinical and Laboratory Standards Institute
950 West Valley Road, Suite 2500
Wayne, PA 19087 USA
P: +1.610.688.0100
F: +1.610.688.0700
www.clsi.org
standard@clsi.org

Preparation and Validation of Commutable Frozen Human Serum Pools as Secondary Reference Materials for Cholesterol Measurement Procedures; Approved Guideline

Abstract

Preparation and Validation of Commutable Frozen Human Serum Pools as Secondary Reference Materials for Cholesterol Measurement Procedures; Approved Guideline (C37-A) outlines procedures for selecting, processing, and combining donor units to prepare frozen serum pools which are commutable among multiple methods for serum cholesterol measurement. The guideline also addresses issues related to the evaluation of the pooled materials. The manufacture and evaluation of two levels of serum cholesterol pools prepared according to this guideline are described. The appendix includes a summary of the results from the pilot study conducted to evaluate the scientific basis of this guideline.

This guideline will provide information to develop reference materials which can be useful to manufacturers of *in vitro* diagnostic (IVD) reagents, systems, and quality control products in establishment of calibration (or assigned values) for their products. Additionally, these reference materials can be beneficial to proficiency testing agencies, as well as end users in the clinical laboratory as tools to independently assess performance (and trueness) of procedures for the routine measurement of serum cholesterol.

Clinical and Laboratory Standards Institute (CLSI). *Preparation and Validation of Commutable Frozen Human Serum Pools as Secondary Reference Materials for Cholesterol Measurement Procedures; Approved Guideline*. CLSI document C37-A (ISBN 1-56238-392-2). Clinical and Laboratory Standards Institute, 950 West Valley Road, Suite 2500, Wayne, Pennsylvania 19087 USA, 1999.

The Clinical and Laboratory Standards Institute consensus process, which is the mechanism for moving a document through two or more levels of review by the health care community, is an ongoing process. Users should expect revised editions of any given document. Because rapid changes in technology may affect the procedures, methods, and protocols in a standard or guideline, users should replace outdated editions with the current editions of CLSI documents. Current editions are listed in the CLSI catalog and posted on our website at www.clsi.org. If your organization is not a member and would like to become one, and to request a copy of the catalog, contact us at: Telephone: 610.688.0100; Fax: 610.688.0700; E-Mail: customerservice@cls.org; Website: www.clsi.org.

November 1999

C37-A
ISBN 1-56238-392-2
ISSN 0273-3099

Preparation and Validation of Commutable Frozen Human
Serum Pools as Secondary Reference Materials for Cholesterol
Measurement Procedures; Approved Guideline

Volume 19 Number 25

Gary L. Myers, Ph.D., Chairholder
John H. Eckfeldt, M.D., Ph.D.
Neil Greenberg, Ph.D.
Jacob B. Levine, M.B.A.
W. Gregory Miller, Ph.D.
Donald A. Wiebe, Ph.D.



**CLINICAL AND
LABORATORY
STANDARDS
INSTITUTE®**

Copyright ©1999 Clinical and Laboratory Standards Institute. Except as stated below, any reproduction of content from a CLSI copyrighted standard, guideline, companion product, or other material requires express written consent from CLSI. All rights reserved. Interested parties may send permission requests to permissions@clsi.org.

CLSI hereby grants permission to each individual member or purchaser to make a single reproduction of this publication for use in its laboratory procedure manual at a single site. To request permission to use this publication in any other manner, e-mail permissions@clsi.org.

Suggested Citation

CLSI. *Preparation and Validation of Commutable Frozen Human Serum Pools as Secondary Reference Materials for Cholesterol Measurement Procedures; Approved Guideline*. CLSI document C37-A. Wayne, PA: Clinical and Laboratory Standards Institute; 1999.

Previous Edition:

May 1998

Reaffirmed:

April 2008

Archived:

January 2019

ISBN 1-56238-392-2

ISSN 0273-3099

Committee Membership

Area Committee on Clinical Chemistry and Toxicology

Basil T. Dumas, Ph.D.
Chairholder

Medical College of Wisconsin
Milwaukee, Wisconsin

W. Gregory Miller, Ph.D.
Vice-Chairholder

Virginia Commonwealth University
Richmond, Virginia

Subcommittee on Cholesterol Reference Material Specifications

Gary L. Myers, Ph.D.
Chairholder

Centers for Disease Control and Prevention
Atlanta, Georgia

John H. Eckfeldt, M.D., Ph.D.

Fairview - University Medical Center
Minneapolis, Minnesota

Neil Greenberg, Ph.D.

Ortho Clinical Diagnostics
Rochester, New York

Bill Haden

Reagents Applications, Inc.
San Diego, California

Arthur Kessner, Ph.D.

Beckman Instruments, Inc.
Brea, California

Jacob B. Levine, M.B.A.

Bayer Corporation
Tarrytown, New York

W. Gregory Miller, Ph.D.

Virginia Commonwealth University
Richmond, Virginia

Nader Rifai, Ph.D.

Children's Hospital
Boston, Massachusetts

Richard C. San George, Ph.D.

Waltham, Massachusetts

Michael J. Welch, Ph.D.

National Institute of Standards and Technology
Gaithersburg, Maryland

Donald A. Wiebe, Ph.D.

University of Wisconsin
Madison, Wisconsin

Advisors

Dennis Bozimowski, Ph.D.

Abbott Laboratories
Abbott Park, Illinois

Eileen Carreiro-Lewandowski

Pawtucket, Rhode Island

Jennifer Colbert

National Institute of Standards and Technology
Gaithersburg, Maryland

Advisors (Continued)

Irene M. Daza	Pennsylvania Department of Health Lionville, Pennsylvania
Robert A. Dean	Indiana University Medical Center Indianapolis, Indiana
Marvin C. Feil, F.A.I.C.	Corning Clinical Laboratories Teterboro, New Jersey
Russell T. Joseph, M.S.	Paramax Chemistry Irvine, California
Norbert Mantor, Ph.D.	St. Thomas Department of Health St. Thomas, Virgin Islands
Alfred E. Peacock, Jr.	Reagents Applications, Inc. San Diego, California
Robert Rej, Ph.D.	New York State Department of Health Albany, New York
Cornelia B. Rooks, M.A.	FDA Center for Devices/Radiological Health Rockville, Maryland
Michael Santulli, M.S.	Bayer Corporation Tarrytown, New York
G. Russell Warnick, M.S., MBA	Pacific Biometrics, Inc. Seattle, Washington
Robert E. Wenk, M.D.	Sinai Hospital of Baltimore Baltimore, Maryland
Wei-Guo Zhao, M.D., M.S.	Dade Behring, Inc. Newark, Delaware
Beth Ann Wise, M.T.(ASCP), M.S.Ed. <i>Staff Liaison</i>	NCCLS Wayne, Pennsylvania
Patrice E. Polgar <i>Editor</i>	NCCLS Wayne, Pennsylvania
Donna M. Wilhelm <i>Assistant Editor</i>	NCCLS Wayne, Pennsylvania

ACTIVE MEMBERSHIP (as of 1 October 1999)

Sustaining Members

Abbott Laboratories
 American Association for
 Clinical Chemistry
 Bayer Corporation
 Beckman Coulter, Inc.
 Becton Dickinson and Company
 bioMérieux, Inc.
 College of American
 Pathologists
 Dade Behring Inc.
 Nippon Becton Dickinson Co,
 Ltd.
 Ortho-Clinical Diagnostics, Inc.
 Pfizer Inc
 Roche Diagnostics, Inc.

Professional Members

American Academy of Family
 Physicians
 American Association of
 Bioanalysts
 American Association of Blood
 Banks
 American Association for
 Clinical Chemistry
 American Association for
 Respiratory Care
 American Chemical Society
 American Medical Technologists
 American Public Health
 Association
 American Society for Clinical
 Laboratory Science
 American Society of
 Hematology
 American Society for
 Microbiology
 American Society of
 Parasitologists, Inc.
 American Type Culture
 Collection, Inc.
 Asociación Española Primera de
 Socorros (Uruguay)
 Asociacion Mexicana de
 Bioquímica Clínica A.C.
 Assn. of Public Health
 Laboratories
 Assoc. Micro. Clinici Italiani-
 A.M.C.L.I.
 Australasian Association of
 Clinical Biochemists
 British Society for Antimicrobial
 Chemotherapy
 Canadian Society for Medical
 Laboratory Science—Société
 Canadienne de Science de
 Laboratoire Médical

Canadian Society of Clinical
 Chemists
 Clinical Laboratory Management
 Association
 College of American
 Pathologists
 College of Medical Laboratory
 Technologists of Ontario
 College of Physicians and
 Surgeons of Saskatchewan
 Commission on Office
 Laboratory Accreditation
 Danish Society of Clinical
 Chemistry
 Fundacion Bioquímica de la
 Provincia (Argentina)
 International Association of
 Medical Laboratory
 Technologists
 International Council for
 Standardization in
 Haematology
 International Federation of
 Clinical Chemistry
 International Society for
 Analytical Cytology
 Italian Society of Clinical
 Biochemistry
 Japan Society of Clinical
 Chemistry
 Japanese Committee for Clinical
 Laboratory Standards
 Joint Commission on
 Accreditation of Healthcare
 Organizations
 National Academy of Clinical
 Biochemistry
 National Society for
 Histotechnology, Inc.
 Ontario Medical Association
 Laboratory Proficiency Testing
 Program
 Ordre professionnel des
 technologistes médicaux du
 Québec
 RCPA Quality Assurance
 Programs PTY Limited
 Sociedade Brasileira de Analises
 Clinicas
 Sociedade Brasileira de
 Patologia Clínica
 Sociedad Espanola de Quimica
 Clínica
 VKCN (The Netherlands)

Government Members

Armed Forces Institute of
 Pathology
 Association of Public Health
 Laboratory Directors

BC Centre for Disease Control
 Centers for Disease Control and
 Prevention
 Chinese Committee for Clinical
 Laboratory Standards
 Commonwealth of Pennsylvania
 Bureau of Laboratories
 Department of Veterans Affairs
 Deutsches Institut für Normung
 (DIN)
 FDA Center for Devices and
 Radiological Health
 FDA Division of Anti-Infective
 Drug Products
 Health Care Financing
 Administration/CLIA Program
 Health Care Financing
 Administration
 Iowa State Hygienic Laboratory
 Massachusetts Department of
 Public Health Laboratories
 Michigan Department of Public
 Health
 National Association of Testing
 Authorities - Australia
 National Center of Infectious
 and Parasitic Diseases
 (Bulgaria)
 National Institute of Standards
 and Technology
 Ohio Department of Health
 Oklahoma State Department of
 Health
 Ontario Ministry of Health
 Saskatchewan Health-
 Provincial Laboratory
 Scientific Institute of Public
 Health; Belgium Ministry of
 Social Affairs, Public Health
 and the Environment
 South African Institute for
 Medical Research
 Swedish Institute for Infectious
 Disease Control
 Thailand Department of Medical
 Sciences

Industry Members

AB Biodisk
 Abbott Laboratories
 Accumetrics, Inc.
 Amersham Pharmacia Biotech
 Ammirati Regulatory Consulting
 Assessor
 AstraZeneca
 Avocet Medical, Inc.
 Bayer Corporation - Elkhart, IN
 Bayer Corporation - Middletown,
 VA

Bayer Corporation - Tarrytown, NY
Bayer Corporation - West Haven, CT
Bayer Medical Ltd.
Beckman Coulter, Inc.
Beckman Coulter, Inc. Primary Care Diagnostics
Beckman Coulter K.K. (Japan)
Becton Dickinson and Company
Becton Dickinson Biosciences
Becton Dickinson Consumer Products
Becton Dickinson Immunocytometry Systems
Becton Dickinson Italia S.P.A.
Becton Dickinson VACUTAINER Systems
Biolog, Inc.
bioMérieux, Inc.
Biometry Consultants
Bio-Rad Laboratories, Inc.
Biotest AG
Bristol-Myers Squibb Company
Canadian Reference Laboratory Ltd.
CASCO•NERL Diagnostics
Checkpoint Development Inc.
Clinical Lab Engineering
COBE Laboratories, Inc.
Combact Diagnostic Systems Ltd.
Community Medical Center (NJ)
Control Lab (Brazil)
Copan Diagnostics Inc.
Cosmetic Ingredient Review
Cubist Pharmaceuticals
Cytometrics, Inc.
Dade Behring Inc. - Deerfield, IL
Dade Behring Inc. - Glasgow, DE
Dade Behring Inc. - Marburg, Germany
Dade Behring Inc. - Miami, FL
Dade Behring Inc. - Sacramento, CA
Dade Behring Inc. - San Jose, CA
DAKO A/S
Diagnostic Products Corporation
DiaSorin
Eiken Chemical Company, Ltd.
Enterprise Analysis Corporation
Fort Dodge Animal Health
Gen-Probe
Glaxo-Wellcome, Inc.
Greiner Mediatech, Inc.
Health Systems Concepts, Inc.
Helena Laboratories
Home Diagnostics, Inc.
Hycor Biomedical Inc.
I-STAT Corporation
Instrumentation Laboratory
Integ, Inc.

International Technidyne Corporation
Johnson City Medical Center
Kendall Sherwood-Davis & Geck
Labtest Diagnostica S.A.
LifeScan, Inc. (a Johnson & Johnson Company)
Lilly Research Laboratories
Medical Device Consultants, Inc.
Medical Laboratory Automation Inc.
MediSense Products (Div. Of Abbott Laboratories)
Medtronic Perfusion Systems
Merck & Company, Inc.
mvi Sciences (MA)
Nabi
Neometrics Inc.
Nichols Institute Diagnostics (Div. of Quest Diagnostics, Inc.)
Nissui Pharmaceutical Co., Ltd.
Nippon Becton Dickinson Co., Ltd.
Norfolk Associates, Inc.
OBC Associates
Olympus Corporation
Optical Sensors, Inc.
Organon Teknika Corporation
Ortho-Clinical Diagnostics, Inc. (England)
Ortho-Clinical Diagnostics, Inc. (Raritan, NJ)
Ortho-Clinical Diagnostics, Inc. (Rochester, NY)
Oxoid Inc.
Pfizer Inc
Pharmacia & Upjohn
Procter & Gamble Pharmaceuticals, Inc.
The Product Development Group
Quest Diagnostics Incorporated
Quintiles, Inc.
Radiometer America, Inc.
Radiometer Medical A/S
David G. Rhoads Associates, Inc.
Rhône-Poulenc Rorer
Roche Diagnostics GmbH
Roche Diagnostics, Inc.
Roche Diagnostic Systems (Div. Hoffmann-La Roche Inc.)
Roche Laboratories (Div. Hoffmann-La Roche Inc.)
The R.W. Johnson
Pharmaceutical Research Institute
Sanofi Diagnostics Pasteur
Sarstedt, Inc.
SARL Laboratoire Carron (France)
Schering Corporation

Schleicher & Schuell, Inc.
Second Opinion
SenDx Medical, Inc.
Showa Yakuhin Kako Company, Ltd.
SmithKline Beecham, S.A.
Streck Laboratories, Inc.
Sysmex Corporation (Japan)
Sysmex Corporation (Long Grove, IL)
The Toledo Hospital (OH)
Trek Diagnostic Systems, Inc.
Vetoquinol S.A.
Vysis, Inc.
Wallac Oy
Warner-Lambert Company
Wyeth-Ayerst
Xyletech Systems, Inc.
YD Consultant

Trade Associations

Association of Medical Diagnostic Manufacturers
Health Industry Manufacturers Association
Japan Association Clinical Reagents Ind. (Tokyo, Japan)
Medical Industry Association of Australia

Associate Active Members

67th CSH Wuerzburg, GE (NY)
121st General Hospital (CA)
Acadiana Medical Laboratories, LTD (LA)
Advocate Laboratories (IL)
Allegheny General Hospital (PA)
Allegheny University of the Health Sciences (PA)
Allina Laboratories (MN)
Alton Ochsner Medical Foundation (LA)
American Medical Laboratories (VA)
Anzac House (Australia)
Associated Regional & University Pathologists (UT)
Aurora Consolidated Laboratories (WI)
Baptist St. Anthony's Health Network (TX)
Baystate Medical Center (MA)
Brazileiro De Promocao (Brazil)
Bristol Regional Medical Center (TN)
Brookdale Hospital Medical Center (NY)
Brooke Army Medical Center (TX)
Brooks Air Force Base (TX)
Broward General Medical Center (FL)

Calgary Laboratory Services (Calgary, AB, Canada)	Health Network Lab (PA)	MDS-Sciex (Concord, ON, Canada)
Cardinal Glennon Children's Hospital (MO)	Health Sciences Centre (Winnipeg, MB, Canada)	Medical College of Virginia Hospital
Central Kansas Medical Center	Heartland Health System (MO)	Melrose-Wakefield Hospital (MA)
Champlain Valley Physicians Hospital (NY)	Hinsdale Hospital ((L)	Memorial Hospital (CO)
Children's Hospital (LA)	Hoag Memorial Hospital Presbyterian (CA)	Memorial Medical Center (Napoleon Ave., New Orleans,LA)
Children's Hospital Medical Center (Akron, OH)	Holmes Regional Medical Center (FL)	Memorial Medical Center (N. Jefferson Davis Pkwy., New Orleans, LA)
Clendo Lab (Puerto Rico)	Holy Spirit Hospital (PA)	Memorial Medical Center (IL)
CLSI Laboratories (PA)	Holzer Medical Center (OH)	Mercy Health System (PA)
Colorado Mental Health Institute at Pueblo	Hospital for Sick Children (Toronto, ON, Canada)	Mercy Hospital (NC)
Commonwealth of Kentucky	Hospital Israelita Albert Einstein (Brazil)	Methodist Hospital (TX)
CompuNet Clinical Laboratories (OH)	Hotel Dieu Hospital (Windsor, ON, Canada)	Methodist Hospital Indiana
Consolidated Laboratory Services (CA)	Huddinge University Hospital (Sweden)	Methodist Hospitals of Memphis (TN)
Danville Regional Medical Center (VA)	Hurley Medical Center (MI)	Mid Michigan Medical Center - Midland
Dean Medical Center (WI)	Indiana University	Milton S. Hershey Medical Center (PA)
Detroit Health Department (MI)	Instituto Scientifico HS. Raffaele (Italy)	Montreal Children's Hospital (Canada)
Diagnostic Laboratory Services, Inc. (HI)	International Health Management Associates, Inc. (IL)	Montreal General Hospital (Canada)
Duke University Medical Center (NC)	Intermountain Health Care Laboratory Services (UT)	Mount Sinai Hospital (NY)
Durham Regional Hospital (NC)	Jacobi Medical Center (NY)	National University Hospital (Singapore)
Duzen Laboratories (Turkey)	John Peter Smith Hospital (TX)	Naval Surface Warfare Center (IN)
Dynacare Laboratories - Eastern Region (Ottawa, ON, Canada)	John Randolph Hospital (VA)	Nebraska Health System
E.A. Conway Medical Center (LA)	Johns Hopkins Medical Institutions (MD)	New Britain General Hospital (CT)
Elmhurst Memorial Hospital (IL)	Johnson City Medical Center (IN)	New England Medical Center Hospital (MA)
Elyria Memorial Hospital (OH)	Kaiser Permanente (CA)	The New York Blood Center
Emory University Hospital (GA)	Kenora-Rainy River Regional Laboratory Program (Dryden, Ontario, Canada)	The New York Hospital Medical Center of Queens
Fairfax Hospital (VA)	Klinicni Center (Slovenia)	New York State Department of Health
Fairview-University Medical Center (MN)	LabCorp (NC)	NorDx (ME)
Foothills Hospital (Calgary, AB, Canada)	Laboratoire de Santé Publique du Quebec (Canada)	North Carolina Laboratory of Public Health
Fox Chase Cancer Center (PA)	Laboratório Fleury S/C Ltda. (Brazil)	North Coast Clinical Laboratory, Inc. (OH)
Franklin Square Hospital Center (MD)	Laboratory Corporation of America (MO)	Northridge Hospital Medical Center (CA)
Fresenius Medical Care/Life Chem (NJ)	Lancaster General Hospital (PA)	Northwestern Memorial Hospital (IL)
Fresno Community Hospital and Medical Center	Langley Air Force Base (VA)	Ohio State University Hospitals
GDS Technology, Inc (IN)	Lewis-Gale Medical Center (VA)	Olin E. Teague Medical Center (TX)
Grady Memorial Hospital (GA)	Libero Instituto Univ. Campus BioMedico (Italy)	Our Lady of Lourdes Hospital (NJ)
Greater Southeast Community Hospital (DC)	LAC and USC Healthcare Network (CA)	Our Lady of the Resurrection Medical Center (IL)
Guthrie Clinic Laboratories (PA)	Louisiana State University Medical Center	Pathology and Cytology Laboratories, Inc. (KY)
Halifax Medical Center (FL)	Lutheran Hospital (WI)	Permanente Medical Group (CA)
Harris Methodist Fort Worth (TX)	Main Line Clinical Laboratories, Inc. (PA)	
Harris Methodist Northwest (TX)	Massachusetts General Hospital	
Hartford Hospital (CT)	Mayo Clinic Scottsdale (AZ)	
Hays Pathology Laboratories, P.A. (KS)	MDS Metro Laboratory Services (Burnaby, BC, Canada)	
Headwaters Health Authority (High River, AB, Canada)		
Health Alliance Laboratory (OH)		

Presbyterian Hospital of Dallas (TX)	Southern California Permanente Medical Group	VA (San Diego) Medical Center (CA)
Providence Health System (OR)	South Western Area Pathology Service (Australia)	VA (Tuskegee) Medical Center (AL)
Providence Medical Center (WA)	Speciality Laboratories, Inc. (CA)	Vejle Hospital (Denmark)
Queen Elizabeth Hospital (Prince Edward Island, Canada)	Stanford Health Services (CA)	Viridae Clinical Sciences, Inc. (Vancouver, BC, Canada)
Queensland Health Pathology Services (Australia)	Stormont-Vail Regional Medical Center (KS)	ViroLogic, Inc. (CA)
Quintiles Laboratories, Ltd. (GA)	Sun Health-Boswell Hospital (AZ)	Waikato Hospital (New Zealand)
Regions Hospital	Sunrise Hospital and Medical Center (NV)	Walter Reed Army Institute of Research (MD)
Research Medical Center (MO)	Sutter Health (CA)	Warde Medical Laboratory (MI)
Riyadh Armed Forces Hospital (Saudi Arabia)	Tampa General Hospital (FL)	Warren Hospital (NJ)
Robert F. Kennedy Medical Center (CA)	Timmins & District Hospital (Timmons, ON, Canada)	Washoe Medical Center (NV)
Royal Columbian Hospital (New Westminster, BC, Canada)	Tripler Army Medical Center (HI)	Watson Clinic (FL)
Saint Mary's Regional Medical Center (NV)	Trumbull Memorial Hospital (OH)	Wilford Hall Medical Center (TX)
St. Alexius Medical Center (ND)	Tulane Medical Center Hospital & Clinic (LA)	William Beaumont Hospital (MI)
St. Anthony Hospital (CO)	Twin Lake Regional Medical Center	Williamsburg Community Hospital (VA)
St. Boniface General Hospital (Winnipeg, Canada)	UCSF Medical Center (CA)	Wilson Memorial Hospital (NY)
St. Elizabeth Hospital (NJ)	UNC Hospitals (NC)	Winchester Hospital (MA)
St. Francis Medical Center (CA)	Unilab Clinical Laboratories (CA)	Winn Army Community Hospital (GA)
St. John Hospital and Medical Center (MI)	University of Alabama - Birmingham Hospital	Wishard Memorial Hospital (IN)
St. John Regional Hospital (St. John, NB, Canada)	University of Alberta Hospitals (Canada)	Womack Army Medical Center (NC)
St. Joseph Hospital (NE)	University of Chicago Hospitals (IL)	Yonsei University College of Medicine (Korea)
St. Joseph Mercy - Oakland (MI)	University Hospital (IN)	York Hospital (PA)
St. Joseph's Hospital - Marshfield Clinic (WI)	University Hospital (Gent) (Belgium)	Zale Lipshy University Hospital (TX)
St. Luke's Hospital (PA)	University Hospital (London, Ontario, Canada)	
St. Luke's Regional Medical Center (IA)	The University Hospitals (OK)	
St. Mary Hospital (NJ)	University of Massachusetts Lowell	
St. Mary Medical Center (CA)	University of Medicine & Dentistry, NJ University Hospital	
St. Mary Medical Center (IN)	University of Michigan	
St. Mary of the Plains Hospital (TX)	University of the Ryukyus (Japan)	
San Francisco General Hospital (CA)	University of Texas Medical School at Houston	
Santa Cabrini Hospital (Montreal, PQ Canada)	University of Virginia Medical Center	
Santa Clara Valley Medical Center (CA)	University of Washington	
Seoul Nat'l University Hospital (Korea)	UPMC Bedford Memorial (PA)	
Shanghai Center for the Clinical Laboratory (China)	USAF Medical Center (OH)	
Shands Healthcare (FL)	UZ-KUL Medical Center (Belgium)	
SmithKline Beecham Clinical Laboratories (GA)	VA (Dayton) Medical Center (OH)	
SmithKline Beecham Clinical Laboratories (WA)	VA (Denver) Medical Center (CO)	
South Bend Medical Foundation (IN)	VA Outpatient Clinic (OH)	

OFFICERS

William F. Koch, Ph.D.,
President
National Institute of Standards
and Technology

F. Alan Andersen, Ph.D.,
President Elect
Cosmetic Ingredient Review

Robert F. Moran, Ph.D.,
FCCM, FAIC
Secretary
mvi Sciences

Donna M. Meyer, Ph.D.,
Treasurer
CHRISTUS Health

A. Samuel Koenig, III, M.D.,
Past President
Family Medical Care

John V. Bergen, Ph.D.,
Executive Director

BOARD OF DIRECTORS

Sharon S. Ehrmeyer, Ph.D.
University of Wisconsin

Robert L. Habig, Ph.D.
Cytometrics, Inc.

Thomas L. Hearn, Ph.D.
Centers for Disease Control and
Prevention

Gerald A. Hoeltge, M.D.
The Cleveland Clinic Foundation

Elizabeth D. Jacobson, Ph.D.
FDA Center for Devices and
Radiological Health

Carolyn D. Jones, J.D., M.P.H.
Health Industry Manufacturers
Association

Hartmut Jung, Ph.D.
Roche Diagnostics GmbH

Tadashi Kawai, M.D., Ph.D.
International Clinical Pathology
Center

Barbara G. Painter, Ph.D.
Bayer Corporation

Marianne C. Watters,
M.T.(ASCP)
Parkland Health & Hospital
System

Ann M. Willey, Ph.D.
New York State Department of
Health

Judith A. Yost, M.A.,
M.T.(ASCP)
Health Care Financing
Administration

Contents

Abstract.....	i
Committee Membership.....	v
Active Membership.....	vii
Foreword.....	xv
1 Introduction.....	1
2 Scope.....	1
3 Standard Precautions	1
4 Definitions	1
5 Design Considerations	2
5.1 Pool Preparation.....	2
5.2 Validation of Performance and Commutability	4
6 Raw Materials.....	5
6.1 Donor Selection	5
6.2 General Exclusion Criteria for Donor Units and Pools	5
7 Blood Collection and Processing Procedure.....	5
References	7
Appendix A. Quality Control Sampling Test for Pool Homogeneity.....	9
A1 Test Objective	9
A2 Test Assumptions	9
A3 Test Plan	9
A4 Data Analysis	9
Appendix B. Verification of Material Specifications.....	10
B1 Validation Study Materials	10
B2 Donor Selection	10
B3 Materials and Apparatus.....	12
B4 Blood Collection and Processing	13
B5 Testing Procedures Relevant to Production of Cholesterol Pools	14
B6 Materials Validation Results	15
B7 Certificates of Analysis for Cholesterol Pools.....	17
Summary of Comments and Subcommittee Responses	21
Related NCCLS Publications	26

Foreword

Numerous epidemiological and clinical investigations across the world have documented that elevated levels of blood cholesterol are a major risk factor for coronary heart disease (CHD) and that lowering blood cholesterol levels decreases morbidity and mortality from CHD.¹⁻⁵ Reduction in the prevalence of elevated blood cholesterol has been a primary objective of the National Cholesterol Education Program (NCEP), established in 1985 by the National Heart, Lung and Blood Institute.^{6,7} The efforts of NCEP and other similar international cholesterol education programs to reduce high blood cholesterol are based on the need for reliable blood cholesterol measurements for proper patient risk classification.⁸ An important milestone in the NCEP effort was the establishment of the National Reference System for Cholesterol (NRS/CHOL) by NCCLS, which provides the accuracy target for cholesterol measurement.⁹ The NCEP's Laboratory Standardization Panel recommended that all cholesterol measurements be standardized and traceable to the NRS/CHOL.⁸ It is extremely important that manufacturers of clinical instruments, reagents, and calibrator materials, as well as clinical laboratories, have available to them reference materials that will establish traceability to the NRS/CHOL and assure true and precise results on human subjects, in all field methods being used worldwide. In order to achieve this goal, reference materials with long-term stability and appropriate analyte concentrations are required.

An additional very important requirement is commutability, termed "transmutability" by some investigators. Even though often prepared from human sources, processed reference materials in the past have often taken on properties, loosely termed "matrix effects," that make them behave differently from patient specimens in some reagent/instrument systems.¹⁰ Matrix effects differ from more typical analytical interferences in that the latter are effects of exogenous (e.g., drug) or endogenous (e.g., bilirubin) substances that have similar effects on both patient and manufactured control and calibrator materials. In contrast, matrix effects are properties that are unique to the processed reference materials (and control materials). Since matrix effects lead to underrecovery, or less commonly overrecovery, of the analyte of interest in some field methods, knowing very true and precise analyte concentrations on a reference material which demonstrates matrix effects with certain field methods does little to transfer trueness to those methods.

The currently accepted approach to document traceability to the NRS/CHOL for many field methods is via a split-sample comparison with the reference method using fresh human serum specimens.¹¹ The presence or absence of matrix-effect interferences with field methods in a given sample can be confirmed by direct comparison of field method measurements with measurements from a reference method which is not subject to matrix effects. Protocols have been developed which can identify matrix effects in particular materials.¹² However, the use of commutable reference materials is the preferred approach and, therefore, a need exists to develop reference materials for cholesterol measurement procedures that do not demonstrate matrix effects. The few materials-formulation guidelines that exist are of a generic nature, not specifically addressing the problem of modern cholesterol methods and the newer, high-throughput, microsampling analytical equipment.¹³⁻¹⁵ It is therefore clear that there is a significant gap in the NRS/CHOL and that an immediate need exists for agreement on material standards applicable to the establishment of the calibration and trueness of serum cholesterol measurement systems.

Key Words

Cholesterol, commutable, matrix effects, NRS/CHOL, reference materials, serum pools, traceability

Preparation and Validation of Commutable Frozen Human Serum Pools as Secondary Reference Materials for Cholesterol Measurement Procedures; Approved Guideline

1 Introduction

This guideline describes the procedures for preparing frozen human serum pools of high quality suitable for cholesterol measurement. Although prepared for cholesterol, the concepts described in this written guideline may be applied to the preparation of frozen human serum pools for other analytes.

2 Scope

These specifications are designed to enable laboratory scientists, *in vitro* diagnostic (IVD) manufacturers, proficiency testing providers, and suppliers of clinical laboratory reference materials^a to prepare frozen human serum pools which demonstrate minimal matrix effects, are commutable across different cholesterol measurement procedures, and, as such, are suitable for assessing the trueness of these procedures. These serum pools, when assigned target values by the definitive method and/or reference method of the National Reference System for Cholesterol (NRS/CHOL), may be used to properly calibrate and/or assess trueness of field methods for serum cholesterol measurement, and provide an alternative method to split-sample comparisons with human specimens for establishing traceability of calibration of field methods to the NRS/CHOL. This document provides guidance for collecting and processing raw materials to manufacture frozen serum pools and for performing quality assurance of the final product. As part of the guideline development, two frozen human serum pools using a large number of selected individual donor units of "off-the-clot" serum were prepared and characterized. The performance of the pools was assessed in terms of the degree of commutability (observed response versus predicted response) relative to the individual sera which comprised the pools.

^a Examples of suppliers of reference materials in the U.S are the National Institute of Standards and Technology, the College of American Pathologists, and the Centers for Disease Control and Prevention.

3 Standard Precautions

Because it is often impossible to know what might be infectious, all human blood specimens are to be treated as infectious and handled according to "standard precautions." Standard precautions are new guidelines that combine the major features of "universal precautions and body substance isolation" practices. Standard precautions cover the transmission of any pathogen and thus are more comprehensive than universal precautions which are intended to apply only to transmission of blood-borne pathogens. Standard precaution and universal precaution guidelines are available from the U.S. Centers for Disease Control and Prevention (Guideline for Isolation Precautions in Hospitals, Infection Control and Hospital Epidemiology, CDC, Vol 17;1:53-80.), [MMWR 1987;36(suppl 2S):2S-18S] and (MMWR 1988;37:377-382, 387-388). For specific precautions for preventing the laboratory transmission of blood-borne infection from laboratory instruments and materials; and recommendations for the management of blood-borne exposure, refer to NCCLS document M29—*Protection of Laboratory Workers from Instrument Biohazards and Infectious Disease Transmitted by Blood, Body Fluids, and Tissue*.

4 Definitions^b

Accuracy, *n* - Closeness of the agreement between the result of a measurement and a true value of the measurand. **NOTE:** The concept of accuracy of measurement is described by trueness of measurement and precision of measurement. Thus, accuracy is not a synonym for trueness or for precision (VIM: 1993, 3.5¹⁶).

Aseptic, *adj* - Environmental conditions which minimize microbial contamination.

^b Some of these definitions are found in NCCLS document NRSL8—*Terminology and Definitions for Use in NCCLS Documents*. For complete definitions and detailed source information, please refer to the most current edition of that document.