


Sundhedsinformatik – Interoperabilitet mellem enheder – Del 20601: Kommunikation med personligt sundhedsudstyr – Anvendelsesprofil – Optimeret udvekslingsprotokol

Health informatics – Device interoperability –
Part 20601: Personal health device communication –
Application profile – Optimized exchange protocol



DANSK STANDARD
Danish Standards Association

Göteborg Plads 1
DK-2150 Nordhavn

Tel: +45 39 96 61 01
dansk.standard@ds.dk
www.ds.dk

DS projekt: M363485

ICS: 35.240.80

Første del af denne publikations betegnelse er:

DS/ISO, hvilket betyder, at det er en international standard, der har status som dansk standard.

Denne publikations overensstemmelse er:

IDT med: ISO/IEEE 11073-20601:2022

DS-publikationen er på engelsk.

DS-publikationstyper

Dansk Standard udgiver forskellige publikationstyper.

Typen på denne publikation fremgår af forsiden.

Der kan være tale om:

Dansk standard

- standard, der er udarbejdet på nationalt niveau, eller som er baseret på et andet lands nationale standard, eller
- standard, der er udarbejdet på internationalt og/eller europæisk niveau, og som har fået status som dansk standard

DS-information

- publikation, der er udarbejdet på nationalt niveau, og som ikke har opnået status som standard, eller
- publikation, der er udarbejdet på internationalt og/eller europæisk niveau, og som ikke har fået status som standard, fx en teknisk rapport, eller
- europæisk præstandard

DS-håndbog

- samling af standarder, eventuelt suppleret med informativt materiale

DS-hæfte

- publikation med informativt materiale

Til disse publikationstyper kan endvidere udgives

- tillæg og rettelsesblade

DS-publikationsform

Publikationstyperne udgives i forskellig form som henholdsvis

- fuldtekstpublikation (publikationen er trykt i sin helhed)
- godkendelsesblad (publikationen leveres i kopi med et trykt DS-omslag)
- elektronisk (publikationen leveres på et elektronisk medie)

DS-betegnelse

Alle DS-publikationers betegnelse begynder med DS efterfulgt af et eller flere præfikser og et nr., fx **DS 383**, **DS/EN 5414** osv. Hvis der efter nr. er angivet et **A** eller **Cor**, betyder det, enten at det er et **tillæg** eller et **rettelsesblad** til hovedstandard, eller at det er indført i hovedstandard.

DS-betegnelse angives på forsiden.

Overensstemmelse med anden publikation:

Overensstemmelse kan enten være IDT, EQV, NEQ eller MOD

- **IDT:** Når publikationen er identisk med en given publikation.
- **EQV:** Når publikationen teknisk er i overensstemmelse med en given publikation, men præsentationen er ændret.
- **NEQ:** Når publikationen teknisk eller præsentationsmæssigt ikke er i overensstemmelse med en given standard, men udarbejdet på baggrund af denne.
- **MOD:** Når publikationen er modificeret i forhold til en given publikation.

Third edition
2022-12

Health informatics — Device interoperability —

Part 20601: Personal health device communication — Application profile — Optimized exchange protocol

Informatique de santé — Interopérabilité des dispositifs —

*Partie 20601: Communication entre dispositifs de santé personnels —
Profil d'application — Protocole d'échange optimisé*



Reference number
ISO/IEEE 11073-20601:2022(E)

© IEEE 2019



COPYRIGHT PROTECTED DOCUMENT

© IEEE 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from IEEE at the address below.

Institute of Electrical and Electronics Engineers, Inc
3 Park Avenue, New York
NY 10016-5997, USA

Email: stds.ipr@ieee.org
Website: www.ieee.org

Published in Switzerland

This is a preview of "DS/ISO/IEEE 11073-20...". Click here to purchase the full version from the ANSI store.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted (see www.iso.org/directives).

IEEE Standards documents are developed within the IEEE Societies and the Standards Coordinating Committees of the IEEE Standards Association (IEEE-SA) Standards Board. The IEEE develops its standards through a consensus development process, approved by the American National Standards Institute, which brings together volunteers representing varied viewpoints and interests to achieve the final product. Volunteers are not necessarily members of the Institute and serve without compensation. While the IEEE administers the process and establishes rules to promote fairness in the consensus development process, the IEEE does not independently evaluate, test, or verify the accuracy of any of the information contained in its standards.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

ISO/IEEE 11073-20601 was prepared by the *IEEE 11073 Standards Committee of the IEEE Engineering in Medicine and Biology Society* (as IEEE Std 11073-20601-2019) and drafted in accordance with its editorial rules. It was adopted, under the "fast-track procedure" defined in the Partner Standards Development Organization cooperation agreement between ISO and IEEE, by Technical Committee ISO/TC 215, *Health informatics*.

This third edition cancels and replaces the second edition (ISO/IEEE 11073-20601:2016), which has been technically revised. It also incorporates the Technical Corrigendum ISO/IEEE 11073-20601:2016/Cor 1:2016.

A list of all parts in the ISO/IEEE 11073 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This is a preview of "DS/ISO/IEEE 11073-20...". Click [here](#) to purchase the full version from the ANSI store.

This is a preview of "DS/ISO/IEEE 11073-20...". [Click here to purchase the full version from the ANSI store.](#)

(Revision of IEEE Std 11073-20601-2014)

Health informatics—Personal health device communication

Part 20601: Application profile— Optimized Exchange Protocol

Developed by the

IEEE 11073™ Standards Committee
of the
IEEE Engineering in Medicine and Biology Society

Approved 5 September 2019

IEEE SA Standards Board

Abstract: Within the context of the ISO/IEEE 11073 family of standards for device communication, a common framework for making an abstract model of personal health data available in transport-independent transfer syntax required to establish logical connections between systems and to provide presentation capabilities and services needed to perform communication tasks is described in this standard. The protocol is optimized to personal health usage requirements and leverages commonly used methods and tools wherever possible.

Keywords: IEEE 11073-20601™, medical device communication, personal health devices

The Institute of Electrical and Electronics Engineers, Inc.
3 Park Avenue, New York, NY 10016-5997, USA

Copyright © 2019 by The Institute of Electrical and Electronics Engineers, Inc.
All rights reserved. Published 20 December 2019. Printed in the United States of America.

IEEE and IEEE 802 are registered trademarks in the U.S. Patent & Trademark Office, owned by The Institute of Electrical and Electronics Engineers, Incorporated.

PDF: ISBN 978-1-5044-6379-9 STD24014
Print: ISBN 978-1-5044-6380-5 STD24014

IEEE prohibits discrimination, harassment, and bullying.

For more information, visit <http://www.ieee.org/web/aboutus/whatis/policies/p9-26.html>.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without the prior written permission of the publisher.

This is a preview of "DS/ISO/IEEE 11073-20...". Click here to purchase the full version from the ANSI store.

Important Notices and Disclaimers Concerning IEEE Standards Documents

IEEE documents are made available for use subject to important notices and legal disclaimers. These notices and disclaimers, or a reference to this page, appear in all standards and may be found under the heading "Important Notices and Disclaimers Concerning IEEE Standards Documents." They can also be obtained on request from IEEE or viewed at <http://standards.ieee.org/ipr/disclaimers.html>.

Notice and Disclaimer of Liability Concerning the Use of IEEE Standards Documents

IEEE Standards documents (standards, recommended practices, and guides), both full-use and trial-use, are developed within IEEE Societies and the Standards Coordinating Committees of the IEEE Standards Association ("IEEE SA") Standards Board. IEEE ("the Institute") develops its standards through a consensus development process, approved by the American National Standards Institute ("ANSI"), which brings together volunteers representing varied viewpoints and interests to achieve the final product. IEEE Standards are documents developed through scientific, academic, and industry-based technical working groups. Volunteers in IEEE working groups are not necessarily members of the Institute and participate without compensation from IEEE. While IEEE administers the process and establishes rules to promote fairness in the consensus development process, IEEE does not independently evaluate, test, or verify the accuracy of any of the information or the soundness of any judgments contained in its standards.

IEEE Standards do not guarantee or ensure safety, security, health, or environmental protection, or ensure against interference with or from other devices or networks. Implementers and users of IEEE Standards documents are responsible for determining and complying with all appropriate safety, security, environmental, health, and interference protection practices and all applicable laws and regulations.

IEEE does not warrant or represent the accuracy or content of the material contained in its standards, and expressly disclaims all warranties (express, implied and statutory) not included in this or any other document relating to the standard, including, but not limited to, the warranties of: merchantability; fitness for a particular purpose; non-infringement; and quality, accuracy, effectiveness, currency, or completeness of material. In addition, IEEE disclaims any and all conditions relating to: results; and workmanlike effort. IEEE standards documents are supplied "AS IS" and "WITH ALL FAULTS."

Use of an IEEE standard is wholly voluntary. The existence of an IEEE standard does not imply that there are no other ways to produce, test, measure, purchase, market, or provide other goods and services related to the scope of the IEEE standard. Furthermore, the viewpoint expressed at the time a standard is approved and issued is subject to change brought about through developments in the state of the art and comments received from users of the standard.

In publishing and making its standards available, IEEE is not suggesting or rendering professional or other services for, or on behalf of, any person or entity nor is IEEE undertaking to perform any duty owed by any other person or entity to another. Any person utilizing any IEEE Standards document, should rely upon his or her own independent judgment in the exercise of reasonable care in any given circumstances or, as appropriate, seek the advice of a competent professional in determining the appropriateness of a given IEEE standard.

IN NO EVENT SHALL IEEE BE LIABLE FOR ANY DIRECT, INDIRECT, INCIDENTAL, SPECIAL, EXEMPLARY, OR CONSEQUENTIAL DAMAGES (INCLUDING, BUT NOT LIMITED TO: PROCUREMENT OF SUBSTITUTE GOODS OR SERVICES; LOSS OF USE, DATA, OR PROFITS; OR BUSINESS INTERRUPTION) HOWEVER CAUSED AND ON ANY THEORY OF LIABILITY, WHETHER IN CONTRACT, STRICT LIABILITY, OR TORT (INCLUDING NEGLIGENCE OR OTHERWISE) ARISING IN ANY WAY OUT OF THE PUBLICATION, USE OF, OR RELIANCE UPON ANY STANDARD, EVEN IF ADVISED OF THE POSSIBILITY OF SUCH DAMAGE AND REGARDLESS OF WHETHER SUCH DAMAGE WAS FORESEEABLE.

Translations

The IEEE consensus development process involves the review of documents in English only. In the event that an IEEE standard is translated, only the English version published by IEEE should be considered the approved IEEE standard.

Official statements

A statement, written or oral, that is not processed in accordance with the IEEE SA Standards Board Operations Manual shall not be considered or inferred to be the official position of IEEE or any of its committees and shall not be considered to be, or be relied upon as, a formal position of IEEE. At lectures, symposia, seminars, or educational courses, an individual presenting information on IEEE standards shall make it clear that his or her views should be considered the personal views of that individual rather than the formal position of IEEE.

Comments on standards

Comments for revision of IEEE Standards documents are welcome from any interested party, regardless of membership affiliation with IEEE. However, IEEE does not provide consulting information or advice pertaining to IEEE Standards documents. Suggestions for changes in documents should be in the form of a proposed change of text, together with appropriate supporting comments. Since IEEE standards represent a consensus of concerned interests, it is important that any responses to comments and questions also receive the concurrence of a balance of interests. For this reason, IEEE and the members of its societies and Standards Coordinating Committees are not able to provide an instant response to comments or questions except in those cases where the matter has previously been addressed. For the same reason, IEEE does not respond to interpretation requests. Any person who would like to participate in revisions to an IEEE standard is welcome to join the relevant IEEE working group.

Comments on standards should be submitted to the following address:

Secretary, IEEE SA Standards Board
445 Hoes Lane
Piscataway, NJ 08854 USA

Laws and regulations

Users of IEEE Standards documents should consult all applicable laws and regulations. Compliance with the provisions of any IEEE Standards document does not imply compliance to any applicable regulatory requirements. Implementers of the standard are responsible for observing or referring to the applicable regulatory requirements. IEEE does not, by the publication of its standards, intend to urge action that is not in compliance with applicable laws, and these documents may not be construed as doing so.

Copyrights

IEEE draft and approved standards are copyrighted by IEEE under US and international copyright laws. They are made available by IEEE and are adopted for a wide variety of both public and private uses. These include both use, by reference, in laws and regulations, and use in private self-regulation, standardization, and the promotion of engineering practices and methods. By making these documents available for use and adoption by public authorities and private users, IEEE does not waive any rights in copyright to the documents.

This is a preview of "DS/ISO/IEEE 11073-20...". Click here to purchase the full version from the ANSI store.

Photocopies

Subject to payment of the appropriate fee, IEEE will grant users a limited, non-exclusive license to photocopy portions of any individual standard for company or organizational internal use or individual, non-commercial use only. To arrange for payment of licensing fees, please contact Copyright Clearance Center, Customer Service, 222 Rosewood Drive, Danvers, MA 01923 USA; +1 978 750 8400. Permission to photocopy portions of any individual standard for educational classroom use can also be obtained through the Copyright Clearance Center.

Updating of IEEE Standards documents

Users of IEEE Standards documents should be aware that these documents may be superseded at any time by the issuance of new editions or may be amended from time to time through the issuance of amendments, corrigenda, or errata. An official IEEE document at any point in time consists of the current edition of the document together with any amendments, corrigenda, or errata then in effect.

Every IEEE standard is subjected to review at least every 10 years. When a document is more than 10 years old and has not undergone a revision process, it is reasonable to conclude that its contents, although still of some value, do not wholly reflect the present state of the art. Users are cautioned to check to determine that they have the latest edition of any IEEE standard.

In order to determine whether a given document is the current edition and whether it has been amended through the issuance of amendments, corrigenda, or errata, visit IEEE Xplore at <http://ieeexplore.ieee.org/> or contact IEEE at the address listed previously. For more information about the IEEE SA or IEEE's standards development process, visit the IEEE SA Website at <http://standards.ieee.org>.

Errata

Errata, if any, for IEEE standards can be accessed via <https://standards.ieee.org/standard/index.html>. Search for standard number and year of approval to access the web page of the published standard. Errata links are located under the Additional Resources Details section. Errata are also available in IEEE Xplore: <https://ieeexplore.ieee.org/browse/standards/collection/ieee/>. Users are encouraged to periodically check for errata.

Patents

Attention is called to the possibility that implementation of this standard may require use of subject matter covered by patent rights. By publication of this standard, no position is taken by the IEEE with respect to the existence or validity of any patent rights in connection therewith. If a patent holder or patent applicant has filed a statement of assurance via an Accepted Letter of Assurance, then the statement is listed on the IEEE SA Website at <https://standards.ieee.org/about/sasb/patcom/patents.html>. Letters of Assurance may indicate whether the Submitter is willing or unwilling to grant licenses under patent rights without compensation or under reasonable rates, with reasonable terms and conditions that are demonstrably free of any unfair discrimination to applicants desiring to obtain such licenses.

Essential Patent Claims may exist for which a Letter of Assurance has not been received. The IEEE is not responsible for identifying Essential Patent Claims for which a license may be required, for conducting inquiries into the legal validity or scope of Patents Claims, or determining whether any licensing terms or conditions provided in connection with submission of a Letter of Assurance, if any, or in any licensing agreements are reasonable or non-discriminatory. Users of this standard are expressly advised that determination of the validity of any patent rights, and the risk of infringement of such rights, is entirely their own responsibility. Further information may be obtained from the IEEE Standards Association.

Participants

At the time this standard was completed, the Personal Health Devices Working Group had the following membership:

Daidi Zhong, *Co-Chair*
Michael J. Kirwan, *Co-Chair*

Karsten Aalders
Charles R. Abbruscato
Nabil Abujbara
Maher Abuzaid
James Agnew
Haidar Ahmad
Manfred Aigner
Jorge Alberola
Rolf Ambuehl
David Aparisi
Lawrence Arne
Diego B. Arquillo
Serafin Arroyo
Muhammad Asim
Kit August
Doug Baird
David Baker
Anindya Bakshi
Ananth Balasubramanian
Sunlee Bang
M. Jonathan Barkley
Gilberto Barrón
David Bean
John Bell
Rudy Belliardi
Kathryn M. Bennett
Daniel Bernstein
George A. Bertos
Chris Biernacki
Ola Björnsne
Thomas Blackadar
Marc Blanchet
Thomas Bluethner
Douglas P. Bogia
Xavier Boniface
Shannon Boucousis
Julius Broma
Lyle G. Bullock, Jr.
Bernard Burg
Chris Burns
Anthony Butt
Jeremy Byford-Rew
Satya Calloji
Xiaoying Cao
Carole C. Carey
Craig Carlson
Santiago Carot-Nemesio
Randy W. Carroll
Simon Carter
Seungchul Chae
Rahul Chauhan

Wenjuan Chen
James Cheng
Peggy Chien
David Chiu
Jinyong Choi
Chia-Chin Chong
Saeed A. Choudhary
Jinhan Chung
John A. Cogan
John T. Collins
Cory Condek
Todd H. Cooper
David Cornejo
Douglas Coup
Nigel Cox
Hans Crommenacker
Tomio Crosley
Allen Curtis
Eyal Dassau
David Davenport
Russell Davis
Sushil K. Deka
Ciro de la Vega
Pedro de-las-Heras-Quiros
Jim Dello Stritto
Matthew d'Entremont
Kent Dicks
Hyoungdo Do
Alistair Donaldson
Xiaolian Duan
Brian Dubreuil
Sourav Dutta
Jakob Ehrensvar
Fredrik Einberg
Michihiro Enokida
Javier Escayola Calvo
Mark Estes
Leonardo Estevez
Hailing Feng
Bosco T. Fernandes
Christoph Fischer
Morten Flintrup
Joseph W. Forler
Russell Foster
Eric Freudenthal
Matthias Frohner
Ken Fuchs
Jing Gao
Qi Gao
Marcus Garbe
John Garguilo

Rick Geimer
Igor Gejdos
Ferenc Gerbovics
Alan Godfrey
Nicolae Goga
Julian Goldman
Raul Gonzalez Gomez
Chris Gough
Channa Gowda
Charles M. Gropper
Amit Gupta
Jeff Guttmacher
Rasmus Haahr
Christian Habermann
Michael Hagerty
Jerry Hahn
Robert Hall
Shu Han
Nathaniel Hamming
Rickey L. Hampton
Sten Hanke
Aki Harma
Jordan Hartmann
Kai Hassing
Wolfgang Heck
Nathaniel Heintzman
Charles Henderson
Jun-Ho Her
Helen B. Hernandez
Takashi Hibino
Timothy L. Hirou
Allen Hobbs
Alex Holland
Arto Holopainen
Kris Holtzclaw
Xinyi Hong
Robert Hoy
Frank Hsu
Anne Huang
Sen-Der Huang
Zhiyong Huang
Ron Huby
David Hughes
Robert D. Hughes
Jiyoung Huh
Hugh Hunter
Hitoshi Ikeda
Yutaka Ikeda
Philip O. Isaacson
Atsushi Ito
Michael Jaffe

This is a preview of "DS/ISO/IEEE 11073-20...". Click here to purchase the full version from the ANSI store.

Praduman Jain	Jim McCain	Jose A. Santos-Cadenas
Danny Jochelson	László Meleg	Stefan Sauermann
Phaneeth Junga	Alexander Mense	John Sawyer
Akiyoshi Kabe	Behnaz Minaei	Guillaume Schatz
Steve Kahle	Jinsei Miyazaki	Alois Schloegl
Tomio Kamioka	Erik Moll	Paul S. Schluter
James J. Kang	Darr Moore	Lars Schmitt
Kei Kariya	Carsten Mueglitz	Mark G. Schnell
Andy Kaschl	Piotr Murawski	Richard A. Schrenker
Junzo Kashiwara	Soundharya Nagasubramanian	Antonio Scorpiniti
Kohichi Kashiwagi	Jae-Wook Nah	KwangSeok Seo
Ralph Kent	Alex Neefus	Riccardo Serafin
Laurie M. Kermes	Trong-Nghia Nguyen-Dobinsky	Sid Shaw
Ikuo Keshi	Michael E. Nidd	Frank Shen
Ahmad Kheirandish	Jim Niswander	Min Shih
Junhyung Kim	Hongliang Niu	Mazen Shihabi
Minho Kim	Hiroaki Niwamoto	Redmond Shouldice
Min-Joon Kim	Thomas Norgall	Sternly K. Simon
Taekon Kim	Anand Noubade	Marjorie Skubic
Tetsuya Kimura	Yoshiteru Nozoe	Robert Smith
Alfred Kloos	Abraham Ofek	Ivan Soh
Jeongmee Koh	Brett Olive	Motoki Sone
Jean-Marc Koller	Begonya Otal	Emily Sopensky
John Koon	Marco Paleari	Rajagopalan Srinivasan
Patty Krantz	Bud Panjwani	Nicholas Steblay
Raymond Krasinski	Carl Pantiskas	Lars Steubesand
Alexander Kraus	Harry P. Pappas	John (Ivo) Stivorice
Ramesh Krishna	Hanna Park	Raymond A. Strickland
Geoffrey Kruse	Jong-Tae Park	Chandrasekaran Subramaniam
Falko Kuester	Myungeun Park	Hermann Suominen
Rafael Lajara	Soojun Park	Lee Surprenant
Pierre Landau	Phillip E. Pash	Ravi Swami
Jaechul Lee	TongBi Pei	Ray Sweidan
JongMuk Lee	Lucian Pestritu	Yi Tang
Kyong Ho Lee	Soren Petersen	Haruyuyki Tatsumi
Rami Lee	James Petisce	Isabel Tejero
Sungkee Lee	Peter Piction	John W. Thomas
Woojae Lee	Michael Pliskin	Tom Thompson
Yonghee Lee	Varshney Prabodh	Jonas Tirén
Joe Lenart	Jeff Price	Alexandra Todiruta
Kathryn A. Lesh	Harald Prinzhorn	Janet Traub
Qiong Li	Harry Qiu	Jesús Daniel Trigo
Xiangchen Li	Arif Rahman	Gary Tschautscher
Yingsong Li	Tanzilur Rahman	Masato Tsuchid
Zhuofang Li	Steve Ray	Ken Tubman
Patrick Lichter	Phillip Raymond	Yoshihiro Uchida
Jisoon Lim	Terrie Reed	Akib Uddin
Joon-Ho Lim	Tim Reilly	Sunil Unadkat
John Lin	Barry Reinhold	Fabio Urbani
Xiaoming Liu	Brian Reinhold	Philipp Urbauer
Wei-Jung Lo	Melvin I. Reynolds	Laura Vanzago
Charles Lowe	John G. Rhoads	Alpo Värri
Don Ludolph	Jeffrey S. Robbins	Andrei Vasilateanu
Christian Luszick	Chris Roberts	Dalimar Velez
Bob MacWilliams	Moskowitz Robert	Martha Velezis
Srikkanth Madhurbootheswaran	Scott M. Robertson	Rudi Voon
Miriam L. Makhoul	Timothy Robertson	Barry Vornbrock
Romain Marmot	Patricia Order	Isobel Walker
Sandra Martinez	David Rosales	David Wang
Miguel Martínez de Espronceda	Fatemeh Saki	Jerry P. Wang
Cámara	Bill Saltzstein	Yao Wang
Peter Mayhew	Giovanna Sannino	Yi Wang

Steve Warren
Fujio Watanabe
Toru Watsuji
Mike Weng
Yuefeng Weng
Kathleen Wible
Paul Williamson
Jan Wittenber

Jia-Rong Wu
Will Wykeham
Ariton Xhafa
Ricky Yang
Shaoqin Ye
Melanie S. Yeung
Qiang Yin
Done-Sik Yoo

Zhi Yu
Jianchao Zeng
Jason Zhang
Zhiqiang Zhang
Thomas Zhao
Hongyuan Zhong
Miha Zoubek
Szymon Zyskoter

The following members of the individual Standards Association balloting group voted on this standard. Balloters may have voted for approval, disapproval, or abstention.

Juan Carreon
Randy W. Carroll
Lawrence Catchpole
Keith Chow
Malcolm Clarke
Russell Davis
David Fuschi
Randall Groves
Nathaniel Hamming

Werner Hoelzl
Noriyuki Ikeuchi
Atsushi Ito
Stefan Karl
Piotr Karocki
Stuart Kerry
Joerg-Uwe Meyer
Michael Newman

Beth Pumo
Bartien Sayogo
Paul S. Schluter
Carl Singer
Walter Struppler
Lai King (Anna) Tee
Jan Wittenber
Oren Yuen
Daidi Zhong

When the IEEE SA Standards Board approved this standard on 5 September 2019, it had the following membership:

Gary Hoffman, *Chair*
Ted Burse, *Vice Chair*
Jean-Philippe Faure, *Past Chair*
Konstantinos Karachalios, *Secretary*

Masayuki Ariyoshi
Stephen D. Dukes
J. Travis Griffith
Guido Hiertz
Christel Hunter
Joseph L. Koepfinger*
Thomas Koshy
John D. Kulick

David J. Law
Joseph Levy
Howard Li
Xiaohui Liu
Kevin Lu
Daleep Mohla
Andrew Myles

Annette Reilly
Dorothy Stanley
Sha Wei
Phil Wennblom
Philip Winston
Howard Wolfman
Feng Wu
Jingyi Zhou

*Member Emeritus

This is a preview of "DS/ISO/IEEE 11073-20...". Click here to purchase the full version from the ANSI store.

Introduction

This introduction is not part of IEEE Std 11073-20601-2019, Health informatics—Personal health device communication—Part 20601: Application profile—Optimized Exchange Protocol.

ISO and IEEE 11073 standards enable communication between medical devices and external computer systems. This standard and corresponding IEEE 11073-104zz standards address a need for a simplified and optimized communication approach for personal health devices, which may or may not be regulated devices. These standards align with, and draw upon, the existing clinically focused standards to provide easy management of data from either a clinical or personal health device.

This document addresses a need for an openly defined, independent standard for converting the collected information into an interoperable transmission format so the information can be exchanged between agents and managers.

Other closely related standards include the following:

- IEEE Std 11073-00103™ [B6]¹ provides an overview of the personal health space and defines the underlying use cases and usage models.
- ISO/IEEE 11073-10201:2004 [B17] documents the extensive domain information model (DIM) leveraged by this standard.
- ISO/IEEE 11073-104zz standards define specific device specializations. For example, ISO/IEEE 11073-10404 [B18] defines how interoperable pulse oximeters work.
- ISO/IEEE 11073-20101:2004 [B21] defines the medical device encoding rules (MDER) used in this standard.

¹ The numbers in brackets correspond to the numbers of the bibliography in Annex M.

Contents

1. Overview	13
1.1 Scope	13
1.2 Purpose	13
1.3 Context	13
1.4 Word usage	17
2. Normative references.....	17
3. Definitions, acronyms, and abbreviations	18
3.1 Definitions	18
3.2 Acronyms and abbreviations	19
4. Guiding principles	20
5. Introduction to IEEE 11073 personal health devices.....	21
5.1 General	21
5.2 Domain information model (DIM)	21
5.3 Service model	21
5.4 Communication model.....	22
5.5 Compliance with other standards.....	22
5.6 Security.....	22
6. Personal health device DIM	22
6.1 General	22
6.2 Nomenclature usage	23
6.3 Personal health object class definitions	24
6.4 Information model extensibility rules.....	83
7. Personal health device service model.....	83
7.1 General	83
7.2 Association service	84
7.3 Object access services.....	84
7.4 Specific application of object access EVENT REPORT services for personal health devices.....	85
8. Communication model	93
8.1 General	93
8.2 System context.....	93
8.3 Communications characteristics	94
8.4 State machines	97
8.5 Connected procedure	103
8.6 Unassociated procedure	103
8.7 Associating procedure	104
8.8 Configuring procedure.....	111
8.9 Operating procedure	115
8.10 Disassociating procedure	130
8.11 Message encoding.....	131
8.12 Time coordination.....	131
9. Conformance model	135
9.1 Applicability	135
9.2 Conformance specification	135
9.3 Implementation conformance statements (ICSs).....	136
9.4 General conformance.....	136
9.5 Device additions/extensions ICS	140

This is a preview of "DS/ISO/IEEE 11073-20...". Click here to purchase the full version from the ANSI store.

Annex A (normative) ASN.1 definitions.....	144
A.1 General.....	144
A.2 Common data types	144
A.3 Attribute data types.....	150
A.4 ACTION-method-related data types.....	158
A.5 Message-related data types	160
A.6 Other	160
A.7 Personal health device protocol frame.....	160
A.8 Association protocol definitions	161
A.9 Presentation protocol definitions	164
A.10 Data protocol definitions	164
A.11 Data types for new object attributes and object services	169
Annex B (informative) Scale and range specification example.....	185
B.1 General.....	185
B.2 Thermometer example	185
Annex C (informative) The PM-store concept	187
C.1 General.....	187
C.2 Persistent metric (PM) store object hierarchy	188
Annex D (informative) Transport profile types	192
D.1 General.....	192
D.2 Type 1	192
D.3 Type 2	193
D.4 Type 3	193
D.5 Summary.....	194
Annex E (normative) State tables.....	195
E.1 General.....	195
E.2 Events.....	195
E.3 Agent state table.....	197
E.4 Manager state table	206
Annex F (normative) Medical device encoding rules (MDER).....	215
F.1 General	215
F.2 Supported ASN.1 syntax	215
F.3 Byte order.....	216
F.4 Encodings	218
F.5 Floating point numbers.....	224
F.6 Floating point data structure—FLOAT-Type.....	224
F.7 Short floating point data structure—SFLOAT-Type.....	225
F.8 Expression of precision of floating point numbers.....	226
Annex G (informative) Encoded data type definitions	227
Annex H (informative) Examples.....	249
H.1 General.....	249
H.2 Weighing scale.....	249
H.3 Pulse oximeter	255
H.4 PM-store and PM-segment transactions	256
Annex I (normative) Nomenclature codes.....	265
Annex J (informative) Derivation and modification history.....	272
J.1 General.....	272
J.2 ASN.1 structures	272
J.3 Medical device encoding rules (MDER).....	272
J.4 Nomenclature codes	272

This is a preview of "DS/ISO/IEEE 11073-20...". Click [here](#) to purchase the full version from the ANSI store.

Annex K (informative) The schedule-store concept.....	275
K.1 General.....	275
K.2 Schedule-store object hierarchy.....	276
Annex L (informative) Revision history.....	279
Annex M (informative) Bibliography.....	280

Health informatics—Personal health device communication

Part 20601: Application profile— Optimized Exchange Protocol

1. Overview

1.1 Scope

Within the context of the ISO/IEEE 11073 personal health device standard family, this standard defines an optimized exchange protocol and modeling techniques to be used by implementers of personal health devices to create interoperability between device types and vendors. This standard establishes a common framework for an abstract model of personal health data available in transport-independent transfer syntax required to establish logical connections between systems and to provide presentation capabilities and services needed to perform communication tasks. The protocol is optimized to personal health usage requirements and leverages commonly used methods and tools wherever possible.

1.2 Purpose

This standard addresses a need for an openly defined, independent standard for controlling information exchange to and from personal health devices and managers (e.g., cell phones, personal computers, personal health appliances, and set top boxes). Interoperability is key to growing the potential market for these devices and enabling people to be better-informed participants in the management of their health.

1.3 Context

Figure 1 shows categories and typical devices supporting the personal health space. Agents (e.g., blood pressure monitors, weighing scales, and pedometers) collect information about a person (or persons) and transfer the information to a manager (e.g., cell phone, health appliance, or personal computer) for collection, display, and possible later transmission. The manager may also forward the data to remote support services for further analysis. The information is available from a range of domains including disease management, health and fitness, or aging independently applications.

Part 20601: Application profile—Optimized Exchange Protocol

The communication path between agent and manager is assumed to be a logical point-to-point connection. Generally, an agent communicates with a single manager at any point in time. A manager may communicate with multiple agents simultaneously using separate point-to-point connections.




Figure 1—Overall context of work

The overlay shows the focus area of the IEEE 11073™ Personal Health Devices Working Group. The primary concentration is the interface and data exchange between the agents and manager. However, this interface cannot be created in isolation by ignoring the remainder of the solution space. Remaining cognizant of the entire system helps to move data reasonably from the agents all the way to the remote support services when necessary. This path may include converting the data format, exchange protocols, and transport protocols across different interfaces. Much of the standardization effort is outside of the scope of the Personal Health Devices Working Group; however, aligning all standardization efforts allows data to flow seamlessly through the overall set of systems.

Figure 2 shows a hierarchical view of the architecture of an agent or manager superimposed with a view of the related standards. The application layers are, for the most part, not specific to any particular transport. Where necessary, this standard identifies assumptions that require direct support by a transport or a “shim” layer above the transport. This approach allows support for various transports. The definition of the transports is outside the scope of this standard and the working group.

This is a preview of "DS/ISO/IEEE 11073-20...". Click here to purchase the full version from the ANSI store.

Part 20601: Application profile—Optimized Exchange Protocol




Figure 2—Document map

Above the transport layer is the Optimized Exchange Protocol. This protocol consists of two aspects: the application layer services and the definition of the data-exchange protocol between agents and managers. The application layer services provide the protocol for connection management and reliable transfer of actions and data between agent and manager. The data-exchange protocol defines the commands, agent configuration information, data format, and overall protocol. The Optimized Exchange Protocol provides the basis to support any type of agent. For a specific device type, the reader is directed to the device specialization for that agent to understand the capabilities of the device and its implementation according to this standard. The device specialization indicates which aspects of this standard to comprehend and where further information to implement the device is found.

Above the exchange protocol are device specializations that describe specific details relative to the particular agent (e.g., blood pressure monitor, weighing scale, or pedometer). The specializations describe the details of how these agents work and act as a detailed description for creating a specific type of agent. Additionally, they provide reference to a related standard for further details. The standard numbers reserved for device specializations range from IEEE Std 11073-10401 through IEEE Std 11073-10499, inclusive. When the collection of standards is being referenced, the term *IEEE 11073-104zz* is used where *zz* could be any number in the range from 01 to 99, inclusive.

Some device specializations describe broad categories of device types (e.g., the IEEE 11073-10441™ model device types that promote cardiovascular activity such as step counters or exercise cycles). Other device specializations have a narrow focus on a single device type (e.g., IEEE 11073-10408™ model thermometers). Specializations that address one or more device types may also define *profiles*. A profile further constrains the model defined in a specialization to increase interoperability (e.g., the step counter profile utilizes a limited portion of IEEE 11073-10441 modeling).

Two device specializations are defined by this document itself, “hydra” and “generic”. These two specializations are not covered in a separate IEEE 11073-104zz standard. These specializations are defined to support signaling of multi-function or generic agents and managers at the transport level. See 7.4.3.4 for more details.

Note that, in the medical device system (MDS) object defined in this specification, an agent can indicate the full list of supported specializations.

Part 20601: Application profile—Optimized Exchange Protocol

IEEE Std 11073-00103™ [B6]² technical report describes the overall personal health space with further definition of the underlying use cases and usage models.

The personal health device specializations are not being created independently of all other standards. There are a number of existing standards generated for clinical environments upon which these standards draw. Figure 3 shows the relationship to the remainder of the IEEE 11073 documents. There are two types of relationships:

- Drawing ideas and/or content from the other documents (dashed lines).
- Leveraging information from the other document and introducing new content into that document to support this standard (solid lines).




Figure 3—Relationship to other IEEE 11073 documents

This standard imports information from ISO/IEEE 11073-10201:2004 [B17] and ISO/IEEE 11073-20101:2004 [B21] as normative annexes. If there is a discrepancy between these standards, this standard

² The numbers in brackets correspond to the numbers of the bibliography in Annex M.

takes priority. Because of the reuse of constructs from these standards, some of the names appear to be more clinically focused (e.g., MDS instead of personal health device system); however, to maintain consistency, the traditional names have been preserved.

This standard reproduces relevant portions of IEEE Std 11073-1010. ³ Between this standard, IEEE Std 11073-10101, and IEEE 11073-104zz, all required nomenclature codes for implementation are documented. New codes may be defined in newer versions/revisions of each of these documents. In the case of a conflict, where one term code has been assigned to two separate semantic concepts with different RefIDs, in general the oldest definition that is in actual use should take precedence. The same policy applies when one RefID has two different code values assigned in different specifications. The resolution of such conflicts will be determined through joint action by the responsible working groups and other stakeholders, and any corrective actions will be published as corrigenda.

1.4 Word usage

The word *shall* indicates mandatory requirements strictly to be followed in order to conform to the standard and from which no deviation is permitted (*shall* equals *is required to*).^{4,5}

The word *should* indicates that among several possibilities one is recommended as particularly suitable, without mentioning or excluding others; or that a certain course of action is preferred but not necessarily required (*should* equals *is recommended that*).

The word *may* is used to indicate a course of action permissible within the limits of the standard (*may* equals *is permitted to*).

The word *can* is used for statements of possibility and capability, whether material, physical, or causal (*can* equals *is able to*).

2. Normative references

The following referenced documents are indispensable for the application of this document (i.e., they must be understood and used, so each referenced document is cited in text and its relationship to this document is explained). For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments or corrigenda) applies.

IEEE Std 802®-2014, IEEE Standard for Local and Metropolitan Area Networks: Overview and Architecture.^{6,7}

IEEE Std 1541™-2002 (Reaff 2008), IEEE Standard for Prefixes for Binary Multiples.

IEEE Std 11073-10101™, Health informatics—Point-of-care medical device communication—Part 10101: Nomenclature.

³ Information about normative references can be found in Clause 2.

⁴ The use of the word *must* is deprecated and cannot be used when stating mandatory requirements; *must* is used only to describe unavoidable situations.

⁵ The use of *will* is deprecated and cannot be used when stating mandatory requirements; *will* is only used in statements of fact.

⁶ IEEE publications are available from The Institute of Electrical and Electronics Engineers (<http://standards.ieee.org>).

⁷ The IEEE standards or products referred to in this clause are trademarks of The Institute of Electrical and Electronics Engineers, Inc.

This is a preview of "DS/ISO/IEEE 11073-20...". Click here to purchase the full version from the ANSI store.

Part 20601: Application profile—Optimized Exchange Protocol

ISO/IEC 80000-13:2008, Quantities and units—Part 13: Information science and technology.^{8, 9}

ITU-T Rec. X.667 (09/2004), Information technology—Open Systems Interconnection—Procedures for the operation of OSI Registration Authorities: Generation and registration of universally unique identifiers (UUIDs) and their use as ASN.1 object identifier components.¹⁰

⁸ This standard cancels and replaces sections 3.8 and 3.9 of IEC 60027-2 (2005).

⁹ ISO/IEC publications are available from the International Organization for Standardization (<http://www.iso.ch/>). ISO/IEC publications are also available in the United States from Global Engineering Documents (<http://global.ihs.com/>). Electronic copies are available in the United States from the American National Standards Institute (<http://www.ansi.org/>).

¹⁰ ITU-T publications are available from the International Telecommunications Union (<http://www.itu.int/>).