

This is a preview of "IEC/TS 60318-7 Ed. 1...". Click here to purchase the full version from the ANSI store.



Edition 1.0 2011-02

# TECHNICAL SPECIFICATION

---

**Electroacoustics – Simulators of human head and ear –  
Part 7: Head and torso simulator for the measurement of hearing aids**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

PRICE CODE



---

ICS 17.140.50

ISBN 978-2-88912-379-7

## CONTENTS

FOREWORD.....	3
INTRODUCTION.....	5
1 Scope.....	6
2 Normative references.....	6
3 Terms and definitions.....	6
4 Construction.....	11
4.1 General.....	11
4.2 Geometrical dimensions of the manikin.....	11
4.2.1 Torso.....	11
4.2.2 Head.....	18
4.2.3 Pinna.....	24
4.2.4 Ear simulator.....	28
4.2.5 Materials.....	29
4.2.6 Markings.....	29
4.2.7 Tolerances.....	29
4.3 Acoustical characteristics of the manikin.....	29
4.3.1 Free field frequency response.....	29
4.3.2 Tolerances.....	29
4.3.3 Openings.....	30
5 Calibration.....	30
5.1 Atmospheric reference conditions.....	30
5.2 Calibration method.....	30
5.2.1 General.....	30
5.2.2 Test space and measurement equipment.....	30
5.2.3 Measurement of sound pressure level.....	31
5.2.4 Alignment of manikin azimuth and elevation.....	31
5.2.5 Test for sound leakage.....	31
6 Maximum permitted expanded uncertainty of measurements.....	32
Annex A (informative) Ear canal sound pressure ratio.....	34
Bibliography.....	35
Figure 1 – Manikin geometrical references.....	9
Figure 2 – Coordinate scheme for azimuth and elevation angles.....	10
Figure 3 – Illustration of manikin head and torso dimensions.....	12
Figure 4 – Cross-sections of torso (tolerance $\pm 4$ mm).....	17
Figure 5 – Cross-sections of head (tolerance $\pm 2,5$ mm).....	23
Figure 6 – Illustration of manikin pinna dimensions.....	24
Figure 7 – Cross-sections of right pinna (tolerance $\pm 1,5$ mm).....	28
Table 1 – Head and torso dimensions for the manikin and average human data.....	12
Table 2 – Pinna dimensions for the manikin.....	24
Table 3 – Free field frequency response of the manikin.....	32
Table 4 – Values of $U_{\max}$ for basic measurements.....	33
Table A.1 – Ear canal sound pressure ratio.....	34

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

### **ELECTROACOUSTICS – SIMULATORS OF HUMAN HEAD AND EAR –**

#### **Part 7: Head and torso simulator for the measurement of hearing aids**

#### FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. In exceptional circumstances, a technical committee may propose the publication of a technical specification when

- the required support cannot be obtained for the publication of an International Standard, despite repeated efforts, or
- the subject is still under technical development or where, for any other reason, there is the future but no immediate possibility of an agreement on an International Standard.

Technical specifications are subject to review within three years of publication to decide whether they can be transformed into International Standards.

IEC/TS 60318-7, which is a technical specification, has been prepared by IEC technical committee 29: Electroacoustics.

This is a preview of "IEC/TS 60318-7 Ed. 1...". [Click here to purchase the full version from the ANSI store.](#)

The text of this technical specification is based on the following documents:

Enquiry draft	Report on voting
29/716/DTS	29/729A/RVC

Full information on the voting for the approval of this technical specification can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts of the IEC 60318 series, published under the general title *Electroacoustics – Simulators of human head and ear*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- transformed into an International standard,
- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

A bilingual version of this document may be issued at a later date.

This is a preview of "IEC/TS 60318-7 Ed. 1...". [Click here to purchase the full version from the ANSI store.](#)

## INTRODUCTION

This technical specification describes a head and torso simulator for hearing aid measurements. It has been developed as a revision of IEC/TR 60959 (1990). The main changes, as compared to the technical report, are the insertion of maximum permitted expanded measurement uncertainties and additional references in the Bibliography. A future IEC Standard which will include additional models of head and torso simulators for hearing aid measurements is planned.

## **ELECTROACOUSTICS – SIMULATORS OF HUMAN HEAD AND EAR –**

### **Part 7: Head and torso simulator for the measurement of hearing aids**

#### **1 Scope**

This part of IEC 60318 describes a head and torso simulator (manikin) intended for the measurement of air-conduction hearing aids in the frequency range from 100 Hz to 10 000 Hz. The device consists of a head mounted on a torso that extends to the waist. The head is equipped with simulated pinnae and with cylindrical cavities having acoustic impedance terminations and microphones located at positions corresponding to those of the eardrums in a median human adult. It has been designed to provide acoustic diffraction similar to that encountered around the median human head and torso.

The device with its present pinna simulator, however, is not suitable for the measurement of all types of hearing aids. For example, most in-the-ear (ITE) and completely-in-the-canal (CIC) hearing aids cannot be measured correctly.

The manikin is specified in terms of both, its geometrical dimensions and its acoustical properties.

NOTE 1 Measurement results obtained with a manikin may differ substantially from similar results obtained on an individual person, due to anatomical variations.

NOTE 2 The median values of the human head and torso were drawn from the population samples described in [5]<sup>1</sup>.

NOTE 3 It is acknowledged that devices conforming to this part of IEC 60318 are also used as the basis for applications extending beyond this Scope. In such cases it is recommended that any design variations that are necessary are documented, and that a statistical analysis of the measurement data be carried out to determine the level of repeatability that can be achieved. It will also be necessary to consider the relevance of the measurements made with the head and torso simulator to the application in question.

#### **2 Normative references**

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60318-4, *Simulators of human head and ear – Part 4: Occluded-ear simulator for the measurement of earphones coupled to the ear by means of ear inserts*

ISO/IEC Guide 98-3, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

#### **3 Terms and definitions**

For the purposes of this document, the following terms and definitions apply. Reference is also made to Figure 1 and Figure 2.

---

<sup>1</sup> Numbers in square brackets refer to the Bibliography.