



INTERNATIONAL STANDARD

**Radiation protection instrumentation - Equipment for continuous monitoring of
radioactivity in gaseous effluents -
Part 2: Specific requirements for radioactive aerosol monitors including
transuranic aerosols**

This is a preview of IEC 60761-2 Ed. 3.0 en:2026. [Click here to purchase the full version from the ANSI store.](#)

FOREWORD	3
INTRODUCTION	5
1 Scope	6
2 Normative references	6
3 Terms and definitions, abbreviated terms, quantities and units	6
3.1 Terms and definitions	6
3.2 Abbreviated terms	7
3.3 Quantities and units	7
3.4 Classification of aerosol effluent monitors	8
4 General characteristics	8
4.1 General	8
4.2 Alarm characteristics	8
4.3 Explosive atmospheres	8
5 General requirements	9
5.1 Nature of tests	9
5.2 Uncertainties	9
5.3 Air flow	9
5.3.1 General	9
5.3.2 Sampled volume correctness	9
5.3.3 Flow-rate robustness	10
5.3.4 Leak	10
5.4 Aerosol sampling efficiency	10
5.5 Particle collection efficiency	11
5.5.1 General	11
5.5.2 Requirement	11
5.5.3 Method of test	11
5.6 Sampling inlet and delivery line	13
5.7 Expression of measurement	13
5.8 Reference sources	13
5.9 Radiation detector	14
5.10 Ease of decontamination	14
6 Radiation detection requirements	14
6.1 General	14
6.2 Reference response	15
6.2.1 Requirements	15
6.2.2 Method of test	15
6.3 Variation of the detection efficiency as a function of activity - Linearity	15
6.3.1 General	15
6.3.2 Requirements	15
6.3.3 Method of test	15
6.4 Variation of the detection efficiency as a function of radiation energy	16
6.4.1 General	16
6.4.2 Requirements	16
6.4.3 Method of test	16
6.5 Cross efficiency	16
6.5.1 Requirements	16

This is a preview of IEC 60761-2 Ed. 3.0 en:2026. [Click here to purchase the full version from the ANSI store.](#)

6.6	Response to radioactive gases	17
6.6.1	Requirements	17
6.6.2	Method of test	17
6.7	Response to ^{222}Rn and ^{220}Rn daughters	17
6.8	Compensation of natural activity	18
6.8.1	General	18
6.8.2	Requirements	18
6.8.3	Method of test	18
6.9	Reference response with natural activity	18
6.9.1	Requirements	18
6.9.2	Method of test	18
6.10	Response to aerosol granulometry variation (optional)	19
6.10.1	General	19
6.10.2	Requirement	19
6.10.3	Method of test	19
7	Environmental requirements	19
8	Mechanical requirements	19
9	Electromagnetic requirements	19
10	Documentation	19
10.1	Operation and maintenance manual	19
10.2	Type test report	20
Annex A (informative) List of radionuclides suitable for tests with variation of beta energy		24
Bibliography		26
Table 1 – Reference conditions and standard test conditions		20
Table 2 – Tests performed under standard test conditions		21
Table 3 – Tests performed with variation of influence quantities		21
Table 4 – Tests of air circuit		23
Table A.1 – List of radionuclides suitable for tests with variation of beta energy		24
Table A.2 – List of radionuclides suitable for tests with alpha emissions		25

**Radiation protection instrumentation - Equipment for continuous
monitoring of radioactivity in gaseous effluents -
Part 2: Specific requirements for radioactive aerosol
monitors including transuranic aerosols**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 60761-2 has been prepared by subcommittee 45B: Radiation protection instrumentation, of IEC technical committee 45: Nuclear instrumentation. It is an International Standard.

This International Standard is to be used in conjunction with IEC 60761-1:2002.

This third edition cancels and replaces the second edition published in 2002. This edition constitutes a technical revision.

This is a preview of IEC 60761-2 Ed. 3.0 en:2026. Click here to purchase the full version from the ANSI store.

edition:

- a) more precise tests for air-flow were added:
 - 1) sampled volume correctness;
 - 2) flow-rate robustness;
- b) uncertainties have been taken into account for the reference response test;
- c) addition of tests against aerosol granulometry variation;
- d) creating a uniform functionality test for all environmental, electromagnetic and mechanical tests and a requirement for the coefficient of variation of each nominal mean reading.

The text of this International Standard is based on the following documents:

Draft	Report on voting
45B/1109/FDIS	45B/1114/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

A list of all parts in the IEC 60761 series, published under the general title *Radiation protection instrumentation - Equipment for continuous monitoring of radioactivity in gaseous effluents*, can be found on the IEC website.

Future documents in this series will carry the new general title as cited above. Titles of existing documents in this series will be updated at the time of the next edition.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

This is a preview of IEC 60761-2 Ed. 3.0 en:2026. [Click here to purchase the full version from the ANSI store.](#)

IEC 60761 consists of the following parts, under the general title: *Radiation protection instrumentation - Equipment for continuous monitoring of radioactivity in gaseous effluents.*

Part 1: General requirements

Part 2: Specific requirements for radioactive aerosol monitors including transuranic aerosols

Part 3: Specific requirements for radioactive noble gas monitors

Part 4: Specific requirements for radioactive iodine monitors

Part 5: Specific requirements for tritium monitors

This is a preview of IEC 60761-2 Ed. 3.0 en:2026. [Click here to purchase the full version from the ANSI store.](#)

This part of IEC 60761 is applicable to equipment intended for simultaneous, delayed or discrete sequential measurement of aerosols in gaseous effluents discharged into the environment.

It is applicable to equipment designed to fulfil the following functions:

- the measurement of the volumetric activity (Bq/m^3) of the aerosols in either gaseous effluents or the released total activity of aerosols (Bq), or both;
- the actuation of an alarm signal when either a predetermined volumetric activity or a predetermined total released activity of aerosols is exceeded.

This equipment is intended for measurement over a wide range of activity, including very small quantities in the presence of a much larger natural background. The daughters of ^{222}Rn (radon) and ^{220}Rn (thoron) are naturally occurring aerosols contributing to the natural background. The discrimination against natural activity can be an important problem in monitoring low level activity. In order to provide more and better information, complementary or retrospective laboratory analysis of the filters after collection can be performed.

The objective of this document is to establish specific standard requirements, including technical characteristics and general test conditions, and to give examples of acceptable methods for aerosol effluent monitors.

The general requirements, technical characteristics, test procedures, radiation characteristics, electrical, mechanical, safety and environmental characteristics are given in IEC 60761-1. Unless otherwise stated, these requirements apply to this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60050-395, *International Electrotechnical Vocabulary - Part 395: Nuclear instrumentation: Physical phenomena, basic concepts, instruments, systems, equipment and detectors*

IEC 60761-1:2002, *Equipment for continuous monitoring of radioactivity in gaseous effluents - Part 1: General requirements*

IEC 61578:1997, *Radiation protection instrumentation - Calibration and verification of the effectiveness of radon compensation for alpha and/or beta aerosol measuring instruments - Test methods*

This is a preview of IEC 60761-2 Ed. 3.0 en:2026. [Click here to purchase the full version from the ANSI store.](#)

IEC 60068-2-27, *Environmental testing - Part 2-27: Tests - Test Ea and guidance: Shock*

IEC 60079-11:2023, *Explosive atmospheres - Part 11: Equipment protection by intrinsic safety “i”*

IEC 61000 (all parts), *Electromagnetic compatibility (EMC)*

IEC 61000-4-4, *Electromagnetic compatibility (EMC) - Part 4-4: Testing and measurement techniques – Electrical fast transient/burst immunity test*

EN 481:1993, *Workplace atmospheres - Size fraction definitions for measurement of airborne particles*

EN 55022, *Information technology equipment - Radio disturbance characteristics - Limits and methods of measurement*

HOARAU, G., DOUGNIAUX, G. et al. Impact of the coarse indoor non-radioactive aerosols on the background radon progenies compensation of continuous air monitor, *Health Physics*, 2022, 122 (5), pp.563-574. 10.1097/HP.0000000000001530

KORDAS, J.F., *The Transuranic Aerosol Measurement System and Its Field Results*, Lawrence Livermore National Laboratory, Livermore, CA, UCRL-83533 (1979)

MERCER, T.T. and STOWE, W.A., *Deposition of Unattached Radon Decay Product in an Impactor Stage*, *Health Physic.*, 17, 259 (1959)

RUSH, G.K. and McDOWEL, W.P., *The ZPR-9 Airborne Plutonium Monitoring System*, *IEEE Trans. Nucle. Sci. NS-23*, p. 690 (1976)

WOOLAM, P.B., *An On-Line Actinide in Air Monitor to Operate at Concentrations below 0.1*, ICRP MPCa, CEGB BNL TPRD/B 0731/N85, November 1985