

ANSI/IESNA RP-27.2-00



Recommended
Practice *for*
Photobiological Safety
for Lamps &
Lamp Systems-
Measurement
Techniques

IES

The
LIGHTING
AUTHORITY®

Prepared by the IESNA Photobiology Committee

ANSI/IESNA RP-27.2-00

Recommended Practice for Photobiological Safety for Lamps and Lamp Systems-Measurement Techniques

Publication of this Committee Report has been approved by the IESNA. Suggestions for revisions should be directed to the IESNA.

Prepared by:

The IESNA Photobiology Committee

Copyright 2001 by the Illuminating Engineering Society of North America.

Approved by the IESNA Board of Directors August 3, 2000, as a Transaction of the Illuminating Engineering Society of North America.

Approved by the American National Standards Institute as an American National Standard on October 6, 2000.

All rights reserved. No part of this publication may be reproduced in any form, in any electronic retrieval system or otherwise, without prior written permission of the IESNA.

Published by the Illuminating Engineering Society of North America, 120 Wall Street, New York, New York 10005.

IESNA Standards and Guides are developed through committee consensus and produced by the IESNA Office in New York. Careful attention is given to style and accuracy. If any errors are noted in this document, please forward them to Rita Harrold, Director Educational and Technical Development, at the above address for verification and correction. The IESNA welcomes and urges feedback and comments.

ISBN # 0-87995-164-8

Printed in the United States of America.

Prepared by the IESNA Photobiology Committee

George C. Brainard, *Chair*

Sharon Miller, *Vice Chair*

D. S. Berger
K. H. Canning
D. Ciolkosz
A. M. Clarke
R. G. Collins*
R. O. Daubach*
H. D. Edmunds*
H. Edwards
D. H. Epley
P. D. Forbes
D.C. Gross*
H. L. Hoover
R. Kaiser*
R. Landry*
R. E. Levin
A. L. Lewis*
R. D. Liddle*
S. L. Martel
S. L. Mintz
C. E. Moss
C. C. Mpelkas
A. C. Rousseau
R. M. Sayre
R. A. Schiele*
J. B. Sheehy
D. H. Sliney
A. Thompson
F. Urbach
R. Vincent
M. Waxler*
D. E. Wiedrich
M. L. Wolbarsht
A. M. Zielinsky
S. Zigman*

* *Advisory*

Special thanks to David Gross, Chair of the authoring task force and contributors: Dan Berger, Rolf Bergman, Kevin Carr, David Ellis, Anne Georgalas, Herbert Hoover, Ralph Kaiser, Robert Landry, Robert Levin, Teresa Mott West, David Sliney, Ambler Thompson, Richard Young.

This is a preview of "ANSI/IESNA RP-27.2-0...". [Click here to purchase the full version from the ANSI store.](#)

Contents

1.0 Introduction	1
2.0 Scope	1
3.0 References	1
3.1 Normative References	1
3.2 General References	2
4.0 Measurement Conditions	2
4.1 Lamp Seasoning	2
4.2 Test Environment	2
4.3 Temperature	2
4.4 Drafts	2
4.5 Extraneous Radiation	2
4.6 Lamp and Lamp System Operation	3
4.6.1 Lamp Operation	3
4.6.2 Lamp System Operation	3
4.6.3 Safety	3
5.0 Measurement Instrumentation	3
5.1 Recommended Measurement Instrumentation - Double Monochromator	3
5.1.1 Instrument Spectral Response	3
5.1.2 Wavelength Accuracy	4
5.1.3 Stray Radiant Power	4
5.1.4 Input Optics Recommendation: Integrating Sphere	4
5.1.4.1 Polarization	5
5.1.4.2 Angular Response	5
5.1.4.3 Monochromator Input	5
5.1.5 Calibration Sources	5
5.2 Broadband Detector	5
5.2.1 An Example of Combined Instrument Testing	5
5.3 Instrument Limitations	5
5.3.1 An Example of a Spectroradiometer System	6
5.4 Linearity	6
5.5 Source Size Measurement	6
5.6 Pulse Width Measurement	6
6.0 Idealized Measurement Procedure	6
6.1 Irradiance Measurement	6
6.1.1 Ideal Instrument	6
6.1.2 Conceptual Realization	7
6.1.3 Application	7
6.2 Spectral Irradiance Measurement	7
6.2.1 Geometry	7
6.2.2 Detector	7
6.2.3 Application	7
6.3 Radiance Measurement	7
6.3.1 Ideal Instrument	7
6.3.2 Conceptual Realization	8
6.3.3 Application	8
6.4 Spectral Radiance Measurement	8
6.4.1 Geometry	8
6.4.2 Detector	8
6.4.3 Application	8

Contents

6.5 Subtended Angle of the Source	8
6.5.1 Circular Uniform Radiance Source	9
6.5.2 Non-Circular Uniform Radiance Source	9
6.5.3 Non-Uniform Radiance Source	9
7.0 Analysis of Results	9
7.1 Weighting Curve Interpolations	9
7.2 Calculations	9
7.3 Measurement Uncertainty	10
Annex A -Uncertainty Analysis	11
Annex B -Stray Radiant Power	12
Annex C -Application Notes and Examples	15
Annex D -Summary Table: RP-27.1 Criteria	28
Annex E -Table of Weighting Functions Interpolated	30

Recommended Practice for Photobiological Safety of Lamps and Lamp Systems—Measurement Techniques

1.0 INTRODUCTION

This standard is the second in a series of standards relating to the photobiological safety of lamps and lamp systems and is devoted to measurement of sources for the purpose of hazard evaluation.

Following the recommendations of this standard, an experienced practitioner with adequate equipment and time will achieve precision and accuracy necessary for the classification of lamps and lamp systems and the verification of results.

The measurement of optical radiation for the purpose of computing photobiological effective quantities as used in the ANSI/IESNA RP-27 series poses a significant challenge. Photobiological action spectra, such as the UV Hazard Weighting Function, $S(\lambda)$, have rapidly changing values with slight changes in wavelength. Furthermore, sources such as lamps with glass envelopes have rapidly increasing output within the same ultraviolet wavelength band where the UV Hazard Weighting Function $S(\lambda)$ is rapidly decreasing. This can be seen in **Annex C, Figure C-5**. Hence, substantial inaccuracies in weighted results can arise from small measurement uncertainties.

The testing done for this series of standards shall include a full analysis of the uncertainty in the results. This requirement leads to several corollary requirements.

- The testing shall be done by persons experienced in radiometry.
- The equipment used shall be fully characterized.
- The testing shall be scrutinized for all influences and sources of error.

This standard recommends a double monochromator system for measurements used in classifying sources, although such instrumentation may not be practical for some types of testing. The standard therefore provides guidance on the use of other methods and when they may be appropriate. Alternative measurement methods described shall be used with full understanding of the limitations of each, and the method selected should be traceable back to spectral measurements. Further, the testing shall include a full analysis of the uncertainty of the results. The equipment used shall be fully characterized and all sources of error documented.

2.0 SCOPE

The ANSI/IESNA RP-27 series of standards applies to all electrically powered sources of optical radiation that emit in the wavelength range from 200 nm to 3,000 nm. The standards do not apply to any lasers or to those light emitting diodes used in optical fiber communication systems.

This standard is to be used by the radiometrist for guidance regarding special problems related to photobiological hazard measurements. Additionally, specific recommendations are included to provide consistency and to reduce test design time and effort.

It is impractical for this standard to teach all of the concepts or provide all experience needed to make accurate photobiological safety measurements.

3.0 REFERENCES

3.1 Normative References

The following documents contain provisions which, through reference in this text, constitute provisions of the American National Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this American National Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below.

ANSI/IESNA RP-27.1-1996, *Photobiological Safety of Lamps and Lamp Systems - General Requirements*. New York: Illuminating Engineering Society of North America.

ANSI/IESNA RP-16-1996, *Nomenclature and Definitions for Illuminating Engineering*. New York: Illuminating Engineering Society of North America.

IESNA LM-9-2000, *Electrical and Photometric Measurements of Fluorescent Lamps*. New York: Illuminating Engineering Society of North America.

IESNA LM-20-1994, *Method for Photometric Testing of Reflector Type Lamps*. New York: Illuminating Engineering Society of North America.

IESNA LM-45-2000, *Electrical and Photometric Measurements of General Service Incandescent Filament Lamps*. New York: Illuminating Engineering Society of North America.