



Approved Method: **Electrical and  
Photometric Measurements  
of Solid-State Lighting  
Products**

*Copyright 2008 by the Illuminating Engineering Society.*

*Approved by the IES Board of Directors December 31, 2007, as a Transaction of the Illuminating Engineering Society.*

*All rights reserved.* No part of this publication may be reproduced in any form, in any electronic retrieval system or otherwise, without prior written permission of the IES.

Published by the Illuminating Engineering Society, 120 Wall Street, New York, New York 10005.

IES Standards and Guides are developed through committee consensus and produced by the IES Office in New York. Careful attention is given to style and accuracy. If any errors are noted in this document, please forward them to Rita Harrold, Director Educational and Technical Development, at the above address for verification and correction. The IES welcomes and urges feedback and comments.

ISBN # 978-0-87995-226-6

Printed in the United States of America.

#### **DISCLAIMER**

IES publications are developed through the consensus standards development process approved by the American National Standards Institute. This process brings together volunteers representing varied viewpoints and interests to achieve consensus on lighting recommendations. While the IES administers the process and establishes policies and procedures to promote fairness in the development of consensus, it makes no guaranty or warranty as to the accuracy or completeness of any information published herein. The IES disclaims liability for any injury to persons or property or other damages of any nature whatsoever, whether special, indirect, consequential or compensatory, directly or indirectly resulting from the publication, use of, or reliance on this document

In issuing and making this document available, the IES is not undertaking to render professional or other services for or on behalf of any person or entity. Nor is the IES undertaking to perform any duty owed by any person or entity to someone else. Anyone using this document should rely on his or her own independent judgment or, as appropriate, seek the advice of a competent professional in determining the exercise of reasonable care in any given circumstances.

The IES has no power, nor does it undertake, to police or enforce compliance with the contents of this document. Nor does the IES list, certify, test or inspect products, designs, or installations for compliance with this document. Any certification or statement of compliance with the requirements of this document shall not be attributable to the IES and is solely the responsibility of the certifier or maker of the statement.

**Prepared by the Subcommittee on Solid-State Lighting of the IES Testing Procedures Committee**

**Solid-State Lighting Subcommittee**

Kevin Dowling, Chair

Yoshi Ohno, Technical Coordinator

R. C. Berger	P. F. Keebler*
R. S. Bergman	M. Kotrebai
E. Bretschneider*	K. K. Krueger
J. R. Cyre	J. P. Marella
M. T. Dyle LC*	M. J. Mayer
S. D. Ellersick*	D. M. Mesh*
D. Ellis*	C. C. Miller
M. Grather	Y. Ohno
P. J. Havens*	M. L. Riebling*
A. Jackson*	M. B. Sapcoe*
J. Jiao*	L. Stafford*
C. F. Jones*	G. Trott*
M. A. Kalkas*	R. C. Tuttle
D. Karambelas	J. W. Yon *
H. S. Kashani*	J. X. Zhang

**IES Testing Procedures Committee**

Michael Grather, Chair

C. K. Andersen	J. Lawton*
D. V. Andreyev*	L. E. Leetzow*
J. B. Arens	K. C. Lerbs*
L. M. Ayers	R. E. Levin*
W.E. Beakes	I. Lewin
R. C. Berger	R. Low*
R. P. Bergin*	J. P. Marella
R. S. Bergman	G. McKee
J. R. Cyre*	S. W. McKnight*
R. C. Dahl*	D. C. Mertz**
R. O. Daubach*	C. C. Miller
K. J. Dowling	B. Mosher
D. Ellis	W. A. Newland
A.M. Foy*	Y. Ohno*
P. J. Franck*	D. W. Parkansky*
R. V. Heinisch	D. N. Randolph
T. T. Hernandez*	D. Rector
R. E. Horan	M. B. Sapcoe
D. E. Husby**	D. C. Smith*
J. Hospodarsky*	R. C. Speck**
J. Jiao*	L. Stafford*
M. A. Kalkas	G. A. Steinberg
D. Karambelas	N. Stuffer**
M. Kotrebai	T. G. Yahraus*
K.K. Krueger	J. X. Zhang
B. Kuebler*	
E. Ladouceur*	

\* Advisory Member  
\*\* Honorary Member

**This approved method has been developed in collaboration with the ANSI Solid State Lighting Joint Working Group C78-09 and C82-04.**

**Special recognition to Yoshi Ohno for his technical assistance, review and collaboration.**

## Contents

<b>1.0</b>	<b>Introduction</b>	<b>1</b>
1.1	Scope	1
1.2	General	1
1.3	Nomenclature and Definitions	1
<b>2.0</b>	<b>Ambient Conditions</b>	<b>2</b>
2.1	General	2
2.2	Air Temperature	2
2.3	Thermal Conditions for Mounting SSL Products	2
2.4	Air Movement	2
<b>3.0</b>	<b>Power Supply Characteristics</b>	<b>2</b>
3.1	Waveshape of AC power supply	2
3.2	Voltage Regulation	2
<b>4.0</b>	<b>Seasoning of SSL Product</b>	<b>3</b>
<b>5.0</b>	<b>Stabilization of SSL Product</b>	<b>3</b>
<b>6.0</b>	<b>Operating Orientation</b>	<b>3</b>
<b>7.0</b>	<b>Electrical Settings</b>	<b>3</b>
<b>8.0</b>	<b>Electrical Instrumentation</b>	<b>3</b>
8.1	Circuits	3
8.2	Uncertainties	3
<b>9.0</b>	<b>Test Methods for Total Luminous Flux measurement</b>	<b>4</b>
9.1	<b>Integrating sphere with a spectroradiometer (Sphere-spectroradiometer system)</b>	<b>4</b>
9.1.1	Integrating sphere	4
9.1.2	Sphere geometry	5
9.1.3	Principle of measurement	6
9.1.4	Spectroradiometer	7
9.1.5	Self-absorption correction	7
9.1.6	Calibration	7
9.2	<b>Integrating sphere with a photometer head (Sphere-photometer system)</b>	<b>7</b>
9.2.1	Integrating sphere	7
9.2.2	Sphere geometry	7
9.2.3	Principle of measurement	8
9.2.4	Photometer head	8
9.2.5	Self-absorption correction	8
9.2.6	Determination of $f'_1$ and Spectral Mismatch Correction Factor	9
9.2.7	Calibration	9
9.3	<b>Goniophotometer</b>	<b>9</b>
9.3.1	Type of Goniometer	9
9.3.2	Principle of Total Luminous Flux Measurement	9
9.3.3	Scanning resolution	10
9.3.4	Angle coverage	10
9.3.5	Polarization	10
9.3.6	Photometer head	10
9.3.7	Calibration	10
<b>10.0</b>	<b>Luminous Intensity Distribution</b>	<b>10</b>
<b>11.0</b>	<b>Luminous Efficacy</b>	<b>11</b>

<b>12.0</b>	<b>Test Methods for Color Characteristics of SSL Products</b> .....	<b>11</b>
12.1	Method using a sphere-spectroradiometer system .....	11
12.2	Method using a spectroradiometer or colorimeter spatially scanned .....	11
12.3	Spectroradiometer parameters impacting measured color characteristics .....	12
12.4	Colorimetric calculations .....	12
12.5	Spatial non-uniformity of chromaticity .....	12
<b>13.0</b>	<b>Uncertainty statement</b> .....	<b>13</b>
<b>14.0</b>	<b>Test Report</b> .....	<b>13</b>
	<b>References</b> .....	<b>14</b>
	<b>Annex (Informative)</b> .....	<b>15</b>

This is a preview of "IESNA LM-79-08". [Click here to purchase the full version from the ANSI store.](#)

# IES Approved Method for the Electrical and Photometric Measurements of Solid-State Lighting Products

## Foreword

This approved method is a guide developed for the measurement of solid state lighting (SSL) products. While many other standards for photometric measurements of light sources and luminaires are available, these standards are separated for measurement of lamps or luminaires. Since the current SSL products are in the forms of luminaires or lamps, and LED light sources in the luminaires are not easily separated as replaceable lamps, these existing standards cannot be applied directly to SSL products. This necessitates the use of absolute photometry. See the annex in this document for a description of how absolute photometry varies from relative photometry, which has historically been the lighting industry standard. Thus, this standard provides test methods addressing the requirements for measurement of SSL products. Since SSL technologies are still at their early stages, requirements for measurement conditions and appropriate measurement techniques may be subject to change at any time as the SSL technologies advance.

---

## 1.0 INTRODUCTION

---

### 1.1 Scope

This approved method describes the procedures to be followed and precautions to be observed in performing reproducible measurements of total luminous flux, electrical power, luminous intensity distribution, and chromaticity, of solid-state lighting (SSL) products for illumination purposes, under standard conditions. This approved method covers LED-based SSL products with control electronics and heat sinks incorporated, that is, those devices that require only AC mains power or a DC voltage power supply to operate. This document does not cover SSL products that require external operating circuits or external heat sinks (e.g., LED chips, LED packages, and LED modules). This document covers SSL products in a form of luminaires (fixtures incorporating light sources) as well as integrated LED lamps (see **section 1.3 f**). This document does not cover fixtures designed for SSL products sold without a light source. This document describes test methods for individual SSL products, and does not cover the determination of the performance rating of products, in which individual variations among the products should be considered.

### 1.2 General

SSL products as defined in this document utilize LEDs (including inorganic and organic LEDs) as the optical radiation sources to generate light for illumination purposes. An LED is a  $p-n$  junction semiconductor device that emits incoherent optical radiation when biased in the forward direction. White light is produced by LEDs using two methods: visible spectra of two or more colors produced by LEDs are mixed, or the emission (in the blue or ultraviolet region) from LEDs is used to excite one or more phosphors to produce broadband emission in the visible region (Stokes emission). A general description of LEDs and lighting is available in Ref. 1. Although constant current control is typical for stand alone LEDs, this document deals with integrated SSL products incorporating the semiconductor device level current control, thus the electrical parameters of interest are the SSL product's input electrical parameters.

For special purposes, it may be useful to determine the characteristics of SSL products when they are operated at other than the standard conditions described in this approved method. Where this is done, such results are meaningful only for the particular conditions under which they were obtained and these conditions shall be stated in the test report.

The photometric information typically required for SSL products is total luminous flux (lumens), luminous efficacy (lm/W), luminous intensity (candelas) in one or more directions, chromaticity coordinates, correlated color temperature, and color rendering index. For the purpose of this approved method, the determination of these data will be considered photometric measurements.

The electrical characteristics measured for AC-powered SSL products are input RMS AC voltage, input RMS AC current, input AC power, input voltage frequency and power factor. For DC-powered SSL products, measured electrical characteristics are input DC voltage, input DC current, and input power. For the purpose of this approved method, the determination of these data will be considered electrical measurements.

### 1.3 Nomenclature and Definitions

- a) Units of electrical measurement are the volt, the ampere and the watt.
- b) Units of photometric measurement are the lumen and the candela<sup>2</sup>. Chromaticity coordinates are specified in terms of the CIE recommended systems<sup>3</sup>, the  $(x, y)$  or  $(u', v')$  chromaticity coordinates. To specify tolerance of chromaticity independent