

**INSTITUTE OF
ENVIRONMENTAL
SCIENCES AND
TECHNOLOGY**

**Contamination Control Division
Technical Guide 1002**

IEST-G-CC1002

**Determination of the
Concentration of
Airborne Ultrafine Particles**

INSTITUTE OF ENVIRONMENTAL SCIENCES AND TECHNOLOGY

Arlington Place One
2340 S. Arlington Heights Road, Suite 620
Arlington Heights, IL 60005-4510
Phone: (847) 981-0100 • Fax: (847) 981-4130
E-mail: iest@iest.org • Web: www.iest.org



This Guide is published by the INSTITUTE OF ENVIRONMENTAL SCIENCES AND TECHNOLOGY to advance contamination control technology and the technical and engineering sciences. Its use is entirely voluntary, and determination of its applicability and suitability for any particular use is solely the responsibility of the user.

Copyright © 1999 by the **INSTITUTE OF ENVIRONMENTAL SCIENCES AND TECHNOLOGY**

ISBN 978-1-877862-75-5

No part of this publication may be copied or reproduced in any form, or by any means, including electronic, mechanical, photocopying, or otherwise, without the prior written permission of the Institute of Environmental Sciences and Technology, 2340 S. Arlington Heights Road, Arlington Heights, IL 60005-4516.

INSTITUTE OF ENVIRONMENTAL SCIENCES AND TECHNOLOGY

Arlington Place One

2340 S. Arlington Heights Road, Suite 620

Arlington Heights, IL 60005-4510

Phone: (847) 981-0100 • Fax: (847) 981-4130

E-mail: iest@iest.org • Web: www.iest.org

Determination of the Concentration of Airborne Ultrafine Particles IEST-G-CC1002

CONTENTS

SECTION

1	Application	5
2	Definitions	5
3	Counting apparatus for ultrafine particles	6
4	Counting procedure	7
5	Data reporting	8
6	References	8

APPENDIX

APPENDIX A; Data Report Example	9
---------------------------------------	---

FIGURE

Figure 1. Envelope of acceptable U descriptor counting efficiency	7
---	---

INSTITUTE OF ENVIRONMENTAL SCIENCES AND TECHNOLOGY
Contamination Control Division
Guide 1002

Determination of the Concentration of Airborne Ultrafine Particles

IEST-G-CC1002

1 Application

This guide describes standardized procedures for determination of the concentration of airborne particles in the ultrafine size range (i.e., particles with a size distribution having a threshold [lower limit] particle size smaller than 0.1 μm). The numerical value of the concentration of such particles, expressed in terms of particles per cubic meter, is the *U descriptor*. This term may refer to a specified concentration of ultrafine particles or to measurement of the concentration of ultrafine particles in a given sample of air.

The procedures described herein are suitable for use in conjunction with related cleanroom standards, such as ISO 14644-1 for determination of airborne particulate cleanliness, and ISO 14644-2 for monitoring. The procedures are applicable to cleanrooms and clean zones in any of three occupancy states, as defined in ISO 14644-1.

This document is most appropriate for use with cleanroom and clean zone environments that qualify as ISO Class 4 or cleaner, as defined in ISO 14644-1.

2 Definitions

2.1 condensation nucleus counter (CNC)

An instrument for counting airborne particles, in the nanometer size range and larger, by optically detecting droplets formed by condensation of a vapor upon those particles.

2.2 discrete-particle counter (DPC)

An instrument, such as an optical particle counter or a modified condensation nucleus counter, capable of counting and sizing individual airborne particles within a particle population. A DPC used for counting ultrafine particles shall be able to discriminate threshold sizes smaller than 0.1 μm equivalent diameter.

2.3 inlet device

A size cutoff device which, when attached to the sample inlet of a DPC or CNC, will remove particles smaller than the defined ultrafine particle size. The device is chosen so that its particle retention by diffusion effects will meet the defined par-