

INSTITUTE OF ENVIRONMENTAL SCIENCES AND TECHNOLOGY

Contamination Control Division Technical Guide 1003

IEST-G-CC1003

Measurement of Airborne Macroparticles

INSTITUTE OF ENVIRONMENTAL SCIENCES AND TECHNOLOGY

Arlington Place One
2340 S. Arlington Heights Road, Suite 620
Arlington Heights, IL 60005-4510
Phone: (847) 981-0100 • Fax: (847) 981-4130
E-mail: iest@iest.org • Web: www.iest.org



This is a preview of "IEST-G-CC1003". [Click here to purchase the full version from the ANSI store.](#)

This Guide is published by the INSTITUTE OF ENVIRONMENTAL SCIENCES AND TECHNOLOGY to advance contamination control technology and the technical and engineering sciences. Its use is entirely voluntary, and determination of its applicability and suitability for any particular use is solely the responsibility of the user.

Copyright © 1999 by the **INSTITUTE OF ENVIRONMENTAL SCIENCES AND TECHNOLOGY**

ISBN 978-1-877862-76-2

No part of this publication may be copied or reproduced in any form, or by any means, including electronic, mechanical, photocopying, or otherwise, without the prior written permission of the Institute of Environmental Sciences and Technology, 2340 S. Arlington Heights Road, Arlington Heights, IL 60005-4516.

INSTITUTE OF ENVIRONMENTAL SCIENCES AND TECHNOLOGY

Arlington Place One
2340 S. Arlington Heights Road, Suite 620
Arlington Heights, IL 60005-4510
Phone: (847) 981-0100 • Fax: (847) 981-4130
E-mail: iest@iest.org • Web: www.iest.org

This is a preview of "IEST-G-CC1003". [Click here to purchase the full version from the ANSI store.](#)

Measurement of Airborne Macroparticles

IEST-G-CC1003

CONTENTS

SECTION

1	Application	7
2	Scope	7
3	Sample handling procedures	8
4	Measurement procedures	10
5	Procedure for determination of the M descriptor	12
6	Data reporting	12
7	References	13

APPENDIX

APPENDIX A; Data Report Example	14
---------------------------------------	----

FIGURE

Figure 1. Sampling probe inlet areas for isokinetic sampling	8
Figure 2. Vertical sampling probe inlet efficiencies in turbulent flow	9

This is a preview of "IEST-G-CC1003". [Click here to purchase the full version from the ANSI store.](#)

INSTITUTE OF ENVIRONMENTAL SCIENCES AND TECHNOLOGY
Contamination Control Division
Guide 1003

Measurement of Airborne Macroparticles

IEST-G-CC1003

1 Application

This guide describes procedures for measurement of the concentration and size of airborne *macroparticles* (i.e., particles that are larger than 5 μm in diameter). The numerical value of the concentration of such particles, expressed in terms of macroparticles per cubic meter, is the *M descriptor*. This term may refer to a specified concentration of macroparticles or to measurement of the concentration of macroparticles in a given sample of air.

The procedures contained herein are suitable for use in conjunction with related cleanroom standards such as ISO 14644-1, for defining air cleanliness, and ISO 14644-2, for monitoring. The procedures are applicable to cleanrooms and clean zones in any of three occupancy states, as defined in ISO 14644-1.

In relation to standard air cleanliness classifications, as defined in ISO 14644-1, this document is typically most appropriate for use with cleanroom and clean zone environments that qualify as ISO Class 5 or less clean. It may also be used in any situation where process or product considerations require monitoring or control of airborne contaminants in the macroparticle size range.

NOTE: For further definitions and explanations of terms in this document, see IEST-RD-CC011.

2 Scope

There are numerous procedures available for counting macroparticles. Since different procedures may not respond to the same particle size parameter(s), the data produced by a given procedure may not correlate with data from another procedure. Therefore, a variety of methods will be discussed herein.

Two general categories of macroparticle measurement operations are considered:

- a) Collection by filtration or inertial effects, followed by microscopic measurement of the number and size, or measurement of the mass of collected particles.
- b) In-situ measurement of the concentration and size of macroparticles with a time-of-flight particle counter or a discrete-particle counter.

The importance of proper sample acquisition and handling in order to minimize losses of macroparticles in the sample handling operations is emphasized. Procedures are provided for calculating macroparticle losses when ideal sample handling procedures cannot be accomplished.