This is a preview of ISO 24477:2025. Click here to purchase the full version from the ANSI store.

International

ISO 24477

Vacuum technology — Vacuum gauges — Specifications, calibration and measurement uncertainties for spinning rotor gauges

Technique du vide — Manomètres à vide — Spécifications, étalonnage et incertitudes de mesure pour manomètres à rotor

Second edition
2025-03

Reference number
ISO 24477:2025(en)

© ISO 2025

This is a preview of ISO 24477:2025. Click here to purchase the full version from the ANSI store.



COPYRIGHT PROTECTED DOCUMENT

© ISO 2025

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

This is a preview of ISO 24477:2025. Click here to purchase the full version from the ANSI store.

Foreword ..... **iv**
Introduction ..... **v**
1 Scope **1**
2 Normative references **1**
3 Terms and definitions **1**
3.1 Terms related to components 2
3.2 Terms related to physical parameters 2
4 Symbols and abbreviated terms **4**
5 Principle of a spinning rotor gauge **4**
6 Specifications for spinning rotor gauge **5**
6.1 Diameter and density of rotor 5
6.2 Materials of rotor and thimble 5
6.3 Connecting flange of thimble 6
6.4 Leak rate of thimble with flange 6
6.5 Suspension head positioning 6
6.6 Limits of rotor frequency 6
6.7 Warm-up period 6
6.8 Baking temperature 6
6.9 Measurement range 6
6.10 Sampling time 6
6.11 Internal volume 6
6.12 Interface and pin connections of controller 6
6.13 Dimensions and weights of suspension head and controller 6
6.14 Display and signal output 6
6.15 Nominal operating (environment) conditions 7
6.16 Storage and transportation conditions 7
6.17 Input power and electrical requirements 7
6.18 Number of display digits 7
7 Additional (optional) specifications for spinning rotor gauge **7**
7.1 Long-term instability 7
7.2 Expected measurement uncertainty 7
7.3 Compatibility between a suspension head and a rotor 7
7.4 Inspection record and calibration certificate 7
7.5 Allowed vibration level 7
7.6 Cable length 8
7.7 Photograph 8
8 Calibration **8**
8.1 Parameters to be calibrated 8
8.2 Calibration procedures 8
8.2.1 General 8
8.2.2 Calibration method A for $\sigma_{eff,0}$ 9
8.2.3 Calibration method B for $\sigma_{eff,0}$ 9
8.3 Calibration uncertainty 10
8.4 Calibration certificate 10
9 Measurement uncertainty at use **10**
Bibliography **11**

© ISO 2025 – All rights reserved

This is a preview of ISO 24477:2025. Click here to purchase the full version from the ANSI store.

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 112, *Vacuum technology*.

This second edition cancels and replaces the first edition (ISO 24477:2022), which has been technically revised.

The main changes are as follows:

— 8.2.3 Calibration method B for $\sigma_{\mathrm{eff},0}$ has been technically revised

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

© ISO 2025 – All rights reserved

This is a preview of ISO 24477:2025. Click here to purchase the full version from the ANSI store.

This document complements ISO 3567 and ISO 27893 when characterizing, calibrating or using spinning rotor gauges (SRGs) as reference gauges.

SRGs are used to measure pressure in the high and medium vacuum. For the dissemination of the pressure scale and measurement of high and medium vacuum pressures by this gauge, the relevant parameters, calibration guidelines and uncertainties should be given, which are described in this document.

© ISO 2025 – All rights reserved