
Water quality — Determination of fluoride —

Part 1:

Electrochemical probe method for potable and
lightly polluted water

Qualité de l'eau — Dosage des fluorures —

*Partie 1: Méthode de la sonde électrochimique pour l'eau potable et
faiblement polluée*



Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 10359-1 was prepared by Technical Committee ISO/TC 147, *Water quality*, Sub-Committee SC 2, *Physical, chemical, biochemical methods*.

ISO 10359 consists of the following parts, under the general title *Water quality — Determination of fluoride*:

- *Part 1: Electrochemical probe method for potable and lightly polluted water*
- *Part 2: Determination of inorganically bound total fluoride after digestion and distillation*

Annex A of this part of ISO 10359 is for information only.

© ISO 1992

All rights reserved. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

This is a preview of "ISO 10359-1:1992". Click [here](#) to purchase the full version from the ANSI store.

Introduction

Fluoride ions occur in almost all ground and surface waters. Their concentration depends primarily on the hydrogeological conditions and is generally below 1 mg/l.

Certain industrial waste waters may also contain fluoride ions in higher concentrations.

The fluoride value is also dependant on the type and concentration of cations present at the same time in water, such as Ca^{2+} , Mg^{2+} , Al^{3+} or Fe^{3+} , which may form sparingly soluble compounds with fluoride ions or complexes of low dissociation constant.

Apart from these compounds, stable boron-fluoride complexes exist.

Several different methods are available for determining fluoride and the choice of method depends on the type of problem posed as follows.

- a) Direct measurement using fluoride ion selective electrodes. This method is suitable for the determination of fluoride in potable and surface water.

It is specified in this part of ISO 10359.

- b) Determination of the total inorganically bound fluoride using decomposition, distillation and potentiometric measurement.

This method will be specified in ISO 10359-2.

This is a preview of "ISO 10359-1:1992". Click here to purchase the full version from the ANSI store.

Water quality — Determination of fluoride —

Part 1:

Electrochemical probe method for potable and lightly polluted water

1 Scope

1.1 Field of application

This part of ISO 10359 specifies a method for the determination of dissolved fluoride in fresh, potable and low contaminated water, and some surface waters, using an electrochemical technique.

The method is directly suitable for measuring fluoride concentrations from 0,2 mg/l to 2,0 g/l.

After the addition of a known amount of fluoride, concentrations as low as 0,02 mg/l can be detected (see 7.3).

The method is not suitable for waste waters and industrial effluents; this determination will be the subject of ISO 10359-2.

1.2 Interferences

The electrode will respond directly to hydroxide ions. The formation of HF under acidic conditions will reduce the measured fluoride concentration. Therefore, buffer all test aliquots to a pH between 5 and 7 to prevent such interference. Cations such as calcium, magnesium, iron and aluminium form complexes with fluoride or precipitates to which the electrode does not respond. Therefore the buffer solution also contains *trans*-1,2-diaminocyclohexane-*N,N,N',N'*-tetraacetic acid (CDTA) as a decomplexing agent to free bound fluoride. The boron tetrafluoride anion, BF_4^- , is not decomplexed by the addition of buffer.

2 Normative reference

The following standard contains provisions which, through reference in this text, constitute provisions of this part of ISO 10359. At the time of publication, the edition indicated was valid. All standards are subject to revision, and parties to agreements based on this part of ISO 10359 are encouraged to investigate the possibility of applying the most recent edition of the standard indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 5667-3:—¹⁾, *Water quality — Sampling — Part 3: Guidance on the preservation and handling of samples*.

3 Principle

When a fluoride ion-selective electrode comes into contact with an aqueous solution containing fluoride ions, a potential difference develops between the measuring electrode and the reference electrode. The value of this potential difference is proportional to the logarithm of the value of the fluoride ion activity in accordance with the Nernst equation.

Temperature and ionic strength may influence the potential difference. Accordingly, these parameters shall be the same during calibration and measurement and shall be kept constant throughout the procedure.

The activity of the fluoride ions is also pH-dependant. Values of pH between 5 and 7 have proved favorable for measurement. Special buffer solutions are used to fix the pH and the activity coefficient.

1) To be published. (Revision of ISO 5667-3:1985)