

This is a preview of "ISO 11070:2014". [Click here to purchase the full version from the ANSI store.](#)

Second edition
2014-11-01

Sterile single-use intravascular introducers, dilators and guidewires

Introducteurs, dilatateurs et guides intravasculaires stériles non réutilisables



Reference number
ISO 11070:2014(E)

© ISO 2014

This is a preview of "ISO 11070:2014". [Click here to purchase the full version from the ANSI store.](#)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2014

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

This is a preview of "ISO 11070:2014". Click here to purchase the full version from the ANSI store.

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General requirements	5
4.1 Sterilization	5
4.2 Biocompatibility	5
4.3 Surface	5
4.4 Corrosion resistance	5
4.5 Radio-detectability	5
4.6 Information to be supplied by the manufacturer	5
5 Additional requirements for introducer needles	6
5.1 General	6
5.2 Size designation	6
5.3 Needle point	6
5.4 Hub	6
5.5 Information to be supplied by the manufacturer	6
6 Additional requirements for introducer catheters	6
6.1 General	6
6.2 Tip	7
6.3 Peak tensile force	7
6.4 Hub	7
6.5 Size designation	7
6.6 Information to be supplied by the manufacturer	8
7 Additional requirements for sheath introducers	8
7.1 General	8
7.2 Size designation	8
7.3 Freedom from leakage from sheath introducer	8
7.4 Freedom from leakage through haemostasis valve	8
7.5 Hub	8
7.6 Peak tensile force	8
7.7 Information to be supplied by the manufacturer	8
8 Additional requirements for guidewires	8
8.1 General	8
8.2 Size designation	9
8.3 Safety wire	9
8.4 Fracture test	9
8.5 Flexing test	9
8.6 Peak tensile force of guidewire	9
8.7 Information to be supplied by the manufacturer	10
9 Additional requirements for dilators	10
9.1 General	10
9.2 Size designation	10
9.3 Hub	10
9.4 Information to be supplied by the manufacturer	10
10 Additional requirements for kits containing combinations of devices specified in this International Standard	10
Annex A (informative) Guidance on materials and design	12

This is a preview of "ISO 11070:2014". [Click here to purchase the full version from the ANSI store.](#)

Annex B (normative) Test method for corrosion resistance	13
Annex C (normative) Method for determining peak tensile force of introducer catheters, sheath introducers, and dilators	14
Annex D (normative) Test method for liquid leakage from sheath introducers under pressure	16
Annex E (normative) Test method for liquid leakage through haemostasis valves of sheath introducers	18
Annex F (normative) Test method for fracture of guidewires	19
Annex G (normative) Test method for resistance of guidewires to damage by flexing	21
Annex H (normative) Method for determining peak tensile force of guidewires	23
Annex I (normative) Determination of strength of union of needle hub and needle	25
Bibliography	26

This is a preview of "ISO 11070:2014". [Click here to purchase the full version from the ANSI store.](#)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword — Supplementary information](#).

The committee responsible for this document is ISO/TC 84, *Devices for administration of medicinal products and catheters*.

This second edition cancels and replaces the first edition (ISO 11070:1998), which has been technically revised.

This is a preview of "ISO 11070:2014". [Click here to purchase the full version from the ANSI store.](#)

Introduction

The purpose of this International Standard is to

- update requirements and test methods to support the function of the guidewire, and
- update size designation.