

This is a preview of "ISO 11500:2022". [Click here to purchase the full version from the ANSI store.](#)

Third edition
2022-12

Hydraulic fluid power — Determination of the particulate contamination level of a liquid sample by automatic particle counting using the light-extinction principle

*Transmissions hydrauliques — Détermination du niveau de pollution
particulaire d'un échantillon liquide par comptage automatique des
particules par absorption de lumière*



Reference number
ISO 11500:2022(E)

© ISO 2022



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

This is a preview of "ISO 11500:2022". [Click here to purchase the full version from the ANSI store.](#)

Contents

	Page
Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 Materials and equipment.....	2
5 Diluent liquid.....	3
6 Pre-test requirements and procedures.....	4
6.1 Precautions.....	4
6.1.1 Chemicals.....	4
6.1.2 Electrical interference.....	4
6.1.3 Use of magnetic stirrer.....	4
6.1.4 Relative humidity.....	4
6.1.5 Sample storage.....	4
6.2 Glassware cleaning procedure.....	4
6.3 APC calibration procedure.....	5
6.4 APC operation.....	5
6.5 Sample inspection and preparation before counting.....	5
6.5.1 Outline.....	5
6.5.2 Initial preparation and inspection.....	5
6.5.3 Preparation of sample of excessive volume.....	6
6.5.4 Determination of presence of water.....	7
6.6 Determination of need for sample dilution.....	7
7 Procedure for determining the particulate contamination level by automatic counting.....	7
7.1 Outline.....	7
7.2 Sample dilution.....	8
7.2.1 Use of sample dilution.....	8
7.2.2 Precautions.....	9
7.2.3 Volumetric dilution method.....	9
7.2.4 Mass dilution method.....	10
7.3 Analysis procedure.....	10
8 Test report.....	12
9 Identification statement (reference to this document).....	13
Annex A (normative) Procedure for performing a statistical check of an automatic particle counter.....	14
Annex B (informative) Acceptable diluents.....	15
Annex C (informative) Method for pre-cleaning diluent and method for incorporating additives into the diluent to eliminate the influence of static electricity on particle counts.....	16
Annex D (informative) Reporting particulate contamination level in a hydraulic liquid sample as determined by automatic counting.....	18
Bibliography.....	20

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 131, *Fluid power systems*, Subcommittee SC 6, *Contamination control*.

This third edition cancels and replaces the second edition (ISO 11500:2008), which has been technically revised.

The main changes are as follows:

- the cleanliness requirements for sample containers have been updated to match ISO 11171;
- the instructions for flushing and diluting solutions using propan-2-ol (2-propanol) and demineralised water have been removed;
- Annex E has been deleted.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This is a preview of "ISO 11500:2022". [Click here to purchase the full version from the ANSI store.](#)

Introduction

In hydraulic fluid power systems, power is transmitted and controlled through a liquid under pressure within an enclosed circuit. The liquid is both a lubricant and a power-transmitting medium.

The presence of solid contaminant particles in the liquid interferes with the ability of the hydraulic fluid to lubricate and causes wear to the components. The extent of contamination in the fluid has a direct bearing on the performance and reliability of the system, and it is essential to control solid contaminant particles to levels that are considered appropriate for the system concerned.

A quantitative determination of particulate contamination involves precision in obtaining the sample and in determining the extent of contamination. The liquid automatic particle counter (APC), which works on the light-extinction principle, has become an accepted means of determining the extent of contamination. The accuracy of particle count data can be affected by the techniques used to obtain such data.

This document details procedures for the analysis of contaminated liquid samples using an automatic particle counter. Correct use of an automatic particle counter helps to reduce errors and enhances the accuracy of reproducibility in data.